



284 Sheffield Street, Mountainside, New Jersey 07092, Phone : 908 789 8900,
Fax : 908 789 8922

Initial and Continuing Calibration Verification

Client: VERINA CONSULTING GROUP, LLC

SDG No.: Q2512

Project: Rotor Clip NJ WTD - 2025

RunNo.: LB136370

Analyte	Units	Result	True Value	% Recovery	Acceptance Window (%R)	Analysis Date
Sample ID: ICV Residual Chlorine	mg/L	0.419	0.4	105	90-110	07/03/2025
Sample ID: CCV1 Residual Chlorine	mg/L	0.378	0.4	94	90-110	07/03/2025
Sample ID: CCV2 Residual Chlorine	mg/L	0.388	0.4	97	90-110	07/03/2025

Initial and Continuing Calibration Verification

Client: VERINA CONSULTING GROUP, LLC

SDG No.: Q2512

Project: Rotor Clip NJ WTD - 2025

RunNo.: LB136405

Analyte	Units	Result	True Value	% Recovery	Acceptance Window (%R)	Analysis Date
Sample ID: ICV1 Ammonia as N	mg/L	0.99	1	99	90-110	07/09/2025
Sample ID: CCV1 Ammonia as N	mg/L	0.94	1	94	90-110	07/09/2025
Sample ID: CCV2 Ammonia as N	mg/L	0.96	1	96	90-110	07/09/2025
Sample ID: CCV3 Ammonia as N	mg/L	1	1	100	90-110	07/09/2025
Sample ID: CCV4 Ammonia as N	mg/L	0.96	1	96	90-110	07/09/2025
Sample ID: CCV5 Ammonia as N	mg/L	0.99	1	99	90-110	07/09/2025