

Duplicate Sample Summary

Client:	VERINA CONSULTING GROUP, LLC	SDG No.:	Q2512
Project:	Rotor Clip NJ WTD - 2025	Sample ID:	Q2512-01
Client ID:	WATER-TREATMENT DISCHARGEDUP	Percent Solids for Spike Sample:	0

Analyte	Units	Acceptance Limit	Sample Result	Conc. Qualifier	Duplicate Result	Conc. Qualifier	Dilution Factor	RPD/AD	Qual	Analysis Date
Residual Chlorine	mg/L	+/-20	0.023	U	0.023	U	1	0		07/03/2025
Ammonia as N	mg/L	+/-20	2.20	OR	2.30	OR	1	4		07/09/2025
Ammonia as N	mg/L	+/-20	2.40	D	2.40	D	2	0		07/09/2025

Duplicate Sample Summary

Client: VERINA CONSULTING GROUP, LLC	SDG No.: Q2512
Project: Rotor Clip NJ WTD - 2025	Sample ID: Q2512-01
Client ID: WATER-TREATMENT DISCHARGEMSD	Percent Solids for Spike Sample: 0

Analyte	Units	Acceptance Limit	Sample Result	Conc. Qualifier	Duplicate Result	Conc. Qualifier	Dilution Factor	RPD/ AD	Qual	Analysis Date
Residual Chlorine	mg/L	+/-20	0.37		0.38		1	2.68		07/03/2025
Ammonia as N	mg/L	+/-20	3.20	OR	3.10	OR	1	3		07/09/2025