

Laboratory Control Sample Summary

Client:	VERINA CONSULTING GROUP, LLC	SDG No.:	Q2512
Project:	Rotor Clip NJ WTD - 2025	Run No.:	LB136370

Analyte	Units	True Value	Result	Conc. Qualifier	% Recovery	Dilution Factor	Acceptance Limit %R	Analysis Date
Sample ID	LB136370BS							
Residual Chlorine	mg/L	0.4	0.41		102	1	90-110	07/03/2025

Laboratory Control Sample Summary

Client:	VERINA CONSULTING GROUP, LLC	SDG No.:	Q2512
Project:	Rotor Clip NJ WTD - 2025	Run No.:	LB136405

Analyte	Units	True Value	Result	Conc. Qualifier	% Recovery	Dilution Factor	Acceptance Limit %R	Analysis Date
Sample ID	PB168758BS							
Ammonia as N	mg/L	1	0.99		99	1	90-110	07/09/2025