



284 Sheffield Street, Mountainside, New Jersey 07092, Phone : 908 789 8900,
Fax : 908 789 8922

Metals

- 3a -

INITIAL AND CONTINUING CALIBRATION BLANK SUMMARY

Client: VERINA CONSULTING GROUP, LLC

SDG No.: Q2512

Contract: VERI01

Lab Code: ACE

Sample ID	Analyte	Result ug/L	Acceptance Limit	Conc Qual	CRQL	M	Analysis Date	Analysis Time	Run Number
ICB01	Chromium	2.12	+/-5	U	10.0	P	07/10/2025	18:55	LB136434
	Copper	4.60	+/-10	U	20.0	P	07/10/2025	18:55	LB136434
	Nickel	3.06	+/-20	U	40.0	P	07/10/2025	18:55	LB136434
	Zinc	3.50	+/-20	U	40.0	P	07/10/2025	18:55	LB136434

Metals

- 3a -

INITIAL AND CONTINUING CALIBRATION BLANK SUMMARY

Client: VERINA CONSULTING GROUP, LLC

SDG No.: Q2512

Contract: VERI01

Lab Code: ACE

Sample ID	Analyte	Result ug/L	Acceptance Limit	Conc Qual	CRQL	M	Analysis Date	Analysis Time	Run Number
CCB01	Chromium	2.12	+/-5	U	10.0	P	07/10/2025	19:54	LB136434
	Copper	4.60	+/-10	U	20.0	P	07/10/2025	19:54	LB136434
	Nickel	3.06	+/-20	U	40.0	P	07/10/2025	19:54	LB136434
	Zinc	3.50	+/-20	U	40.0	P	07/10/2025	19:54	LB136434
CCB02	Chromium	2.12	+/-5	U	10.0	P	07/10/2025	21:04	LB136434
	Copper	4.60	+/-10	U	20.0	P	07/10/2025	21:04	LB136434
	Nickel	3.06	+/-20	U	40.0	P	07/10/2025	21:04	LB136434
	Zinc	3.50	+/-20	U	40.0	P	07/10/2025	21:04	LB136434
CCB03	Chromium	2.12	+/-5	U	10.0	P	07/10/2025	22:36	LB136434
	Copper	4.60	+/-10	U	20.0	P	07/10/2025	22:36	LB136434
	Nickel	3.06	+/-20	U	40.0	P	07/10/2025	22:36	LB136434
	Zinc	3.50	+/-20	J	40.0	P	07/10/2025	22:36	LB136434
CCB04	Chromium	2.12	+/-5	U	10.0	P	07/10/2025	23:53	LB136434
	Copper	4.60	+/-10	U	20.0	P	07/10/2025	23:53	LB136434
	Nickel	3.06	+/-20	U	40.0	P	07/10/2025	23:53	LB136434
	Zinc	3.50	+/-20	U	40.0	P	07/10/2025	23:53	LB136434
CCB05	Chromium	2.12	+/-5	U	10.0	P	07/11/2025	00:51	LB136434
	Copper	4.60	+/-10	U	20.0	P	07/11/2025	00:51	LB136434
	Nickel	3.06	+/-20	U	40.0	P	07/11/2025	00:51	LB136434
	Zinc	3.50	+/-20	U	40.0	P	07/11/2025	00:51	LB136434
CCB06	Chromium	2.12	+/-5	U	10.0	P	07/11/2025	02:23	LB136434
	Copper	4.60	+/-10	U	20.0	P	07/11/2025	02:23	LB136434
	Nickel	3.06	+/-20	U	40.0	P	07/11/2025	02:23	LB136434
	Zinc	3.50	+/-20	U	40.0	P	07/11/2025	02:23	LB136434
CCB07	Chromium	2.12	+/-5	U	10.0	P	07/11/2025	03:26	LB136434
	Copper	4.60	+/-10	U	20.0	P	07/11/2025	03:26	LB136434
	Nickel	3.06	+/-20	U	40.0	P	07/11/2025	03:26	LB136434
	Zinc	3.50	+/-20	U	40.0	P	07/11/2025	03:26	LB136434
CCB08	Chromium	2.12	+/-5	U	10.0	P	07/11/2025	04:27	LB136434
	Copper	4.60	+/-10	U	20.0	P	07/11/2025	04:27	LB136434
	Nickel	3.06	+/-20	U	40.0	P	07/11/2025	04:27	LB136434
	Zinc	3.50	+/-20	U	40.0	P	07/11/2025	04:27	LB136434
CCB09	Chromium	2.12	+/-5	U	10.0	P	07/11/2025	05:24	LB136434
	Copper	4.60	+/-10	U	20.0	P	07/11/2025	05:24	LB136434
	Nickel	3.06	+/-20	U	40.0	P	07/11/2025	05:24	LB136434
	Zinc	3.50	+/-20	U	40.0	P	07/11/2025	05:24	LB136434
CCB10	Chromium	2.12	+/-5	U	10.0	P	07/11/2025	06:21	LB136434
	Copper	4.60	+/-10	U	20.0	P	07/11/2025	06:21	LB136434
	Nickel	3.06	+/-20	U	40.0	P	07/11/2025	06:21	LB136434

Metals

- 3a -

INITIAL AND CONTINUING CALIBRATION BLANK SUMMARY

Client: VERINA CONSULTING GROUP, LLC

SDG No.: Q2512

Contract: VERI01

Lab Code: ACE

Sample ID	Analyte	Result ug/L	Acceptance Limit	Conc Qual	CRQL	M	Analysis Date	Analysis Time	Run Number
CCB10	Zinc	3.50	+/-20	U	40.0	P	07/11/2025	06:21	LB136434
CCB11	Chromium	2.12	+/-5	U	10.0	P	07/11/2025	06:47	LB136434
	Copper	4.60	+/-10	U	20.0	P	07/11/2025	06:47	LB136434
	Nickel	3.06	+/-20	U	40.0	P	07/11/2025	06:47	LB136434
	Zinc	3.50	+/-20	U	40.0	P	07/11/2025	06:47	LB136434