

**Metals**  
**- 3b -**  
**PREPARATION BLANK SUMMARY**

**Client:** VERINA CONSULTING GROUP, LLC

**SDG No.:** Q2512

**Instrument:** P5

| Sample ID         | Analyte      | Result<br>(ug/L) | Acceptance<br>Limit | Conc<br>Qual         | CRQL<br>ug/L    | M | Analysis<br>Date  | Analysis<br>Time  | Run      |
|-------------------|--------------|------------------|---------------------|----------------------|-----------------|---|-------------------|-------------------|----------|
| <b>PB168739BL</b> | <b>WATER</b> |                  |                     | <b>Batch Number:</b> | <b>PB168739</b> |   | <b>Prep Date:</b> | <b>07/07/2025</b> |          |
|                   | Chromium     | 1.06             | <2.5                | U                    | 5.00            | P | 07/11/2025        | 00:56             | LB136434 |
|                   | Copper       | 2.30             | <5                  | U                    | 10.0            | P | 07/11/2025        | 00:56             | LB136434 |
|                   | Nickel       | 1.53             | <10                 | U                    | 20.0            | P | 07/11/2025        | 00:56             | LB136434 |
|                   | Zinc         | 1.75             | <10                 | U                    | 20.0            | P | 07/11/2025        | 00:56             | LB136434 |