

Metals

- 4 -

INTERFERENCE CHECK SAMPLE

Client: VERINA CONSULTING GROUP, LLC

SDG No.: Q2512

Contract: VERI01

Lab Code: ACE

ICS Source: EPA

Instrument ID: P5

Sample ID	Analyte	Result ug/L	True Value ug/L	% Recovery	Low Limit (ug/L)	High Limit (ug/L)	Analysis Date	Analysis Time	Run Number
ICSA01	Chromium	48.9	52.0	94	42	62	07/10/2025	19:11	LB136434
	Copper	-9.61	2.0	480	-18	22	07/10/2025	19:11	LB136434
	Nickel	16.4	2.0	820	-38	42	07/10/2025	19:11	LB136434
	Zinc	11.0			-40	40	07/10/2025	19:11	LB136434
ICSAB01	Chromium	548	540	102	460	624	07/10/2025	19:15	LB136434
	Copper	468	510	92	434	588	07/10/2025	19:15	LB136434
	Nickel	971	950	102	810	1100	07/10/2025	19:15	LB136434
	Zinc	983	950	104	809	1095	07/10/2025	19:15	LB136434