

metals
- 5a -
MATRIX SPIKE SUMMARY

client: VERINA CONSULTING GROUP, LLC **level:** low **sdg no.:** Q2512
contract: VERI01 **lab code:** ACE
matrix: Water **sample id:** Q2512-01 **client id:** WATER TREATMENT DISCHARGEMS
Percent Solids for Sample: NA **Spiked ID:** Q2512-01MS **Percent Solids for Spike Sample:** NA

Analyte	Units	Acceptance Limit %R	Spiked Result	C	Sample Result	C	Spike Added	% Recovery	Qual	M
Chromium	ug/L	75 - 125	336		138		200	99		P
Copper	ug/L	75 - 125	142		8.21	J	150	89		P
Nickel	ug/L	75 - 125	298		38.8		250	104		P
Zinc	ug/L	75 - 125	407		319		100	88		P

metals
- 5a -
MATRIX SPIKE DUPLICATE SUMMARY

client: <u>VERINA CONSULTING GROUP, LLC</u>	level: <u>low</u>	sdg no.: <u>Q2512</u>
contract: <u>VERI01</u>		lab code: <u>ACE</u>
matrix: <u>Water</u>	sample id: <u>Q2512-01</u>	client id: <u>WATER TREATMENT DISCHARGEMSD</u>
Percent Solids for Sample: <u>NA</u>	Spiked ID: <u>Q2512-01MSD</u>	Percent Solids for Spike Sample: <u>NA</u>

Analyte	Units	Acceptance Limit %R	MSD Result	C	Sample Result	C	Spike Added	% Recovery	Qual	M
Chromium	ug/L	75 - 125	337		138		200	99		P
Copper	ug/L	75 - 125	145		8.21	J	150	91		P
Nickel	ug/L	75 - 125	304		38.8		250	106		P
Zinc	ug/L	75 - 125	412		319		100	93		P