

Metals

- 7 -

LABORATORY CONTROL SAMPLE SUMMARY

Client: VERINA CONSULTING GROUP, LLC

SDG No.: Q2512

Contract: VERI01

Lab Code: ACE

Analyte	Units	True Value	Result	C	% Recovery	Acceptance Limits	M
PB168739BS							
Chromium	ug/L	200	198		99	80 - 120	P
Copper	ug/L	150	162		108	80 - 120	P
Nickel	ug/L	250	233		93	80 - 120	P
Zinc	ug/L	100	100		100	80 - 120	P