



284 Sheffield Street, Mountainside, New Jersey 07092, Phone : 908 789 8900,
Fax : 908 789 8922

Metals

- 3a -

INITIAL AND CONTINUING CALIBRATION BLANK SUMMARY

Client: CDM Smith

SDG No.: Q2514

Contract: CAMP02

Lab Code: ACE

Sample ID	Analyte	Result ug/L	Acceptance Limit	Conc Qual	CRQL	M	Analysis Date	Analysis Time	Run Number
ICB29	Mercury	0.076	+/-0.2	U	0.20	CV	07/16/2025	11:59	LB136499

Metals

- 3a -

INITIAL AND CONTINUING CALIBRATION BLANK SUMMARY

Client: CDM Smith

SDG No.: Q2514

Contract: CAMP02

Lab Code: ACE

Sample ID	Analyte	Result ug/L	Acceptance Limit	Conc Qual	CRQL	M	Analysis Date	Analysis Time	Run Number
CCB09	Mercury	0.076	+/-0.2	U	0.20	CV	07/16/2025	12:03	LB136499
CCB10	Mercury	0.076	+/-0.2	U	0.20	CV	07/16/2025	12:43	LB136499
CCB11	Mercury	0.076	+/-0.2	U	0.20	CV	07/16/2025	13:14	LB136499
CCB12	Mercury	0.076	+/-0.2	U	0.20	CV	07/16/2025	13:42	LB136499
CCB13	Mercury	0.076	+/-0.2	U	0.20	CV	07/16/2025	14:10	LB136499
CCB14	Mercury	0.076	+/-0.2	U	0.20	CV	07/16/2025	14:28	LB136499

Metals

- 3a -

INITIAL AND CONTINUING CALIBRATION BLANK SUMMARY

Client: CDM Smith

SDG No.: Q2514

Contract: CAMP02

Lab Code: ACE

Sample ID	Analyte	Result ug/L	Acceptance Limit	Conc Qual	CRQL	M	Analysis Date	Analysis Time	Run Number
ICB01	Aluminum	31.1	+/-50	J	100	P	07/09/2025	10:22	LB136407
	Antimony	6.76	+/-25	U	50.0	P	07/09/2025	10:22	LB136407
	Arsenic	5.12	+/-10	U	20.0	P	07/09/2025	10:22	LB136407
	Barium	14.6	+/-50	U	100	P	07/09/2025	10:22	LB136407
	Beryllium	0.56	+/-3	U	6.00	P	07/09/2025	10:22	LB136407
	Cadmium	0.50	+/-3	U	6.00	P	07/09/2025	10:22	LB136407
	Calcium	234	+/-1000	U	2000	P	07/09/2025	10:22	LB136407
	Chromium	2.12	+/-5	U	10.0	P	07/09/2025	10:22	LB136407
	Cobalt	2.26	+/-15	U	30.0	P	07/09/2025	10:22	LB136407
	Copper	4.60	+/-10	U	20.0	P	07/09/2025	10:22	LB136407
	Iron	25.4	+/-50	J	100	P	07/09/2025	10:22	LB136407
	Lead	2.30	+/-6	U	12.0	P	07/09/2025	10:22	LB136407
	Magnesium	244	+/-1000	U	2000	P	07/09/2025	10:22	LB136407
	Manganese	7.46	+/-10	J	20.0	P	07/09/2025	10:22	LB136407
	Nickel	3.06	+/-20	U	40.0	P	07/09/2025	10:22	LB136407
	Potassium	918	+/-1000	U	2000	P	07/09/2025	10:22	LB136407
	Selenium	9.64	+/-10	U	20.0	P	07/09/2025	10:22	LB136407
	Silver	1.62	+/-5	U	10.0	P	07/09/2025	10:22	LB136407
	Sodium	868	+/-1000	U	2000	P	07/09/2025	10:22	LB136407
	Thallium	4.38	+/-20	U	40.0	P	07/09/2025	10:22	LB136407
	Vanadium	6.26	+/-20	U	40.0	P	07/09/2025	10:22	LB136407
	Zinc	4.21	+/-20	J	40.0	P	07/09/2025	10:22	LB136407

Metals

- 3a -

INITIAL AND CONTINUING CALIBRATION BLANK SUMMARY

Client: CDM Smith

SDG No.: Q2514

Contract: CAMP02

Lab Code: ACE

Sample ID	Analyte	Result ug/L	Acceptance Limit	Conc Qual	CRQL	M	Analysis Date	Analysis Time	Run Number
CCB01	Aluminum	11.3	+/-50	U	100	P	07/09/2025	10:57	LB136407
	Antimony	6.76	+/-25	U	50.0	P	07/09/2025	10:57	LB136407
	Arsenic	5.12	+/-10	U	20.0	P	07/09/2025	10:57	LB136407
	Barium	14.6	+/-50	U	100	P	07/09/2025	10:57	LB136407
	Beryllium	0.56	+/-3	U	6.00	P	07/09/2025	10:57	LB136407
	Cadmium	0.50	+/-3	U	6.00	P	07/09/2025	10:57	LB136407
	Calcium	234	+/-1000	U	2000	P	07/09/2025	10:57	LB136407
	Chromium	2.12	+/-5	U	10.0	P	07/09/2025	10:57	LB136407
	Cobalt	2.26	+/-15	U	30.0	P	07/09/2025	10:57	LB136407
	Copper	4.60	+/-10	U	20.0	P	07/09/2025	10:57	LB136407
	Iron	23.4	+/-50	U	100	P	07/09/2025	10:57	LB136407
	Lead	2.30	+/-6	U	12.0	P	07/09/2025	10:57	LB136407
	Magnesium	244	+/-1000	U	2000	P	07/09/2025	10:57	LB136407
	Manganese	5.94	+/-10	U	20.0	P	07/09/2025	10:57	LB136407
	Nickel	3.06	+/-20	U	40.0	P	07/09/2025	10:57	LB136407
	Potassium	918	+/-1000	U	2000	P	07/09/2025	10:57	LB136407
	Selenium	9.64	+/-10	U	20.0	P	07/09/2025	10:57	LB136407
	Silver	1.62	+/-5	U	10.0	P	07/09/2025	10:57	LB136407
	Sodium	868	+/-1000	U	2000	P	07/09/2025	10:57	LB136407
	Thallium	4.38	+/-20	U	40.0	P	07/09/2025	10:57	LB136407
	Vanadium	6.26	+/-20	U	40.0	P	07/09/2025	10:57	LB136407
	Zinc	3.50	+/-20	U	40.0	P	07/09/2025	10:57	LB136407
CCB02	Aluminum	27.5	+/-50	J	100	P	07/09/2025	11:56	LB136407
	Antimony	6.76	+/-25	U	50.0	P	07/09/2025	11:56	LB136407
	Arsenic	5.12	+/-10	U	20.0	P	07/09/2025	11:56	LB136407
	Barium	14.6	+/-50	U	100	P	07/09/2025	11:56	LB136407
	Beryllium	0.56	+/-3	U	6.00	P	07/09/2025	11:56	LB136407
	Cadmium	0.50	+/-3	U	6.00	P	07/09/2025	11:56	LB136407
	Calcium	234	+/-1000	U	2000	P	07/09/2025	11:56	LB136407
	Chromium	2.12	+/-5	U	10.0	P	07/09/2025	11:56	LB136407
	Cobalt	2.26	+/-15	U	30.0	P	07/09/2025	11:56	LB136407
	Copper	4.60	+/-10	U	20.0	P	07/09/2025	11:56	LB136407
	Iron	23.4	+/-50	U	100	P	07/09/2025	11:56	LB136407
	Lead	2.30	+/-6	U	12.0	P	07/09/2025	11:56	LB136407
	Magnesium	244	+/-1000	U	2000	P	07/09/2025	11:56	LB136407
	Manganese	5.94	+/-10	U	20.0	P	07/09/2025	11:56	LB136407
	Nickel	3.06	+/-20	U	40.0	P	07/09/2025	11:56	LB136407
	Potassium	918	+/-1000	U	2000	P	07/09/2025	11:56	LB136407
	Selenium	9.64	+/-10	U	20.0	P	07/09/2025	11:56	LB136407

Metals

- 3a -

INITIAL AND CONTINUING CALIBRATION BLANK SUMMARY

Client: CDM Smith

SDG No.: Q2514

Contract: CAMP02

Lab Code: ACE

Sample ID	Analyte	Result ug/L	Acceptance Limit	Conc Qual	CRQL	M	Analysis Date	Analysis Time	Run Number
CCB02	Silver	1.62	+/-5	U	10.0	P	07/09/2025	11:56	LB136407
	Sodium	868	+/-1000	U	2000	P	07/09/2025	11:56	LB136407
	Thallium	4.38	+/-20	U	40.0	P	07/09/2025	11:56	LB136407
	Vanadium	6.26	+/-20	U	40.0	P	07/09/2025	11:56	LB136407
	Zinc	3.50	+/-20	U	40.0	P	07/09/2025	11:56	LB136407
CCB03	Aluminum	65.9	+/-50	J*	100	P	07/09/2025	12:58	LB136407
	Antimony	6.76	+/-25	U	50.0	P	07/09/2025	12:58	LB136407
	Arsenic	5.12	+/-10	U	20.0	P	07/09/2025	12:58	LB136407
	Barium	14.6	+/-50	U	100	P	07/09/2025	12:58	LB136407
	Beryllium	0.56	+/-3	U	6.00	P	07/09/2025	12:58	LB136407
	Cadmium	0.50	+/-3	U	6.00	P	07/09/2025	12:58	LB136407
	Calcium	234	+/-1000	U	2000	P	07/09/2025	12:58	LB136407
	Chromium	2.12	+/-5	U	10.0	P	07/09/2025	12:58	LB136407
	Cobalt	2.26	+/-15	U	30.0	P	07/09/2025	12:58	LB136407
	Copper	4.60	+/-10	U	20.0	P	07/09/2025	12:58	LB136407
	Iron	23.4	+/-50	U	100	P	07/09/2025	12:58	LB136407
	Lead	2.30	+/-6	U	12.0	P	07/09/2025	12:58	LB136407
	Magnesium	244	+/-1000	U	2000	P	07/09/2025	12:58	LB136407
	Manganese	5.94	+/-10	U	20.0	P	07/09/2025	12:58	LB136407
	Nickel	3.06	+/-20	U	40.0	P	07/09/2025	12:58	LB136407
	Potassium	918	+/-1000	U	2000	P	07/09/2025	12:58	LB136407
	Selenium	9.64	+/-10	U	20.0	P	07/09/2025	12:58	LB136407
	Silver	1.62	+/-5	U	10.0	P	07/09/2025	12:58	LB136407
	Sodium	868	+/-1000	U	2000	P	07/09/2025	12:58	LB136407
	Thallium	4.38	+/-20	U	40.0	P	07/09/2025	12:58	LB136407
	Vanadium	6.26	+/-20	U	40.0	P	07/09/2025	12:58	LB136407
	Zinc	3.50	+/-20	U	40.0	P	07/09/2025	12:58	LB136407
CCB04	Aluminum	47.3	+/-50	J	100	P	07/09/2025	14:28	LB136407
	Antimony	6.76	+/-25	U	50.0	P	07/09/2025	14:28	LB136407
	Arsenic	5.12	+/-10	U	20.0	P	07/09/2025	14:28	LB136407
	Barium	14.6	+/-50	U	100	P	07/09/2025	14:28	LB136407
	Beryllium	0.56	+/-3	U	6.00	P	07/09/2025	14:28	LB136407
	Cadmium	0.50	+/-3	U	6.00	P	07/09/2025	14:28	LB136407
	Calcium	234	+/-1000	U	2000	P	07/09/2025	14:28	LB136407
	Chromium	2.12	+/-5	U	10.0	P	07/09/2025	14:28	LB136407
	Cobalt	2.26	+/-15	U	30.0	P	07/09/2025	14:28	LB136407
	Copper	4.60	+/-10	U	20.0	P	07/09/2025	14:28	LB136407
	Iron	23.4	+/-50	U	100	P	07/09/2025	14:28	LB136407
	Lead	2.30	+/-6	U	12.0	P	07/09/2025	14:28	LB136407

Metals

- 3a -

INITIAL AND CONTINUING CALIBRATION BLANK SUMMARY

Client: CDM Smith

SDG No.: Q2514

Contract: CAMP02

Lab Code: ACE

Sample ID	Analyte	Result ug/L	Acceptance Limit	Conc Qual	CRQL	M	Analysis Date	Analysis Time	Run Number
CCB04	Magnesium	244	+/-1000	U	2000	P	07/09/2025	14:28	LB136407
	Manganese	5.94	+/-10	U	20.0	P	07/09/2025	14:28	LB136407
	Nickel	3.06	+/-20	U	40.0	P	07/09/2025	14:28	LB136407
	Potassium	918	+/-1000	U	2000	P	07/09/2025	14:28	LB136407
	Selenium	9.64	+/-10	U	20.0	P	07/09/2025	14:28	LB136407
	Silver	7.86	+/-5	J*	10.0	P	07/09/2025	14:28	LB136407
	Sodium	868	+/-1000	U	2000	P	07/09/2025	14:28	LB136407
	Thallium	4.38	+/-20	U	40.0	P	07/09/2025	14:28	LB136407
	Vanadium	6.26	+/-20	U	40.0	P	07/09/2025	14:28	LB136407
	Zinc	3.50	+/-20	U	40.0	P	07/09/2025	14:28	LB136407
CCB05	Aluminum	26.6	+/-50	J	100	P	07/09/2025	15:33	LB136407
	Antimony	6.76	+/-25	U	50.0	P	07/09/2025	15:33	LB136407
	Arsenic	5.12	+/-10	U	20.0	P	07/09/2025	15:33	LB136407
	Barium	14.6	+/-50	U	100	P	07/09/2025	15:33	LB136407
	Beryllium	0.56	+/-3	U	6.00	P	07/09/2025	15:33	LB136407
	Cadmium	0.50	+/-3	U	6.00	P	07/09/2025	15:33	LB136407
	Calcium	234	+/-1000	U	2000	P	07/09/2025	15:33	LB136407
	Chromium	2.12	+/-5	U	10.0	P	07/09/2025	15:33	LB136407
	Cobalt	2.26	+/-15	U	30.0	P	07/09/2025	15:33	LB136407
	Copper	4.60	+/-10	U	20.0	P	07/09/2025	15:33	LB136407
	Iron	23.4	+/-50	U	100	P	07/09/2025	15:33	LB136407
	Lead	2.30	+/-6	U	12.0	P	07/09/2025	15:33	LB136407
	Magnesium	244	+/-1000	U	2000	P	07/09/2025	15:33	LB136407
	Manganese	5.94	+/-10	U	20.0	P	07/09/2025	15:33	LB136407
	Nickel	3.06	+/-20	U	40.0	P	07/09/2025	15:33	LB136407
	Potassium	918	+/-1000	U	2000	P	07/09/2025	15:33	LB136407
	Selenium	9.64	+/-10	U	20.0	P	07/09/2025	15:33	LB136407
	Silver	6.92	+/-5	J*	10.0	P	07/09/2025	15:33	LB136407
	Sodium	868	+/-1000	U	2000	P	07/09/2025	15:33	LB136407
	Thallium	4.38	+/-20	U	40.0	P	07/09/2025	15:33	LB136407
	Vanadium	6.26	+/-20	U	40.0	P	07/09/2025	15:33	LB136407
	Zinc	3.50	+/-20	U	40.0	P	07/09/2025	15:33	LB136407
CCB06	Aluminum	18.8	+/-50	J	100	P	07/09/2025	16:43	LB136407
	Antimony	6.76	+/-25	U	50.0	P	07/09/2025	16:43	LB136407
	Arsenic	5.12	+/-10	U	20.0	P	07/09/2025	16:43	LB136407
	Barium	14.6	+/-50	U	100	P	07/09/2025	16:43	LB136407
	Beryllium	0.56	+/-3	U	6.00	P	07/09/2025	16:43	LB136407
	Cadmium	0.50	+/-3	U	6.00	P	07/09/2025	16:43	LB136407
	Calcium	234	+/-1000	U	2000	P	07/09/2025	16:43	LB136407

Metals

- 3a -

INITIAL AND CONTINUING CALIBRATION BLANK SUMMARY

Client: CDM Smith

SDG No.: Q2514

Contract: CAMP02

Lab Code: ACE

Sample ID	Analyte	Result ug/L	Acceptance Limit	Conc Qual	CRQL	M	Analysis Date	Analysis Time	Run Number
CCB06	Chromium	2.12	+/-5	U	10.0	P	07/09/2025	16:43	LB136407
	Cobalt	2.26	+/-15	U	30.0	P	07/09/2025	16:43	LB136407
	Copper	4.60	+/-10	U	20.0	P	07/09/2025	16:43	LB136407
	Iron	23.4	+/-50	U	100	P	07/09/2025	16:43	LB136407
	Lead	2.30	+/-6	U	12.0	P	07/09/2025	16:43	LB136407
	Magnesium	244	+/-1000	U	2000	P	07/09/2025	16:43	LB136407
	Manganese	5.94	+/-10	U	20.0	P	07/09/2025	16:43	LB136407
	Nickel	3.06	+/-20	U	40.0	P	07/09/2025	16:43	LB136407
	Potassium	918	+/-1000	U	2000	P	07/09/2025	16:43	LB136407
	Selenium	9.64	+/-10	U	20.0	P	07/09/2025	16:43	LB136407
	Silver	3.51	+/-5	J	10.0	P	07/09/2025	16:43	LB136407
	Sodium	868	+/-1000	U	2000	P	07/09/2025	16:43	LB136407
	Thallium	4.38	+/-20	U	40.0	P	07/09/2025	16:43	LB136407
	Vanadium	6.26	+/-20	U	40.0	P	07/09/2025	16:43	LB136407
	Zinc	3.50	+/-20	U	40.0	P	07/09/2025	16:43	LB136407
CCB07	Aluminum	11.3	+/-50	U	100	P	07/09/2025	17:54	LB136407
	Antimony	6.76	+/-25	U	50.0	P	07/09/2025	17:54	LB136407
	Arsenic	5.12	+/-10	U	20.0	P	07/09/2025	17:54	LB136407
	Barium	14.6	+/-50	U	100	P	07/09/2025	17:54	LB136407
	Beryllium	0.56	+/-3	U	6.00	P	07/09/2025	17:54	LB136407
	Cadmium	0.50	+/-3	U	6.00	P	07/09/2025	17:54	LB136407
	Calcium	234	+/-1000	U	2000	P	07/09/2025	17:54	LB136407
	Chromium	2.12	+/-5	U	10.0	P	07/09/2025	17:54	LB136407
	Cobalt	2.26	+/-15	U	30.0	P	07/09/2025	17:54	LB136407
	Copper	4.60	+/-10	U	20.0	P	07/09/2025	17:54	LB136407
	Iron	23.4	+/-50	U	100	P	07/09/2025	17:54	LB136407
	Lead	2.30	+/-6	U	12.0	P	07/09/2025	17:54	LB136407
	Magnesium	244	+/-1000	U	2000	P	07/09/2025	17:54	LB136407
	Manganese	5.94	+/-10	U	20.0	P	07/09/2025	17:54	LB136407
	Nickel	3.06	+/-20	U	40.0	P	07/09/2025	17:54	LB136407
	Potassium	918	+/-1000	U	2000	P	07/09/2025	17:54	LB136407
	Selenium	9.64	+/-10	U	20.0	P	07/09/2025	17:54	LB136407
	Silver	2.23	+/-5	J	10.0	P	07/09/2025	17:54	LB136407
	Sodium	868	+/-1000	U	2000	P	07/09/2025	17:54	LB136407
	Thallium	4.38	+/-20	U	40.0	P	07/09/2025	17:54	LB136407
	Vanadium	6.26	+/-20	U	40.0	P	07/09/2025	17:54	LB136407
	Zinc	3.50	+/-20	U	40.0	P	07/09/2025	17:54	LB136407
CCB08	Aluminum	44.5	+/-50	J	100	P	07/09/2025	18:58	LB136407
	Antimony	6.76	+/-25	U	50.0	P	07/09/2025	18:58	LB136407

Metals

- 3a -

INITIAL AND CONTINUING CALIBRATION BLANK SUMMARY

Client: CDM Smith

SDG No.: Q2514

Contract: CAMP02

Lab Code: ACE

Sample ID	Analyte	Result ug/L	Acceptance Limit	Conc Qual	CRQL	M	Analysis Date	Analysis Time	Run Number
CCB08	Arsenic	5.12	+/-10	U	20.0	P	07/09/2025	18:58	LB136407
	Barium	14.6	+/-50	U	100	P	07/09/2025	18:58	LB136407
	Beryllium	0.56	+/-3	U	6.00	P	07/09/2025	18:58	LB136407
	Cadmium	0.50	+/-3	U	6.00	P	07/09/2025	18:58	LB136407
	Calcium	234	+/-1000	U	2000	P	07/09/2025	18:58	LB136407
	Chromium	2.12	+/-5	U	10.0	P	07/09/2025	18:58	LB136407
	Cobalt	2.26	+/-15	U	30.0	P	07/09/2025	18:58	LB136407
	Copper	4.60	+/-10	U	20.0	P	07/09/2025	18:58	LB136407
	Iron	23.4	+/-50	U	100	P	07/09/2025	18:58	LB136407
	Lead	3.36	+/-6	J	12.0	P	07/09/2025	18:58	LB136407
	Magnesium	244	+/-1000	U	2000	P	07/09/2025	18:58	LB136407
	Manganese	5.94	+/-10	U	20.0	P	07/09/2025	18:58	LB136407
	Nickel	3.06	+/-20	U	40.0	P	07/09/2025	18:58	LB136407
	Potassium	918	+/-1000	U	2000	P	07/09/2025	18:58	LB136407
	Selenium	9.64	+/-10	U	20.0	P	07/09/2025	18:58	LB136407
	Silver	2.96	+/-5	J	10.0	P	07/09/2025	18:58	LB136407
	Sodium	868	+/-1000	U	2000	P	07/09/2025	18:58	LB136407
	Thallium	4.38	+/-20	U	40.0	P	07/09/2025	18:58	LB136407
	Vanadium	6.26	+/-20	U	40.0	P	07/09/2025	18:58	LB136407
	Zinc	3.50	+/-20	U	40.0	P	07/09/2025	18:58	LB136407
CCB09	Aluminum	11.3	+/-50	U	100	P	07/09/2025	19:52	LB136407
	Antimony	6.76	+/-25	U	50.0	P	07/09/2025	19:52	LB136407
	Arsenic	5.12	+/-10	U	20.0	P	07/09/2025	19:52	LB136407
	Barium	14.6	+/-50	U	100	P	07/09/2025	19:52	LB136407
	Beryllium	0.56	+/-3	U	6.00	P	07/09/2025	19:52	LB136407
	Cadmium	0.50	+/-3	U	6.00	P	07/09/2025	19:52	LB136407
	Calcium	234	+/-1000	U	2000	P	07/09/2025	19:52	LB136407
	Chromium	2.12	+/-5	U	10.0	P	07/09/2025	19:52	LB136407
	Cobalt	2.26	+/-15	U	30.0	P	07/09/2025	19:52	LB136407
	Copper	4.60	+/-10	U	20.0	P	07/09/2025	19:52	LB136407
	Iron	23.4	+/-50	U	100	P	07/09/2025	19:52	LB136407
	Lead	2.75	+/-6	J	12.0	P	07/09/2025	19:52	LB136407
	Magnesium	244	+/-1000	U	2000	P	07/09/2025	19:52	LB136407
	Manganese	5.94	+/-10	U	20.0	P	07/09/2025	19:52	LB136407
	Nickel	3.06	+/-20	U	40.0	P	07/09/2025	19:52	LB136407
	Potassium	918	+/-1000	U	2000	P	07/09/2025	19:52	LB136407
	Selenium	9.64	+/-10	U	20.0	P	07/09/2025	19:52	LB136407
	Silver	2.59	+/-5	J	10.0	P	07/09/2025	19:52	LB136407
	Sodium	868	+/-1000	U	2000	P	07/09/2025	19:52	LB136407

Metals

- 3a -

INITIAL AND CONTINUING CALIBRATION BLANK SUMMARY

Client: CDM Smith

SDG No.: Q2514

Contract: CAMP02

Lab Code: ACE

Sample ID	Analyte	Result ug/L	Acceptance Limit	Conc Qual	CRQL	M	Analysis Date	Analysis Time	Run Number
CCB09	Thallium	4.38	+/-20	U	40.0	P	07/09/2025	19:52	LB136407
	Vanadium	6.26	+/-20	U	40.0	P	07/09/2025	19:52	LB136407
	Zinc	3.50	+/-20	U	40.0	P	07/09/2025	19:52	LB136407
CCB10	Aluminum	11.3	+/-50	U	100	P	07/09/2025	20:45	LB136407
	Antimony	6.76	+/-25	U	50.0	P	07/09/2025	20:45	LB136407
	Arsenic	5.12	+/-10	U	20.0	P	07/09/2025	20:45	LB136407
	Barium	14.6	+/-50	U	100	P	07/09/2025	20:45	LB136407
	Beryllium	0.56	+/-3	U	6.00	P	07/09/2025	20:45	LB136407
	Cadmium	0.50	+/-3	U	6.00	P	07/09/2025	20:45	LB136407
	Calcium	234	+/-1000	U	2000	P	07/09/2025	20:45	LB136407
	Chromium	2.12	+/-5	U	10.0	P	07/09/2025	20:45	LB136407
	Cobalt	2.26	+/-15	U	30.0	P	07/09/2025	20:45	LB136407
	Copper	4.60	+/-10	U	20.0	P	07/09/2025	20:45	LB136407
	Iron	23.4	+/-50	U	100	P	07/09/2025	20:45	LB136407
	Lead	2.30	+/-6	U	12.0	P	07/09/2025	20:45	LB136407
	Magnesium	244	+/-1000	U	2000	P	07/09/2025	20:45	LB136407
	Manganese	5.94	+/-10	U	20.0	P	07/09/2025	20:45	LB136407
	Nickel	3.06	+/-20	U	40.0	P	07/09/2025	20:45	LB136407
	Potassium	918	+/-1000	U	2000	P	07/09/2025	20:45	LB136407
	Selenium	9.64	+/-10	U	20.0	P	07/09/2025	20:45	LB136407
	Silver	1.92	+/-5	J	10.0	P	07/09/2025	20:45	LB136407
	Sodium	868	+/-1000	U	2000	P	07/09/2025	20:45	LB136407
	Thallium	4.38	+/-20	U	40.0	P	07/09/2025	20:45	LB136407
	Vanadium	6.26	+/-20	U	40.0	P	07/09/2025	20:45	LB136407
	Zinc	3.50	+/-20	U	40.0	P	07/09/2025	20:45	LB136407
	Aluminum	23.1	+/-50	J	100	P	07/09/2025	21:16	LB136407
	Antimony	6.76	+/-25	U	50.0	P	07/09/2025	21:16	LB136407
	Arsenic	5.12	+/-10	U	20.0	P	07/09/2025	21:16	LB136407
	Barium	14.6	+/-50	U	100	P	07/09/2025	21:16	LB136407
	Beryllium	0.56	+/-3	U	6.00	P	07/09/2025	21:16	LB136407
	Cadmium	0.50	+/-3	U	6.00	P	07/09/2025	21:16	LB136407
	Calcium	234	+/-1000	U	2000	P	07/09/2025	21:16	LB136407
	Chromium	2.12	+/-5	U	10.0	P	07/09/2025	21:16	LB136407
	Cobalt	2.26	+/-15	U	30.0	P	07/09/2025	21:16	LB136407
	Copper	4.60	+/-10	U	20.0	P	07/09/2025	21:16	LB136407
	Iron	23.4	+/-50	U	100	P	07/09/2025	21:16	LB136407
	Lead	4.79	+/-6	J	12.0	P	07/09/2025	21:16	LB136407
	Magnesium	244	+/-1000	U	2000	P	07/09/2025	21:16	LB136407
	Manganese	5.94	+/-10	U	20.0	P	07/09/2025	21:16	LB136407

Metals

- 3a -

INITIAL AND CONTINUING CALIBRATION BLANK SUMMARY

Client: CDM Smith

SDG No.: Q2514

Contract: CAMP02

Lab Code: ACE

Sample ID	Analyte	Result ug/L	Acceptance Limit	Conc Qual	CRQL	M	Analysis Date	Analysis Time	Run Number
CCB11	Nickel	3.06	+/-20	U	40.0	P	07/09/2025	21:16	LB136407
	Potassium	918	+/-1000	U	2000	P	07/09/2025	21:16	LB136407
	Selenium	9.64	+/-10	U	20.0	P	07/09/2025	21:16	LB136407
	Silver	1.62	+/-5	U	10.0	P	07/09/2025	21:16	LB136407
	Sodium	868	+/-1000	U	2000	P	07/09/2025	21:16	LB136407
	Thallium	4.38	+/-20	U	40.0	P	07/09/2025	21:16	LB136407
	Vanadium	6.26	+/-20	U	40.0	P	07/09/2025	21:16	LB136407
	Zinc	3.50	+/-20	U	40.0	P	07/09/2025	21:16	LB136407

Metals

- 3a -

INITIAL AND CONTINUING CALIBRATION BLANK SUMMARY

Client: CDM Smith

SDG No.: Q2514

Contract: CAMP02

Lab Code: ACE

Sample ID	Analyte	Result ug/L	Acceptance Limit	Conc Qual	CRQL	M	Analysis Date	Analysis Time	Run Number
ICB01	Aluminum	11.3	+/-50	U	100	P	07/10/2025	18:55	LB136434
	Antimony	6.76	+/-25	U	50.0	P	07/10/2025	18:55	LB136434
	Arsenic	5.12	+/-10	U	20.0	P	07/10/2025	18:55	LB136434
	Barium	14.6	+/-50	U	100	P	07/10/2025	18:55	LB136434
	Beryllium	0.56	+/-3	U	6.00	P	07/10/2025	18:55	LB136434
	Cadmium	0.50	+/-3	U	6.00	P	07/10/2025	18:55	LB136434
	Calcium	234	+/-1000	U	2000	P	07/10/2025	18:55	LB136434
	Chromium	2.12	+/-5	U	10.0	P	07/10/2025	18:55	LB136434
	Cobalt	2.26	+/-15	U	30.0	P	07/10/2025	18:55	LB136434
	Copper	4.60	+/-10	U	20.0	P	07/10/2025	18:55	LB136434
	Iron	23.4	+/-50	U	100	P	07/10/2025	18:55	LB136434
	Lead	2.30	+/-6	U	12.0	P	07/10/2025	18:55	LB136434
	Magnesium	244	+/-1000	U	2000	P	07/10/2025	18:55	LB136434
	Manganese	5.94	+/-10	U	20.0	P	07/10/2025	18:55	LB136434
	Nickel	3.06	+/-20	U	40.0	P	07/10/2025	18:55	LB136434
	Potassium	918	+/-1000	U	2000	P	07/10/2025	18:55	LB136434
	Selenium	9.64	+/-10	U	20.0	P	07/10/2025	18:55	LB136434
	Silver	1.62	+/-5	U	10.0	P	07/10/2025	18:55	LB136434
	Sodium	868	+/-1000	U	2000	P	07/10/2025	18:55	LB136434
	Thallium	4.38	+/-20	U	40.0	P	07/10/2025	18:55	LB136434
	Vanadium	6.26	+/-20	U	40.0	P	07/10/2025	18:55	LB136434
	Zinc	3.50	+/-20	U	40.0	P	07/10/2025	18:55	LB136434

Metals

- 3a -

INITIAL AND CONTINUING CALIBRATION BLANK SUMMARY

Client: CDM Smith

SDG No.: Q2514

Contract: CAMP02

Lab Code: ACE

Sample ID	Analyte	Result ug/L	Acceptance Limit	Conc Qual	CRQL	M	Analysis Date	Analysis Time	Run Number
CCB01	Aluminum	11.3	+/-50	U	100	P	07/10/2025	19:54	LB136434
	Antimony	6.76	+/-25	U	50.0	P	07/10/2025	19:54	LB136434
	Arsenic	5.12	+/-10	U	20.0	P	07/10/2025	19:54	LB136434
	Barium	14.6	+/-50	U	100	P	07/10/2025	19:54	LB136434
	Beryllium	0.56	+/-3	U	6.00	P	07/10/2025	19:54	LB136434
	Cadmium	0.50	+/-3	U	6.00	P	07/10/2025	19:54	LB136434
	Calcium	234	+/-1000	U	2000	P	07/10/2025	19:54	LB136434
	Chromium	2.12	+/-5	U	10.0	P	07/10/2025	19:54	LB136434
	Cobalt	2.26	+/-15	U	30.0	P	07/10/2025	19:54	LB136434
	Copper	4.60	+/-10	U	20.0	P	07/10/2025	19:54	LB136434
	Iron	23.4	+/-50	U	100	P	07/10/2025	19:54	LB136434
	Lead	2.30	+/-6	U	12.0	P	07/10/2025	19:54	LB136434
	Magnesium	244	+/-1000	U	2000	P	07/10/2025	19:54	LB136434
	Manganese	5.94	+/-10	U	20.0	P	07/10/2025	19:54	LB136434
	Nickel	3.06	+/-20	U	40.0	P	07/10/2025	19:54	LB136434
	Potassium	918	+/-1000	U	2000	P	07/10/2025	19:54	LB136434
	Selenium	9.64	+/-10	U	20.0	P	07/10/2025	19:54	LB136434
	Silver	1.62	+/-5	U	10.0	P	07/10/2025	19:54	LB136434
	Sodium	868	+/-1000	U	2000	P	07/10/2025	19:54	LB136434
	Thallium	4.38	+/-20	U	40.0	P	07/10/2025	19:54	LB136434
	Vanadium	6.26	+/-20	U	40.0	P	07/10/2025	19:54	LB136434
	Zinc	3.50	+/-20	U	40.0	P	07/10/2025	19:54	LB136434
CCB02	Aluminum	11.3	+/-50	U	100	P	07/10/2025	21:04	LB136434
	Antimony	6.76	+/-25	U	50.0	P	07/10/2025	21:04	LB136434
	Arsenic	5.12	+/-10	U	20.0	P	07/10/2025	21:04	LB136434
	Barium	14.6	+/-50	U	100	P	07/10/2025	21:04	LB136434
	Beryllium	0.56	+/-3	U	6.00	P	07/10/2025	21:04	LB136434
	Cadmium	0.50	+/-3	U	6.00	P	07/10/2025	21:04	LB136434
	Calcium	234	+/-1000	U	2000	P	07/10/2025	21:04	LB136434
	Chromium	2.12	+/-5	U	10.0	P	07/10/2025	21:04	LB136434
	Cobalt	2.26	+/-15	U	30.0	P	07/10/2025	21:04	LB136434
	Copper	4.60	+/-10	U	20.0	P	07/10/2025	21:04	LB136434
	Iron	23.4	+/-50	U	100	P	07/10/2025	21:04	LB136434
	Lead	2.30	+/-6	U	12.0	P	07/10/2025	21:04	LB136434
	Magnesium	244	+/-1000	U	2000	P	07/10/2025	21:04	LB136434
	Manganese	5.94	+/-10	U	20.0	P	07/10/2025	21:04	LB136434
	Nickel	3.06	+/-20	U	40.0	P	07/10/2025	21:04	LB136434
	Potassium	918	+/-1000	U	2000	P	07/10/2025	21:04	LB136434
	Selenium	9.64	+/-10	U	20.0	P	07/10/2025	21:04	LB136434

Metals

- 3a -

INITIAL AND CONTINUING CALIBRATION BLANK SUMMARY

Client: CDM Smith

SDG No.: Q2514

Contract: CAMP02

Lab Code: ACE

Sample ID	Analyte	Result ug/L	Acceptance Limit	Conc Qual	CRQL	M	Analysis Date	Analysis Time	Run Number
CCB02	Silver	1.62	+/-5	U	10.0	P	07/10/2025	21:04	LB136434
	Sodium	868	+/-1000	U	2000	P	07/10/2025	21:04	LB136434
	Thallium	4.38	+/-20	U	40.0	P	07/10/2025	21:04	LB136434
	Vanadium	6.26	+/-20	U	40.0	P	07/10/2025	21:04	LB136434
	Zinc	3.50	+/-20	U	40.0	P	07/10/2025	21:04	LB136434
CCB03	Aluminum	11.3	+/-50	U	100	P	07/10/2025	22:36	LB136434
	Antimony	6.76	+/-25	U	50.0	P	07/10/2025	22:36	LB136434
	Arsenic	5.12	+/-10	U	20.0	P	07/10/2025	22:36	LB136434
	Barium	14.6	+/-50	U	100	P	07/10/2025	22:36	LB136434
	Beryllium	0.56	+/-3	U	6.00	P	07/10/2025	22:36	LB136434
	Cadmium	0.50	+/-3	U	6.00	P	07/10/2025	22:36	LB136434
	Calcium	234	+/-1000	U	2000	P	07/10/2025	22:36	LB136434
	Chromium	2.12	+/-5	U	10.0	P	07/10/2025	22:36	LB136434
	Cobalt	2.26	+/-15	U	30.0	P	07/10/2025	22:36	LB136434
	Copper	4.60	+/-10	U	20.0	P	07/10/2025	22:36	LB136434
	Iron	23.4	+/-50	U	100	P	07/10/2025	22:36	LB136434
	Lead	2.30	+/-6	U	12.0	P	07/10/2025	22:36	LB136434
	Magnesium	244	+/-1000	U	2000	P	07/10/2025	22:36	LB136434
	Manganese	5.94	+/-10	U	20.0	P	07/10/2025	22:36	LB136434
	Nickel	3.06	+/-20	U	40.0	P	07/10/2025	22:36	LB136434
	Potassium	918	+/-1000	U	2000	P	07/10/2025	22:36	LB136434
	Selenium	9.64	+/-10	U	20.0	P	07/10/2025	22:36	LB136434
	Silver	1.62	+/-5	U	10.0	P	07/10/2025	22:36	LB136434
	Sodium	868	+/-1000	U	2000	P	07/10/2025	22:36	LB136434
	Thallium	4.38	+/-20	U	40.0	P	07/10/2025	22:36	LB136434
	Vanadium	6.26	+/-20	U	40.0	P	07/10/2025	22:36	LB136434
	Zinc	3.50	+/-20	J	40.0	P	07/10/2025	22:36	LB136434
CCB04	Aluminum	11.3	+/-50	U	100	P	07/10/2025	23:53	LB136434
	Antimony	6.76	+/-25	U	50.0	P	07/10/2025	23:53	LB136434
	Arsenic	5.12	+/-10	U	20.0	P	07/10/2025	23:53	LB136434
	Barium	14.6	+/-50	U	100	P	07/10/2025	23:53	LB136434
	Beryllium	0.56	+/-3	U	6.00	P	07/10/2025	23:53	LB136434
	Cadmium	0.50	+/-3	U	6.00	P	07/10/2025	23:53	LB136434
	Calcium	234	+/-1000	U	2000	P	07/10/2025	23:53	LB136434
	Chromium	2.12	+/-5	U	10.0	P	07/10/2025	23:53	LB136434
	Cobalt	2.26	+/-15	U	30.0	P	07/10/2025	23:53	LB136434
	Copper	4.60	+/-10	U	20.0	P	07/10/2025	23:53	LB136434
	Iron	23.4	+/-50	U	100	P	07/10/2025	23:53	LB136434
	Lead	2.30	+/-6	U	12.0	P	07/10/2025	23:53	LB136434

Metals

- 3a -

INITIAL AND CONTINUING CALIBRATION BLANK SUMMARY

Client: CDM Smith

SDG No.: Q2514

Contract: CAMP02

Lab Code: ACE

Sample ID	Analyte	Result ug/L	Acceptance Limit	Conc Qual	CRQL	M	Analysis Date	Analysis Time	Run Number
CCB04	Magnesium	244	+/-1000	U	2000	P	07/10/2025	23:53	LB136434
	Manganese	5.94	+/-10	U	20.0	P	07/10/2025	23:53	LB136434
	Nickel	3.06	+/-20	U	40.0	P	07/10/2025	23:53	LB136434
	Potassium	918	+/-1000	U	2000	P	07/10/2025	23:53	LB136434
	Selenium	9.64	+/-10	U	20.0	P	07/10/2025	23:53	LB136434
	Silver	1.62	+/-5	U	10.0	P	07/10/2025	23:53	LB136434
	Sodium	868	+/-1000	U	2000	P	07/10/2025	23:53	LB136434
	Thallium	4.38	+/-20	U	40.0	P	07/10/2025	23:53	LB136434
	Vanadium	6.26	+/-20	U	40.0	P	07/10/2025	23:53	LB136434
	Zinc	3.50	+/-20	U	40.0	P	07/10/2025	23:53	LB136434
CCB05	Aluminum	11.3	+/-50	U	100	P	07/11/2025	00:51	LB136434
	Antimony	6.76	+/-25	U	50.0	P	07/11/2025	00:51	LB136434
	Arsenic	5.12	+/-10	U	20.0	P	07/11/2025	00:51	LB136434
	Barium	14.6	+/-50	U	100	P	07/11/2025	00:51	LB136434
	Beryllium	0.56	+/-3	U	6.00	P	07/11/2025	00:51	LB136434
	Cadmium	0.50	+/-3	U	6.00	P	07/11/2025	00:51	LB136434
	Calcium	234	+/-1000	U	2000	P	07/11/2025	00:51	LB136434
	Chromium	2.12	+/-5	U	10.0	P	07/11/2025	00:51	LB136434
	Cobalt	2.26	+/-15	U	30.0	P	07/11/2025	00:51	LB136434
	Copper	4.60	+/-10	U	20.0	P	07/11/2025	00:51	LB136434
	Iron	23.4	+/-50	U	100	P	07/11/2025	00:51	LB136434
	Lead	2.30	+/-6	U	12.0	P	07/11/2025	00:51	LB136434
	Magnesium	244	+/-1000	U	2000	P	07/11/2025	00:51	LB136434
	Manganese	5.94	+/-10	U	20.0	P	07/11/2025	00:51	LB136434
	Nickel	3.06	+/-20	U	40.0	P	07/11/2025	00:51	LB136434
	Potassium	918	+/-1000	U	2000	P	07/11/2025	00:51	LB136434
	Selenium	9.64	+/-10	U	20.0	P	07/11/2025	00:51	LB136434
	Silver	1.62	+/-5	U	10.0	P	07/11/2025	00:51	LB136434
	Sodium	868	+/-1000	U	2000	P	07/11/2025	00:51	LB136434
	Thallium	4.38	+/-20	U	40.0	P	07/11/2025	00:51	LB136434
	Vanadium	6.26	+/-20	U	40.0	P	07/11/2025	00:51	LB136434
	Zinc	3.50	+/-20	U	40.0	P	07/11/2025	00:51	LB136434
CCB06	Aluminum	11.3	+/-50	U	100	P	07/11/2025	02:23	LB136434
	Antimony	6.76	+/-25	U	50.0	P	07/11/2025	02:23	LB136434
	Arsenic	5.12	+/-10	U	20.0	P	07/11/2025	02:23	LB136434
	Barium	14.6	+/-50	U	100	P	07/11/2025	02:23	LB136434
	Beryllium	0.56	+/-3	U	6.00	P	07/11/2025	02:23	LB136434
	Cadmium	0.50	+/-3	U	6.00	P	07/11/2025	02:23	LB136434
	Calcium	234	+/-1000	U	2000	P	07/11/2025	02:23	LB136434

Metals

- 3a -

INITIAL AND CONTINUING CALIBRATION BLANK SUMMARY

Client: CDM Smith

SDG No.: Q2514

Contract: CAMP02

Lab Code: ACE

Sample ID	Analyte	Result ug/L	Acceptance Limit	Conc Qual	CRQL	M	Analysis Date	Analysis Time	Run Number
CCB06	Chromium	2.12	+/-5	U	10.0	P	07/11/2025	02:23	LB136434
	Cobalt	2.26	+/-15	U	30.0	P	07/11/2025	02:23	LB136434
	Copper	4.60	+/-10	U	20.0	P	07/11/2025	02:23	LB136434
	Iron	23.4	+/-50	U	100	P	07/11/2025	02:23	LB136434
	Lead	2.30	+/-6	U	12.0	P	07/11/2025	02:23	LB136434
	Magnesium	244	+/-1000	U	2000	P	07/11/2025	02:23	LB136434
	Manganese	8.75	+/-10	J	20.0	P	07/11/2025	02:23	LB136434
	Nickel	3.06	+/-20	U	40.0	P	07/11/2025	02:23	LB136434
	Potassium	918	+/-1000	U	2000	P	07/11/2025	02:23	LB136434
	Selenium	9.64	+/-10	U	20.0	P	07/11/2025	02:23	LB136434
	Silver	4.87	+/-5	J	10.0	P	07/11/2025	02:23	LB136434
	Sodium	868	+/-1000	U	2000	P	07/11/2025	02:23	LB136434
	Thallium	4.38	+/-20	U	40.0	P	07/11/2025	02:23	LB136434
	Vanadium	6.26	+/-20	U	40.0	P	07/11/2025	02:23	LB136434
	Zinc	3.50	+/-20	U	40.0	P	07/11/2025	02:23	LB136434
CCB07	Aluminum	11.3	+/-50	U	100	P	07/11/2025	03:26	LB136434
	Antimony	6.76	+/-25	U	50.0	P	07/11/2025	03:26	LB136434
	Arsenic	5.12	+/-10	U	20.0	P	07/11/2025	03:26	LB136434
	Barium	14.6	+/-50	U	100	P	07/11/2025	03:26	LB136434
	Beryllium	0.56	+/-3	U	6.00	P	07/11/2025	03:26	LB136434
	Cadmium	0.50	+/-3	U	6.00	P	07/11/2025	03:26	LB136434
	Calcium	234	+/-1000	U	2000	P	07/11/2025	03:26	LB136434
	Chromium	2.12	+/-5	U	10.0	P	07/11/2025	03:26	LB136434
	Cobalt	2.26	+/-15	U	30.0	P	07/11/2025	03:26	LB136434
	Copper	4.60	+/-10	U	20.0	P	07/11/2025	03:26	LB136434
	Iron	23.4	+/-50	U	100	P	07/11/2025	03:26	LB136434
	Lead	2.30	+/-6	U	12.0	P	07/11/2025	03:26	LB136434
	Magnesium	244	+/-1000	U	2000	P	07/11/2025	03:26	LB136434
	Manganese	9.63	+/-10	J	20.0	P	07/11/2025	03:26	LB136434
	Nickel	3.06	+/-20	U	40.0	P	07/11/2025	03:26	LB136434
	Potassium	918	+/-1000	U	2000	P	07/11/2025	03:26	LB136434
	Selenium	9.64	+/-10	U	20.0	P	07/11/2025	03:26	LB136434
	Silver	4.57	+/-5	J	10.0	P	07/11/2025	03:26	LB136434
	Sodium	868	+/-1000	U	2000	P	07/11/2025	03:26	LB136434
	Thallium	4.38	+/-20	U	40.0	P	07/11/2025	03:26	LB136434
	Vanadium	6.26	+/-20	U	40.0	P	07/11/2025	03:26	LB136434
	Zinc	3.50	+/-20	U	40.0	P	07/11/2025	03:26	LB136434
CCB08	Aluminum	11.3	+/-50	U	100	P	07/11/2025	04:27	LB136434
	Antimony	6.76	+/-25	U	50.0	P	07/11/2025	04:27	LB136434

Metals

- 3a -

INITIAL AND CONTINUING CALIBRATION BLANK SUMMARY

Client: CDM Smith

SDG No.: Q2514

Contract: CAMP02

Lab Code: ACE

Sample ID	Analyte	Result ug/L	Acceptance Limit	Conc Qual	CRQL	M	Analysis Date	Analysis Time	Run Number
CCB08	Arsenic	5.12	+/-10	U	20.0	P	07/11/2025	04:27	LB136434
	Barium	14.6	+/-50	U	100	P	07/11/2025	04:27	LB136434
	Beryllium	0.56	+/-3	U	6.00	P	07/11/2025	04:27	LB136434
	Cadmium	0.50	+/-3	U	6.00	P	07/11/2025	04:27	LB136434
	Calcium	234	+/-1000	U	2000	P	07/11/2025	04:27	LB136434
	Chromium	2.12	+/-5	U	10.0	P	07/11/2025	04:27	LB136434
	Cobalt	2.26	+/-15	U	30.0	P	07/11/2025	04:27	LB136434
	Copper	4.60	+/-10	U	20.0	P	07/11/2025	04:27	LB136434
	Iron	23.4	+/-50	U	100	P	07/11/2025	04:27	LB136434
	Lead	2.30	+/-6	U	12.0	P	07/11/2025	04:27	LB136434
	Magnesium	244	+/-1000	U	2000	P	07/11/2025	04:27	LB136434
	Manganese	5.94	+/-10	U	20.0	P	07/11/2025	04:27	LB136434
	Nickel	3.06	+/-20	U	40.0	P	07/11/2025	04:27	LB136434
	Potassium	918	+/-1000	U	2000	P	07/11/2025	04:27	LB136434
	Selenium	9.64	+/-10	U	20.0	P	07/11/2025	04:27	LB136434
	Silver	6.06	+/-5	J*	10.0	P	07/11/2025	04:27	LB136434
	Sodium	868	+/-1000	U	2000	P	07/11/2025	04:27	LB136434
	Thallium	4.38	+/-20	U	40.0	P	07/11/2025	04:27	LB136434
	Vanadium	6.26	+/-20	U	40.0	P	07/11/2025	04:27	LB136434
	Zinc	3.50	+/-20	U	40.0	P	07/11/2025	04:27	LB136434
CCB09	Aluminum	11.3	+/-50	U	100	P	07/11/2025	05:24	LB136434
	Antimony	6.76	+/-25	U	50.0	P	07/11/2025	05:24	LB136434
	Arsenic	5.12	+/-10	U	20.0	P	07/11/2025	05:24	LB136434
	Barium	14.6	+/-50	U	100	P	07/11/2025	05:24	LB136434
	Beryllium	0.56	+/-3	U	6.00	P	07/11/2025	05:24	LB136434
	Cadmium	0.50	+/-3	U	6.00	P	07/11/2025	05:24	LB136434
	Calcium	234	+/-1000	U	2000	P	07/11/2025	05:24	LB136434
	Chromium	2.12	+/-5	U	10.0	P	07/11/2025	05:24	LB136434
	Cobalt	2.26	+/-15	U	30.0	P	07/11/2025	05:24	LB136434
	Copper	4.60	+/-10	U	20.0	P	07/11/2025	05:24	LB136434
	Iron	23.4	+/-50	U	100	P	07/11/2025	05:24	LB136434
	Lead	2.30	+/-6	U	12.0	P	07/11/2025	05:24	LB136434
	Magnesium	244	+/-1000	U	2000	P	07/11/2025	05:24	LB136434
	Manganese	5.94	+/-10	U	20.0	P	07/11/2025	05:24	LB136434
	Nickel	3.06	+/-20	U	40.0	P	07/11/2025	05:24	LB136434
	Potassium	918	+/-1000	U	2000	P	07/11/2025	05:24	LB136434
	Selenium	9.64	+/-10	U	20.0	P	07/11/2025	05:24	LB136434
	Silver	2.97	+/-5	J	10.0	P	07/11/2025	05:24	LB136434
	Sodium	868	+/-1000	U	2000	P	07/11/2025	05:24	LB136434

Metals

- 3a -

INITIAL AND CONTINUING CALIBRATION BLANK SUMMARY

Client: CDM Smith

SDG No.: Q2514

Contract: CAMP02

Lab Code: ACE

Sample ID	Analyte	Result ug/L	Acceptance Limit	Conc Qual	CRQL	M	Analysis Date	Analysis Time	Run Number
CCB09	Thallium	4.38	+/-20	U	40.0	P	07/11/2025	05:24	LB136434
	Vanadium	6.26	+/-20	U	40.0	P	07/11/2025	05:24	LB136434
	Zinc	3.50	+/-20	U	40.0	P	07/11/2025	05:24	LB136434
CCB10	Aluminum	11.3	+/-50	U	100	P	07/11/2025	06:21	LB136434
	Antimony	6.76	+/-25	U	50.0	P	07/11/2025	06:21	LB136434
	Arsenic	5.12	+/-10	U	20.0	P	07/11/2025	06:21	LB136434
	Barium	14.6	+/-50	U	100	P	07/11/2025	06:21	LB136434
	Beryllium	0.56	+/-3	U	6.00	P	07/11/2025	06:21	LB136434
	Cadmium	0.50	+/-3	U	6.00	P	07/11/2025	06:21	LB136434
	Calcium	234	+/-1000	U	2000	P	07/11/2025	06:21	LB136434
	Chromium	2.12	+/-5	U	10.0	P	07/11/2025	06:21	LB136434
	Cobalt	2.26	+/-15	U	30.0	P	07/11/2025	06:21	LB136434
	Copper	4.60	+/-10	U	20.0	P	07/11/2025	06:21	LB136434
	Iron	23.4	+/-50	U	100	P	07/11/2025	06:21	LB136434
	Lead	2.30	+/-6	U	12.0	P	07/11/2025	06:21	LB136434
	Magnesium	244	+/-1000	U	2000	P	07/11/2025	06:21	LB136434
	Manganese	5.94	+/-10	U	20.0	P	07/11/2025	06:21	LB136434
	Nickel	3.06	+/-20	U	40.0	P	07/11/2025	06:21	LB136434
	Potassium	918	+/-1000	U	2000	P	07/11/2025	06:21	LB136434
	Selenium	9.64	+/-10	U	20.0	P	07/11/2025	06:21	LB136434
	Silver	1.62	+/-5	U	10.0	P	07/11/2025	06:21	LB136434
	Sodium	868	+/-1000	U	2000	P	07/11/2025	06:21	LB136434
	Thallium	4.38	+/-20	U	40.0	P	07/11/2025	06:21	LB136434
	Vanadium	6.26	+/-20	U	40.0	P	07/11/2025	06:21	LB136434
	Zinc	3.50	+/-20	U	40.0	P	07/11/2025	06:21	LB136434
	Aluminum	11.3	+/-50	U	100	P	07/11/2025	06:47	LB136434
	Antimony	6.76	+/-25	U	50.0	P	07/11/2025	06:47	LB136434
	Arsenic	5.12	+/-10	U	20.0	P	07/11/2025	06:47	LB136434
	Barium	14.6	+/-50	U	100	P	07/11/2025	06:47	LB136434
	Beryllium	0.56	+/-3	U	6.00	P	07/11/2025	06:47	LB136434
	Cadmium	0.50	+/-3	U	6.00	P	07/11/2025	06:47	LB136434
	Calcium	234	+/-1000	U	2000	P	07/11/2025	06:47	LB136434
	Chromium	2.12	+/-5	U	10.0	P	07/11/2025	06:47	LB136434
	Cobalt	2.26	+/-15	U	30.0	P	07/11/2025	06:47	LB136434
	Copper	4.60	+/-10	U	20.0	P	07/11/2025	06:47	LB136434
	Iron	23.4	+/-50	U	100	P	07/11/2025	06:47	LB136434
	Lead	2.30	+/-6	U	12.0	P	07/11/2025	06:47	LB136434
	Magnesium	244	+/-1000	U	2000	P	07/11/2025	06:47	LB136434
	Manganese	5.94	+/-10	U	20.0	P	07/11/2025	06:47	LB136434

Metals

- 3a -

INITIAL AND CONTINUING CALIBRATION BLANK SUMMARY

Client: CDM Smith

SDG No.: Q2514

Contract: CAMP02

Lab Code: ACE

Sample ID	Analyte	Result ug/L	Acceptance Limit	Conc Qual	CRQL	M	Analysis Date	Analysis Time	Run Number
CCB11	Nickel	3.06	+/-20	U	40.0	P	07/11/2025	06:47	LB136434
	Potassium	918	+/-1000	U	2000	P	07/11/2025	06:47	LB136434
	Selenium	9.64	+/-10	U	20.0	P	07/11/2025	06:47	LB136434
	Silver	1.62	+/-5	U	10.0	P	07/11/2025	06:47	LB136434
	Sodium	868	+/-1000	U	2000	P	07/11/2025	06:47	LB136434
	Thallium	4.38	+/-20	U	40.0	P	07/11/2025	06:47	LB136434
	Vanadium	6.26	+/-20	U	40.0	P	07/11/2025	06:47	LB136434
	Zinc	3.50	+/-20	U	40.0	P	07/11/2025	06:47	LB136434

Metals

- 3a -

INITIAL AND CONTINUING CALIBRATION BLANK SUMMARY

Client: CDM Smith

SDG No.: Q2514

Contract: CAMP02

Lab Code: ACE

Sample ID	Analyte	Result ug/L	Acceptance Limit	Conc Qual	CRQL	M	Analysis Date	Analysis Time	Run Number
ICB01	Aluminum	44.1	+/-50	J	100	P	07/17/2025	14:29	LB136534
	Antimony	6.76	+/-25	U	50.0	P	07/17/2025	14:29	LB136534
	Arsenic	5.12	+/-10	U	20.0	P	07/17/2025	14:29	LB136534
	Barium	14.6	+/-50	U	100	P	07/17/2025	14:29	LB136534
	Beryllium	0.56	+/-3	U	6.00	P	07/17/2025	14:29	LB136534
	Cadmium	0.50	+/-3	U	6.00	P	07/17/2025	14:29	LB136534
	Calcium	234	+/-1000	U	2000	P	07/17/2025	14:29	LB136534
	Chromium	2.12	+/-5	U	10.0	P	07/17/2025	14:29	LB136534
	Cobalt	2.26	+/-15	U	30.0	P	07/17/2025	14:29	LB136534
	Copper	4.60	+/-10	U	20.0	P	07/17/2025	14:29	LB136534
	Iron	23.4	+/-50	U	100	P	07/17/2025	14:29	LB136534
	Lead	2.30	+/-6	U	12.0	P	07/17/2025	14:29	LB136534
	Magnesium	244	+/-1000	U	2000	P	07/17/2025	14:29	LB136534
	Manganese	5.94	+/-10	U	20.0	P	07/17/2025	14:29	LB136534
	Nickel	3.06	+/-20	U	40.0	P	07/17/2025	14:29	LB136534
	Potassium	918	+/-1000	U	2000	P	07/17/2025	14:29	LB136534
	Selenium	9.64	+/-10	U	20.0	P	07/17/2025	14:29	LB136534
	Silver	1.62	+/-5	U	10.0	P	07/17/2025	14:29	LB136534
	Sodium	868	+/-1000	U	2000	P	07/17/2025	14:29	LB136534
	Thallium	4.38	+/-20	U	40.0	P	07/17/2025	14:29	LB136534
	Vanadium	6.26	+/-20	U	40.0	P	07/17/2025	14:29	LB136534
	Zinc	3.50	+/-20	U	40.0	P	07/17/2025	14:29	LB136534

Metals

- 3a -

INITIAL AND CONTINUING CALIBRATION BLANK SUMMARY

Client: CDM Smith

SDG No.: Q2514

Contract: CAMP02

Lab Code: ACE

Sample ID	Analyte	Result ug/L	Acceptance Limit	Conc Qual	CRQL	M	Analysis Date	Analysis Time	Run Number
CCB01	Aluminum	34.3	+/-50	J	100	P	07/17/2025	15:01	LB136534
	Antimony	6.76	+/-25	U	50.0	P	07/17/2025	15:01	LB136534
	Arsenic	5.12	+/-10	U	20.0	P	07/17/2025	15:01	LB136534
	Barium	14.6	+/-50	U	100	P	07/17/2025	15:01	LB136534
	Beryllium	0.56	+/-3	U	6.00	P	07/17/2025	15:01	LB136534
	Cadmium	0.50	+/-3	U	6.00	P	07/17/2025	15:01	LB136534
	Calcium	234	+/-1000	U	2000	P	07/17/2025	15:01	LB136534
	Chromium	2.12	+/-5	U	10.0	P	07/17/2025	15:01	LB136534
	Cobalt	2.26	+/-15	U	30.0	P	07/17/2025	15:01	LB136534
	Copper	4.60	+/-10	U	20.0	P	07/17/2025	15:01	LB136534
	Iron	23.4	+/-50	U	100	P	07/17/2025	15:01	LB136534
	Lead	2.30	+/-6	U	12.0	P	07/17/2025	15:01	LB136534
	Magnesium	244	+/-1000	U	2000	P	07/17/2025	15:01	LB136534
	Manganese	5.94	+/-10	U	20.0	P	07/17/2025	15:01	LB136534
	Nickel	3.06	+/-20	U	40.0	P	07/17/2025	15:01	LB136534
	Potassium	918	+/-1000	U	2000	P	07/17/2025	15:01	LB136534
	Selenium	9.64	+/-10	U	20.0	P	07/17/2025	15:01	LB136534
	Silver	1.62	+/-5	U	10.0	P	07/17/2025	15:01	LB136534
	Sodium	868	+/-1000	U	2000	P	07/17/2025	15:01	LB136534
	Thallium	4.38	+/-20	U	40.0	P	07/17/2025	15:01	LB136534
	Vanadium	6.26	+/-20	U	40.0	P	07/17/2025	15:01	LB136534
	Zinc	3.50	+/-20	U	40.0	P	07/17/2025	15:01	LB136534
CCB02	Aluminum	20.6	+/-50	J	100	P	07/17/2025	15:49	LB136534
	Antimony	6.76	+/-25	U	50.0	P	07/17/2025	15:49	LB136534
	Arsenic	5.12	+/-10	U	20.0	P	07/17/2025	15:49	LB136534
	Barium	14.6	+/-50	U	100	P	07/17/2025	15:49	LB136534
	Beryllium	0.56	+/-3	U	6.00	P	07/17/2025	15:49	LB136534
	Cadmium	0.50	+/-3	U	6.00	P	07/17/2025	15:49	LB136534
	Calcium	234	+/-1000	U	2000	P	07/17/2025	15:49	LB136534
	Chromium	2.12	+/-5	U	10.0	P	07/17/2025	15:49	LB136534
	Cobalt	2.26	+/-15	U	30.0	P	07/17/2025	15:49	LB136534
	Copper	4.60	+/-10	U	20.0	P	07/17/2025	15:49	LB136534
	Iron	23.4	+/-50	U	100	P	07/17/2025	15:49	LB136534
	Lead	2.30	+/-6	U	12.0	P	07/17/2025	15:49	LB136534
	Magnesium	244	+/-1000	U	2000	P	07/17/2025	15:49	LB136534
	Manganese	5.94	+/-10	U	20.0	P	07/17/2025	15:49	LB136534
	Nickel	3.06	+/-20	U	40.0	P	07/17/2025	15:49	LB136534
	Potassium	918	+/-1000	U	2000	P	07/17/2025	15:49	LB136534
	Selenium	9.64	+/-10	U	20.0	P	07/17/2025	15:49	LB136534

Metals

- 3a -

INITIAL AND CONTINUING CALIBRATION BLANK SUMMARY

Client: CDM Smith

SDG No.: Q2514

Contract: CAMP02

Lab Code: ACE

Sample ID	Analyte	Result ug/L	Acceptance Limit	Conc Qual	CRQL	M	Analysis Date	Analysis Time	Run Number
CCB02	Silver	1.62	+/-5	U	10.0	P	07/17/2025	15:49	LB136534
	Sodium	868	+/-1000	U	2000	P	07/17/2025	15:49	LB136534
	Thallium	4.38	+/-20	U	40.0	P	07/17/2025	15:49	LB136534
	Vanadium	6.26	+/-20	U	40.0	P	07/17/2025	15:49	LB136534
	Zinc	3.50	+/-20	U	40.0	P	07/17/2025	15:49	LB136534
CCB03	Aluminum	30.1	+/-50	J	100	P	07/17/2025	17:36	LB136534
	Antimony	6.76	+/-25	U	50.0	P	07/17/2025	17:36	LB136534
	Arsenic	5.12	+/-10	U	20.0	P	07/17/2025	17:36	LB136534
	Barium	14.6	+/-50	U	100	P	07/17/2025	17:36	LB136534
	Beryllium	0.56	+/-3	U	6.00	P	07/17/2025	17:36	LB136534
	Cadmium	0.50	+/-3	U	6.00	P	07/17/2025	17:36	LB136534
	Calcium	234	+/-1000	U	2000	P	07/17/2025	17:36	LB136534
	Chromium	2.12	+/-5	U	10.0	P	07/17/2025	17:36	LB136534
	Cobalt	2.26	+/-15	U	30.0	P	07/17/2025	17:36	LB136534
	Copper	4.60	+/-10	U	20.0	P	07/17/2025	17:36	LB136534
	Iron	23.4	+/-50	U	100	P	07/17/2025	17:36	LB136534
	Lead	2.30	+/-6	U	12.0	P	07/17/2025	17:36	LB136534
	Magnesium	244	+/-1000	U	2000	P	07/17/2025	17:36	LB136534
	Manganese	5.94	+/-10	U	20.0	P	07/17/2025	17:36	LB136534
	Nickel	3.06	+/-20	U	40.0	P	07/17/2025	17:36	LB136534
	Potassium	918	+/-1000	U	2000	P	07/17/2025	17:36	LB136534
	Selenium	9.64	+/-10	U	20.0	P	07/17/2025	17:36	LB136534
	Silver	1.62	+/-5	U	10.0	P	07/17/2025	17:36	LB136534
	Sodium	868	+/-1000	U	2000	P	07/17/2025	17:36	LB136534
	Thallium	4.38	+/-20	U	40.0	P	07/17/2025	17:36	LB136534
	Vanadium	6.26	+/-20	U	40.0	P	07/17/2025	17:36	LB136534
	Zinc	3.50	+/-20	U	40.0	P	07/17/2025	17:36	LB136534
CCB04	Aluminum	27.5	+/-50	J	100	P	07/17/2025	18:50	LB136534
	Antimony	6.76	+/-25	U	50.0	P	07/17/2025	18:50	LB136534
	Arsenic	5.12	+/-10	U	20.0	P	07/17/2025	18:50	LB136534
	Barium	14.6	+/-50	U	100	P	07/17/2025	18:50	LB136534
	Beryllium	0.56	+/-3	U	6.00	P	07/17/2025	18:50	LB136534
	Cadmium	0.50	+/-3	U	6.00	P	07/17/2025	18:50	LB136534
	Calcium	234	+/-1000	U	2000	P	07/17/2025	18:50	LB136534
	Chromium	2.12	+/-5	U	10.0	P	07/17/2025	18:50	LB136534
	Cobalt	2.26	+/-15	U	30.0	P	07/17/2025	18:50	LB136534
	Copper	4.60	+/-10	U	20.0	P	07/17/2025	18:50	LB136534
	Iron	23.4	+/-50	U	100	P	07/17/2025	18:50	LB136534
	Lead	2.30	+/-6	U	12.0	P	07/17/2025	18:50	LB136534

Metals

- 3a -

INITIAL AND CONTINUING CALIBRATION BLANK SUMMARY

Client: CDM Smith

SDG No.: Q2514

Contract: CAMP02

Lab Code: ACE

Sample ID	Analyte	Result ug/L	Acceptance Limit	Conc Qual	CRQL	M	Analysis Date	Analysis Time	Run Number
CCB04	Magnesium	244	+/-1000	U	2000	P	07/17/2025	18:50	LB136534
	Manganese	5.94	+/-10	U	20.0	P	07/17/2025	18:50	LB136534
	Nickel	3.06	+/-20	U	40.0	P	07/17/2025	18:50	LB136534
	Potassium	918	+/-1000	U	2000	P	07/17/2025	18:50	LB136534
	Selenium	9.64	+/-10	U	20.0	P	07/17/2025	18:50	LB136534
	Silver	1.62	+/-5	U	10.0	P	07/17/2025	18:50	LB136534
	Sodium	868	+/-1000	U	2000	P	07/17/2025	18:50	LB136534
	Thallium	4.38	+/-20	U	40.0	P	07/17/2025	18:50	LB136534
	Vanadium	6.26	+/-20	U	40.0	P	07/17/2025	18:50	LB136534
	Zinc	3.50	+/-20	U	40.0	P	07/17/2025	18:50	LB136534
CCB05	Aluminum	19.5	+/-50	J	100	P	07/17/2025	20:16	LB136534
	Antimony	6.76	+/-25	U	50.0	P	07/17/2025	20:16	LB136534
	Arsenic	5.12	+/-10	U	20.0	P	07/17/2025	20:16	LB136534
	Barium	14.6	+/-50	U	100	P	07/17/2025	20:16	LB136534
	Beryllium	0.56	+/-3	U	6.00	P	07/17/2025	20:16	LB136534
	Cadmium	0.50	+/-3	U	6.00	P	07/17/2025	20:16	LB136534
	Calcium	234	+/-1000	U	2000	P	07/17/2025	20:16	LB136534
	Chromium	2.12	+/-5	U	10.0	P	07/17/2025	20:16	LB136534
	Cobalt	2.26	+/-15	U	30.0	P	07/17/2025	20:16	LB136534
	Copper	4.60	+/-10	U	20.0	P	07/17/2025	20:16	LB136534
	Iron	23.4	+/-50	U	100	P	07/17/2025	20:16	LB136534
	Lead	2.30	+/-6	U	12.0	P	07/17/2025	20:16	LB136534
	Magnesium	244	+/-1000	U	2000	P	07/17/2025	20:16	LB136534
	Manganese	5.94	+/-10	U	20.0	P	07/17/2025	20:16	LB136534
	Nickel	3.06	+/-20	U	40.0	P	07/17/2025	20:16	LB136534
	Potassium	918	+/-1000	U	2000	P	07/17/2025	20:16	LB136534
	Selenium	9.64	+/-10	U	20.0	P	07/17/2025	20:16	LB136534
	Silver	1.62	+/-5	U	10.0	P	07/17/2025	20:16	LB136534
	Sodium	868	+/-1000	U	2000	P	07/17/2025	20:16	LB136534
	Thallium	4.38	+/-20	U	40.0	P	07/17/2025	20:16	LB136534
	Vanadium	6.26	+/-20	U	40.0	P	07/17/2025	20:16	LB136534
	Zinc	3.50	+/-20	U	40.0	P	07/17/2025	20:16	LB136534
CCB06	Aluminum	12.5	+/-50	J	100	P	07/17/2025	20:51	LB136534
	Antimony	6.76	+/-25	U	50.0	P	07/17/2025	20:51	LB136534
	Arsenic	5.12	+/-10	U	20.0	P	07/17/2025	20:51	LB136534
	Barium	14.6	+/-50	U	100	P	07/17/2025	20:51	LB136534
	Beryllium	0.56	+/-3	U	6.00	P	07/17/2025	20:51	LB136534
	Cadmium	0.50	+/-3	U	6.00	P	07/17/2025	20:51	LB136534
	Calcium	234	+/-1000	U	2000	P	07/17/2025	20:51	LB136534

Metals

- 3a -

INITIAL AND CONTINUING CALIBRATION BLANK SUMMARY

Client: CDM Smith

SDG No.: Q2514

Contract: CAMP02

Lab Code: ACE

Sample ID	Analyte	Result ug/L	Acceptance Limit	Conc Qual	CRQL	M	Analysis Date	Analysis Time	Run Number
CCB06	Chromium	2.12	+/-5	U	10.0	P	07/17/2025	20:51	LB136534
	Cobalt	2.26	+/-15	U	30.0	P	07/17/2025	20:51	LB136534
	Copper	4.60	+/-10	U	20.0	P	07/17/2025	20:51	LB136534
	Iron	23.4	+/-50	U	100	P	07/17/2025	20:51	LB136534
	Lead	2.30	+/-6	U	12.0	P	07/17/2025	20:51	LB136534
	Magnesium	244	+/-1000	U	2000	P	07/17/2025	20:51	LB136534
	Manganese	5.94	+/-10	U	20.0	P	07/17/2025	20:51	LB136534
	Nickel	3.06	+/-20	U	40.0	P	07/17/2025	20:51	LB136534
	Potassium	918	+/-1000	U	2000	P	07/17/2025	20:51	LB136534
	Selenium	9.64	+/-10	U	20.0	P	07/17/2025	20:51	LB136534
	Silver	1.62	+/-5	U	10.0	P	07/17/2025	20:51	LB136534
	Sodium	868	+/-1000	U	2000	P	07/17/2025	20:51	LB136534
	Thallium	4.38	+/-20	U	40.0	P	07/17/2025	20:51	LB136534
	Vanadium	6.26	+/-20	U	40.0	P	07/17/2025	20:51	LB136534
	Zinc	3.50	+/-20	U	40.0	P	07/17/2025	20:51	LB136534