

DIESEL RANGE ORGANICS CONTINUING CALIBRATION SUMMARY

50 PPM TRPH STD

Lab Name: Alliane Contract: CAMP02
ProjectID: South River WM Replacement
Lab Code: ACE SDG No.: Q2514
DataFile: FG016287.D Analyst Name: YP\AJ Analyst Date: 07-10-2025

Conc. (PPM)	Area Count	RF	Average RF	%D
500	62962280	125925	115010	9.49