

Data Path : Z:\pestpcbsrv\HPCHEM1\FID_G\Data\FG070825\
 Data File : FG016253.D
 Signal(s) : FID1A.ch
 Acq On : 08 Jul 2025 18:54
 Operator : YP\AJ
 Sample : 50 PPM TRPH STD
 Misc :
 ALS Vial : 3 Sample Multiplier: 1

Instrument :
 FID_G
 ClientSampleId :
 50 PPM TRPH STD

Integration File: Sample.e
 Quant Time: Jul 09 03:15:29 2025
 Quant Method : Z:\pestpcbsrv\HPCHEM1\FID_G\Method\FG061325.M
 Quant Title :
 QLast Update : Fri Jun 13 12:34:28 2025
 Response via : Initial Calibration
 Integrator: ChemStation

Volume Inj. : 1uL
 Signal Phase : Rxi-1ms
 Signal Info : 20mx0.18mmx0.18um

Compound	R.T.	Response	Conc Units

System Monitoring Compounds			
9) S TETRACOSANE-d50 (SURR...	15.205	5505273	46.933 ug/ml
Target Compounds			
2) N-DECANE	4.674	6046993	48.198 ug/ml
3) N-DODECANE	6.857	6172923	48.376 ug/ml
4) N-TETRADECANE	8.697	6209107	48.113 ug/ml
5) N-HEXADECANE	10.317	6263234	47.986 ug/ml
6) N-OCTADECANE	11.772	6401431	47.800 ug/ml
7) N-EICOSANE	13.092	6434644	47.566 ug/ml
8) N-DOCOSANE	14.300	6244972	47.369 ug/ml
10) N-TETRACOSANE	15.411	6194202	47.022 ug/ml
11) N-HEXACOSANE	16.439	6084518	46.480 ug/ml
12) N-OCTACOSANE	17.395	5937111	45.958 ug/ml

(f)=RT Delta > 1/2 Window

(m)=manual int.

Data Path : Z:\pestpcbsrv\HPCHEM1\FID_G\Data\FG070825\
Data File : FG016253.D
Signal(s) : FID1A.ch
Acq On : 08 Jul 2025 18:54
Operator : YP\AJ
Sample : 50 PPM TRPH STD
Misc :
ALS Vial : 3 Sample Multiplier: 1

Instrument :
FID_G
ClientSampleId :
50 PPM TRPH STD

Integration File: Sample.e
Quant Time: Jul 09 03:15:29 2025
Quant Method : Z:\pestpcbsrv\HPCHEM1\FID_G\Method\FG061325.M
Quant Title :
QLast Update : Fri Jun 13 12:34:28 2025
Response via : Initial Calibration
Integrator: ChemStation

Volume Inj. : 1uL
Signal Phase : Rxi-1ms
Signal Info : 20mx0.18mmx0.18um

