

Instrument ID: FID_B

Daily Analysis Runlog For Sequence/QC Batch ID # FB062325

Review By	yogesh	Review On	6/23/2025 3:07:11 PM
Supervise By	mohammad	Supervise On	6/25/2025 2:41:46 AM
SubDirectory	FB062325	HP Acquire Method	HP Processing Method
			FB062325
STD. NAME	STD REF.#		
Tune/Reschk Initial Calibration Stds	PP24110,PP24656,PP24658,PP24659,PP24660,PP24661		
CCC Internal Standard/PEM ICV/I.BLK Surrogate Standard MS/MSD Standard LCS Standard	PP24111,PP24662		

Sr#	SampleID	ClientID	Data File Name	Date-Time	Comment	Operator	Status
1	5 GRO STD		FB031909.D	23 Jun 2025 12:17		YP/AJ	Ok
2	10 GRO STD		FB031910.D	23 Jun 2025 12:44		YP/AJ	Ok
3	20 GRO STD		FB031911.D	23 Jun 2025 13:12		YP/AJ	Ok
4	50 GRO STD		FB031912.D	23 Jun 2025 13:40		YP/AJ	Ok
5	100 GRO STD		FB031913.D	23 Jun 2025 14:07		YP/AJ	Ok
6	FB062325GROICV		FB031914.D	23 Jun 2025 14:48		YP/AJ	Ok

M : Manual Integration

Instrument ID: FID_B

Daily Analysis Runlog For Sequence/QC Batch ID # FB070925

Review By	yogesh	Review On	7/9/2025 3:32:13 PM
Supervise By	mohammad	Supervise On	7/11/2025 1:36:50 AM
SubDirectory	FB070925	HP Acquire Method	HP Processing Method
			FB062325

STD. NAME	STD REF.#
Tune/Reschk Initial Calibration Stds	PP24110,PP24656,PP24658,PP24659,PP24660,PP24661
CCC Internal Standard/PEM ICV/I.BLK Surrogate Standard MS/MSD Standard LCS Standard	PP24705,PP24706,PP24707 PP24111,PP24662

Sr#	SampleID	ClientID	Data File Name	Date-Time	Comment	Operator	Status
1	20 PPB GRO STD		FB032033.D	9 Jul 2025 8:12		YP/AJ	Ok
2	VBF0709S1		FB032034.D	9 Jul 2025 9:18		YP/AJ	Ok
3	VBF0709S2		FB032035.D	9 Jul 2025 9:45		YP/AJ	Ok
4	BSF0709S1		FB032036.D	9 Jul 2025 10:13		YP/AJ	Ok
5	Q2487-09		FB032037.D	9 Jul 2025 10:58	Vial -C	YP/AJ	Ok
6	Q2487-10		FB032038.D	9 Jul 2025 11:26	Vial -C	YP/AJ	Ok
7	Q2487-11		FB032039.D	9 Jul 2025 11:53	Vial -C	YP/AJ	Ok
8	Q2487-12		FB032040.D	9 Jul 2025 12:20	Vial-B	YP/AJ	Ok
9	Q2487-13		FB032041.D	9 Jul 2025 12:48	Vial-B, not perged	YP/AJ	Not Ok
10	Q2487-14		FB032042.D	9 Jul 2025 13:15	Vial-B, Not Purged	YP/AJ	Not Ok
11	BSF0709S2		FB032043.D	9 Jul 2025 13:42		YP/AJ	Ok
12	20 PPB GRO STD		FB032044.D	9 Jul 2025 15:30		YP/AJ	Ok,M
13	Q2514-02		FB032045.D	9 Jul 2025 16:15	Vial-B	YP/AJ	Ok
14	Q2514-01		FB032046.D	9 Jul 2025 16:42	Vial-B	YP/AJ	Ok
15	Q2514-08		FB032047.D	9 Jul 2025 17:10	Vial-A	YP/AJ	Ok
16	Q2514-07		FB032048.D	9 Jul 2025 17:37	Vial-A	YP/AJ	Ok
17	Q2514-03		FB032049.D	9 Jul 2025 18:05	Vial-B	YP/AJ	Ok
18	Q2514-04		FB032050.D	9 Jul 2025 18:32	Vial-B	YP/AJ	Ok

Instrument ID: FID_B

Daily Analysis Runlog For Sequence/QC Batch ID # FB070925

Review By	yogesh	Review On	7/9/2025 3:32:13 PM
Supervise By	mohammad	Supervise On	7/11/2025 1:36:50 AM
SubDirectory	FB070925	HP Acquire Method	HP Processing Method
			FB062325
STD. NAME	STD REF.#		
Tune/Reschk Initial Calibration Stds	PP24110,PP24656,PP24658,PP24659,PP24660,PP24661		
CCC	PP24705,PP24706,PP24707		
Internal Standard/PEM			
ICV/I.BLK	PP24111,PP24662		
Surrogate Standard			
MS/MSD Standard			
LCS Standard			

19	Q2514-05		FB032051.D	9 Jul 2025 19:00	Vial-B	YP/AJ	Ok
20	Q2514-06		FB032052.D	9 Jul 2025 19:27	Vial-A	YP/AJ	Ok,M
21	Q2514-10		FB032053.D	9 Jul 2025 19:54	Vial-A	YP/AJ	Ok
22	Q2514-09		FB032054.D	9 Jul 2025 20:21	Vial-A	YP/AJ	Ok
23	Q2487-13		FB032055.D	9 Jul 2025 20:49	Vial -C	YP/AJ	Ok
24	Q2487-14		FB032056.D	9 Jul 2025 21:16	Vial -C	YP/AJ	Ok
25	BSF0709S3		FB032057.D	9 Jul 2025 21:43		YP/AJ	Ok
26	20 PPB GRO STD		FB032058.D	9 Jul 2025 22:10		YP/AJ	Ok

M : Manual Integration