

COMPOUND DETECTION SUMMARY

CLIENT SAMPLE NO.

PB168753BS

Lab Name: Alliance

Contract: CAMP02

Lab Code: ACE

SDG NO.: Q2514

Lab Sample ID: PB168753BS

Date(s) Analyzed: 07/09/2025 07/09/2025

Instrument ID (1): ECD_S

Instrument ID (2): ECD_S

GC Column: (1): RTX-CLP

ID: 0.32 (mm)

GC Column:(2): RTX-CLP2

ID: 0.32 (mm)

ANALYTE	COL	RT	RT WINDOW		CONCENTRATION	%RPD
			FROM	TO		
DICHLORPROP	1	8.24	8.19	8.29	163	2.5
	2	8.69	8.64	8.74	159	
2,4-D	1	8.47	8.42	8.52	170	10.5
	2	9.03	8.98	9.08	153	
2,4,5-TP(Silvex)	1	9.36	9.31	9.41	176	5.2
	2	9.93	9.88	9.98	167	
2,4,5-T	1	9.65	9.60	9.70	185	9.6
	2	10.36	10.31	10.41	168	
2,4-DB	1	10.23	10.18	10.28	193	18.1
	2	10.93	10.88	10.98	161	
Dinoseb	1	11.45	11.40	11.50	168	9.3
	2	11.31	11.26	11.36	153	
DICAMBA	1	7.53	7.48	7.58	172	6
	2	7.97	7.92	8.02	162	

COMPOUND DETECTION SUMMARY

CLIENT SAMPLE NO.

WC-11MS

Lab Name: Alliance

Contract: CAMP02

Lab Code: ACE

SDG NO.: Q2514

Lab Sample ID: Q2493-01MS

Date(s) Analyzed: 07/09/2025 07/09/2025

Instrument ID (1): ECD_S

Instrument ID (2): ECD_S

GC Column: (1): RTX-CLP

ID: 0.32 (mm)

GC Column:(2): RTX-CLP2

ID: 0.32 (mm)

ANALYTE	COL	RT	RT WINDOW		CONCENTRATION	%RPD
			FROM	TO		
DICHLORPROP	1	8.24	8.19	8.29	144	1.4
	2	8.69	8.64	8.74	142	
2,4-D	1	8.46	8.41	8.51	545	125.9
	2	9.03	8.98	9.08	124	
2,4,5-TP(Silvex)	1	9.36	9.31	9.41	143	0.7
	2	9.93	9.88	9.98	142	
2,4,5-T	1	9.65	9.60	9.70	155	10.2
	2	10.36	10.31	10.41	140	
2,4-DB	1	10.23	10.18	10.28	134	13.5
	2	10.93	10.88	10.98	117	
DICAMBA	1	7.53	7.48	7.58	137	3
	2	7.97	7.92	8.02	133	

COMPOUND DETECTION SUMMARY

CLIENT SAMPLE NO.

WC-11MSD

Lab Name: Alliance

Contract: CAMP02

Lab Code: ACE

SDG NO.: Q2514

Lab Sample ID: Q2493-01MSD

Date(s) Analyzed: 07/09/2025 07/09/2025

Instrument ID (1): ECD_S

Instrument ID (2): ECD_S

GC Column: (1): RTX-CLP

ID: 0.32 (mm)

GC Column:(2): RTX-CLP2

ID: 0.32 (mm)

ANALYTE	COL	RT	RT WINDOW		CONCENTRATION	%RPD
			FROM	TO		
DICAMBA	1	7.53	7.48	7.58	138	3.7
	2	7.97	7.92	8.02	133	
DICHLORPROP	1	8.24	8.19	8.29	133	7.2
	2	8.69	8.64	8.74	143	
2,4-D	1	8.46	8.41	8.51	562	125.8
	2	9.03	8.98	9.08	128	
2,4,5-TP(Silvex)	1	9.36	9.31	9.41	144	0.7
	2	9.93	9.88	9.98	143	
2,4,5-T	1	9.65	9.60	9.70	159	11.3
	2	10.36	10.31	10.41	142	
2,4-DB	1	10.23	10.18	10.28	126	5.7
	2	10.93	10.88	10.98	119	