

CALIBRATION VERIFICATION SUMMARY

Lab Name: Alliance **Contract:** CAMP02
Lab Code: ACE **SDG NO.:** Q2514
GC Column: RTX-CLP **ID:** 0.32 (mm) **Initi. Calib. Date(s):** 07/11/2025 07/11/2025

Client Sample No.: CCAL06 **Date Analyzed:** 07/11/2025
Lab Sample No.: HSTDCCC750 **Data File :** PS031019.D **Time Analyzed:** 22:02

COMPOUND	RT	RT WINDOW FROM TO		CALC AMOUNT(ng)	NOM AMOUNT(ng)	%D
2,4,5-T	9.646	9.547	9.747	798.790	712.500	12.1
2,4,5-TP(Silvex)	9.351	9.253	9.453	750.150	712.500	5.3
2,4-D	8.466	8.367	8.567	747.650	705.000	6.0
2,4-DB	10.222	10.125	10.325	823.210	712.500	15.5
2,4-DCAA	7.332	7.233	7.433	740.320	750.000	-1.3
DICAMBA	7.521	7.422	7.622	693.960	705.000	-1.6
DICHLORPROP	8.233	8.135	8.335	699.300	705.000	-0.8
Dinoseb	11.438	11.339	11.539	758.250	705.000	7.6

CALIBRATION VERIFICATION SUMMARY

Lab Name: Alliance **Contract:** CAMP02
Lab Code: ACE **SDG NO.:** Q2514
GC Column: RTX-CLP2 **ID:** 0.32 (mm) **Initi. Calib. Date(s):** 07/11/2025 07/11/2025

Client Sample No.: CCAL06 **Date Analyzed:** 07/11/2025
Lab Sample No.: HSTDCCC750 **Data File :** PS031019.D **Time Analyzed:** 22:02

COMPOUND	RT	RT WINDOW FROM TO		CALC AMOUNT(ng)	NOM AMOUNT(ng)	%D
2,4,5-T	10.358	10.258	10.458	690.990	712.500	-3.0
2,4,5-TP(Silvex)	9.932	9.832	10.032	689.220	712.500	-3.3
2,4-D	9.026	8.927	9.127	669.050	705.000	-5.1
2,4-DB	10.926	10.826	11.026	677.520	712.500	-4.9
2,4-DCAA	7.769	7.669	7.869	711.250	750.000	-5.2
DICAMBA	7.971	7.871	8.071	684.630	705.000	-2.9
DICHLORPROP	8.690	8.591	8.791	663.820	705.000	-5.8
Dinoseb	11.310	11.210	11.410	673.080	705.000	-4.5