

Data Path : Z:\pestpcbsrv\HPCHEM1\ECD_S\Data\PS071025\
 Data File : PS031003.D
 Signal(s) : Signal #1: ECD1A.ch Signal #2: ECD2B.ch
 Acq On : 10 Jul 2025 23:00
 Operator : AR\AJ
 Sample : HSTDCCC750
 Misc :
 ALS Vial : 23 Sample Multiplier: 1

Instrument :
 ECD_S
 ClientSampleId :
 HSTDCCC750

Manual Integrations
 APPROVED

Reviewed By :Abdul Mirza 07/11/2025
 Supervised By :mohammad ahmed 07/11/2025

Integration File signal 1: autoint1.e
 Integration File signal 2: autoint2.e
 Quant Time: Jul 11 07:17:59 2025
 Quant Method : Z:\pestpcbsrv\HPCHEM1\ECD_S\Method\PS061825.M
 Quant Title : 8080.M
 QLast Update : Thu Jun 19 05:11:26 2025
 Response via : Initial Calibration
 Integrator: ChemStation

Volume Inj. : 1 µl
 Signal #1 Phase : ZB-MR1 Signal #2 Phase: ZB-MR2
 Signal #1 Info : 30M x 0.32mm x0.5 Signal #2 Info : 30M x 0.32mm x 0.25µm

	Compound	RT#1	RT#2	Resp#1	Resp#2	ng/ml	ng/ml

System Monitoring Compounds							
4) S	2,4-DCAA	7.333	7.768	2447.3E6	676.7E6	666.035m	631.256
Target Compounds							
1) T	Dalapon	2.701	2.710	3392.6E6	1545.3E6	568.688	563.637
2) T	3,5-DICHL...	6.495	6.714	3358.6E6	940.6E6	624.561	583.076
3) T	4-Nitroph...	7.133	7.300	998.9E6	1027.3E6	633.506	592.607
5) T	DICAMBA	7.523	7.970	9819.3E6	4033.2E6	645.367m	615.456
6) T	MCPD	7.704	8.069	632.3E6	135.6E6	67.348m	64.664
7) T	MCPA	7.854	8.318	747.8E6	196.0E6	63.750	62.305
8) T	DICHLORPROP	8.236	8.690	2274.1E6	941.8E6	607.100	598.926
9) T	2,4-D	8.468	9.026	2278.9E6	1014.3E6	661.628	596.877
10) T	Pentachlo...	8.774	9.551	35273.5E6	24779.3E6	666.476m	649.999
11) T	2,4,5-TP ...	9.354	9.932	13476.5E6	9176.0E6	672.801	635.956
12) T	2,4,5-T	9.649	10.358	12102.9E6	8772.4E6	735.600	641.980
13) T	2,4-DB	10.226	10.926	1844.6E6	728.3E6	804.275	615.697
14) T	DINOSEB	11.442	11.311	9151.5E6	6521.1E6	659.655m	604.855
15) T	Picloram	11.254	12.421	12562.9E6	15209.2E6	781.461m	654.401
16) T	DCPA	11.739	12.355	17935.8E6	14303.6E6	700.173	648.900

(f)=RT Delta > 1/2 Window (#)=Amounts differ by > 25% (m)=manual int.

Data Path : Z:\pestpcbsrv\HPCHEM1\ECD_S\Data\PS071025\
Data File : PS031003.D
Signal(s) : Signal #1: ECD1A.ch Signal #2: ECD2B.ch
Acq On : 10 Jul 2025 23:00
Operator : AR\AJ
Sample : HSTDCCC750
Misc :
ALS Vial : 23 Sample Multiplier: 1

Instrument :
ECD_S
ClientSampleId :
HSTDCCC750

Manual Integrations
APPROVED

Reviewed By :Abdul Mirza 07/11/2025
Supervised By :mohammad ahmed 07/11/2025

Integration File signal 1: autoint1.e
Integration File signal 2: autoint2.e
Quant Time: Jul 11 07:17:59 2025
Quant Method : Z:\pestpcbsrv\HPCHEM1\ECD_S\Method\PS061825.M
Quant Title : 8080.M
QLast Update : Thu Jun 19 05:11:26 2025
Response via : Initial Calibration
Integrator: ChemStation

Volume Inj. : 1 µl
Signal #1 Phase : ZB-MR1 Signal #2 Phase: ZB-MR2
Signal #1 Info : 30M x 0.32mm x0.5 Signal #2 Info : 30M x 0.32mm x 0.25µm

