

Matrix Spike/Matrix Spike Duplicate Summary

SW-846

SDG No.:	<u>Q2514</u>	Analytical Method:	<u>8151A</u>
Client:	<u>CDM Smith</u>	DataFile :	<u>PS030961.D</u>

		Sample				Rec		RPD		Limits		
	Parameter	Spike	Result	Result	Units	Rec	Qual	RPD	Qual	Low	High	RPD
Lab Sample ID:	Q2493-01MS	Client Sample ID:	WC-11MS									
	(Column 1)											
	DICAMBA		187.5	0	137	ug/Kg	73				10	112
	DICHLORPROP		187.5	0	144	ug/Kg	77				10	113
	2,4-D		187.5	0	545	ug/Kg	291	*			10	144
	2,4,5-TP(Silvex)		187.5	0	143	ug/Kg	76				10	114
	2,4,5-T		187.5	0	155	ug/Kg	83				10	115
	2,4-DB		187.5	0	134	ug/Kg	71				10	140
	Dinoseb		187.5	0	0	ug/Kg	0	*			10	118
Lab Sample ID:	Q2493-01MS	Client Sample ID:	WC-11MS									
	(Column 2)											
	DICAMBA		187.5	0	133	ug/Kg	71				10	112
	DICHLORPROP		187.5	0	142	ug/Kg	76				10	113
	2,4-D		187.5	0	124	ug/Kg	66	*			10	144
	2,4,5-TP(Silvex)		187.5	0	142	ug/Kg	76				10	114
	2,4,5-T		187.5	0	140	ug/Kg	75				10	115
	2,4-DB		187.5	0	117	ug/Kg	62				10	140
	Dinoseb		187.5	0	0	ug/Kg	0	*			10	118