

CALIBRATION VERIFICATION SUMMARY

Lab Name: Alliance **Contract:** CAMP02
Lab Code: ACE **SDG NO.:** Q2514
GC Column: ZB-MR1 **ID:** 0.32 (mm) **Initi. Calib. Date(s):** 06/11/2025 06/11/2025

Client Sample No.: CCAL02 **Date Analyzed:** 07/08/2025
Lab Sample No.: AR1660CCC500 **Data File :** PO112067.D **Time Analyzed:** 03:54

COMPOUND	RT	RT WINDOW		CALC AMOUNT(ng)	NOM AMOUNT(ng)	%D
		FROM	TO			
Aroclor-1016-1	4.761	4.666	4.866	528.310	500.000	5.7
Aroclor-1016-2	4.780	4.684	4.884	545.080	500.000	9.0
Aroclor-1016-3	4.837	4.741	4.941	530.690	500.000	6.1
Aroclor-1016-4	4.956	4.861	5.061	532.770	500.000	6.6
Aroclor-1016-5	5.213	5.118	5.318	522.850	500.000	4.6
Aroclor-1260-1	6.252	6.157	6.357	525.380	500.000	5.1
Aroclor-1260-2	6.441	6.346	6.546	598.920	500.000	19.8
Aroclor-1260-3	6.808	6.713	6.913	541.620	500.000	8.3
Aroclor-1260-4	7.067	6.972	7.172	543.400	500.000	8.7
Aroclor-1260-5	7.309	7.216	7.416	584.810	500.000	17.0
Decachlorobiphenyl	8.703	8.612	8.812	43.520	50.000	-13.0
Tetrachloro-m-xylene	3.675	3.577	3.777	60.770	50.000	21.5

CALIBRATION VERIFICATION SUMMARY

Lab Name: Alliance **Contract:** CAMP02
Lab Code: ACE **SDG NO.:** Q2514
GC Column: ZB-MR2 **ID:** 0.32 (mm) **Initi. Calib. Date(s):** 06/11/2025 06/11/2025

Client Sample No.: CCAL02 **Date Analyzed:** 07/08/2025
Lab Sample No.: AR1660CCC500 **Data File :** PO112067.D **Time Analyzed:** 03:54

COMPOUND	RT	RT WINDOW		CALC AMOUNT(ng)	NOM AMOUNT(ng)	%D
		FROM	TO			
Aroclor-1016-1	4.744	4.651	4.851	552.450	500.000	10.5
Aroclor-1016-2	4.762	4.669	4.869	563.100	500.000	12.6
Aroclor-1016-3	4.937	4.844	5.044	541.710	500.000	8.3
Aroclor-1016-4	4.980	4.887	5.087	478.710	500.000	-4.3
Aroclor-1016-5	5.192	5.099	5.299	596.930	500.000	19.4
Aroclor-1260-1	6.221	6.129	6.329	591.390	500.000	18.3
Aroclor-1260-2	6.409	6.316	6.516	591.380	500.000	18.3
Aroclor-1260-3	6.560	6.469	6.669	531.150	500.000	6.2
Aroclor-1260-4	7.031	6.939	7.139	498.690	500.000	-0.3
Aroclor-1260-5	7.272	7.181	7.381	507.110	500.000	1.4
Decachlorobiphenyl	8.650	8.561	8.761	48.100	50.000	-3.8
Tetrachloro-m-xylene	3.668	3.573	3.773	54.860	50.000	9.7