

CALIBRATION VERIFICATION SUMMARY

Lab Name: Alliance **Contract:** CAMP02
Lab Code: ACE **SDG NO.:** Q2514
GC Column: ZB-MR1 **ID:** 0.32 (mm) **Initi. Calib. Date(s):** 06/11/2025 06/11/2025

Client Sample No.: CCAL03 **Date Analyzed:** 07/08/2025
Lab Sample No.: AR1660CCC500 **Data File :** PO112083.D **Time Analyzed:** 10:09

COMPOUND	RT	RT WINDOW		CALC AMOUNT(ng)	NOM AMOUNT(ng)	%D
		FROM	TO			
Aroclor-1016-1	4.762	4.666	4.866	597.960	500.000	19.6
Aroclor-1016-2	4.780	4.684	4.884	599.020	500.000	19.8
Aroclor-1016-3	4.837	4.741	4.941	561.710	500.000	12.3
Aroclor-1016-4	4.957	4.861	5.061	574.730	500.000	14.9
Aroclor-1016-5	5.215	5.118	5.318	566.580	500.000	13.3
Aroclor-1260-1	6.252	6.157	6.357	563.120	500.000	12.6
Aroclor-1260-2	6.442	6.346	6.546	599.800	500.000	20.0
Aroclor-1260-3	6.808	6.713	6.913	597.780	500.000	19.6
Aroclor-1260-4	7.067	6.972	7.172	548.080	500.000	9.6
Aroclor-1260-5	7.311	7.216	7.416	627.280	500.000	25.5
Decachlorobiphenyl	8.705	8.612	8.812	52.570	50.000	5.1
Tetrachloro-m-xylene	3.676	3.577	3.777	71.620	50.000	43.2

CALIBRATION VERIFICATION SUMMARY

Lab Name: Alliance **Contract:** CAMP02
Lab Code: ACE **SDG NO.:** Q2514
GC Column: ZB-MR2 **ID:** 0.32 (mm) **Initi. Calib. Date(s):** 06/11/2025 06/11/2025

Client Sample No.: CCAL03 **Date Analyzed:** 07/08/2025
Lab Sample No.: AR1660CCC500 **Data File :** PO112083.D **Time Analyzed:** 10:09

COMPOUND	RT	RT WINDOW		CALC AMOUNT(ng)	NOM AMOUNT(ng)	%D
		FROM	TO			
Aroclor-1016-1	4.743	4.651	4.851	466.410	500.000	-6.7
Aroclor-1016-2	4.761	4.669	4.869	474.240	500.000	-5.2
Aroclor-1016-3	4.937	4.844	5.044	455.480	500.000	-8.9
Aroclor-1016-4	4.979	4.887	5.087	452.330	500.000	-9.5
Aroclor-1016-5	5.191	5.099	5.299	441.610	500.000	-11.7
Aroclor-1260-1	6.221	6.129	6.329	441.990	500.000	-11.6
Aroclor-1260-2	6.409	6.316	6.516	464.790	500.000	-7.0
Aroclor-1260-3	6.560	6.469	6.669	435.190	500.000	-13.0
Aroclor-1260-4	7.031	6.939	7.139	432.900	500.000	-13.4
Aroclor-1260-5	7.272	7.181	7.381	431.230	500.000	-13.8
Decachlorobiphenyl	8.650	8.561	8.761	40.190	50.000	-19.6
Tetrachloro-m-xylene	3.668	3.573	3.773	46.280	50.000	-7.4