

Instrument ID: ECD_O

Daily Analysis Runlog For Sequence/QC Batch ID # PO061125

Review By	yogesh	Review On	6/11/2025 3:41:50 PM
Supervise By	mohammad	Supervise On	6/13/2025 1:40:28 AM
SubDirectory	PO061125	HP Acquire Method	HP Processing Method PO061125
STD. NAME	STD REF.#		
Tune/Reschk			
Initial Calibration Stds	PP24330,PP24331,PP24332,PP24333,PP24334,PP24335,PP24336,PP24337,PP24338,PP24339,PP24340,PP24341,PP24342,PP24343,PP24344,PP24345,PP24346,PP24347,PP24348,PP24349,PP24350,PP24351,PP24352,PP24353,PP24354,PP24355,PP24356,PP24357,PP24358,PP24359,PP24360,PP24361,PP24362,PP24363,PP24364,PP24365,PP24366,PP24367,PP24368,PP24369		
CCC	PP24332,PP24347,PP24352,PP24357		
Internal Standard/PEM			
ICV/I.BLK	PP24370,PP24371,PP24372,PP24373,PP24374,PP24375,PP24376,PP24377,PP24378,PP24379,PP24380,PP24381,PP24382,PP24384,PP24385,PP24386,PP2		
Surrogate Standard			
MS/MSD Standard			
LCS Standard			

Sr#	Sampleld	Data File Name	Date-Time	Operator	Status
1	HEXANE	PO111585.D	11 Jun 2025 10:03	YP/AJ	Ok
2	I.BLK	PO111586.D	11 Jun 2025 10:21	YP/AJ	Ok
3	AR1660ICC1000	PO111587.D	11 Jun 2025 10:40	YP/AJ	Ok
4	AR1660ICC750	PO111588.D	11 Jun 2025 10:58	YP/AJ	Ok
5	AR1660ICC500	PO111589.D	11 Jun 2025 11:17	YP/AJ	Ok
6	AR1660ICC250	PO111590.D	11 Jun 2025 11:35	YP/AJ	Ok
7	AR1660ICC050	PO111591.D	11 Jun 2025 11:53	YP/AJ	Ok,M
8	AR1221ICC500	PO111592.D	11 Jun 2025 12:12	YP/AJ	Ok
9	AR1232ICC500	PO111593.D	11 Jun 2025 12:30	YP/AJ	Ok
10	AR1242ICC1000	PO111594.D	11 Jun 2025 12:48	YP/AJ	Ok
11	AR1242ICC750	PO111595.D	11 Jun 2025 13:07	YP/AJ	Ok
12	AR1242ICC500	PO111596.D	11 Jun 2025 13:25	YP/AJ	Ok
13	AR1242ICC250	PO111597.D	11 Jun 2025 13:44	YP/AJ	Ok
14	AR1242ICC050	PO111598.D	11 Jun 2025 14:02	YP/AJ	Ok,M
15	AR1248ICC1000	PO111599.D	11 Jun 2025 14:20	YP/AJ	Ok
16	AR1248ICC750	PO111600.D	11 Jun 2025 14:39	YP/AJ	Ok
17	AR1248ICC500	PO111601.D	11 Jun 2025 15:14	YP/AJ	Ok
18	AR1248ICC250	PO111602.D	11 Jun 2025 15:32	YP/AJ	Ok
19	AR1248ICC050	PO111603.D	11 Jun 2025 15:49	YP/AJ	Ok,M
20	AR1254ICC1000	PO111604.D	11 Jun 2025 16:06	YP/AJ	Ok
21	AR1254ICC750	PO111605.D	11 Jun 2025 16:25	YP/AJ	Ok,M

Instrument ID: ECD_O

Daily Analysis Runlog For Sequence/QC Batch ID # PO061125

Review By	yogesh	Review On	6/11/2025 3:41:50 PM
Supervise By	mohammad	Supervise On	6/13/2025 1:40:28 AM
SubDirectory	PO061125	HP Acquire Method	HP Processing Method PO061125
STD. NAME	STD REF.#		
Tune/Reschk			
Initial Calibration Stds	PP24330,PP24331,PP24332,PP24333,PP24334,PP24335,PP24336,PP24337,PP24338,PP24339,PP24340,PP24341,PP24342,PP24343,PP24344,PP24345,PP24346,PP24347,PP24348,PP24349,PP24350,PP24351,PP24352,PP24353,PP24354,PP24355,PP24356,PP24357,PP24358,PP24359,PP24360,PP24361,PP24362,PP24363,PP24364,PP24365,PP24366,PP24367,PP24368,PP24369		
CCC	PP24332,PP24347,PP24352,PP24357		
Internal Standard/PEM			
ICV/I.BLK	PP24370,PP24371,PP24372,PP24373,PP24374,PP24375,PP24376,PP24377,PP24378,PP24379,PP24380,PP24381,PP24382,PP24384,PP24385,PP24386,PP2		
Surrogate Standard			
MS/MSD Standard			
LCS Standard			

22	AR1254ICC500	PO111606.D	11 Jun 2025 16:43	YP/AJ	Ok
23	AR1254ICC250	PO111607.D	11 Jun 2025 17:00	YP/AJ	Ok
24	AR1254ICC050	PO111608.D	11 Jun 2025 17:18	YP/AJ	Ok
25	AR1262ICC500	PO111609.D	11 Jun 2025 17:36	YP/AJ	Ok
26	AR1268ICC1000	PO111610.D	11 Jun 2025 17:55	YP/AJ	Ok,M
27	AR1268ICC750	PO111611.D	11 Jun 2025 18:13	YP/AJ	Ok
28	AR1268ICC500	PO111612.D	11 Jun 2025 18:31	YP/AJ	Ok
29	AR1268ICC250	PO111613.D	11 Jun 2025 18:50	YP/AJ	Ok,M
30	AR1268ICC050	PO111614.D	11 Jun 2025 19:07	YP/AJ	Ok
31	PO061125ICV500	PO111615.D	11 Jun 2025 19:25	YP/AJ	Ok
32	AR1242ICV500	PO111616.D	11 Jun 2025 20:02	YP/AJ	Ok
33	AR1248ICV500	PO111617.D	11 Jun 2025 20:39	YP/AJ	Ok
34	AR1254ICV500	PO111618.D	11 Jun 2025 21:16	YP/AJ	Ok
35	AR1268ICV500	PO111619.D	11 Jun 2025 21:52	YP/AJ	Ok,M

M : Manual Integration

Instrument ID: ECD_O

Daily Analysis Runlog For Sequence/QC Batch ID # PO070725

Review By	yogesh	Review On	7/7/2025 11:17:57 AM
Supervise By		Supervise On	
SubDirectory	PO070725	HP Acquire Method	HP Processing Method PO061125
STD. NAME	STD REF.#		
Tune/Reschk			
Initial Calibration Stds	PP24330,PP24331,PP24332,PP24333,PP24334,PP24335,PP24336,PP24337,PP24338,PP24339,PP24340,PP24341,PP24342,PP24343,PP24344,PP24345,PP24346,PP24347,PP24348,PP24349,PP24350,PP24351,PP24352,PP24353,PP24354,PP24355,PP24356,PP24357,PP24358,PP24359,PP24360,PP24361,PP24362,PP24363,PP24364,PP24365,PP24366,PP24367,PP24368,PP24369		
CCC	PP24332,PP24347,PP24352,PP24357		
Internal Standard/PEM			
ICV/I.BLK	PP24370,PP24371,PP24372,PP24373,PP24374,PP24375,PP24376,PP24377,PP24378,PP24379,PP24380,PP24381,PP24382,PP24384,PP24385,PP24386,PP2		
Surrogate Standard			
MS/MSD Standard			
LCS Standard			

Sr#	SampleId	Data File Name	Date-Time	Operator	Status
1	HEXANE	PO112021.D	07 Jul 2025 08:40	YP/AJ	Ok
2	AR1660CCC500	PO112022.D	07 Jul 2025 08:58	YP/AJ	Ok,NS
3	AR1242CCC500	PO112023.D	07 Jul 2025 09:17	YP/AJ	Ok
4	AR1248CCC500	PO112024.D	07 Jul 2025 09:34	YP/AJ	Ok,NS
5	AR1254CCC500	PO112025.D	07 Jul 2025 09:53	YP/AJ	Ok,NS
6	I.BLK	PO112026.D	07 Jul 2025 10:10	YP/AJ	Ok
7	DDT ANALOGUE	PO112027.D	07 Jul 2025 10:27	YP/AJ	Ok
8	PB168704BL	PO112028.D	07 Jul 2025 10:46	YP/AJ	Ok
9	PB168710BL	PO112029.D	07 Jul 2025 11:04	YP/AJ	Ok,NS
10	PB168710BS	PO112030.D	07 Jul 2025 11:22	YP/AJ	Ok
11	Q2466-01	PO112031.D	07 Jul 2025 11:40	YP/AJ	Ok
12	Q2466-02	PO112032.D	07 Jul 2025 11:59	YP/AJ	Ok,NS
13	Q2492-01	PO112033.D	07 Jul 2025 12:17	YP/AJ	Ok,NS
14	Q2492-02	PO112034.D	07 Jul 2025 12:34	YP/AJ	Ok,NS
15	Q2476-01	PO112035.D	07 Jul 2025 12:53	YP/AJ	Not Ok
16	Q2481-10	PO112036.D	07 Jul 2025 13:10	YP/AJ	Ok,NS
17	AR1660CCC500	PO112037.D	07 Jul 2025 14:47	YP/AJ	Ok,NS
18	AR1242CCC500	PO112038.D	07 Jul 2025 15:22	YP/AJ	Ok,NS
19	AR1248CCC500	PO112039.D	07 Jul 2025 15:40	YP/AJ	Ok,NS
20	AR1254CCC500	PO112040.D	07 Jul 2025 15:58	YP/AJ	Ok
21	I.BLK	PO112041.D	07 Jul 2025 16:16	YP/AJ	Ok,NS

Instrument ID: ECD_O

Daily Analysis Runlog For Sequence/QC Batch ID # PO070725

Review By	yogesh	Review On	7/7/2025 11:17:57 AM
Supervise By		Supervise On	
SubDirectory	PO070725	HP Acquire Method	HP Processing Method PO061125
STD. NAME	STD REF.#		
Tune/Reschk			
Initial Calibration Stds	PP24330,PP24331,PP24332,PP24333,PP24334,PP24335,PP24336,PP24337,PP24338,PP24339,PP24340,PP24341,PP24342,PP24343,PP24344,PP24345,PP24346,PP24347,PP24348,PP24349,PP24350,PP24351,PP24352,PP24353,PP24354,PP24355,PP24356,PP24357,PP24358,PP24359,PP24360,PP24361,PP24362,PP24363,PP24364,PP24365,PP24366,PP24367,PP24368,PP24369		
CCC	PP24332,PP24347,PP24352,PP24357		
Internal Standard/PEM			
ICV/I.BLK	PP24370,PP24371,PP24372,PP24373,PP24374,PP24375,PP24376,PP24377,PP24378,PP24379,PP24380,PP24381,PP24382,PP24384,PP24385,PP24386,PP2		
Surrogate Standard			
MS/MSD Standard			
LCS Standard			

22	PB168734BL	PO112042.D	07 Jul 2025 16:34	YP/AJ	Ok
23	PB168734BS	PO112043.D	07 Jul 2025 16:51	YP/AJ	Ok
24	Q2505-01	PO112044.D	07 Jul 2025 17:09	YP/AJ	Not Ok
25	Q2505-02	PO112045.D	07 Jul 2025 17:26	YP/AJ	Not Ok
26	Q2505-03	PO112046.D	07 Jul 2025 17:44	YP/AJ	Ok
27	Q2508-01	PO112047.D	07 Jul 2025 18:03	YP/AJ	Ok,NS
28	Q2508-02	PO112048.D	07 Jul 2025 18:20	YP/AJ	Not Ok
29	Q2508-03	PO112049.D	07 Jul 2025 18:39	YP/AJ	Ok
30	Q2516-01	PO112050.D	07 Jul 2025 18:56	YP/AJ	Ok,NS
31	Q2509-02	PO112051.D	07 Jul 2025 19:14	YP/AJ	Not Ok
32	AR1660CCC500	PO112052.D	07 Jul 2025 20:15	YP/AJ	Ok,NS
33	AR1242CCC500	PO112053.D	07 Jul 2025 20:51	YP/AJ	Ok,NS
34	AR1248CCC500	PO112054.D	07 Jul 2025 21:08	YP/AJ	Ok,NS
35	AR1254CCC500	PO112055.D	07 Jul 2025 21:25	YP/AJ	Ok,NS
36	I.BLK	PO112056.D	07 Jul 2025 21:44	YP/AJ	Ok
37	PB168735BL	PO112057.D	07 Jul 2025 22:01	YP/AJ	Ok,NS
38	PB168735BS	PO112058.D	07 Jul 2025 22:20	YP/AJ	Ok,NS
39	Q2507-01	PO112059.D	07 Jul 2025 22:38	YP/AJ	Ok,NS
40	Q2510-01	PO112060.D	07 Jul 2025 22:57	YP/AJ	Ok,NS
41	Q2513-01	PO112061.D	07 Jul 2025 23:15	YP/AJ	Ok,NS
42	Q2513-03	PO112062.D	07 Jul 2025 23:33	YP/AJ	Ok,NS
43	Q2513-03MS	PO112063.D	07 Jul 2025 23:52	YP/AJ	Ok
44	Q2513-03MSD	PO112064.D	08 Jul 2025 00:10	YP/AJ	Ok

Instrument ID: ECD_O

Daily Analysis Runlog For Sequence/QC Batch ID # PO070725

Review By	yogesh	Review On	7/7/2025 11:17:57 AM
Supervise By		Supervise On	
SubDirectory	PO070725	HP Acquire Method	HP Processing Method PO061125
STD. NAME	STD REF.#		
Tune/Reschk			
Initial Calibration Stds	PP24330,PP24331,PP24332,PP24333,PP24334,PP24335,PP24336,PP24337,PP24338,PP24339,PP24340,PP24341,PP24342,PP24343,PP24344,PP24345,PP24346,PP24347,PP24348,PP24349,PP24350,PP24351,PP24352,PP24353,PP24354,PP24355,PP24356,PP24357,PP24358,PP24359,PP24360,PP24361,PP24362,PP24363,PP24364,PP24365,PP24366,PP24367,PP24368,PP24369		
CCC	PP24332,PP24347,PP24352,PP24357		
Internal Standard/PEM			
ICV/I.BLK	PP24370,PP24371,PP24372,PP24373,PP24374,PP24375,PP24376,PP24377,PP24378,PP24379,PP24380,PP24381,PP24382,PP24384,PP24385,PP24386,PP2		
Surrogate Standard			
MS/MSD Standard			
LCS Standard			

45	Q2514-01	PO112065.D	08 Jul 2025 00:29	YP/AJ	Ok,NS
46	Q2514-02	PO112066.D	08 Jul 2025 00:46	YP/AJ	Ok,NS
47	AR1660CCC500	PO112067.D	08 Jul 2025 03:54	YP/AJ	Ok,NS
48	AR1242CCC500	PO112068.D	08 Jul 2025 04:12	YP/AJ	Ok,NS
49	AR1248CCC500	PO112069.D	08 Jul 2025 04:30	YP/AJ	Ok,NS
50	AR1254CCC500	PO112070.D	08 Jul 2025 04:49	YP/AJ	Ok,NS
51	AR1268CCC500	PO112071.D	08 Jul 2025 05:07	YP/AJ	Not Ok
52	I.BLK	PO112072.D	08 Jul 2025 05:26	YP/AJ	Ok,NS
53	Q2514-03	PO112073.D	08 Jul 2025 05:44	YP/AJ	Ok,NS
54	Q2514-04	PO112074.D	08 Jul 2025 06:02	YP/AJ	Ok,NS
55	Q2514-05	PO112075.D	08 Jul 2025 06:21	YP/AJ	Ok,NS
56	Q2514-06	PO112076.D	08 Jul 2025 06:39	YP/AJ	Ok,NS
57	Q2514-07	PO112077.D	08 Jul 2025 06:58	YP/AJ	Ok,NS
58	Q2514-08	PO112078.D	08 Jul 2025 07:16	YP/AJ	Ok,NS
59	Q2514-09	PO112079.D	08 Jul 2025 07:35	YP/AJ	Ok,NS
60	Q2514-10	PO112080.D	08 Jul 2025 07:53	YP/AJ	Ok,NS
61	Q2515-01	PO112081.D	08 Jul 2025 08:11	YP/AJ	Ok,NS
62	Q2487-15	PO112082.D	08 Jul 2025 08:30	YP/AJ	Not Ok
63	AR1660CCC500	PO112083.D	08 Jul 2025 10:09	YP/AJ	Ok,NS
64	AR1242CCC500	PO112084.D	08 Jul 2025 10:46	YP/AJ	Not Ok
65	I.BLK	PO112085.D	08 Jul 2025 11:26	YP/AJ	Ok,NS

M : Manual Integration