

Cover Page

Order ID : Q2514

Project ID : South River WM Replacement

Client : CDM Smith

Lab Sample Number

Q2514-01
Q2514-02
Q2514-03
Q2514-04
Q2514-05
Q2514-06
Q2514-07
Q2514-08
Q2514-09
Q2514-10

Client Sample Number

TP-92
TP-93
TP-94
TP-96
TP-97
TP-103
TP-36
TP-78
TP-81
TP-90

I certify that the data package is in compliance with the terms and conditions of the contract, both technically and for completeness, for other than the conditions detailed above. Release of the data contained in this hard copy data package has been authorized by the laboratory manager or his designee, as verified by the following signature.

Signature : _____

Date: 7/11/2025

DATA REPORTING QUALIFIERS- ORGANIC

For reporting results, the following “Results Qualifiers” are used:

Value	If the result is a value greater than or equal to the detection limit, report the value
U	Indicates the compound was analyzed for but was not detected. Report the minimum detection limit for the sample with the U, i.e. “10 U”. This is not necessarily the instrument detection limit attainable for this particular sample based on any concentration or dilution that may have been required.
ND	Indicates the analyte was analyzed for, but not detected
J	Indicates an estimated value. This flag is used: (1) When estimating a concentration for a tentatively identified compound (library search hits, where a 1:1 response is assumed.) (2) When the mass spectral data indicated the identification, however the result was less than the specified detection limit greater than zero. If the detection limit was 10ug/L and a concentration of 3 ug/L was calculated report as 3 J. This flag is used when similar situation arise on any organic parameter i.e. Pest, PCB and others.
B	Indicates the analyte was found in the blank as well as the sample report as “12 B”.
E	Indicates the analyte ‘s concentration exceeds the calibrated range of the instrument for that specific analysis.
D	This flag identifies all compounds identified in an analysis at a secondary dilution factor.
P	This flag is used for Pesticide/PCB target analyte when there is >25% difference for detected concentrations between the two GC columns. The lower of the two values is reported on Form 1 and flagged with a “P”.
N	This flag indicates presumptive evidence of a compound. This is only used for tentatively identified compounds (TICs), where the identification is based on a mass spectral library search. It applies to all TIC results. For generic characterization of a TIC, such as chlorinated hydrocarbon, the flag is not used.
A	This flag indicates that a Tentatively Identified Compound is a suspected aldol-condensation product.
Q	Indicates the LCS did not meet the control limits requirements

APPENDIX A

QA REVIEW GENERAL DOCUMENTATION

Project #: Q2514

Completed

For thorough review, the report must have the following:

GENERAL:

Are all original paperwork present (chain of custody, record of communication,airbill, sample management lab chronicle, login page)

✓

Check chain-of-custody for proper relinquish/return of samples

✓

Is the chain of custody signed and complete

✓

Check internal chain-of-custody for proper relinquish/return of samples /sample extracts

✓

Collect information for each project id from server. Were all requirements followed

✓

COVER PAGE:

Do numbers of samples correspond to the number of samples in the Chain of Custody on login page

✓

Do lab numbers and client Ids on cover page agree with the Chain of Custody

✓

CHAIN OF CUSTODY:

Do requested analyses on Chain of Custody agree with form I results

✓

Do requested analyses on Chain of Custody agree with the log-in page

✓

Were the correct method log-in for analysis according to the Analytical Request and Chain of Castody

✓

Were the samples received within hold time

✓

Were any problems found with the samples at arrival recorded in the Sample Management Laboratory Chronicle

✓

ANALYTICAL:

Was method requirement followed?

✓

Was client requirement followed?

✓

Does the case narrative summarize all QC failure?

✓

All runlogs and manual integration are reviewed for requirements

✓

All manual calculations and /or hand notations verified

✓

QA Review Signature: PRATHA PARGHI

Date: 07/11/2025

LAB CHRONICLE

OrderID:	Q2514	OrderDate:	7/3/2025 1:29:00 PM
Client:	CDM Smith	Project:	South River WM Replacement
Contact:	Marcie Ann Encinas	Location:	O21,O22,VOA Lab

LabID	ClientID	Matrix	Test	Method	Sample Date	Prep Date	Anal Date	Received
Q2514-01	TP-92	SOIL			07/02/25			07/03/25
			PCB	8082A		07/07/25	07/08/25	
			Pesticide-TCL	8081B		07/07/25	07/07/25	
Q2514-02	TP-93	SOIL			07/02/25			07/03/25
			PCB	8082A		07/07/25	07/08/25	
			Pesticide-TCL	8081B		07/07/25	07/07/25	
Q2514-03	TP-94	SOIL			07/02/25			07/03/25
			PCB	8082A		07/07/25	07/08/25	
			Pesticide-TCL	8081B		07/07/25	07/07/25	
Q2514-04	TP-96	SOIL			07/02/25			07/03/25
			PCB	8082A		07/07/25	07/08/25	
			Pesticide-TCL	8081B		07/07/25	07/07/25	
Q2514-05	TP-97	SOIL			07/02/25			07/03/25
			PCB	8082A		07/07/25	07/08/25	
			Pesticide-TCL	8081B		07/07/25	07/07/25	
Q2514-06	TP-103	SOIL			07/02/25			07/03/25
			PCB	8082A		07/07/25	07/08/25	
			Pesticide-TCL	8081B		07/07/25	07/07/25	
Q2514-07	TP-36	SOIL			07/03/25			07/03/25
			PCB	8082A		07/07/25	07/08/25	
			Pesticide-TCL	8081B		07/07/25	07/07/25	
Q2514-08	TP-78	SOIL			07/03/25			07/03/25
			PCB	8082A		07/07/25	07/08/25	
			Pesticide-TCL	8081B		07/07/25	07/07/25	
Q2514-09	TP-81	SOIL			07/03/25			07/03/25

LAB CHRONICLE

Q2514-10	TP-90	SOIL	PCB	8082A		07/07/25	07/08/25
			Pesticide-TCL	8081B		07/07/25	07/07/25
					07/03/25		
							07/03/25
			PCB	8082A		07/07/25	07/08/25
			Pesticide-TCL	8081B		07/07/25	07/07/25



284 Sheffield Street, Mountainside, New Jersey 07092, Phone : 908 789 8900,
Fax : 908 789 8922

Hit Summary Sheet
SW-846

SDG No.: Q2514

Order ID: Q2514

Client: CDM Smith

Project ID: South River WM Replacement

Sample ID	Client ID	Matrix	Parameter	Concentration	C	MDL	RDL	Units
-----------	-----------	--------	-----------	---------------	---	-----	-----	-------

Client ID :

Total Concentration: 0.000



QC SUMMARY

Surrogate Summary

SDG No.: **Q2514**

Client: **CDM Smith**

Analytical Method: **8082A**

Lab Sample ID	Client ID	Parameter	Column	Spike	Result	Recovery(%)	Qual	Limits(%)	
								Low	High
I.BLK-PO111586.D	PIBLK-PO111586.D	Tetrachloro-m-xyl	1	20	20.3	102		60	140
		Decachlorobiphen	1	20	20.7	104		60	140
		Tetrachloro-m-xyl	2	20	19.0	95		60	140
		Decachlorobiphen	2	20	20.5	103		60	140
I.BLK-PO112056.D	PIBLK-PO112056.D	Tetrachloro-m-xyl	1	20	18.8	94		60	140
		Decachlorobiphen	1	20	16.8	84		60	140
		Tetrachloro-m-xyl	2	20	16.9	85		60	140
		Decachlorobiphen	2	20	17.1	86		60	140
PB168735BL	PB168735BL	Tetrachloro-m-xyl	1	20	20.5	103		32	144
		Decachlorobiphen	1	20	18.0	90		32	175
		Tetrachloro-m-xyl	2	20	18.6	93		32	144
		Decachlorobiphen	2	20	18.5	92		32	175
PB168735BS	PB168735BS	Tetrachloro-m-xyl	1	20	19.5	98		32	144
		Decachlorobiphen	1	20	18.3	92		32	175
		Tetrachloro-m-xyl	2	20	17.5	87		32	144
		Decachlorobiphen	2	20	18.5	92		32	175
Q2513-03MS	HR-3-070325MS	Tetrachloro-m-xyl	1	20	15.0	75		32	144
		Decachlorobiphen	1	20	9.84	49		32	175
		Tetrachloro-m-xyl	2	20	13.8	69		32	144
		Decachlorobiphen	2	20	12.3	61		32	175
Q2513-03MSD	HR-3-070325MSD	Tetrachloro-m-xyl	1	20	14.9	74		32	144
		Decachlorobiphen	1	20	10.3	52		32	175
		Tetrachloro-m-xyl	2	20	13.6	68		32	144
		Decachlorobiphen	2	20	13.0	65		32	175
Q2514-01	TP-92	Tetrachloro-m-xyl	1	20	20.4	102		32	144
		Decachlorobiphen	1	20	8.01	40		32	175
		Tetrachloro-m-xyl	2	20	17.8	89		32	144
		Decachlorobiphen	2	20	10.4	52		32	175
Q2514-02	TP-93	Tetrachloro-m-xyl	1	20	18.2	91		32	144
		Decachlorobiphen	1	20	10.4	52		32	175
		Tetrachloro-m-xyl	2	20	16.4	82		32	144
		Decachlorobiphen	2	20	12.8	64		32	175
I.BLK-PO112072.D	PIBLK-PO112072.D	Tetrachloro-m-xyl	1	20	20.6	103		60	140
		Decachlorobiphen	1	20	16.1	80		60	140
		Tetrachloro-m-xyl	2	20	18.3	92		60	140
		Decachlorobiphen	2	20	17.4	87		60	140
Q2514-03	TP-94	Tetrachloro-m-xyl	1	20	21.5	108		32	144
		Decachlorobiphen	1	20	13.1	65		32	175
		Tetrachloro-m-xyl	2	20	18.8	94		32	144
		Decachlorobiphen	2	20	13.8	69		32	175
Q2514-04	TP-96	Tetrachloro-m-xyl	1	20	20.5	102		32	144

Surrogate Summary

SDG No.: Q2514

Client: CDM Smith

Analytical Method: 8082A

Lab Sample ID	Client ID	Parameter	Column	Spike	Result	Recovery(%)	Qual	Limits(%)	
								Low	High
Q2514-04	TP-96	Decachlorobiphen	1	20	12.9	65		32	175
		Tetrachloro-m-xyl	2	20	18.0	90		32	144
Q2514-05	TP-97	Decachlorobiphen	2	20	14.3	72		32	175
		Tetrachloro-m-xyl	1	20	20.1	100		32	144
		Decachlorobiphen	1	20	12.2	61		32	175
		Tetrachloro-m-xyl	2	20	17.5	87		32	144
		Decachlorobiphen	2	20	13.3	66		32	175
Q2514-06	TP-103	Tetrachloro-m-xyl	1	20	20.1	101		32	144
		Decachlorobiphen	1	20	12.4	62		32	175
		Tetrachloro-m-xyl	2	20	17.6	88		32	144
		Decachlorobiphen	2	20	13.2	66		32	175
Q2514-07	TP-36	Tetrachloro-m-xyl	1	20	21.6	108		32	144
		Decachlorobiphen	1	20	15.4	77		32	175
		Tetrachloro-m-xyl	2	20	18.9	94		32	144
		Decachlorobiphen	2	20	15.7	79		32	175
Q2514-08	TP-78	Tetrachloro-m-xyl	1	20	18.9	94		32	144
		Decachlorobiphen	1	20	10.9	54		32	175
		Tetrachloro-m-xyl	2	20	16.7	84		32	144
		Decachlorobiphen	2	20	10.9	54		32	175
Q2514-09	TP-81	Tetrachloro-m-xyl	1	20	19.4	97		32	144
		Decachlorobiphen	1	20	9.85	49		32	175
		Tetrachloro-m-xyl	2	20	17.2	86		32	144
		Decachlorobiphen	2	20	10.0	50		32	175
Q2514-10	TP-90	Tetrachloro-m-xyl	1	20	22.2	111		32	144
		Decachlorobiphen	1	20	16.5	83		32	175
		Tetrachloro-m-xyl	2	20	19.6	98		32	144
		Decachlorobiphen	2	20	17.0	85		32	175
I.BLK-PO112085.D	PIBLK-PO112085.D	Tetrachloro-m-xyl	1	20	17.8	89		60	140
		Decachlorobiphen	1	20	14.5	72		60	140
		Tetrachloro-m-xyl	2	20	11.4	57	*	60	140
		Decachlorobiphen	2	20	11.1	55	*	60	140

Matrix Spike/Matrix Spike Duplicate Summary

SW-846

SDG No.:	<u>Q2514</u>	Analytical Method:	<u>8082A</u>
Client:	<u>CDM Smith</u>	DataFile :	<u>PO112063.D</u>

		Sample				Rec		RPD		Limits	
Parameter	Spike	Result	Result	Units	Rec	Qual	RPD	Qual	Low	High	RPD
Lab Sample ID:	Q2513-03MS	Client Sample ID:		HR-3-070325MS							
(Column 1)											
AR1016	176.2	0	118	ug/kg	67				55	146	
AR1260	176.2	0	105	ug/kg	60				54	119	
Lab Sample ID:	Q2513-03MS	Client Sample ID:		HR-3-070325MS							
(Column 2)											
AR1016	176.2	0	115	ug/kg	65				55	146	
AR1260	176.2	0	112	ug/kg	64				54	119	

Matrix Spike/Matrix Spike Duplicate Summary

SW-846

SDG No.:	<u>Q2514</u>	Analytical Method:	<u>8082A</u>
Client:	<u>CDM Smith</u>	DataFile :	<u>PO112064.D</u>

	Parameter	Spike	Sample Result	Result	Units	Rec Qual	RPD Qual	Low	Limits High	RPD
Lab Sample ID:	Q2513-03MSD (Column 1)		Client Sample ID:	HR-3-070325MSD						
	AR1016	176.4	0	109	ug/kg	62	8	55	146	15
	AR1260	176.4	0	107	ug/kg	61	2	54	119	15
Lab Sample ID:	Q2513-03MSD (Column 2)		Client Sample ID:	HR-3-070325MSD						
	AR1016	176.4	0	119	ug/kg	67	3	55	146	15
	AR1260	176.4	0	114	ug/kg	65	2	54	119	15



284 Sheffield Street, Mountainside, New Jersey 07092, Phone : 908 789 8900,
Fax : 908 789 8922

Laboratory Control Sample/Laboratory Control Sample Duplicate Summary

SW-846

SDG No.: Q2514

Analytical Method: 8082A

Client: CDM Smith

Datafile : PO112058.D

Lab Sample ID	Parameter	Spike	Result	Units	Rec	RPD	Qual	RPD	Limits		RPD
								Qual	Low	High	
PB168735BS (Column 1)	AR1016	166.6	145	ug/kg	87				71	120	
	AR1260	166.6	163	ug/kg	98				65	130	
PB168735BS (Column 2)	AR1016	166.6	150	ug/kg	90				71	120	
	AR1260	166.6	152	ug/kg	91				65	130	

4C

PESTICIDE METHOD BLANK SUMMARY

Client ID

PB168735BL

Lab Name: Alliance

Contract: CAMP02

Lab Code: ACE

SDG NO.: Q2514

Lab Sample ID: PB168735BL

Lab File ID: PO112057.D

Matrix: (soil/water) Solid

Extraction: (Type) SOXH

Sulfur Cleanup: (Y/N) N

Date Extracted: 07/07/2025

Date Analyzed (1): 07/07/2025

Date Analyzed (2): 07/07/2025

Time Analyzed (1): 22:01

Time Analyzed (2): 22:01

Instrument ID (1): ECD_O

Instrument ID (2): ECD_O

GC Column (1): ZB-MR1 ID: 0.32 (mm)

GC Column (2): ZB-MR2 ID: 0.32 (mm)

THIS METHOD BLANK APPLIES TO THE FOLLOWING SAMPLES, MS AND MSD:

EPA SAMPLE NO.	LAB SAMPLE ID	LAB FILE ID	DATE ANALYZED 1	DATE ANALYZED 2
PB168735BS	PB168735BS	PO112058.D	07/07/2025	07/07/2025
HR-3-070325MS	Q2513-03MS	PO112063.D	07/07/2025	07/07/2025
HR-3-070325MSD	Q2513-03MSD	PO112064.D	07/08/2025	07/08/2025
TP-92	Q2514-01	PO112065.D	07/08/2025	07/08/2025
TP-93	Q2514-02	PO112066.D	07/08/2025	07/08/2025
TP-94	Q2514-03	PO112073.D	07/08/2025	07/08/2025
TP-96	Q2514-04	PO112074.D	07/08/2025	07/08/2025
TP-97	Q2514-05	PO112075.D	07/08/2025	07/08/2025
TP-103	Q2514-06	PO112076.D	07/08/2025	07/08/2025
TP-36	Q2514-07	PO112077.D	07/08/2025	07/08/2025
TP-78	Q2514-08	PO112078.D	07/08/2025	07/08/2025
TP-81	Q2514-09	PO112079.D	07/08/2025	07/08/2025
TP-90	Q2514-10	PO112080.D	07/08/2025	07/08/2025

COMMENTS: _____



SAMPLE DATA

Report of Analysis

Client:	CDM Smith	Date Collected:	07/02/25
Project:	South River WM Replacement	Date Received:	07/03/25
Client Sample ID:	TP-92	SDG No.:	Q2514
Lab Sample ID:	Q2514-01	Matrix:	SOIL
Analytical Method:	8082A	% Solid:	87.4
Sample Wt/Vol:	30.07	Units:	g
Soil Aliquot Vol:			uL
Extraction Type:		Test:	PCB
GPC Factor :	1.0	Injection Volume :	
Prep Method :	SW3541B		

File ID/Qc Batch:	Dilution:	Prep Date	Date Analyzed	Prep Batch ID
PO112065.D	1	07/07/25 08:30	07/08/25 00:29	PB168735

CAS Number	Parameter	Conc.	Qualifier	MDL	LOQ / CRQL	Units(Dry Weight)
TARGETS						
12674-11-2	Aroclor-1016	4.50	U	4.50	19.4	ug/kg
11104-28-2	Aroclor-1221	4.60	U	4.60	19.4	ug/kg
11141-16-5	Aroclor-1232	4.20	U	4.20	19.4	ug/kg
53469-21-9	Aroclor-1242	4.60	U	4.60	19.4	ug/kg
12672-29-6	Aroclor-1248	6.80	U	6.80	19.4	ug/kg
11097-69-1	Aroclor-1254	3.70	U	3.70	19.4	ug/kg
37324-23-5	Aroclor-1262	5.70	U	5.70	19.4	ug/kg
11100-14-4	Aroclor-1268	4.10	U	4.10	19.4	ug/kg
11096-82-5	Aroclor-1260	3.70	U	3.70	19.4	ug/kg
SURROGATES						
877-09-8	Tetrachloro-m-xylene	20.4		32 - 144	102%	SPK: 20
2051-24-3	Decachlorobiphenyl	10.4		32 - 175	52%	SPK: 20

Comments:

U = Not Detected

LOQ = Limit of Quantitation

MDL = Method Detection Limit

LOD = Limit of Detection

E = Value Exceeds Calibration Range

P = Indicates >25% difference for detected concentrations between the two GC columns

Q = indicates LCS control criteria did not meet requirements

M = MS/MSD acceptance criteria did not meet requirements

J = Estimated Value

B = Analyte Found in Associated Method Blank

N = Presumptive Evidence of a Compound

* = Values outside of QC limits

D = Dilution

S = Indicates estimated value where valid five-point calibration was not performed prior to analyte detection in sample.

() = Laboratory InHouse Limit

Data Path : Z:\pestpcbsrv\HPCHEM1\ECD_O\Data\P0070725\
 Data File : PO112065.D
 Signal(s) : Signal #1: ECD1A.ch Signal #2: ECD2B.ch
 Acq On : 08 Jul 2025 00:29
 Operator : YP/AJ
 Sample : Q2514-01
 Misc :
 ALS Vial : 35 Sample Multiplier: 1

Instrument :
 ECD_O
 ClientSampleId :
 TP-92

Manual Integrations
APPROVED

Reviewed By :Yogesh Patel 07/08/2025
 Supervised By :mohammad ahmed 07/09/2025

Integration File signal 1: autoint1.e
 Integration File signal 2: autoint2.e
 Quant Time: Jul 08 02:00:33 2025
 Quant Method : Z:\pestpcbsrv\HPCHEM1\ECD_O\methods\P0061125.M
 Quant Title : GC EXTRACTABLES
 QLast Update : Thu Jun 12 06:25:26 2025
 Response via : Initial Calibration
 Integrator: ChemStation

Volume Inj. : 2 µl
 Signal #1 Phase : ZB-MR1 Signal #2 Phase: ZB-MR2
 Signal #1 Info : 30Mx0.32mmx 0.50µ Signal #2 Info : 30M x 0.32mm x 0.25µm

Compound	RT#1	RT#2	Resp#1	Resp#2	ng/ml	ng/ml

System Monitoring Compounds						
1) SA Tetrachlo...	3.674	3.667	117.2E6	99788387	20.356	17.765
2) SA Decachlor...	8.705	8.651	42050204	18472429	8.011	10.391m#

Target Compounds

(f)=RT Delta > 1/2 Window (#)=Amounts differ by > 25% (m)=manual int.

Data Path : Z:\pestpcbsrv\HPCHEM1\ECD_O\Data\PO070725\
Data File : PO112065.D
Signal(s) : Signal #1: ECD1A.ch Signal #2: ECD2B.ch
Acq On : 08 Jul 2025 00:29
Operator : YP/AJ
Sample : Q2514-01
Misc :
ALS Vial : 35 Sample Multiplier: 1



Instrument :
ECD_O
ClientSampleId :
TP-92


Manual Integrations
APPROVED

Reviewed By :Yogesh Patel 07/08/2025
Supervised By :mohammad ahmed 07/09/2025

Integration File signal 1: autoint1.e
Integration File signal 2: autoint2.e
Quant Time: Jul 08 02:00:33 2025
Quant Method : Z:\pestpcbsrv\HPCHEM1\ECD_O\methods\PO061125.M
Quant Title : GC EXTRACTABLES
QLast Update : Thu Jun 12 06:25:26 2025
Response via : Initial Calibration
Integrator: ChemStation

Volume Inj. : 2 µl
Signal #1 Phase : ZB-MR1 Signal #2 Phase: ZB-MR2
Signal #1 Info : 30Mx0.32mmx 0.50µm Signal #2 Info : 30M x 0.32mm x 0.25µm






#1 Tetrachloro-m-xylene

R.T.: 3.674 min
Delta R.T.: -0.003 min
Response: 117239942
Conc: 20.36 ng/ml

Instrument :
ECD_O
ClientSampleId :
TP-92


Manual Integrations
APPROVED

Reviewed By :Yogesh Patel 07/08/2025
Supervised By :mohammad ahmed 07/09/2025




#1 Tetrachloro-m-xylene

R.T.: 3.667 min
Delta R.T.: -0.006 min
Response: 99788387
Conc: 17.76 ng/ml



#2 Decachlorobiphenyl

R.T.: 8.705 min
Delta R.T.: -0.008 min
Response: 42050204
Conc: 8.01 ng/ml



#2 Decachlorobiphenyl

R.T.: 8.651 min
Delta R.T.: -0.011 min
Response: 18472429
Conc: 10.39 ng/ml m

Report of Analysis

Client:	CDM Smith	Date Collected:	07/02/25
Project:	South River WM Replacement	Date Received:	07/03/25
Client Sample ID:	TP-93	SDG No.:	Q2514
Lab Sample ID:	Q2514-02	Matrix:	SOIL
Analytical Method:	8082A	% Solid:	87.7
Sample Wt/Vol:	30.01	Units:	g
Soil Aliquot Vol:			uL
Extraction Type:		Test:	PCB
GPC Factor :	1.0	Injection Volume :	
Prep Method :	SW3541B		

File ID/Qc Batch:	Dilution:	Prep Date	Date Analyzed	Prep Batch ID
PO112066.D	1	07/07/25 08:30	07/08/25 00:46	PB168735

CAS Number	Parameter	Conc.	Qualifier	MDL	LOQ / CRQL	Units(Dry Weight)
TARGETS						
12674-11-2	Aroclor-1016	4.50	U	4.50	19.4	ug/kg
11104-28-2	Aroclor-1221	4.60	U	4.60	19.4	ug/kg
11141-16-5	Aroclor-1232	4.20	U	4.20	19.4	ug/kg
53469-21-9	Aroclor-1242	4.60	U	4.60	19.4	ug/kg
12672-29-6	Aroclor-1248	6.70	U	6.70	19.4	ug/kg
11097-69-1	Aroclor-1254	3.70	U	3.70	19.4	ug/kg
37324-23-5	Aroclor-1262	5.70	U	5.70	19.4	ug/kg
11100-14-4	Aroclor-1268	4.10	U	4.10	19.4	ug/kg
11096-82-5	Aroclor-1260	3.70	U	3.70	19.4	ug/kg
SURROGATES						
877-09-8	Tetrachloro-m-xylene	18.2		32 - 144	91%	SPK: 20
2051-24-3	Decachlorobiphenyl	12.8		32 - 175	64%	SPK: 20

Comments:

U = Not Detected

LOQ = Limit of Quantitation

MDL = Method Detection Limit

LOD = Limit of Detection

E = Value Exceeds Calibration Range

P = Indicates >25% difference for detected concentrations between the two GC columns

Q = indicates LCS control criteria did not meet requirements

M = MS/MSD acceptance criteria did not meet requirements

J = Estimated Value

B = Analyte Found in Associated Method Blank

N = Presumptive Evidence of a Compound

* = Values outside of QC limits

D = Dilution

S = Indicates estimated value where valid five-point calibration was not performed prior to analyte detection in sample.

() = Laboratory InHouse Limit

Data Path : Z:\pestpcbsrv\HPCHEM1\ECD_O\Data\P0070725\
 Data File : PO112066.D
 Signal(s) : Signal #1: ECD1A.ch Signal #2: ECD2B.ch
 Acq On : 08 Jul 2025 00:46
 Operator : YP/AJ
 Sample : Q2514-02
 Misc :
 ALS Vial : 36 Sample Multiplier: 1

Instrument :
 ECD_O
 ClientSampleId :
 TP-93

Manual Integrations
APPROVED

Reviewed By :Yogesh Patel 07/08/2025
 Supervised By :mohammad ahmed 07/09/2025

Integration File signal 1: autoint1.e
 Integration File signal 2: autoint2.e
 Quant Time: Jul 08 02:00:52 2025
 Quant Method : Z:\pestpcbsrv\HPCHEM1\ECD_O\methods\P0061125.M
 Quant Title : GC EXTRACTABLES
 QLast Update : Thu Jun 12 06:25:26 2025
 Response via : Initial Calibration
 Integrator: ChemStation

Volume Inj. : 2 µl
 Signal #1 Phase : ZB-MR1 Signal #2 Phase: ZB-MR2
 Signal #1 Info : 30Mx0.32mmx 0.50µ Signal #2 Info : 30M x 0.32mm x 0.25µm

Compound	RT#1	RT#2	Resp#1	Resp#2	ng/ml	ng/ml

System Monitoring Compounds						
1) SA Tetrachlo...	3.671	3.667	105.0E6	91817673	18.239	16.346
2) SA Decachlor...	8.702	8.649	54871792	22738766	10.453	12.791m

Target Compounds

(f)=RT Delta > 1/2 Window (#)=Amounts differ by > 25% (m)=manual int.

Data Path : Z:\pestpcbsrv\HPCHEM1\ECD_O\Data\PO070725\
Data File : PO112066.D
Signal(s) : Signal #1: ECD1A.ch Signal #2: ECD2B.ch
Acq On : 08 Jul 2025 00:46
Operator : YP/AJ
Sample : Q2514-02
Misc :
ALS Vial : 36 Sample Multiplier: 1



Instrument :
ECD_O
ClientSampleId :
TP-93


Manual Integrations
APPROVED

Reviewed By :Yogesh Patel 07/08/2025
Supervised By :mohammad ahmed 07/09/2025

Integration File signal 1: autoint1.e
Integration File signal 2: autoint2.e
Quant Time: Jul 08 02:00:52 2025
Quant Method : Z:\pestpcbsrv\HPCHEM1\ECD_O\methods\PO061125.M
Quant Title : GC EXTRACTABLES
QLast Update : Thu Jun 12 06:25:26 2025
Response via : Initial Calibration
Integrator: ChemStation

Volume Inj. : 2 µl
Signal #1 Phase : ZB-MR1 Signal #2 Phase: ZB-MR2
Signal #1 Info : 30Mx0.32mmx 0.50µ Signal #2 Info : 30M x 0.32mm x 0.25µm






#1 Tetrachloro-m-xylene

R.T.: 3.671 min
Delta R.T.: -0.006 min
Response: 105044379
Conc: 18.24 ng/ml

Instrument :
ECD_O
ClientSampleId :
TP-93


Manual Integrations
APPROVED

Reviewed By :Yogesh Patel 07/08/2025
Supervised By :mohammad ahmed 07/09/2025




#1 Tetrachloro-m-xylene

R.T.: 3.667 min
Delta R.T.: -0.006 min
Response: 91817673
Conc: 16.35 ng/ml



#2 Decachlorobiphenyl

R.T.: 8.702 min
Delta R.T.: -0.010 min
Response: 54871792
Conc: 10.45 ng/ml



#2 Decachlorobiphenyl

R.T.: 8.649 min
Delta R.T.: -0.012 min
Response: 22738766
Conc: 12.79 ng/ml m

Report of Analysis

Client:	CDM Smith	Date Collected:	07/02/25
Project:	South River WM Replacement	Date Received:	07/03/25
Client Sample ID:	TP-94	SDG No.:	Q2514
Lab Sample ID:	Q2514-03	Matrix:	SOIL
Analytical Method:	8082A	% Solid:	88.1
Sample Wt/Vol:	30.08	Units:	g
Soil Aliquot Vol:			uL
Extraction Type:		Test:	PCB
GPC Factor :	1.0	PH :	
Prep Method :	SW3541B	Injection Volume :	

File ID/Qc Batch:	Dilution:	Prep Date	Date Analyzed	Prep Batch ID
PO112073.D	1	07/07/25 08:30	07/08/25 05:44	PB168735

CAS Number	Parameter	Conc.	Qualifier	MDL	LOQ / CRQL	Units(Dry Weight)
TARGETS						
12674-11-2	Aroclor-1016	4.50	U	4.50	19.2	ug/kg
11104-28-2	Aroclor-1221	4.60	U	4.60	19.2	ug/kg
11141-16-5	Aroclor-1232	4.20	U	4.20	19.2	ug/kg
53469-21-9	Aroclor-1242	4.50	U	4.50	19.2	ug/kg
12672-29-6	Aroclor-1248	6.70	U	6.70	19.2	ug/kg
11097-69-1	Aroclor-1254	3.60	U	3.60	19.2	ug/kg
37324-23-5	Aroclor-1262	5.70	U	5.70	19.2	ug/kg
11100-14-4	Aroclor-1268	4.10	U	4.10	19.2	ug/kg
11096-82-5	Aroclor-1260	3.70	U	3.70	19.2	ug/kg
SURROGATES						
877-09-8	Tetrachloro-m-xylene	21.5		32 - 144	108%	SPK: 20
2051-24-3	Decachlorobiphenyl	13.8		32 - 175	69%	SPK: 20

Comments:

U = Not Detected

LOQ = Limit of Quantitation

MDL = Method Detection Limit

LOD = Limit of Detection

E = Value Exceeds Calibration Range

P = Indicates >25% difference for detected concentrations between the two GC columns

Q = indicates LCS control criteria did not meet requirements

M = MS/MSD acceptance criteria did not meet requirements

J = Estimated Value

B = Analyte Found in Associated Method Blank

N = Presumptive Evidence of a Compound

* = Values outside of QC limits

D = Dilution

S = Indicates estimated value where valid five-point calibration was not performed prior to analyte detection in sample.

() = Laboratory InHouse Limit

Data Path : Z:\pestpcbsrv\HPCHEM1\ECD_O\Data\P0070725\
 Data File : PO112073.D
 Signal(s) : Signal #1: ECD1A.ch Signal #2: ECD2B.ch
 Acq On : 08 Jul 2025 05:44
 Operator : YP/AJ
 Sample : Q2514-03
 Misc :
 ALS Vial : 37 Sample Multiplier: 1

Instrument :
 ECD_O
 ClientSampleId :
 TP-94

Manual Integrations
APPROVED

Reviewed By :Yogesh Patel 07/08/2025
 Supervised By :mohammad ahmed 07/09/2025

Integration File signal 1: autoint1.e
 Integration File signal 2: autoint2.e
 Quant Time: Jul 08 08:08:16 2025
 Quant Method : Z:\pestpcbsrv\HPCHEM1\ECD_O\methods\P0061125.M
 Quant Title : GC EXTRACTABLES
 QLast Update : Thu Jun 12 06:25:26 2025
 Response via : Initial Calibration
 Integrator: ChemStation

Volume Inj. : 2 µl
 Signal #1 Phase : ZB-MR1 Signal #2 Phase: ZB-MR2
 Signal #1 Info : 30Mx0.32mmx 0.50µ Signal #2 Info : 30M x 0.32mm x 0.25µm

Compound	RT#1	RT#2	Resp#1	Resp#2	ng/ml	ng/ml

System Monitoring Compounds						
1) SA Tetrachlo...	3.674	3.667	123.9E6	105.8E6	21.518	18.827
2) SA Decachlor...	8.704	8.650	68556389	24591824	13.060	13.834m

Target Compounds

(f)=RT Delta > 1/2 Window (#)=Amounts differ by > 25% (m)=manual int.

Data Path : Z:\pestpcbsrv\HPCHEM1\ECD_O\Data\PO070725\
Data File : PO112073.D
Signal(s) : Signal #1: ECD1A.ch Signal #2: ECD2B.ch
Acq On : 08 Jul 2025 05:44
Operator : YP/AJ
Sample : Q2514-03
Misc :
ALS Vial : 37 Sample Multiplier: 1



Instrument :
ECD_O
ClientSampleId :
TP-94


Manual Integrations
APPROVED

Reviewed By :Yogesh Patel 07/08/2025
Supervised By :mohammad ahmed 07/09/2025

Integration File signal 1: autoint1.e
Integration File signal 2: autoint2.e
Quant Time: Jul 08 08:08:16 2025
Quant Method : Z:\pestpcbsrv\HPCHEM1\ECD_O\methods\PO061125.M
Quant Title : GC EXTRACTABLES
QLast Update : Thu Jun 12 06:25:26 2025
Response via : Initial Calibration
Integrator: ChemStation

Volume Inj. : 2 µl
Signal #1 Phase : ZB-MR1 Signal #2 Phase: ZB-MR2
Signal #1 Info : 30Mx0.32mmx 0.50µ Signal #2 Info : 30M x 0.32mm x 0.25µm






#1 Tetrachloro-m-xylene

R.T.: 3.674 min
Delta R.T.: -0.003 min
Response: 123931786
Conc: 21.52 ng/ml

Instrument :
ECD_O
ClientSampleId :
TP-94


Manual Integrations
APPROVED

Reviewed By :Yogesh Patel 07/08/2025
Supervised By :mohammad ahmed 07/09/2025




#1 Tetrachloro-m-xylene

R.T.: 3.667 min
Delta R.T.: -0.006 min
Response: 105755094
Conc: 18.83 ng/ml



#2 Decachlorobiphenyl

R.T.: 8.704 min
Delta R.T.: -0.008 min
Response: 68556389
Conc: 13.06 ng/ml



#2 Decachlorobiphenyl

R.T.: 8.650 min
Delta R.T.: -0.012 min
Response: 24591824
Conc: 13.83 ng/ml m

Report of Analysis

Client:	CDM Smith	Date Collected:	07/02/25
Project:	South River WM Replacement	Date Received:	07/03/25
Client Sample ID:	TP-96	SDG No.:	Q2514
Lab Sample ID:	Q2514-04	Matrix:	SOIL
Analytical Method:	8082A	% Solid:	85.8
Sample Wt/Vol:	30.03	Units:	g
Soil Aliquot Vol:			uL
Extraction Type:		Test:	PCB
GPC Factor :	1.0	Injection Volume :	
Prep Method :	SW3541B		

File ID/Qc Batch:	Dilution:	Prep Date	Date Analyzed	Prep Batch ID
PO112074.D	1	07/07/25 08:30	07/08/25 06:02	PB168735

CAS Number	Parameter	Conc.	Qualifier	MDL	LOQ / CRQL	Units(Dry Weight)
TARGETS						
12674-11-2	Aroclor-1016	4.60	U	4.60	19.8	ug/kg
11104-28-2	Aroclor-1221	4.70	U	4.70	19.8	ug/kg
11141-16-5	Aroclor-1232	4.30	U	4.30	19.8	ug/kg
53469-21-9	Aroclor-1242	4.70	U	4.70	19.8	ug/kg
12672-29-6	Aroclor-1248	6.90	U	6.90	19.8	ug/kg
11097-69-1	Aroclor-1254	3.70	U	3.70	19.8	ug/kg
37324-23-5	Aroclor-1262	5.80	U	5.80	19.8	ug/kg
11100-14-4	Aroclor-1268	4.20	U	4.20	19.8	ug/kg
11096-82-5	Aroclor-1260	3.80	U	3.80	19.8	ug/kg
SURROGATES						
877-09-8	Tetrachloro-m-xylene	20.5		32 - 144	102%	SPK: 20
2051-24-3	Decachlorobiphenyl	14.3		32 - 175	72%	SPK: 20

Comments:

U = Not Detected

LOQ = Limit of Quantitation

MDL = Method Detection Limit

LOD = Limit of Detection

E = Value Exceeds Calibration Range

P = Indicates >25% difference for detected concentrations between the two GC columns

Q = indicates LCS control criteria did not meet requirements

M = MS/MSD acceptance criteria did not meet requirements

J = Estimated Value

B = Analyte Found in Associated Method Blank

N = Presumptive Evidence of a Compound

* = Values outside of QC limits

D = Dilution

S = Indicates estimated value where valid five-point calibration was not performed prior to analyte detection in sample.

() = Laboratory InHouse Limit

Data Path : Z:\pestpcbsrv\HPCHEM1\ECD_O\Data\P0070725\
 Data File : PO112074.D
 Signal(s) : Signal #1: ECD1A.ch Signal #2: ECD2B.ch
 Acq On : 08 Jul 2025 06:02
 Operator : YP/AJ
 Sample : Q2514-04
 Misc :
 ALS Vial : 38 Sample Multiplier: 1

Instrument :
 ECD_O
 ClientSampleId :
 TP-96

Manual Integrations
APPROVED

Reviewed By :Yogesh Patel 07/08/2025
 Supervised By :mohammad ahmed 07/09/2025

Integration File signal 1: autoint1.e
 Integration File signal 2: autoint2.e
 Quant Time: Jul 08 08:08:30 2025
 Quant Method : Z:\pestpcbsrv\HPCHEM1\ECD_O\methods\P0061125.M
 Quant Title : GC EXTRACTABLES
 QLast Update : Thu Jun 12 06:25:26 2025
 Response via : Initial Calibration
 Integrator: ChemStation

Volume Inj. : 2 µl
 Signal #1 Phase : ZB-MR1 Signal #2 Phase: ZB-MR2
 Signal #1 Info : 30Mx0.32mmx 0.50µ Signal #2 Info : 30M x 0.32mm x 0.25µm

Compound	RT#1	RT#2	Resp#1	Resp#2	ng/ml	ng/ml

System Monitoring Compounds						
1) SA Tetrachlo...	3.675	3.668	117.8E6	101.3E6	20.458	18.026
2) SA Decachlor...	8.704	8.650	67763700	25445473	12.909	14.314m

Target Compounds

(f)=RT Delta > 1/2 Window (#)=Amounts differ by > 25% (m)=manual int.

Data Path : Z:\pestpcbsrv\HPCHEM1\ECD_O\Data\PO070725\
Data File : PO112074.D
Signal(s) : Signal #1: ECD1A.ch Signal #2: ECD2B.ch
Acq On : 08 Jul 2025 06:02
Operator : YP/AJ
Sample : Q2514-04
Misc :
ALS Vial : 38 Sample Multiplier: 1



Instrument :
ECD_O
ClientSampleId :
TP-96


Manual Integrations
APPROVED

Reviewed By :Yogesh Patel 07/08/2025
Supervised By :mohammad ahmed 07/09/2025

Integration File signal 1: autoint1.e
Integration File signal 2: autoint2.e
Quant Time: Jul 08 08:08:30 2025
Quant Method : Z:\pestpcbsrv\HPCHEM1\ECD_O\methods\PO061125.M
Quant Title : GC EXTRACTABLES
QLast Update : Thu Jun 12 06:25:26 2025
Response via : Initial Calibration
Integrator: ChemStation

Volume Inj. : 2 µl
Signal #1 Phase : ZB-MR1 Signal #2 Phase: ZB-MR2
Signal #1 Info : 30Mx0.32mmx 0.50µ Signal #2 Info : 30M x 0.32mm x 0.25µm






#1 Tetrachloro-m-xylene

R.T.: 3.675 min
Delta R.T.: -0.002 min
Response: 117825736
Conc: 20.46 ng/ml

Instrument :
ECD_O
ClientSampleId :
TP-96


Manual Integrations
APPROVED

Reviewed By :Yogesh Patel 07/08/2025
Supervised By :mohammad ahmed 07/09/2025




#1 Tetrachloro-m-xylene

R.T.: 3.668 min
Delta R.T.: -0.005 min
Response: 101254635
Conc: 18.03 ng/ml



#2 Decachlorobiphenyl

R.T.: 8.704 min
Delta R.T.: -0.008 min
Response: 67763700
Conc: 12.91 ng/ml



#2 Decachlorobiphenyl

R.T.: 8.650 min
Delta R.T.: -0.011 min
Response: 25445473
Conc: 14.31 ng/ml m

Report of Analysis

Client:	CDM Smith	Date Collected:	07/02/25
Project:	South River WM Replacement	Date Received:	07/03/25
Client Sample ID:	TP-97	SDG No.:	Q2514
Lab Sample ID:	Q2514-05	Matrix:	SOIL
Analytical Method:	8082A	% Solid:	85
Sample Wt/Vol:	30.05	Units:	g
Soil Aliquot Vol:			uL
Extraction Type:		Test:	PCB
GPC Factor :	1.0	Injection Volume :	
Prep Method :	SW3541B		

File ID/Qc Batch:	Dilution:	Prep Date	Date Analyzed	Prep Batch ID
PO112075.D	1	07/07/25 08:30	07/08/25 06:21	PB168735

CAS Number	Parameter	Conc.	Qualifier	MDL	LOQ / CRQL	Units(Dry Weight)
TARGETS						
12674-11-2	Aroclor-1016	4.60	U	4.60	20.0	ug/kg
11104-28-2	Aroclor-1221	4.70	U	4.70	20.0	ug/kg
11141-16-5	Aroclor-1232	4.40	U	4.40	20.0	ug/kg
53469-21-9	Aroclor-1242	4.70	U	4.70	20.0	ug/kg
12672-29-6	Aroclor-1248	7.00	U	7.00	20.0	ug/kg
11097-69-1	Aroclor-1254	3.80	U	3.80	20.0	ug/kg
37324-23-5	Aroclor-1262	5.90	U	5.90	20.0	ug/kg
11100-14-4	Aroclor-1268	4.20	U	4.20	20.0	ug/kg
11096-82-5	Aroclor-1260	3.80	U	3.80	20.0	ug/kg
SURROGATES						
877-09-8	Tetrachloro-m-xylene	20.1		32 - 144	100%	SPK: 20
2051-24-3	Decachlorobiphenyl	13.3		32 - 175	66%	SPK: 20

Comments:

U = Not Detected

LOQ = Limit of Quantitation

MDL = Method Detection Limit

LOD = Limit of Detection

E = Value Exceeds Calibration Range

P = Indicates >25% difference for detected concentrations between the two GC columns

Q = indicates LCS control criteria did not meet requirements

M = MS/MSD acceptance criteria did not meet requirements

J = Estimated Value

B = Analyte Found in Associated Method Blank

N = Presumptive Evidence of a Compound

* = Values outside of QC limits

D = Dilution

S = Indicates estimated value where valid five-point calibration was not performed prior to analyte detection in sample.

() = Laboratory InHouse Limit

Data Path : Z:\pestpcbsrv\HPCHEM1\ECD_O\Data\P0070725\
 Data File : PO112075.D
 Signal(s) : Signal #1: ECD1A.ch Signal #2: ECD2B.ch
 Acq On : 08 Jul 2025 06:21
 Operator : YP/AJ
 Sample : Q2514-05
 Misc :
 ALS Vial : 39 Sample Multiplier: 1

Instrument :
 ECD_O
 ClientSampleId :
 TP-97

Manual Integrations
APPROVED

Reviewed By :Yogesh Patel 07/08/2025
 Supervised By :mohammad ahmed 07/09/2025

Integration File signal 1: autoint1.e
 Integration File signal 2: autoint2.e
 Quant Time: Jul 08 08:08:47 2025
 Quant Method : Z:\pestpcbsrv\HPCHEM1\ECD_O\methods\P0061125.M
 Quant Title : GC EXTRACTABLES
 QLast Update : Thu Jun 12 06:25:26 2025
 Response via : Initial Calibration
 Integrator: ChemStation

Volume Inj. : 2 µl
 Signal #1 Phase : ZB-MR1 Signal #2 Phase: ZB-MR2
 Signal #1 Info : 30Mx0.32mmx 0.50µ Signal #2 Info : 30M x 0.32mm x 0.25µm

Compound	RT#1	RT#2	Resp#1	Resp#2	ng/ml	ng/ml

System Monitoring Compounds						
1) SA Tetrachlo...	3.675	3.668	115.5E6	98085841	20.052	17.461
2) SA Decachlor...	8.705	8.650	63867864	23616971	12.167	13.285m

Target Compounds

(f)=RT Delta > 1/2 Window (#)=Amounts differ by > 25% (m)=manual int.

Data Path : Z:\pestpcbsrv\HPCHEM1\ECD_0\Data\PO070725\
Data File : PO112075.D
Signal(s) : Signal #1: ECD1A.ch Signal #2: ECD2B.ch
Acq On : 08 Jul 2025 06:21
Operator : YP/AJ
Sample : Q2514-05
Misc :
ALS Vial : 39 Sample Multiplier: 1



Instrument :
ECD_O
ClientSampleId :
TP-97

Manual Integrations
APPROVED


Reviewed By :Yogesh Patel 07/08/2025
Supervised By :mohammad ahmed 07/09/2025

Integration File signal 1: autoint1.e
Integration File signal 2: autoint2.e
Quant Time: Jul 08 08:08:47 2025
Quant Method : Z:\pestpcbsrv\HPCHEM1\ECD_0\methods\PO061125.M
Quant Title : GC EXTRACTABLES
QLast Update : Thu Jun 12 06:25:26 2025
Response via : Initial Calibration
Integrator: ChemStation

Volume Inj. : 2 µl
Signal #1 Phase : ZB-MR1 Signal #2 Phase: ZB-MR2
Signal #1 Info : 30Mx0.32mmx 0.50µ Signal #2 Info : 30M x 0.32mm x 0.25µm



Response_ Signal: PO112075.D\ECD1A.ch



#1 Tetrachloro-m-xylene


R.T.: 3.675 min
Delta R.T.: -0.002 min
Response: 115490170
Conc: 20.05 ng/ml

Instrument :
ECD_O
ClientSampleId :
TP-97

Manual Integrations
APPROVED

Reviewed By :Yogesh Patel 07/08/2025
Supervised By :mohammad ahmed 07/09/2025


Time Response_ Signal: PO112075.D\ECD2B.ch



#1 Tetrachloro-m-xylene

R.T.: 3.668 min
Delta R.T.: -0.006 min
Response: 98085841
Conc: 17.46 ng/ml


Time Response_ Signal: PO112075.D\ECD1A.ch



#2 Decachlorobiphenyl

R.T.: 8.705 min
Delta R.T.: -0.008 min
Response: 63867864
Conc: 12.17 ng/ml

Time Response_ Signal: PO112075.D\ECD2B.ch



#2 Decachlorobiphenyl

R.T.: 8.650 min
Delta R.T.: -0.011 min
Response: 23616971
Conc: 13.29 ng/ml m

Report of Analysis

Client:	CDM Smith	Date Collected:	07/02/25
Project:	South River WM Replacement	Date Received:	07/03/25
Client Sample ID:	TP-103	SDG No.:	Q2514
Lab Sample ID:	Q2514-06	Matrix:	SOIL
Analytical Method:	8082A	% Solid:	86.5
Sample Wt/Vol:	30.04	Units:	g
Soil Aliquot Vol:			uL
Extraction Type:		Test:	PCB
GPC Factor :	1.0	PH :	
Prep Method :	SW3541B	Injection Volume :	

File ID/Qc Batch:	Dilution:	Prep Date	Date Analyzed	Prep Batch ID
PO112076.D	1	07/07/25 08:30	07/08/25 06:39	PB168735

CAS Number	Parameter	Conc.	Qualifier	MDL	LOQ / CRQL	Units(Dry Weight)
TARGETS						
12674-11-2	Aroclor-1016	4.60	U	4.60	19.6	ug/kg
11104-28-2	Aroclor-1221	4.70	U	4.70	19.6	ug/kg
11141-16-5	Aroclor-1232	4.30	U	4.30	19.6	ug/kg
53469-21-9	Aroclor-1242	4.60	U	4.60	19.6	ug/kg
12672-29-6	Aroclor-1248	6.80	U	6.80	19.6	ug/kg
11097-69-1	Aroclor-1254	3.70	U	3.70	19.6	ug/kg
37324-23-5	Aroclor-1262	5.80	U	5.80	19.6	ug/kg
11100-14-4	Aroclor-1268	4.20	U	4.20	19.6	ug/kg
11096-82-5	Aroclor-1260	3.70	U	3.70	19.6	ug/kg
SURROGATES						
877-09-8	Tetrachloro-m-xylene	20.1		32 - 144	101%	SPK: 20
2051-24-3	Decachlorobiphenyl	13.2		32 - 175	66%	SPK: 20

Comments:

U = Not Detected
 LOQ = Limit of Quantitation
 MDL = Method Detection Limit
 LOD = Limit of Detection
 E = Value Exceeds Calibration Range
 P = Indicates >25% difference for detected concentrations between the two GC columns
 Q = indicates LCS control criteria did not meet requirements
 M = MS/MSD acceptance criteria did not meet requirements

J = Estimated Value
 B = Analyte Found in Associated Method Blank
 N = Presumptive Evidence of a Compound
 * = Values outside of QC limits
 D = Dilution
 S = Indicates estimated value where valid five-point calibration was not performed prior to analyte detection in sample.
 () = Laboratory InHouse Limit

Data Path : Z:\pestpcbsrv\HPCHEM1\ECD_O\Data\P0070725\
 Data File : PO112076.D
 Signal(s) : Signal #1: ECD1A.ch Signal #2: ECD2B.ch
 Acq On : 08 Jul 2025 06:39
 Operator : YP/AJ
 Sample : Q2514-06
 Misc :
 ALS Vial : 40 Sample Multiplier: 1

Instrument :
 ECD_O
 ClientSampleId :
 TP-103

Manual Integrations
APPROVED

Reviewed By :Yogesh Patel 07/08/2025
 Supervised By :mohammad ahmed 07/09/2025

Integration File signal 1: autoint1.e
 Integration File signal 2: autoint2.e
 Quant Time: Jul 08 08:09:03 2025
 Quant Method : Z:\pestpcbsrv\HPCHEM1\ECD_O\methods\P0061125.M
 Quant Title : GC EXTRACTABLES
 QLast Update : Thu Jun 12 06:25:26 2025
 Response via : Initial Calibration
 Integrator: ChemStation

Volume Inj. : 2 µl
 Signal #1 Phase : ZB-MR1 Signal #2 Phase: ZB-MR2
 Signal #1 Info : 30Mx0.32mmx 0.50µ Signal #2 Info : 30M x 0.32mm x 0.25µm

Compound	RT#1	RT#2	Resp#1	Resp#2	ng/ml	ng/ml

System Monitoring Compounds						
1) SA Tetrachlo...	3.674	3.668	115.8E6	99085365	20.103	17.639
2) SA Decachlor...	8.703	8.649	65196222	23399295	12.420	13.163m

Target Compounds

(f)=RT Delta > 1/2 Window (#)=Amounts differ by > 25% (m)=manual int.

Data Path : Z:\pestpcbsrv\HPCHEM1\ECD_O\Data\PO070725\
Data File : PO112076.D
Signal(s) : Signal #1: ECD1A.ch Signal #2: ECD2B.ch
Acq On : 08 Jul 2025 06:39
Operator : YP/AJ
Sample : Q2514-06
Misc :
ALS Vial : 40 Sample Multiplier: 1



Instrument :
ECD_O
ClientSampleId :
TP-103


Manual Integrations
APPROVED

Reviewed By :Yogesh Patel 07/08/2025
Supervised By :mohammad ahmed 07/09/2025

Integration File signal 1: autoint1.e
Integration File signal 2: autoint2.e
Quant Time: Jul 08 08:09:03 2025
Quant Method : Z:\pestpcbsrv\HPCHEM1\ECD_O\methods\PO061125.M
Quant Title : GC EXTRACTABLES
QLast Update : Thu Jun 12 06:25:26 2025
Response via : Initial Calibration
Integrator: ChemStation

Volume Inj. : 2 µl
Signal #1 Phase : ZB-MR1 Signal #2 Phase: ZB-MR2
Signal #1 Info : 30Mx0.32mmx 0.50µ Signal #2 Info : 30M x 0.32mm x 0.25µm






#1 Tetrachloro-m-xylene

R.T.: 3.674 min
Delta R.T.: -0.003 min
Response: 115780217
Conc: 20.10 ng/ml

Instrument :
ECD_O
ClientSampleId :
TP-103


Manual Integrations
APPROVED

Reviewed By :Yogesh Patel 07/08/2025
Supervised By :mohammad ahmed 07/09/2025




#1 Tetrachloro-m-xylene

R.T.: 3.668 min
Delta R.T.: -0.006 min
Response: 99085365
Conc: 17.64 ng/ml



#2 Decachlorobiphenyl

R.T.: 8.703 min
Delta R.T.: -0.009 min
Response: 65196222
Conc: 12.42 ng/ml



#2 Decachlorobiphenyl

R.T.: 8.649 min
Delta R.T.: -0.012 min
Response: 23399295
Conc: 13.16 ng/ml m

Report of Analysis

Client:	CDM Smith	Date Collected:	07/03/25
Project:	South River WM Replacement	Date Received:	07/03/25
Client Sample ID:	TP-36	SDG No.:	Q2514
Lab Sample ID:	Q2514-07	Matrix:	SOIL
Analytical Method:	8082A	% Solid:	90.3
Sample Wt/Vol:	30.06	Units:	g
Soil Aliquot Vol:			uL
Extraction Type:		Test:	PCB
GPC Factor :	1.0	PH :	
Prep Method :	SW3541B	Injection Volume :	

File ID/Qc Batch:	Dilution:	Prep Date	Date Analyzed	Prep Batch ID
PO112077.D	1	07/07/25 08:30	07/08/25 06:58	PB168735

CAS Number	Parameter	Conc.	Qualifier	MDL	LOQ / CRQL	Units(Dry Weight)
TARGETS						
12674-11-2	Aroclor-1016	4.40	U	4.40	18.8	ug/kg
11104-28-2	Aroclor-1221	4.50	U	4.50	18.8	ug/kg
11141-16-5	Aroclor-1232	4.10	U	4.10	18.8	ug/kg
53469-21-9	Aroclor-1242	4.40	U	4.40	18.8	ug/kg
12672-29-6	Aroclor-1248	6.50	U	6.50	18.8	ug/kg
11097-69-1	Aroclor-1254	3.50	U	3.50	18.8	ug/kg
37324-23-5	Aroclor-1262	5.50	U	5.50	18.8	ug/kg
11100-14-4	Aroclor-1268	4.00	U	4.00	18.8	ug/kg
11096-82-5	Aroclor-1260	3.60	U	3.60	18.8	ug/kg
SURROGATES						
877-09-8	Tetrachloro-m-xylene	21.6		32 - 144	108%	SPK: 20
2051-24-3	Decachlorobiphenyl	15.7		32 - 175	79%	SPK: 20

Comments:

U = Not Detected

LOQ = Limit of Quantitation

MDL = Method Detection Limit

LOD = Limit of Detection

E = Value Exceeds Calibration Range

P = Indicates >25% difference for detected concentrations between the two GC columns

Q = indicates LCS control criteria did not meet requirements

M = MS/MSD acceptance criteria did not meet requirements

J = Estimated Value

B = Analyte Found in Associated Method Blank

N = Presumptive Evidence of a Compound

* = Values outside of QC limits

D = Dilution

S = Indicates estimated value where valid five-point calibration was not performed prior to analyte detection in sample.

() = Laboratory InHouse Limit

Data Path : Z:\pestpcbsrv\HPCHEM1\ECD_O\Data\P0070725\
 Data File : PO112077.D
 Signal(s) : Signal #1: ECD1A.ch Signal #2: ECD2B.ch
 Acq On : 08 Jul 2025 06:58
 Operator : YP/AJ
 Sample : Q2514-07
 Misc :
 ALS Vial : 41 Sample Multiplier: 1

Instrument :
 ECD_O
 ClientSampleId :
 TP-36

Manual Integrations
APPROVED

Reviewed By :Yogesh Patel 07/08/2025
 Supervised By :mohammad ahmed 07/09/2025

Integration File signal 1: autoint1.e
 Integration File signal 2: autoint2.e
 Quant Time: Jul 08 08:09:20 2025
 Quant Method : Z:\pestpcbsrv\HPCHEM1\ECD_O\methods\P0061125.M
 Quant Title : GC EXTRACTABLES
 QLast Update : Thu Jun 12 06:25:26 2025
 Response via : Initial Calibration
 Integrator: ChemStation

Volume Inj. : 2 µl
 Signal #1 Phase : ZB-MR1 Signal #2 Phase: ZB-MR2
 Signal #1 Info : 30Mx0.32mmx 0.50µ Signal #2 Info : 30M x 0.32mm x 0.25µm

Compound	RT#1	RT#2	Resp#1	Resp#2	ng/ml	ng/ml

System Monitoring Compounds						
1) SA Tetrachlo...	3.673	3.667	124.6E6	106.1E6	21.627	18.894
2) SA Decachlor...	8.704	8.648	80636043	27949117	15.361	15.722m

Target Compounds

(f)=RT Delta > 1/2 Window (#)=Amounts differ by > 25% (m)=manual int.

Data Path : Z:\pestpcbsrv\HPCHEM1\ECD_O\Data\PO070725\
Data File : PO112077.D
Signal(s) : Signal #1: ECD1A.ch Signal #2: ECD2B.ch
Acq On : 08 Jul 2025 06:58
Operator : YP/AJ
Sample : Q2514-07
Misc :
ALS Vial : 41 Sample Multiplier: 1



Instrument :
ECD_O
ClientSampleId :
TP-36


Manual Integrations
APPROVED

Reviewed By :Yogesh Patel 07/08/2025
Supervised By :mohammad ahmed 07/09/2025

Integration File signal 1: autoint1.e
Integration File signal 2: autoint2.e
Quant Time: Jul 08 08:09:20 2025
Quant Method : Z:\pestpcbsrv\HPCHEM1\ECD_O\methods\PO061125.M
Quant Title : GC EXTRACTABLES
QLast Update : Thu Jun 12 06:25:26 2025
Response via : Initial Calibration
Integrator: ChemStation

Volume Inj. : 2 µl
Signal #1 Phase : ZB-MR1 Signal #2 Phase: ZB-MR2
Signal #1 Info : 30Mx0.32mmx 0.50µ Signal #2 Info : 30M x 0.32mm x 0.25µm






#1 Tetrachloro-m-xylene

R.T.: 3.673 min
Delta R.T.: -0.004 min
Response: 124556125
Conc: 21.63 ng/ml

Instrument :
ECD_O
ClientSampleId :
TP-36


Manual Integrations
APPROVED

Reviewed By :Yogesh Patel 07/08/2025
Supervised By :mohammad ahmed 07/09/2025




#1 Tetrachloro-m-xylene

R.T.: 3.667 min
Delta R.T.: -0.007 min
Response: 106133325
Conc: 18.89 ng/ml



#2 Decachlorobiphenyl

R.T.: 8.704 min
Delta R.T.: -0.008 min
Response: 80636043
Conc: 15.36 ng/ml



#2 Decachlorobiphenyl

R.T.: 8.648 min
Delta R.T.: -0.013 min
Response: 27949117
Conc: 15.72 ng/ml m

Report of Analysis

Client:	CDM Smith	Date Collected:	07/03/25
Project:	South River WM Replacement	Date Received:	07/03/25
Client Sample ID:	TP-78	SDG No.:	Q2514
Lab Sample ID:	Q2514-08	Matrix:	SOIL
Analytical Method:	8082A	% Solid:	86.3
Sample Wt/Vol:	30.09	Units:	g
Soil Aliquot Vol:			uL
Extraction Type:		Test:	PCB
GPC Factor :	1.0	Injection Volume :	
Prep Method :	SW3541B		

File ID/Qc Batch:	Dilution:	Prep Date	Date Analyzed	Prep Batch ID
PO112078.D	1	07/07/25 08:30	07/08/25 07:16	PB168735

CAS Number	Parameter	Conc.	Qualifier	MDL	LOQ / CRQL	Units(Dry Weight)
TARGETS						
12674-11-2	Aroclor-1016	4.60	U	4.60	19.6	ug/kg
11104-28-2	Aroclor-1221	4.70	U	4.70	19.6	ug/kg
11141-16-5	Aroclor-1232	4.30	U	4.30	19.6	ug/kg
53469-21-9	Aroclor-1242	4.60	U	4.60	19.6	ug/kg
12672-29-6	Aroclor-1248	6.80	U	6.80	19.6	ug/kg
11097-69-1	Aroclor-1254	3.70	U	3.70	19.6	ug/kg
37324-23-5	Aroclor-1262	5.80	U	5.80	19.6	ug/kg
11100-14-4	Aroclor-1268	4.20	U	4.20	19.6	ug/kg
11096-82-5	Aroclor-1260	3.70	U	3.70	19.6	ug/kg
SURROGATES						
877-09-8	Tetrachloro-m-xylene	18.9		32 - 144	94%	SPK: 20
2051-24-3	Decachlorobiphenyl	10.9		32 - 175	54%	SPK: 20

Comments:

U = Not Detected

LOQ = Limit of Quantitation

MDL = Method Detection Limit

LOD = Limit of Detection

E = Value Exceeds Calibration Range

P = Indicates >25% difference for detected concentrations between the two GC columns

Q = indicates LCS control criteria did not meet requirements

M = MS/MSD acceptance criteria did not meet requirements

J = Estimated Value

B = Analyte Found in Associated Method Blank

N = Presumptive Evidence of a Compound

* = Values outside of QC limits

D = Dilution

S = Indicates estimated value where valid five-point calibration was not performed prior to analyte detection in sample.

() = Laboratory InHouse Limit

Data Path : Z:\pestpcbsrv\HPCHEM1\ECD_O\Data\P0070725\
 Data File : PO112078.D
 Signal(s) : Signal #1: ECD1A.ch Signal #2: ECD2B.ch
 Acq On : 08 Jul 2025 07:16
 Operator : YP/AJ
 Sample : Q2514-08
 Misc :
 ALS Vial : 42 Sample Multiplier: 1

Instrument :
 ECD_O
 ClientSampleId :
 TP-78

Manual Integrations
APPROVED

Reviewed By :Yogesh Patel 07/08/2025
 Supervised By :mohammad ahmed 07/09/2025

Integration File signal 1: autoint1.e
 Integration File signal 2: autoint2.e
 Quant Time: Jul 08 08:09:37 2025
 Quant Method : Z:\pestpcbsrv\HPCHEM1\ECD_O\methods\P0061125.M
 Quant Title : GC EXTRACTABLES
 QLast Update : Thu Jun 12 06:25:26 2025
 Response via : Initial Calibration
 Integrator: ChemStation

Volume Inj. : 2 µl
 Signal #1 Phase : ZB-MR1 Signal #2 Phase: ZB-MR2
 Signal #1 Info : 30Mx0.32mmx 0.50µ Signal #2 Info : 30M x 0.32mm x 0.25µm

Compound	RT#1	RT#2	Resp#1	Resp#2	ng/ml	ng/ml

System Monitoring Compounds						
1) SA Tetrachlo...	3.674	3.667	108.8E6	93837297	18.892	16.705
2) SA Decachlor...	8.704	8.650	57070448	19339622	10.872	10.879m

Target Compounds

(f)=RT Delta > 1/2 Window (#)=Amounts differ by > 25% (m)=manual int.

Data Path : Z:\pestpcbsrv\HPCHEM1\ECD_O\Data\PO070725\
Data File : PO112078.D
Signal(s) : Signal #1: ECD1A.ch Signal #2: ECD2B.ch
Acq On : 08 Jul 2025 07:16
Operator : YP/AJ
Sample : Q2514-08
Misc :
ALS Vial : 42 Sample Multiplier: 1



Instrument :
ECD_O
ClientSampleId :
TP-78


Manual Integrations
APPROVED

Reviewed By :Yogesh Patel 07/08/2025
Supervised By :mohammad ahmed 07/09/2025

Integration File signal 1: autoint1.e
Integration File signal 2: autoint2.e
Quant Time: Jul 08 08:09:37 2025
Quant Method : Z:\pestpcbsrv\HPCHEM1\ECD_O\methods\PO061125.M
Quant Title : GC EXTRACTABLES
QLast Update : Thu Jun 12 06:25:26 2025
Response via : Initial Calibration
Integrator: ChemStation

Volume Inj. : 2 µl
Signal #1 Phase : ZB-MR1 Signal #2 Phase: ZB-MR2
Signal #1 Info : 30Mx0.32mmx 0.50µm Signal #2 Info : 30M x 0.32mm x 0.25µm






#1 Tetrachloro-m-xylene

R.T.: 3.674 min
Delta R.T.: -0.003 min
Response: 108805608
Conc: 18.89 ng/ml

Instrument :
ECD_O
ClientSampleId :
TP-78


Manual Integrations
APPROVED

Reviewed By :Yogesh Patel 07/08/2025
Supervised By :mohammad ahmed 07/09/2025




#1 Tetrachloro-m-xylene

R.T.: 3.667 min
Delta R.T.: -0.006 min
Response: 93837297
Conc: 16.71 ng/ml



#2 Decachlorobiphenyl

R.T.: 8.704 min
Delta R.T.: -0.009 min
Response: 57070448
Conc: 10.87 ng/ml



#2 Decachlorobiphenyl

R.T.: 8.650 min
Delta R.T.: -0.011 min
Response: 19339622
Conc: 10.88 ng/ml m

Report of Analysis

Client:	CDM Smith	Date Collected:	07/03/25
Project:	South River WM Replacement	Date Received:	07/03/25
Client Sample ID:	TP-81	SDG No.:	Q2514
Lab Sample ID:	Q2514-09	Matrix:	SOIL
Analytical Method:	8082A	% Solid:	86.3
Sample Wt/Vol:	30.02	Units:	g
Soil Aliquot Vol:			uL
Extraction Type:		Test:	PCB
GPC Factor :	1.0	PH :	
Prep Method :	SW3541B	Injection Volume :	

File ID/Qc Batch:	Dilution:	Prep Date	Date Analyzed	Prep Batch ID
PO112079.D	1	07/07/25 08:30	07/08/25 07:35	PB168735

CAS Number	Parameter	Conc.	Qualifier	MDL	LOQ / CRQL	Units(Dry Weight)
TARGETS						
12674-11-2	Aroclor-1016	4.60	U	4.60	19.7	ug/kg
11104-28-2	Aroclor-1221	4.70	U	4.70	19.7	ug/kg
11141-16-5	Aroclor-1232	4.30	U	4.30	19.7	ug/kg
53469-21-9	Aroclor-1242	4.60	U	4.60	19.7	ug/kg
12672-29-6	Aroclor-1248	6.90	U	6.90	19.7	ug/kg
11097-69-1	Aroclor-1254	3.70	U	3.70	19.7	ug/kg
37324-23-5	Aroclor-1262	5.80	U	5.80	19.7	ug/kg
11100-14-4	Aroclor-1268	4.20	U	4.20	19.7	ug/kg
11096-82-5	Aroclor-1260	3.70	U	3.70	19.7	ug/kg
SURROGATES						
877-09-8	Tetrachloro-m-xylene	19.4		32 - 144	97%	SPK: 20
2051-24-3	Decachlorobiphenyl	10.0		32 - 175	50%	SPK: 20

Comments:

U = Not Detected

LOQ = Limit of Quantitation

MDL = Method Detection Limit

LOD = Limit of Detection

E = Value Exceeds Calibration Range

P = Indicates >25% difference for detected concentrations between the two GC columns

Q = indicates LCS control criteria did not meet requirements

M = MS/MSD acceptance criteria did not meet requirements

J = Estimated Value

B = Analyte Found in Associated Method Blank

N = Presumptive Evidence of a Compound

* = Values outside of QC limits

D = Dilution

S = Indicates estimated value where valid five-point calibration was not performed prior to analyte detection in sample.

() = Laboratory InHouse Limit

Data Path : Z:\pestpcbsrv\HPCHEM1\ECD_O\Data\P0070725\
 Data File : PO112079.D
 Signal(s) : Signal #1: ECD1A.ch Signal #2: ECD2B.ch
 Acq On : 08 Jul 2025 07:35
 Operator : YP/AJ
 Sample : Q2514-09
 Misc :
 ALS Vial : 43 Sample Multiplier: 1

Instrument :
 ECD_O
 ClientSampleId :
 TP-81

Manual Integrations
APPROVED

Reviewed By :Yogesh Patel 07/08/2025
 Supervised By :mohammad ahmed 07/09/2025

Integration File signal 1: autoint1.e
 Integration File signal 2: autoint2.e
 Quant Time: Jul 08 08:09:54 2025
 Quant Method : Z:\pestpcbsrv\HPCHEM1\ECD_O\methods\P0061125.M
 Quant Title : GC EXTRACTABLES
 QLast Update : Thu Jun 12 06:25:26 2025
 Response via : Initial Calibration
 Integrator: ChemStation

Volume Inj. : 2 µl
 Signal #1 Phase : ZB-MR1 Signal #2 Phase: ZB-MR2
 Signal #1 Info : 30Mx0.32mmx 0.50µ Signal #2 Info : 30M x 0.32mm x 0.25µm

Compound	RT#1	RT#2	Resp#1	Resp#2	ng/ml	ng/ml

System Monitoring Compounds						
1) SA Tetrachlo...	3.674	3.667	111.9E6	96471903	19.429	17.174
2) SA Decachlor...	8.703	8.650	51680138	17844108	9.845	10.038m

Target Compounds

(f)=RT Delta > 1/2 Window (#)=Amounts differ by > 25% (m)=manual int.

Data Path : Z:\pestpcbsrv\HPCHEM1\ECD_O\Data\PO070725\
Data File : PO112079.D
Signal(s) : Signal #1: ECD1A.ch Signal #2: ECD2B.ch
Acq On : 08 Jul 2025 07:35
Operator : YP/AJ
Sample : Q2514-09
Misc :
ALS Vial : 43 Sample Multiplier: 1



Instrument :
ECD_O
ClientSampleId :
TP-81

Manual Integrations
APPROVED


Reviewed By :Yogesh Patel 07/08/2025
Supervised By :mohammad ahmed 07/09/2025

Integration File signal 1: autoint1.e
Integration File signal 2: autoint2.e
Quant Time: Jul 08 08:09:54 2025
Quant Method : Z:\pestpcbsrv\HPCHEM1\ECD_O\methods\PO061125.M
Quant Title : GC EXTRACTABLES
QLast Update : Thu Jun 12 06:25:26 2025
Response via : Initial Calibration
Integrator: ChemStation

Volume Inj. : 2 µl
Signal #1 Phase : ZB-MR1 Signal #2 Phase: ZB-MR2
Signal #1 Info : 30Mx0.32mmx 0.50µ Signal #2 Info : 30M x 0.32mm x 0.25µm



Response_ Signal: PO112079.D\ECD1A.ch



#1 Tetrachloro-m-xylene


R.T.: 3.674 min
Delta R.T.: -0.003 min
Response: 111902513
Conc: 19.43 ng/ml

Instrument :
ECD_O
ClientSampleId :
TP-81

Manual Integrations
APPROVED

Reviewed By :Yogesh Patel 07/08/2025
Supervised By :mohammad ahmed 07/09/2025


Response_ Signal: PO112079.D\ECD2B.ch



#1 Tetrachloro-m-xylene

R.T.: 3.667 min
Delta R.T.: -0.006 min
Response: 96471903
Conc: 17.17 ng/ml


Response_ Signal: PO112079.D\ECD1A.ch



#2 Decachlorobiphenyl

R.T.: 8.703 min
Delta R.T.: -0.009 min
Response: 51680138
Conc: 9.85 ng/ml

Response_ Signal: PO112079.D\ECD2B.ch



#2 Decachlorobiphenyl

R.T.: 8.650 min
Delta R.T.: -0.012 min
Response: 17844108
Conc: 10.04 ng/ml m

Report of Analysis

Client:	CDM Smith	Date Collected:	07/03/25
Project:	South River WM Replacement	Date Received:	07/03/25
Client Sample ID:	TP-90	SDG No.:	Q2514
Lab Sample ID:	Q2514-10	Matrix:	SOIL
Analytical Method:	8082A	% Solid:	91.6
Sample Wt/Vol:	30.04	Units:	g
Soil Aliquot Vol:			uL
Extraction Type:		Test:	PCB
GPC Factor :	1.0	PH :	
Prep Method :	SW3541B	Injection Volume :	

File ID/Qc Batch:	Dilution:	Prep Date	Date Analyzed	Prep Batch ID
PO112080.D	1	07/07/25 08:30	07/08/25 07:53	PB168735

CAS Number	Parameter	Conc.	Qualifier	MDL	LOQ / CRQL	Units(Dry Weight)
TARGETS						
12674-11-2	Aroclor-1016	4.30	U	4.30	18.5	ug/kg
11104-28-2	Aroclor-1221	4.40	U	4.40	18.5	ug/kg
11141-16-5	Aroclor-1232	4.10	U	4.10	18.5	ug/kg
53469-21-9	Aroclor-1242	4.40	U	4.40	18.5	ug/kg
12672-29-6	Aroclor-1248	6.50	U	6.50	18.5	ug/kg
11097-69-1	Aroclor-1254	3.50	U	3.50	18.5	ug/kg
37324-23-5	Aroclor-1262	5.50	U	5.50	18.5	ug/kg
11100-14-4	Aroclor-1268	3.90	U	3.90	18.5	ug/kg
11096-82-5	Aroclor-1260	3.50	U	3.50	18.5	ug/kg
SURROGATES						
877-09-8	Tetrachloro-m-xylene	22.2		32 - 144	111%	SPK: 20
2051-24-3	Decachlorobiphenyl	17.0		32 - 175	85%	SPK: 20

Comments:

U = Not Detected

LOQ = Limit of Quantitation

MDL = Method Detection Limit

LOD = Limit of Detection

E = Value Exceeds Calibration Range

P = Indicates >25% difference for detected concentrations between the two GC columns

Q = indicates LCS control criteria did not meet requirements

M = MS/MSD acceptance criteria did not meet requirements

J = Estimated Value

B = Analyte Found in Associated Method Blank

N = Presumptive Evidence of a Compound

* = Values outside of QC limits

D = Dilution

S = Indicates estimated value where valid five-point calibration was not performed prior to analyte detection in sample.

() = Laboratory InHouse Limit

Data Path : Z:\pestpcbsrv\HPCHEM1\ECD_O\Data\P0070725\
 Data File : PO112080.D
 Signal(s) : Signal #1: ECD1A.ch Signal #2: ECD2B.ch
 Acq On : 08 Jul 2025 07:53
 Operator : YP/AJ
 Sample : Q2514-10
 Misc :
 ALS Vial : 44 Sample Multiplier: 1

Instrument :
 ECD_O
 ClientSampleId :
 TP-90

Manual Integrations
APPROVED

Reviewed By :Yogesh Patel 07/08/2025
 Supervised By :mohammad ahmed 07/09/2025

Integration File signal 1: autoint1.e
 Integration File signal 2: autoint2.e
 Quant Time: Jul 08 08:54:26 2025
 Quant Method : Z:\pestpcbsrv\HPCHEM1\ECD_O\methods\P0061125.M
 Quant Title : GC EXTRACTABLES
 QLast Update : Thu Jun 12 06:25:26 2025
 Response via : Initial Calibration
 Integrator: ChemStation

Volume Inj. : 2 µl
 Signal #1 Phase : ZB-MR1 Signal #2 Phase: ZB-MR2
 Signal #1 Info : 30Mx0.32mmx 0.50µ Signal #2 Info : 30M x 0.32mm x 0.25µm

Compound	RT#1	RT#2	Resp#1	Resp#2	ng/ml	ng/ml

System Monitoring Compounds						
1) SA Tetrachlo...	3.674	3.667	127.9E6	110.1E6	22.210	19.605
2) SA Decachlor...	8.703	8.649	86672613	30217683	16.511	16.999m

Target Compounds

(f)=RT Delta > 1/2 Window (#)=Amounts differ by > 25% (m)=manual int.

Data Path : Z:\pestpcbsrv\HPCHEM1\ECD_O\Data\P0070725\
Data File : P0112080.D
Signal(s) : Signal #1: ECD1A.ch Signal #2: ECD2B.ch
Acq On : 08 Jul 2025 07:53
Operator : YP/AJ
Sample : Q2514-10
Misc :
ALS Vial : 44 Sample Multiplier: 1



Instrument :
ECD_O
ClientSampleId :
TP-90


Manual Integrations
APPROVED

Reviewed By :Yogesh Patel 07/08/2025
Supervised By :mohammad ahmed 07/09/2025

Integration File signal 1: autoint1.e
Integration File signal 2: autoint2.e
Quant Time: Jul 08 08:54:26 2025
Quant Method : Z:\pestpcbsrv\HPCHEM1\ECD_O\methods\P0061125.M
Quant Title : GC EXTRACTABLES
QLast Update : Thu Jun 12 06:25:26 2025
Response via : Initial Calibration
Integrator: ChemStation

Volume Inj. : 2 µl
Signal #1 Phase : ZB-MR1 Signal #2 Phase: ZB-MR2
Signal #1 Info : 30Mx0.32mmx 0.50µ Signal #2 Info : 30M x 0.32mm x 0.25µm






#1 Tetrachloro-m-xylene

R.T.: 3.674 min
Delta R.T.: -0.003 min
Response: 127917958
Conc: 22.21 ng/ml

Instrument :
ECD_O
ClientSampleId :
TP-90


Manual Integrations
APPROVED

Reviewed By :Yogesh Patel 07/08/2025
Supervised By :mohammad ahmed 07/09/2025




#1 Tetrachloro-m-xylene

R.T.: 3.667 min
Delta R.T.: -0.006 min
Response: 110126220
Conc: 19.60 ng/ml



#2 Decachlorobiphenyl

R.T.: 8.703 min
Delta R.T.: -0.009 min
Response: 86672613
Conc: 16.51 ng/ml



#2 Decachlorobiphenyl

R.T.: 8.649 min
Delta R.T.: -0.012 min
Response: 30217683
Conc: 17.00 ng/ml m



CALIBRATION SUMMARY

RETENTION TIMES OF INITIAL CALIBRATION

Lab Name:	<u>Alliance</u>	Contract:	<u>CAMP02</u>
Lab Code:	<u>ACE</u>	SDG NO.:	<u>Q2514</u>
Instrument ID:	<u>ECD_O</u>	Calibration Date(s):	<u>06/11/2025</u> <u>06/11/2025</u>
		Calibration Times:	<u>10:40</u> <u>19:07</u>

 GC Column: ZB-MR1 ID: 0.32 (mm)

LAB FILE ID:	RT 1000 = <u>PO111587.D</u>	RT 750 = <u>PO111588.D</u>
RT 500 = <u>PO111589.D</u>	RT 250 = <u>PO111590.D</u>	RT 050 = <u>PO111591.D</u>

COMPOUND		RT 1000	RT 750	RT 500	RT 250	RT 050	MEAN RT	RT WINDOW	
								FROM	TO
Aroclor-1016-1	(1)	4.77	4.77	4.77	4.77	4.77	4.77	4.67	4.87
Aroclor-1016-2	(2)	4.78	4.79	4.78	4.79	4.78	4.78	4.68	4.88
Aroclor-1016-3	(3)	4.84	4.84	4.84	4.84	4.84	4.84	4.74	4.94
Aroclor-1016-4	(4)	4.96	4.96	4.96	4.96	4.96	4.96	4.86	5.06
Aroclor-1016-5	(5)	5.22	5.22	5.22	5.22	5.22	5.22	5.12	5.32
Aroclor-1260-1	(1)	6.26	6.26	6.26	6.26	6.26	6.26	6.16	6.36
Aroclor-1260-2	(2)	6.45	6.45	6.45	6.45	6.45	6.45	6.35	6.55
Aroclor-1260-3	(3)	6.81	6.81	6.81	6.81	6.81	6.81	6.71	6.91
Aroclor-1260-4	(4)	7.07	7.07	7.07	7.07	7.07	7.07	6.97	7.17
Aroclor-1260-5	(5)	7.32	7.32	7.32	7.32	7.32	7.32	7.22	7.42
Decachlorobiphenyl		8.71	8.71	8.71	8.71	8.71	8.71	8.61	8.81
Tetrachloro-m-xylene		3.68	3.68	3.68	3.68	3.68	3.68	3.58	3.78
Aroclor-1242-1	(1)	4.77	4.77	4.77	4.76	4.77	4.77	4.67	4.87
Aroclor-1242-2	(2)	4.79	4.79	4.78	4.78	4.78	4.78	4.68	4.88
Aroclor-1242-3	(3)	4.84	4.84	4.84	4.84	4.84	4.84	4.74	4.94
Aroclor-1242-4	(4)	4.96	4.96	4.96	4.96	4.96	4.96	4.86	5.06
Aroclor-1242-5	(5)	5.61	5.61	5.61	5.61	5.61	5.61	5.51	5.71
Decachlorobiphenyl		8.71	8.71	8.71	8.71	8.71	8.71	8.61	8.81
Tetrachloro-m-xylene		3.68	3.68	3.68	3.68	3.68	3.68	3.58	3.78
Aroclor-1248-1	(1)	4.77	4.77	4.76	4.76	4.76	4.76	4.66	4.86
Aroclor-1248-2	(2)	5.01	5.00	5.00	5.00	5.00	5.00	4.90	5.10
Aroclor-1248-3	(3)	5.22	5.22	5.22	5.22	5.22	5.22	5.12	5.32
Aroclor-1248-4	(4)	5.57	5.57	5.57	5.57	5.57	5.57	5.47	5.67
Aroclor-1248-5	(5)	5.61	5.61	5.61	5.61	5.61	5.61	5.51	5.71
Decachlorobiphenyl		8.71	8.71	8.71	8.71	8.71	8.71	8.61	8.81
Tetrachloro-m-xylene		3.68	3.68	3.68	3.68	3.68	3.68	3.58	3.78
Aroclor-1254-1	(1)	5.57	5.57	5.57	5.57	5.57	5.57	5.47	5.67
Aroclor-1254-2	(2)	5.72	5.72	5.72	5.72	5.72	5.72	5.62	5.82
Aroclor-1254-3	(3)	6.12	6.13	6.13	6.13	6.12	6.13	6.03	6.23
Aroclor-1254-4	(4)	6.35	6.36	6.36	6.35	6.35	6.35	6.25	6.45
Aroclor-1254-5	(5)	6.77	6.78	6.78	6.77	6.77	6.77	6.67	6.87
Decachlorobiphenyl		8.71	8.71	8.71	8.71	8.71	8.71	8.61	8.81
Tetrachloro-m-xylene		3.68	3.68	3.68	3.68	3.68	3.68	3.58	3.78
Aroclor-1268-1	(1)	7.60	7.60	7.60	7.60	7.60	7.60	7.50	7.70
Aroclor-1268-2	(2)	7.67	7.67	7.67	7.67	7.66	7.67	7.57	7.77
Aroclor-1268-3	(3)	7.87	7.87	7.87	7.87	7.87	7.87	7.77	7.97
Aroclor-1268-4	(4)	8.16	8.16	8.16	8.16	8.16	8.16	8.06	8.26
Aroclor-1268-5	(5)	8.46	8.46	8.46	8.46	8.45	8.46	8.36	8.56

[illegible]

Calibration Times: **10:40** **19:07**

LAB FILE ID: RT 1000 = PO111587.D RT 750 = PO111588.D
RT 500 = PO111589.D RT 250 = PO111590.D RT 050 = PO111591.D

[illegible]

[illegible]



284 Sheffield Street, Mountainside, New Jersey 07092, Phone : 908 789 8900,
Fax : 908 789 8922

CALIBRATION FACTOR OF INITIAL CALIBRATION

Lab Name: Alliance
Lab Code: ACE
Instrument ID: ECD_O

Contract: CAMP02
SDG NO.: Q2514

Calibration Date(s): 06/11/2025 06/11/2025
Calibration Times: 10:40 19:07

GC Column: ZB-MR1 ID: 0.32 (mm)

LAB FILE ID:		CF 1000 = <u>PO111587.D</u>		CF 750 = <u>PO111588.D</u>			
CF 500 = <u>PO111589.D</u>		CF 250 = <u>PO111590.D</u>		CF 050 = <u>PO111591.D</u>			
COMPOUND		CF 1000	CF 750	CF 500	CF 250	CF 050	% RSD
Aroclor-1016-1	(1)	220511854	230179169	244104994	255923844	242047420	238553456
Aroclor-1016-2	(2)	319525390	333752768	345607420	360231916	304600680	332743635
Aroclor-1016-3	(3)	218699222	228973452	241664208	255477324	205889780	230140797
Aroclor-1016-4	(4)	174886969	182974663	191580278	204290796	173581000	185462741
Aroclor-1016-5	(5)	177464979	184697736	193390828	207039800	170624200	186643509
Aroclor-1260-1	(1)	321794228	337454645	358028956	426721964	332609080	355321775
Aroclor-1260-2	(2)	433743230	452724105	474832224	506755684	494307920	472472633
Aroclor-1260-3	(3)	393522712	410790572	429624930	441759844	447671720	424673956
Aroclor-1260-4	(4)	282329615	299358140	315388464	327883556	347275780	314447111
Aroclor-1260-5	(5)	781713901	810719289	832144010	867226572	826835640	823727882
Decachlorobiphenyl		4901811500	5109294200	5313242720	5581051480	5340821400	5249244260
Tetrachloro-m-xylene		5558158730	5743696387	5927517640	6035511000	5532196200	5759415991
Aroclor-1242-1	(1)	189891715	199375873	208390472	218835528	190726000	201443918
Aroclor-1242-2	(2)	269264014	283815699	295038528	310860236	249854560	281766607
Aroclor-1242-3	(3)	185208795	198314185	205655032	216442696	188497160	198823574
Aroclor-1242-4	(4)	148912444	156458187	164284520	169958380	151372200	158197146
Aroclor-1242-5	(5)	154468521	166978497	176155118	183078792	155400820	167216350
Decachlorobiphenyl		4880711810	5081655560	5262786220	5522085600	4928788200	5135205478
Tetrachloro-m-xylene		5456812090	5681549133	5847202300	6031977960	5214216400	5646351577
Aroclor-1248-1	(1)	145884945	152954188	162370132	170480356	174637940	161265512
Aroclor-1248-2	(2)	193392956	203021513	214142790	221601180	245868320	215605352
Aroclor-1248-3	(3)	243691383	258111092	270734920	294861996	281294320	269738742
Aroclor-1248-4	(4)	355904013	376445483	398411334	426549232	441707040	399803420
Aroclor-1248-5	(5)	249802900	267040200	283633312	302535740	301041360	280810702
Decachlorobiphenyl		4929616300	5142061960	5420744260	5685225920	6003686400	5436266968
Tetrachloro-m-xylene		5626812480	5847124520	6076407960	6245856240	6492772200	6057794680
Aroclor-1254-1	(1)	375291496	390445984	409053510	418326128	532236220	425070668
Aroclor-1254-2	(2)	334554268	347730616	365467502	374564136	475325660	379528436
Aroclor-1254-3	(3)	528273778	541799637	566990470	566909168	652766900	571347991
Aroclor-1254-4	(4)	330811190	344690741	356705876	353681912	358524400	348882824
Aroclor-1254-5	(5)	481176015	491854427	511981248	504869196	570769000	512129977
Decachlorobiphenyl		5009210160	5167761240	5403649020	5380598200	6191612600	5430566244
Tetrachloro-m-xylene		5724223470	6030704747	6021777640	5926293200	6653717200	6071343251
Aroclor-1268-1	(1)	1033906302	1058286327	1096846802	1155194040	1288627080	1126572110



284 Sheffield Street, Mountainside, New Jersey 07092, Phone : 908 789 8900,
Fax : 908 789 8922

CALIBRATION FACTOR OF INITIAL CALIBRATION

Aroclor-1268-2	(2)	906036168	918687719	954310100	989229764	1117888140	977230378	9
Aroclor-1268-3	(3)	760320035	776861388	800241936	843634592	917077000	819626990	8
Aroclor-1268-4	(4)	324679419	326739527	344775706	356707332	367012100	343982817	5
Aroclor-1268-5	(5)	2115204224	2135729147	2187245170	2254511260	2409521900	2220442340	5
Decachlorobiphenyl		8820062120	9005455773	9339565760	9886719320	10604055600	9531171715	8
Tetrachloro-m-xylene		5944441710	6079157480	6269391520	6520960080	6795679200	6321925998	5



284 Sheffield Street, Mountainside, New Jersey 07092, Phone : 908 789 8900,
Fax : 908 789 8922

CALIBRATION FACTOR OF INITIAL CALIBRATION

Lab Name: Alliance
Lab Code: ACE
Instrument ID: ECD_O

Contract: CAMP02
SDG NO.: Q2514

Calibration Date(s): 06/11/2025 06/11/2025
Calibration Times: 10:40 19:07

GC Column: ZB-MR2 ID: 0.32 (mm)

LAB FILE ID:		CF 1000 = <u>PO111587.D</u>		CF 750 = <u>PO111588.D</u>			
CF 500 = <u>PO111589.D</u>		CF 250 = <u>PO111590.D</u>		CF 050 = <u>PO111591.D</u>			
COMPOUND		CF 1000	CF 750	CF 500	CF 250	CF 050	% RSD
Aroclor-1016-1	(1)	180839726	189231809	196263600	204676936	212823680	6
Aroclor-1016-2	(2)	269525798	277038755	287808724	296313292	293716580	4
Aroclor-1016-3	(3)	141401189	146814288	152988496	158088464	151904200	4
Aroclor-1016-4	(4)	112384684	117918833	124672358	128074364	125886140	5
Aroclor-1016-5	(5)	145637432	151733320	158595618	164650128	175375940	7
Aroclor-1260-1	(1)	232933493	242290633	253160914	265400132	260795120	5
Aroclor-1260-2	(2)	272930024	283658681	293882326	306210008	313403160	6
Aroclor-1260-3	(3)	248293091	256892847	265761412	279567532	285382000	6
Aroclor-1260-4	(4)	171150736	180595663	190038668	201212768	207483680	8
Aroclor-1260-5	(5)	395557404	413148045	424269490	442012232	450112560	5
Decachlorobiphenyl		1664373530	1742647173	1808169220	1884747160	1788371000	5
Tetrachloro-m-xylene		5585912110	5701804613	5792994800	5795065440	5210609600	4
Aroclor-1242-1	(1)	152168573	159777101	166064310	176069720	159718720	5
Aroclor-1242-2	(2)	225834322	234974032	243469856	251904876	224121580	5
Aroclor-1242-3	(3)	119492390	123624105	129191264	135018820	120620420	5
Aroclor-1242-4	(4)	114585972	119959547	125650318	133140376	119457260	6
Aroclor-1242-5	(5)	141763911	149312545	155170630	161394704	141804460	6
Decachlorobiphenyl		1628957800	1697321880	1768741900	1847187720	1627451400	6
Tetrachloro-m-xylene		5447566380	5592119960	5668881340	5704776560	4755587600	7
Aroclor-1248-1	(1)	118834435	122872489	131279096	140544936	152890960	10
Aroclor-1248-2	(2)	160429628	168292228	179033076	186589356	196432880	8
Aroclor-1248-3	(3)	169865875	177833136	187951664	198291792	213246720	9
Aroclor-1248-4	(4)	200486745	209433661	222430720	237578848	271687720	12
Aroclor-1248-5	(5)	195830172	205941008	216128424	227994268	242930420	8
Decachlorobiphenyl		1644850390	1717850333	1805319940	1883765920	1992763800	8
Tetrachloro-m-xylene		5458685310	5581768067	5723793620	5741246280	5688025000	2
Aroclor-1254-1	(1)	296936934	303927659	319123596	320561628	374491800	9
Aroclor-1254-2	(2)	255453905	262352463	275811104	277251804	325698480	10
Aroclor-1254-3	(3)	392851690	400205872	416928596	411547332	473734340	8
Aroclor-1254-4	(4)	214635691	222689013	230857806	231773900	239050800	4
Aroclor-1254-5	(5)	301414788	308036695	323788090	319823096	374889540	9
Decachlorobiphenyl		1681329490	1723219680	1812787460	1794853560	2027152800	7
Tetrachloro-m-xylene		5554099840	5630982267	5714907100	5492046840	5804310800	2
Aroclor-1268-1	(1)	439189744	444186417	457911336	484649164	566087560	11



284 Sheffield Street, Mountainside, New Jersey 07092, Phone : 908 789 8900,
Fax : 908 789 8922

CALIBRATION FACTOR OF INITIAL CALIBRATION

Aroclor-1268-2	(2)	385991375	390586096	399904992	422505028	483522600	416502018	10
Aroclor-1268-3	(3)	289581994	294897103	304525684	324673600	371079040	316951484	10
Aroclor-1268-4	(4)	108353442	109112276	112870114	119224140	128726300	115657254	7
Aroclor-1268-5	(5)	686403257	694858911	706410288	735443232	800112960	724645730	6
Decachlorobiphenyl		2944164530	2997882747	3096344980	3278501400	3473378600	3158054451	7
Tetrachloro-m-xylene		5557311950	5631320067	5696482640	5807035360	5718379000	5682105803	2



284 Sheffield Street, Mountainside, New Jersey 07092, Phone : 908 789 8900,
Fax : 908 789 8922

INITIAL CALIBRATION OF MULTICOMPONENT ANALYTES

Lab Name: Alliance **Contract:** CAMP02
Lab Code: ACE **SDG NO.:** Q2514
Instrument ID: ECD_O **Date(s) Analyzed:** 06/11/2025 06/11/2025
GC Column: ZB-MR1 **ID:** 0.32 (mm)

COMPOUND	AMOUNT (ng)	PEAK	RT	RT WINDOW		CALIBRATION FACTOR
				FROM	TO	
Aroclor-1221	500	1	3.89	3.79	3.99	87078800
		2	3.98	3.88	4.08	64167800
		3	4.05	3.95	4.15	195036000
		4	0.00			0
		5	0.00			0
Aroclor-1232	500	1	4.05	3.95	4.15	151749000
		2	4.54	4.44	4.64	86221200
		3	4.79	4.69	4.89	162881000
		4	4.96	4.86	5.06	87019200
		5	5.00	4.90	5.10	56580600
Aroclor-1262	500	1	6.82	6.72	6.92	609876000
		2	7.32	7.22	7.42	941402000
		3	7.60	7.50	7.70	396622000
		4	7.66	7.56	7.76	657002000
		5	8.16	8.06	8.26	304668000



284 Sheffield Street, Mountainside, New Jersey 07092, Phone : 908 789 8900,
Fax : 908 789 8922

INITIAL CALIBRATION OF MULTICOMPONENT ANALYTES

Lab Name: Alliance **Contract:** CAMP02
Lab Code: ACE **SDG NO.:** Q2514
Instrument ID: ECD_O **Date(s) Analyzed:** 06/11/2025 06/11/2025
GC Column: ZB-MR2 **ID:** 0.32 (mm)

COMPOUND	AMOUNT (ng)	PEAK	RT	RT WINDOW		CALIBRATION FACTOR
				FROM	TO	
Aroclor-1221	500	1	3.88	3.78	3.98	74673800
		2	3.97	3.87	4.07	55948800
		3	4.04	3.94	4.14	171830000
		4	0.00			0
		5	0.00			0
Aroclor-1232	500	1	4.04	3.94	4.14	132990000
		2	4.77	4.67	4.87	131791000
		3	4.94	4.84	5.04	69118200
		4	5.03	4.93	5.13	60484000
		5	5.20	5.10	5.30	66414600
Aroclor-1262	500	1	6.78	6.68	6.88	337664000
		2	7.28	7.18	7.38	465774000
		3	7.56	7.46	7.66	166285000
		4	7.63	7.53	7.73	277070000
		5	8.12	8.02	8.22	101619000

Data Path : Z:\pestpcbsrv\HPCHEM1\ECD_O\Data\P0061125\
 Data File : P0111587.D
 Signal(s) : Signal #1: ECD1A.ch Signal #2: ECD2B.ch
 Acq On : 11 Jun 2025 10:40
 Operator : YP/AJ
 Sample : AR1660ICC1000
 Misc :
 ALS Vial : 3 Sample Multiplier: 1

Instrument :
 ECD_O
 ClientSampleId :
 AR1660ICC1000

Integration File signal 1: autoint1.e
 Integration File signal 2: autoint2.e
 Quant Time: Jun 11 11:47:01 2025
 Quant Method : Z:\pestpcbsrv\HPCHEM1\ECD_O\methods\P0061125.M
 Quant Title : GC EXTRACTABLES
 QLast Update : Wed Jun 11 11:43:51 2025
 Response via : Initial Calibration
 Integrator: ChemStation

Volume Inj. : 2 µl
 Signal #1 Phase : ZB-MR1 Signal #2 Phase: ZB-MR2
 Signal #1 Info : 30Mx0.32mmx 0.50µ Signal #2 Info : 30M x 0.32mm x 0.25µm

	Compound	RT#1	RT#2	Resp#1	Resp#2	ng/ml	ng/ml

System Monitoring Compounds							
1)	SA Tetrachlo...	3.678	3.674	555.8E6	558.6E6	96.784	98.180
2)	SA Decachlor...	8.713	8.661	490.2E6	166.4E6	95.972	95.859
Target Compounds							
3)	L1 AR-1016-1	4.766	4.751	220.5E6	180.8E6	949.220	959.099
4)	L1 AR-1016-2	4.784	4.769	319.5E6	269.5E6	960.787	967.196
5)	L1 AR-1016-3	4.842	4.944	218.7E6	141.4E6	950.116	960.640
6)	L1 AR-1016-4	4.961	4.987	174.9E6	112.4E6	954.448	948.166
7)	L1 AR-1016-5	5.219	5.199	177.5E6	145.6E6	957.056	957.407
31)	L7 AR-1260-1	6.257	6.229	321.8E6	232.9E6	946.700	958.388
32)	L7 AR-1260-2	6.447	6.417	433.7E6	272.9E6	954.776	963.035
33)	L7 AR-1260-3	6.814	6.569	393.5E6	248.3E6	956.141	966.019
34)	L7 AR-1260-4	7.074	7.040	282.3E6	171.2E6	944.692	947.706
35)	L7 AR-1260-5	7.317	7.281	781.7E6	395.6E6	968.752	964.978



(f)=RT Delta > 1/2 Window (#)=Amounts differ by > 25% (m)=manual int.

Data Path : Z:\pestpcbsrv\HPCHEM1\ECD_O\Data\PO061125\
Data File : PO111587.D
Signal(s) : Signal #1: ECD1A.ch Signal #2: ECD2B.ch
Acq On : 11 Jun 2025 10:40
Operator : YP/AJ
Sample : AR1660ICC1000
Misc :
ALS Vial : 3 Sample Multiplier: 1

Instrument :
ECD_O
ClientSampleId :
AR1660ICC1000

Integration File signal 1: autoint1.e
Integration File signal 2: autoint2.e
Quant Time: Jun 11 11:47:01 2025
Quant Method : Z:\pestpcbsrv\HPCHEM1\ECD_O\methods\PO061125.M
Quant Title : GC EXTRACTABLES
QLast Update : Wed Jun 11 11:43:51 2025
Response via : Initial Calibration
Integrator: ChemStation

Volume Inj. : 2 µl
Signal #1 Phase : ZB-MR1 Signal #2 Phase: ZB-MR2
Signal #1 Info : 30Mx0.32mmx 0.50µm Signal #2 Info : 30M x 0.32mm x 0.25µm



Data Path : Z:\pestpcbsrv\HPCHEM1\ECD_O\Data\P0061125\
 Data File : P0111588.D
 Signal(s) : Signal #1: ECD1A.ch Signal #2: ECD2B.ch
 Acq On : 11 Jun 2025 10:58
 Operator : YP/AJ
 Sample : AR1660ICC750
 Misc :
 ALS Vial : 4 Sample Multiplier: 1

Instrument :
 ECD_O
 ClientSampleId :
 AR1660ICC750

Integration File signal 1: autoint1.e
 Integration File signal 2: autoint2.e
 Quant Time: Jun 11 11:49:22 2025
 Quant Method : Z:\pestpcbsrv\HPCHEM1\ECD_O\methods\P0061125.M
 Quant Title : GC EXTRACTABLES
 QLast Update : Wed Jun 11 11:43:51 2025
 Response via : Initial Calibration
 Integrator: ChemStation

Volume Inj. : 2 µl
 Signal #1 Phase : ZB-MR1 Signal #2 Phase: ZB-MR2
 Signal #1 Info : 30Mx0.32mmx 0.50µ Signal #2 Info : 30M x 0.32mm x 0.25µm

Compound	RT#1	RT#2	Resp#1	Resp#2	ng/ml	ng/ml

System Monitoring Compounds						
1) SA Tetrachlo...	3.677	3.674	430.8E6	427.6E6	75.007	75.108
2) SA Decachlor...	8.713	8.661	383.2E6	130.7E6	75.017	75.183
Target Compounds						
3) L1 AR-1016-1	4.766	4.751	172.6E6	141.9E6	745.403	751.801
4) L1 AR-1016-2	4.785	4.769	250.3E6	207.8E6	751.782	747.072
5) L1 AR-1016-3	4.842	4.944	171.7E6	110.1E6	747.371	748.706
6) L1 AR-1016-4	4.962	4.987	137.2E6	88439125	749.293	747.424
7) L1 AR-1016-5	5.219	5.199	138.5E6	113.8E6	748.029	748.739
31) L7 AR-1260-1	6.257	6.228	253.1E6	181.7E6	746.377	748.442
32) L7 AR-1260-2	6.447	6.417	339.5E6	212.7E6	748.277	750.445
33) L7 AR-1260-3	6.813	6.569	308.1E6	192.7E6	749.048	749.738
34) L7 AR-1260-4	7.073	7.039	224.5E6	135.4E6	750.835	750.003
35) L7 AR-1260-5	7.316	7.280	608.0E6	309.9E6	752.345	753.935



(f)=RT Delta > 1/2 Window (#)=Amounts differ by > 25% (m)=manual int.

Data Path : Z:\pestpcbsrv\HPCHEM1\ECD_O\Data\PO061125\
Data File : PO111588.D
Signal(s) : Signal #1: ECD1A.ch Signal #2: ECD2B.ch
Acq On : 11 Jun 2025 10:58
Operator : YP/AJ
Sample : AR1660ICC750
Misc :
ALS Vial : 4 Sample Multiplier: 1

Instrument :
ECD_O
ClientSampleId :
AR1660ICC750

Integration File signal 1: autoint1.e
Integration File signal 2: autoint2.e
Quant Time: Jun 11 11:49:22 2025
Quant Method : Z:\pestpcbsrv\HPCHEM1\ECD_O\methods\PO061125.M
Quant Title : GC EXTRACTABLES
QLast Update : Wed Jun 11 11:43:51 2025
Response via : Initial Calibration
Integrator: ChemStation

Volume Inj. : 2 µl
Signal #1 Phase : ZB-MR1 Signal #2 Phase: ZB-MR2
Signal #1 Info : 30Mx0.32mmx 0.50µ Signal #2 Info : 30M x 0.32mm x 0.25µm



Data Path : Z:\pestpcbsrv\HPCHEM1\ECD_O\Data\P0061125\
 Data File : P0111589.D
 Signal(s) : Signal #1: ECD1A.ch Signal #2: ECD2B.ch
 Acq On : 11 Jun 2025 11:17
 Operator : YP/AJ
 Sample : AR1660ICC500
 Misc :
 ALS Vial : 5 Sample Multiplier: 1

Instrument :
 ECD_O
 ClientSampleId :
 AR1660ICC500

Integration File signal 1: autoint1.e
 Integration File signal 2: autoint2.e
 Quant Time: Jun 11 11:44:12 2025
 Quant Method : Z:\pestpcbsrv\HPCHEM1\ECD_O\methods\P0061125.M
 Quant Title : GC EXTRACTABLES
 QLast Update : Wed Jun 11 11:43:51 2025
 Response via : Initial Calibration
 Integrator: ChemStation

Volume Inj. : 2 µl
 Signal #1 Phase : ZB-MR1 Signal #2 Phase: ZB-MR2
 Signal #1 Info : 30Mx0.32mmx 0.50µ Signal #2 Info : 30M x 0.32mm x 0.25µm

Compound	RT#1	RT#2	Resp#1	Resp#2	ng/ml	ng/ml

System Monitoring Compounds						
1) SA Tetrachlo...	3.677	3.673	296.4E6	289.6E6	50.000	50.000
2) SA Decachlor...	8.712	8.661	265.7E6	90408461	50.000	50.000
Target Compounds						
3) L1 AR-1016-1	4.766	4.751	122.1E6	98131800	500.000	500.000
4) L1 AR-1016-2	4.784	4.769	172.8E6	143.9E6	500.000	500.000
5) L1 AR-1016-3	4.841	4.944	120.8E6	76494248	500.000	500.000
6) L1 AR-1016-4	4.961	4.987	95790139	62336179	500.000	500.000
7) L1 AR-1016-5	5.218	5.199	96695414	79297809	500.000	500.000
31) L7 AR-1260-1	6.257	6.229	179.0E6	126.6E6	500.000	500.000
32) L7 AR-1260-2	6.446	6.416	237.4E6	146.9E6	500.000	500.000
33) L7 AR-1260-3	6.813	6.569	214.8E6	132.9E6	500.000	500.000
34) L7 AR-1260-4	7.072	7.039	157.7E6	95019334	500.000	500.000
35) L7 AR-1260-5	7.316	7.281	416.1E6	212.1E6	500.000	500.000



(f)=RT Delta > 1/2 Window (#)=Amounts differ by > 25% (m)=manual int.

Data Path : Z:\pestpcbsrv\HPCHEM1\ECD_O\Data\PO061125\
Data File : PO111589.D
Signal(s) : Signal #1: ECD1A.ch Signal #2: ECD2B.ch
Acq On : 11 Jun 2025 11:17
Operator : YP/AJ
Sample : AR1660ICC500
Misc :
ALS Vial : 5 Sample Multiplier: 1

Instrument :
ECD_O
ClientSampleId :
AR1660ICC500

Integration File signal 1: autoint1.e
Integration File signal 2: autoint2.e
Quant Time: Jun 11 11:44:12 2025
Quant Method : Z:\pestpcbsrv\HPCHEM1\ECD_O\methods\PO061125.M
Quant Title : GC EXTRACTABLES
QLast Update : Wed Jun 11 11:43:51 2025
Response via : Initial Calibration
Integrator: ChemStation

Volume Inj. : 2 µl
Signal #1 Phase : ZB-MR1 Signal #2 Phase: ZB-MR2
Signal #1 Info : 30Mx0.32mmx 0.50µm Signal #2 Info : 30M x 0.32mm x 0.25µm



Data Path : Z:\pestpcbsrv\HPCHEM1\ECD_O\Data\P0061125\
 Data File : P0111590.D
 Signal(s) : Signal #1: ECD1A.ch Signal #2: ECD2B.ch
 Acq On : 11 Jun 2025 11:35
 Operator : YP/AJ
 Sample : AR1660ICC250
 Misc :
 ALS Vial : 6 Sample Multiplier: 1

Instrument :
 ECD_O
 ClientSampleId :
 AR1660ICC250

Integration File signal 1: autoint1.e
 Integration File signal 2: autoint2.e
 Quant Time: Jun 11 11:52:26 2025
 Quant Method : Z:\pestpcbsrv\HPCHEM1\ECD_O\methods\P0061125.M
 Quant Title : GC EXTRACTABLES
 QLast Update : Wed Jun 11 11:43:51 2025
 Response via : Initial Calibration
 Integrator: ChemStation

Volume Inj. : 2 µl
 Signal #1 Phase : ZB-MR1 Signal #2 Phase: ZB-MR2
 Signal #1 Info : 30Mx0.32mmx 0.50µ Signal #2 Info : 30M x 0.32mm x 0.25µm

Compound	RT#1	RT#2	Resp#1	Resp#2	ng/ml	ng/ml

System Monitoring Compounds						
1) SA Tetrachlo...	3.678	3.673	150.9E6	144.9E6	25.943	25.333
2) SA Decachlor...	8.713	8.662	139.5E6	47118679	26.697	26.546
Target Compounds						
3) L1 AR-1016-1	4.766	4.751	63980961	51169234	269.190	265.465
4) L1 AR-1016-2	4.785	4.769	90057979	74078323	265.048	262.065
5) L1 AR-1016-3	4.842	4.944	63869331	39522116	270.400	263.792
6) L1 AR-1016-4	4.962	4.987	51072699	32018591	271.039	265.137
7) L1 AR-1016-5	5.219	5.199	51759950	41162532	271.494	265.301
31) L7 AR-1260-1	6.258	6.229	106.7E6	66350033	295.514	267.060
32) L7 AR-1260-2	6.447	6.417	126.7E6	76552502	271.274	264.732
33) L7 AR-1260-3	6.813	6.569	110.4E6	69891883	263.627	266.124
34) L7 AR-1260-4	7.073	7.039	81970889	50303192	267.669	270.812
35) L7 AR-1260-5	7.316	7.280	216.8E6	110.5E6	263.450	263.890



(f)=RT Delta > 1/2 Window (#)=Amounts differ by > 25% (m)=manual int.

Data Path : Z:\pestpcbsrv\HPCHEM1\ECD_O\Data\PO061125\
Data File : PO111590.D
Signal(s) : Signal #1: ECD1A.ch Signal #2: ECD2B.ch
Acq On : 11 Jun 2025 11:35
Operator : YP/AJ
Sample : AR1660ICC250
Misc :
ALS Vial : 6 Sample Multiplier: 1

Instrument :
ECD_O
ClientSampleId :
AR1660ICC250

Integration File signal 1: autoint1.e
Integration File signal 2: autoint2.e
Quant Time: Jun 11 11:52:26 2025
Quant Method : Z:\pestpcbsrv\HPCHEM1\ECD_O\methods\PO061125.M
Quant Title : GC EXTRACTABLES
QLast Update : Wed Jun 11 11:43:51 2025
Response via : Initial Calibration
Integrator: ChemStation

Volume Inj. : 2 µl
Signal #1 Phase : ZB-MR1 Signal #2 Phase: ZB-MR2
Signal #1 Info : 30Mx0.32mmx 0.50µm Signal #2 Info : 30M x 0.32mm x 0.25µm



Data Path : Z:\pestpcbsrv\HPCHEM1\ECD_O\Data\P0061125\
 Data File : P0111591.D
 Signal(s) : Signal #1: ECD1A.ch Signal #2: ECD2B.ch
 Acq On : 11 Jun 2025 11:53
 Operator : YP/AJ
 Sample : AR1660ICC050
 Misc :
 ALS Vial : 7 Sample Multiplier: 1

Instrument :
 ECD_O
 ClientSampleId :
 AR1660ICC050

Manual Integrations
 APPROVED

Reviewed By :Yogesh Patel 06/12/2025
 Supervised By :mohammad ahmed 06/13/2025

Integration File signal 1: autoint1.e
 Integration File signal 2: autoint2.e
 Quant Time: Jun 11 12:07:16 2025
 Quant Method : Z:\pestpcbsrv\HPCHEM1\ECD_O\methods\P0061125.M
 Quant Title : GC EXTRACTABLES
 QLast Update : Wed Jun 11 12:07:04 2025
 Response via : Initial Calibration
 Integrator: ChemStation

Volume Inj. : 2 µl
 Signal #1 Phase : ZB-MR1 Signal #2 Phase: ZB-MR2
 Signal #1 Info : 30Mx0.32mmx 0.50µ Signal #2 Info : 30M x 0.32mm x 0.25µm

Compound	RT#1	RT#2	Resp#1	Resp#2	ng/ml	ng/ml

System Monitoring Compounds						
1) SA Tetrachlo...	3.678	3.673	27660981	26053048	4.803	4.638
2) SA Decachlor...	8.712	8.662	26704107	8941855	5.087	5.030
Target Compounds						
3) L1 AR-1016-1	4.765	4.750	12102371	10641184	51.662m	54.080
4) L1 AR-1016-2	4.784	4.769	15230034	14685829	45.601m	51.551
5) L1 AR-1016-3	4.841	4.944	10294489	7595210	43.965m	50.554
6) L1 AR-1016-4	4.961	4.987	8679050	6294307	46.206m	51.683
7) L1 AR-1016-5	5.218	5.199	8531210	8768797	45.620m	55.081
31) L7 AR-1260-1	6.256	6.229	16630454	13039756	46.467m	51.748m
32) L7 AR-1260-2	6.446	6.416	24715396	15670158	52.210m	53.392m
33) L7 AR-1260-3	6.813	6.569	22383586	14269100	52.708	53.406
34) L7 AR-1260-4	7.072	7.039	17363789	10374184	55.220	54.573
35) L7 AR-1260-5	7.316	7.280	41341782	22505628	50.189	52.952

(f)=RT Delta > 1/2 Window (#)=Amounts differ by > 25% (m)=manual int.

Data Path : Z:\pestpcbsrv\HPCHEM1\ECD_O\Data\PO061125\
Data File : PO111591.D
Signal(s) : Signal #1: ECD1A.ch Signal #2: ECD2B.ch
Acq On : 11 Jun 2025 11:53
Operator : YP/AJ
Sample : AR1660ICC050
Misc :
ALS Vial : 7 Sample Multiplier: 1



Instrument :
ECD_O
ClientSampleId :
AR1660ICC050

Manual Integrations
APPROVED

Reviewed By :Yogesh Patel 06/12/2025
Supervised By :mohammad ahmed 06/13/2025

Integration File signal 1: autoint1.e
Integration File signal 2: autoint2.e
Quant Time: Jun 11 12:07:16 2025
Quant Method : Z:\pestpcbsrv\HPCHEM1\ECD_O\methods\PO061125.M
Quant Title : GC EXTRACTABLES
QLast Update : Wed Jun 11 12:07:04 2025
Response via : Initial Calibration
Integrator: ChemStation

Volume Inj. : 2 µl
Signal #1 Phase : ZB-MR1 Signal #2 Phase: ZB-MR2
Signal #1 Info : 30Mx0.32mmx 0.50µm Signal #2 Info : 30M x 0.32mm x 0.25µm



Data Path : Z:\pestpcbsrv\HPCHEM1\ECD_O\Data\P0061125\
 Data File : P0111592.D
 Signal(s) : Signal #1: ECD1A.ch Signal #2: ECD2B.ch
 Acq On : 11 Jun 2025 12:12
 Operator : YP/AJ
 Sample : AR1221ICC500
 Misc :
 ALS Vial : 8 Sample Multiplier: 1

Instrument :
 ECD_O
 ClientSampleId :
 AR1221ICC500

Integration File signal 1: autoint1.e
 Integration File signal 2: autoint2.e
 Quant Time: Jun 11 12:46:16 2025
 Quant Method : Z:\pestpcbsrv\HPCHEM1\ECD_O\methods\P0061125.M
 Quant Title : GC EXTRACTABLES
 QLast Update : Wed Jun 11 12:46:00 2025
 Response via : Initial Calibration
 Integrator: ChemStation

Volume Inj. : 2 µl
 Signal #1 Phase : ZB-MR1 Signal #2 Phase: ZB-MR2
 Signal #1 Info : 30Mx0.32mmx 0.50µ Signal #2 Info : 30M x 0.32mm x 0.25µm

Compound	RT#1	RT#2	Resp#1	Resp#2	ng/ml	ng/ml

System Monitoring Compounds						
1) SA Tetrachlo...	3.677	3.673	287.0E6	275.6E6	50.000	50.000
2) SA Decachlor...	8.712	8.661	267.2E6	90058921	50.000	50.000
Target Compounds						
8) L2 AR-1221-1	3.889	3.883	43539401	37336939	500.000	500.000
9) L2 AR-1221-2	3.976	3.968	32083882	27974442	500.000	500.000
10) L2 AR-1221-3	4.051	4.043	97518158	85915227	500.000	500.000



(f)=RT Delta > 1/2 Window (#)=Amounts differ by > 25% (m)=manual int.

Data Path : Z:\pestpcbsrv\HPCHEM1\ECD_O\Data\PO061125\
Data File : PO111592.D
Signal(s) : Signal #1: ECD1A.ch Signal #2: ECD2B.ch
Acq On : 11 Jun 2025 12:12
Operator : YP/AJ
Sample : AR1221ICC500
Misc :
ALS Vial : 8 Sample Multiplier: 1

Instrument :
ECD_O
ClientSampleId :
AR1221ICC500

Integration File signal 1: autoint1.e
Integration File signal 2: autoint2.e
Quant Time: Jun 11 12:46:16 2025
Quant Method : Z:\pestpcbsrv\HPCHEM1\ECD_O\methods\PO061125.M
Quant Title : GC EXTRACTABLES
QLast Update : Wed Jun 11 12:46:00 2025
Response via : Initial Calibration
Integrator: ChemStation

Volume Inj. : 2 µl
Signal #1 Phase : ZB-MR1 Signal #2 Phase: ZB-MR2
Signal #1 Info : 30Mx0.32mmx 0.50µm Signal #2 Info : 30M x 0.32mm x 0.25µm



Data Path : Z:\pestpcbsrv\HPCHEM1\ECD_O\Data\P0061125\
 Data File : P0111593.D
 Signal(s) : Signal #1: ECD1A.ch Signal #2: ECD2B.ch
 Acq On : 11 Jun 2025 12:30
 Operator : YP/AJ
 Sample : AR1232ICC500
 Misc :
 ALS Vial : 9 Sample Multiplier: 1

Instrument :
 ECD_O
 ClientSampleId :
 AR1232ICC500

Integration File signal 1: autoint1.e
 Integration File signal 2: autoint2.e
 Quant Time: Jun 11 12:49:36 2025
 Quant Method : Z:\pestpcbsrv\HPCHEM1\ECD_O\methods\P0061125.M
 Quant Title : GC EXTRACTABLES
 QLast Update : Wed Jun 11 12:46:00 2025
 Response via : Initial Calibration
 Integrator: ChemStation

Volume Inj. : 2 µl
 Signal #1 Phase : ZB-MR1 Signal #2 Phase: ZB-MR2
 Signal #1 Info : 30Mx0.32mmx 0.50µ Signal #2 Info : 30M x 0.32mm x 0.25µm

Compound	RT#1	RT#2	Resp#1	Resp#2	ng/ml	ng/ml

System Monitoring Compounds						
1) SA Tetrachlo...	3.678	3.673	281.8E6	280.2E6	50.000	50.000
2) SA Decachlor...	8.712	8.661	258.0E6	86695497	50.000	50.000
Target Compounds						
11) L3 AR-1232-1	4.052	4.043	75874515	66494811	500.000	500.000
12) L3 AR-1232-2	4.544	4.769	43110570	65895506	500.000	500.000
13) L3 AR-1232-3	4.785	4.944	81440353	34559074	500.000	500.000
14) L3 AR-1232-4	4.961	5.029	43509571	30242026	500.000	500.000
15) L3 AR-1232-5	5.004	5.199	28290272	33207345	500.000	500.000



(f)=RT Delta > 1/2 Window (#)=Amounts differ by > 25% (m)=manual int.

Data Path : Z:\pestpcbsrv\HPCHEM1\ECD_O\Data\PO061125\
Data File : PO111593.D
Signal(s) : Signal #1: ECD1A.ch Signal #2: ECD2B.ch
Acq On : 11 Jun 2025 12:30
Operator : YP/AJ
Sample : AR1232ICC500
Misc :
ALS Vial : 9 Sample Multiplier: 1

Instrument :
ECD_O
ClientSampleId :
AR1232ICC500

Integration File signal 1: autoint1.e
Integration File signal 2: autoint2.e
Quant Time: Jun 11 12:49:36 2025
Quant Method : Z:\pestpcbsrv\HPCHEM1\ECD_O\methods\PO061125.M
Quant Title : GC EXTRACTABLES
QLast Update : Wed Jun 11 12:46:00 2025
Response via : Initial Calibration
Integrator: ChemStation

Volume Inj. : 2 µl
Signal #1 Phase : ZB-MR1 Signal #2 Phase: ZB-MR2
Signal #1 Info : 30Mx0.32mmx 0.50µm Signal #2 Info : 30M x 0.32mm x 0.25µm



Data Path : Z:\pestpcbsrv\HPCHEM1\ECD_O\Data\P0061125\
 Data File : P0111594.D
 Signal(s) : Signal #1: ECD1A.ch Signal #2: ECD2B.ch
 Acq On : 11 Jun 2025 12:48
 Operator : YP/AJ
 Sample : AR1242ICC1000
 Misc :
 ALS Vial : 10 Sample Multiplier: 1

Instrument :
 ECD_O
 ClientSampleId :
 AR1242ICC1000

Integration File signal 1: autoint1.e
 Integration File signal 2: autoint2.e
 Quant Time: Jun 11 14:17:20 2025
 Quant Method : Z:\pestpcbsrv\HPCHEM1\ECD_O\methods\P0061125.M
 Quant Title : GC EXTRACTABLES
 QLast Update : Wed Jun 11 14:16:47 2025
 Response via : Initial Calibration
 Integrator: ChemStation

Volume Inj. : 2 µl
 Signal #1 Phase : ZB-MR1 Signal #2 Phase: ZB-MR2
 Signal #1 Info : 30Mx0.32mmx 0.50µ Signal #2 Info : 30M x 0.32mm x 0.25µm

Compound	RT#1	RT#2	Resp#1	Resp#2	ng/ml	ng/ml

System Monitoring Compounds						
1) SA Tetrachlo...	3.678	3.674	545.7E6	544.8E6	96.546	98.009
2) SA Decachlor...	8.713	8.661	488.1E6	162.9E6	96.233	95.886
Target Compounds						
16) L4 AR-1242-1	4.766	4.750	189.9E6	152.2E6	953.554	956.335
17) L4 AR-1242-2	4.785	4.769	269.3E6	225.8E6	954.325	962.422
18) L4 AR-1242-3	4.841	4.944	185.2E6	119.5E6	947.690	960.999
19) L4 AR-1242-4	4.961	5.028	148.9E6	114.6E6	950.919	953.944
20) L4 AR-1242-5	5.613	5.548	154.5E6	141.8E6	934.407	954.850



(f)=RT Delta > 1/2 Window (#)=Amounts differ by > 25% (m)=manual int.

Data Path : Z:\pestpcbsrv\HPCHEM1\ECD_O\Data\PO061125\
Data File : PO111594.D
Signal(s) : Signal #1: ECD1A.ch Signal #2: ECD2B.ch
Acq On : 11 Jun 2025 12:48
Operator : YP/AJ
Sample : AR1242ICC1000
Misc :
ALS Vial : 10 Sample Multiplier: 1

Instrument :
ECD_O
ClientSampleId :
AR1242ICC1000

Integration File signal 1: autoint1.e
Integration File signal 2: autoint2.e
Quant Time: Jun 11 14:17:20 2025
Quant Method : Z:\pestpcbsrv\HPCHEM1\ECD_O\methods\PO061125.M
Quant Title : GC EXTRACTABLES
QLast Update : Wed Jun 11 14:16:47 2025
Response via : Initial Calibration
Integrator: ChemStation

Volume Inj. : 2 µl
Signal #1 Phase : ZB-MR1 Signal #2 Phase: ZB-MR2
Signal #1 Info : 30Mx0.32mmx 0.50µm Signal #2 Info : 30M x 0.32mm x 0.25µm



Data Path : Z:\pestpcbsrv\HPCHEM1\ECD_O\Data\P0061125\
 Data File : P0111595.D
 Signal(s) : Signal #1: ECD1A.ch Signal #2: ECD2B.ch
 Acq On : 11 Jun 2025 13:07
 Operator : YP/AJ
 Sample : AR1242ICC750
 Misc :
 ALS Vial : 11 Sample Multiplier: 1

Instrument :
 ECD_O
 ClientSampleId :
 AR1242ICC750

Integration File signal 1: autoint1.e
 Integration File signal 2: autoint2.e
 Quant Time: Jun 11 14:20:31 2025
 Quant Method : Z:\pestpcbsrv\HPCHEM1\ECD_O\methods\P0061125.M
 Quant Title : GC EXTRACTABLES
 QLast Update : Wed Jun 11 14:19:50 2025
 Response via : Initial Calibration
 Integrator: ChemStation

Volume Inj. : 2 µl
 Signal #1 Phase : ZB-MR1 Signal #2 Phase: ZB-MR2
 Signal #1 Info : 30Mx0.32mmx 0.50µ Signal #2 Info : 30M x 0.32mm x 0.25µm

Compound	RT#1	RT#2	Resp#1	Resp#2	ng/ml	ng/ml

System Monitoring Compounds						
1) SA Tetrachlo...	3.678	3.674	426.1E6	419.4E6	75.261	75.304
2) SA Decachlor...	8.713	8.662	381.1E6	127.3E6	75.098	74.955
Target Compounds						
16) L4 AR-1242-1	4.766	4.751	149.5E6	119.8E6	750.589	752.073
17) L4 AR-1242-2	4.785	4.769	212.9E6	176.2E6	752.944	750.686
18) L4 AR-1242-3	4.841	4.944	148.7E6	92718079	757.338	747.108
19) L4 AR-1242-4	4.962	5.029	117.3E6	89969660	749.552	749.340
20) L4 AR-1242-5	5.613	5.549	125.2E6	112.0E6	755.024	752.841



(f)=RT Delta > 1/2 Window (#)=Amounts differ by > 25% (m)=manual int.

Data Path : Z:\pestpcbsrv\HPCHEM1\ECD_O\Data\PO061125\
Data File : PO111595.D
Signal(s) : Signal #1: ECD1A.ch Signal #2: ECD2B.ch
Acq On : 11 Jun 2025 13:07
Operator : YP/AJ
Sample : AR1242ICC750
Misc :
ALS Vial : 11 Sample Multiplier: 1

Instrument :
ECD_O
ClientSampleId :
AR1242ICC750

Integration File signal 1: autoint1.e
Integration File signal 2: autoint2.e
Quant Time: Jun 11 14:20:31 2025
Quant Method : Z:\pestpcbsrv\HPCHEM1\ECD_O\methods\PO061125.M
Quant Title : GC EXTRACTABLES
QLast Update : Wed Jun 11 14:19:50 2025
Response via : Initial Calibration
Integrator: ChemStation

Volume Inj. : 2 µl
Signal #1 Phase : ZB-MR1 Signal #2 Phase: ZB-MR2
Signal #1 Info : 30Mx0.32mmx 0.50µm Signal #2 Info : 30M x 0.32mm x 0.25µm



Data Path : Z:\pestpcbsrv\HPCHEM1\ECD_O\Data\P0061125\
 Data File : P0111596.D
 Signal(s) : Signal #1: ECD1A.ch Signal #2: ECD2B.ch
 Acq On : 11 Jun 2025 13:25
 Operator : YP/AJ
 Sample : AR1242ICC500
 Misc :
 ALS Vial : 12 Sample Multiplier: 1

Instrument :
 ECD_O
 ClientSampleId :
 AR1242ICC500

Integration File signal 1: autoint1.e
 Integration File signal 2: autoint2.e
 Quant Time: Jun 11 14:13:41 2025
 Quant Method : Z:\pestpcbsrv\HPCHEM1\ECD_O\methods\P0061125.M
 Quant Title : GC EXTRACTABLES
 QLast Update : Wed Jun 11 14:13:27 2025
 Response via : Initial Calibration
 Integrator: ChemStation

Volume Inj. : 2 µl
 Signal #1 Phase : ZB-MR1 Signal #2 Phase: ZB-MR2
 Signal #1 Info : 30Mx0.32mmx 0.50µ Signal #2 Info : 30M x 0.32mm x 0.25µm

Compound	RT#1	RT#2	Resp#1	Resp#2	ng/ml	ng/ml

System Monitoring Compounds						
1) SA Tetrachlo...	3.678	3.673	292.4E6	283.4E6	50.000	50.000
2) SA Decachlor...	8.713	8.662	263.1E6	88437095	50.000	50.000
Target Compounds						
16) L4 AR-1242-1	4.765	4.751	104.2E6	83032155	500.000	500.000
17) L4 AR-1242-2	4.784	4.769	147.5E6	121.7E6	500.000	500.000
18) L4 AR-1242-3	4.841	4.944	102.8E6	64595632	500.000	500.000
19) L4 AR-1242-4	4.960	5.028	82142260	62825159	500.000	500.000
20) L4 AR-1242-5	5.612	5.549	88077559	77585315	500.000	500.000



(f)=RT Delta > 1/2 Window (#)=Amounts differ by > 25% (m)=manual int.

Data Path : Z:\pestpcbsrv\HPCHEM1\ECD_O\Data\PO061125\
Data File : PO111596.D
Signal(s) : Signal #1: ECD1A.ch Signal #2: ECD2B.ch
Acq On : 11 Jun 2025 13:25
Operator : YP/AJ
Sample : AR1242ICC500
Misc :
ALS Vial : 12 Sample Multiplier: 1

Instrument :
ECD_O
ClientSampleId :
AR1242ICC500

Integration File signal 1: autoint1.e
Integration File signal 2: autoint2.e
Quant Time: Jun 11 14:13:41 2025
Quant Method : Z:\pestpcbsrv\HPCHEM1\ECD_O\methods\PO061125.M
Quant Title : GC EXTRACTABLES
QLast Update : Wed Jun 11 14:13:27 2025
Response via : Initial Calibration
Integrator: ChemStation

Volume Inj. : 2 µl
Signal #1 Phase : ZB-MR1 Signal #2 Phase: ZB-MR2
Signal #1 Info : 30Mx0.32mmx 0.50µm Signal #2 Info : 30M x 0.32mm x 0.25µm



Data Path : Z:\pestpcbsrv\HPCHEM1\ECD_O\Data\P0061125\
 Data File : P0111597.D
 Signal(s) : Signal #1: ECD1A.ch Signal #2: ECD2B.ch
 Acq On : 11 Jun 2025 13:44
 Operator : YP/AJ
 Sample : AR1242ICC250
 Misc :
 ALS Vial : 13 Sample Multiplier: 1

Instrument :
 ECD_O
 ClientSampleId :
 AR1242ICC250

Integration File signal 1: autoint1.e
 Integration File signal 2: autoint2.e
 Quant Time: Jun 11 14:23:00 2025
 Quant Method : Z:\pestpcbsrv\HPCHEM1\ECD_O\methods\P0061125.M
 Quant Title : GC EXTRACTABLES
 QLast Update : Wed Jun 11 14:19:50 2025
 Response via : Initial Calibration
 Integrator: ChemStation

Volume Inj. : 2 µl
 Signal #1 Phase : ZB-MR1 Signal #2 Phase: ZB-MR2
 Signal #1 Info : 30Mx0.32mmx 0.50µ Signal #2 Info : 30M x 0.32mm x 0.25µm

Compound	RT#1	RT#2	Resp#1	Resp#2	ng/ml	ng/ml

System Monitoring Compounds						
1) SA Tetrachlo...	3.676	3.672	150.8E6	142.6E6	26.206	25.453
2) SA Decachlor...	8.711	8.661	138.1E6	46179693	26.616	26.608
Target Compounds						
16) L4 AR-1242-1	4.764	4.749	54708882	44017430	268.019	269.187
17) L4 AR-1242-2	4.783	4.768	77715059	62976219	268.219	263.448
18) L4 AR-1242-3	4.840	4.943	54110674	33754705	268.666	266.138
19) L4 AR-1242-4	4.959	5.028	42489595	33285094	265.720	269.878
20) L4 AR-1242-5	5.612	5.548	45769698	40348676	268.964	265.608



(f)=RT Delta > 1/2 Window (#)=Amounts differ by > 25% (m)=manual int.

Data Path : Z:\pestpcbsrv\HPCHEM1\ECD_O\Data\PO061125\
Data File : PO111597.D
Signal(s) : Signal #1: ECD1A.ch Signal #2: ECD2B.ch
Acq On : 11 Jun 2025 13:44
Operator : YP/AJ
Sample : AR1242ICC250
Misc :
ALS Vial : 13 Sample Multiplier: 1

Instrument :
ECD_O
ClientSampleId :
AR1242ICC250

Integration File signal 1: autoint1.e
Integration File signal 2: autoint2.e
Quant Time: Jun 11 14:23:00 2025
Quant Method : Z:\pestpcbsrv\HPCHEM1\ECD_O\methods\PO061125.M
Quant Title : GC EXTRACTABLES
QLast Update : Wed Jun 11 14:19:50 2025
Response via : Initial Calibration
Integrator: ChemStation

Volume Inj. : 2 µl
Signal #1 Phase : ZB-MR2 Signal #2 Phase: ZB-MR2
Signal #1 Info : 30Mx0.32mmx 0.50µm Signal #2 Info : 30M x 0.32mm x 0.25µm



Data Path : Z:\pestpcbsrv\HPCHEM1\ECD_O\Data\P0061125\
 Data File : P0111598.D
 Signal(s) : Signal #1: ECD1A.ch Signal #2: ECD2B.ch
 Acq On : 11 Jun 2025 14:02
 Operator : YP/AJ
 Sample : AR1242ICC050
 Misc :
 ALS Vial : 14 Sample Multiplier: 1

Instrument :
 ECD_O
 ClientSampleId :
 AR1242ICC050

Manual Integrations
 APPROVED

Reviewed By :Yogesh Patel 06/12/2025
 Supervised By :mohammad ahmed 06/13/2025

Integration File signal 1: autoint1.e
 Integration File signal 2: autoint2.e
 Quant Time: Jun 11 14:26:28 2025
 Quant Method : Z:\pestpcbsrv\HPCHEM1\ECD_O\methods\P0061125.M
 Quant Title : GC EXTRACTABLES
 QLast Update : Wed Jun 11 14:26:18 2025
 Response via : Initial Calibration
 Integrator: ChemStation

Volume Inj. : 2 µl
 Signal #1 Phase : ZB-MR1 Signal #2 Phase: ZB-MR2
 Signal #1 Info : 30Mx0.32mmx 0.50µ Signal #2 Info : 30M x 0.32mm x 0.25µm

Compound	RT#1	RT#2	Resp#1	Resp#2	ng/ml	ng/ml

System Monitoring Compounds						
1) SA Tetrachlo...	3.677	3.672	26071082	23777938	4.617	4.376
2) SA Decachlor...	8.712	8.661	24643941	8137257	4.799	4.748
Target Compounds						
16) L4 AR-1242-1	4.765	4.751	9536300	7985936	47.674m	49.066
17) L4 AR-1242-2	4.784	4.769	12492728	11206079	44.428m	47.471
18) L4 AR-1242-3	4.841	4.944	9424858	6031021	47.765m	48.022
19) L4 AR-1242-4	4.961	5.029	7568610	5972863	47.687m	48.735
20) L4 AR-1242-5	5.613	5.549	7770041	7090223	47.727m	47.303

(f)=RT Delta > 1/2 Window (#)=Amounts differ by > 25% (m)=manual int.

Data Path : Z:\pestpcbsrv\HPCHEM1\ECD_O\Data\PO061125\
Data File : PO111598.D
Signal(s) : Signal #1: ECD1A.ch Signal #2: ECD2B.ch
Acq On : 11 Jun 2025 14:02
Operator : YP/AJ
Sample : AR1242ICC050
Misc :
ALS Vial : 14 Sample Multiplier: 1



Instrument :
ECD_O
ClientSampleId :
AR1242ICC050

Manual Integrations
APPROVED

Reviewed By :Yogesh Patel 06/12/2025
Supervised By :mohammad ahmed 06/13/2025

Integration File signal 1: autoint1.e
Integration File signal 2: autoint2.e
Quant Time: Jun 11 14:26:28 2025
Quant Method : Z:\pestpcbsrv\HPCHEM1\ECD_O\methods\PO061125.M
Quant Title : GC EXTRACTABLES
QLast Update : Wed Jun 11 14:26:18 2025
Response via : Initial Calibration
Integrator: ChemStation

Volume Inj. : 2 µl
Signal #1 Phase : ZB-MR1 Signal #2 Phase: ZB-MR2
Signal #1 Info : 30Mx0.32mmx 0.50µ Signal #2 Info : 30M x 0.32mm x 0.25µm



Data Path : Z:\pestpcbsrv\HPCHEM1\ECD_O\Data\P0061125\
 Data File : P0111599.D
 Signal(s) : Signal #1: ECD1A.ch Signal #2: ECD2B.ch
 Acq On : 11 Jun 2025 14:20
 Operator : YP/AJ
 Sample : AR1248ICC1000
 Misc :
 ALS Vial : 15 Sample Multiplier: 1

Instrument :
 ECD_O
 ClientSampleId :
 AR1248ICC1000

Integration File signal 1: autoint1.e
 Integration File signal 2: autoint2.e
 Quant Time: Jun 11 16:22:01 2025
 Quant Method : Z:\pestpcbsrv\HPCHEM1\ECD_O\methods\P0061125.M
 Quant Title : GC EXTRACTABLES
 QLast Update : Wed Jun 11 16:18:41 2025
 Response via : Initial Calibration
 Integrator: ChemStation

Volume Inj. : 2 µl
 Signal #1 Phase : ZB-MR1 Signal #2 Phase: ZB-MR2
 Signal #1 Info : 30Mx0.32mmx 0.50µ Signal #2 Info : 30M x 0.32mm x 0.25µm

Compound	RT#1	RT#2	Resp#1	Resp#2	ng/ml	ng/ml

System Monitoring Compounds						
1) SA Tetrachlo...	3.678	3.673	562.7E6	545.9E6	92.601	95.368
2) SA Decachlor...	8.712	8.662	493.0E6	164.5E6	90.940	91.111
Target Compounds						
21) L5 AR-1248-1	4.766	4.751	145.9E6	118.8E6	898.472	905.205
22) L5 AR-1248-2	5.005	4.987	193.4E6	160.4E6	903.103	896.089
23) L5 AR-1248-3	5.219	5.029	243.7E6	169.9E6	900.111	903.774
24) L5 AR-1248-4	5.572	5.199	355.9E6	200.5E6	893.308	901.345
25) L5 AR-1248-5	5.614	5.591	249.8E6	195.8E6	880.725	906.082



(f)=RT Delta > 1/2 Window (#)=Amounts differ by > 25% (m)=manual int.

Data Path : Z:\pestpcbsrv\HPCHEM1\ECD_O\Data\PO061125\
Data File : PO111599.D
Signal(s) : Signal #1: ECD1A.ch Signal #2: ECD2B.ch
Acq On : 11 Jun 2025 14:20
Operator : YP/AJ
Sample : AR1248ICC1000
Misc :
ALS Vial : 15 Sample Multiplier: 1

Instrument :
ECD_O
ClientSampleId :
AR1248ICC1000

Integration File signal 1: autoint1.e
Integration File signal 2: autoint2.e
Quant Time: Jun 11 16:22:01 2025
Quant Method : Z:\pestpcbsrv\HPCHEM1\ECD_O\methods\PO061125.M
Quant Title : GC EXTRACTABLES
QLast Update : Wed Jun 11 16:18:41 2025
Response via : Initial Calibration
Integrator: ChemStation

Volume Inj. : 2 µl
Signal #1 Phase : ZB-MR1 Signal #2 Phase: ZB-MR2
Signal #1 Info : 30Mx0.32mmx 0.50µm Signal #2 Info : 30M x 0.32mm x 0.25µm



Data Path : Z:\pestpcbsrv\HPCHEM1\ECD_O\Data\P0061125\
 Data File : P0111600.D
 Signal(s) : Signal #1: ECD1A.ch Signal #2: ECD2B.ch
 Acq On : 11 Jun 2025 14:39
 Operator : YP/AJ
 Sample : AR1248ICC750
 Misc :
 ALS Vial : 16 Sample Multiplier: 1

Instrument :
 ECD_O
 ClientSampleId :
 AR1248ICC750

Integration File signal 1: autoint1.e
 Integration File signal 2: autoint2.e
 Quant Time: Jun 11 16:59:25 2025
 Quant Method : Z:\pestpcbsrv\HPCHEM1\ECD_O\methods\P0061125.M
 Quant Title : GC EXTRACTABLES
 QLast Update : Wed Jun 11 16:18:41 2025
 Response via : Initial Calibration
 Integrator: ChemStation

Volume Inj. : 2 µl
 Signal #1 Phase : ZB-MR1 Signal #2 Phase: ZB-MR2
 Signal #1 Info : 30Mx0.32mmx 0.50µ Signal #2 Info : 30M x 0.32mm x 0.25µm

Compound	RT#1	RT#2	Resp#1	Resp#2	ng/ml	ng/ml

System Monitoring Compounds						
1) SA Tetrachlo...	3.678	3.674	438.5E6	418.6E6	74.962	74.915
2) SA Decachlor...	8.713	8.661	385.7E6	128.8E6	74.679	74.790
Target Compounds						
21) L5 AR-1248-1	4.765	4.751	114.7E6	92154367	746.184	741.216
22) L5 AR-1248-2	5.004	4.987	152.3E6	126.2E6	748.166	745.749
23) L5 AR-1248-3	5.218	5.029	193.6E6	133.4E6	751.743	746.988
24) L5 AR-1248-4	5.572	5.199	282.3E6	157.1E6	749.055	745.196
25) L5 AR-1248-5	5.613	5.590	200.3E6	154.5E6	750.604	749.907



(f)=RT Delta > 1/2 Window (#)=Amounts differ by > 25% (m)=manual int.

Data Path : Z:\pestpcbsrv\HPCHEM1\ECD_O\Data\PO061125\
Data File : PO111600.D
Signal(s) : Signal #1: ECD1A.ch Signal #2: ECD2B.ch
Acq On : 11 Jun 2025 14:39
Operator : YP/AJ
Sample : AR1248ICC750
Misc :
ALS Vial : 16 Sample Multiplier: 1

Instrument :
ECD_O
ClientSampleId :
AR1248ICC750

Integration File signal 1: autoint1.e
Integration File signal 2: autoint2.e
Quant Time: Jun 11 16:59:25 2025
Quant Method : Z:\pestpcbsrv\HPCHEM1\ECD_O\methods\PO061125.M
Quant Title : GC EXTRACTABLES
QLast Update : Wed Jun 11 16:18:41 2025
Response via : Initial Calibration
Integrator: ChemStation

Volume Inj. : 2 µl
Signal #1 Phase : ZB-MR1 Signal #2 Phase: ZB-MR2
Signal #1 Info : 30Mx0.32mmx 0.50µm Signal #2 Info : 30M x 0.32mm x 0.25µm



Data Path : Z:\pestpcbsrv\HPCHEM1\ECD_O\Data\P0061125\
 Data File : P0111601.D
 Signal(s) : Signal #1: ECD1A.ch Signal #2: ECD2B.ch
 Acq On : 11 Jun 2025 15:14
 Operator : YP/AJ
 Sample : AR1248ICC500
 Misc :
 ALS Vial : 17 Sample Multiplier: 1

Instrument :
 ECD_O
 ClientSampleId :
 AR1248ICC500

Integration File signal 1: autoint1.e
 Integration File signal 2: autoint2.e
 Quant Time: Jun 11 16:19:54 2025
 Quant Method : Z:\pestpcbsrv\HPCHEM1\ECD_O\methods\P0061125.M
 Quant Title : GC EXTRACTABLES
 QLast Update : Wed Jun 11 16:18:41 2025
 Response via : Initial Calibration
 Integrator: ChemStation

Volume Inj. : 2 µl
 Signal #1 Phase : ZB-MR1 Signal #2 Phase: ZB-MR2
 Signal #1 Info : 30Mx0.32mmx 0.50µ Signal #2 Info : 30M x 0.32mm x 0.25µm

Compound	RT#1	RT#2	Resp#1	Resp#2	ng/ml	ng/ml

System Monitoring Compounds						
1) SA Tetrachlo...	3.676	3.673	303.8E6	286.2E6	50.000	50.000
2) SA Decachlor...	8.710	8.661	271.0E6	90265997	50.000	50.000
Target Compounds						
21) L5 AR-1248-1	4.764	4.750	81185066	65639548	500.000	500.000
22) L5 AR-1248-2	5.002	4.986	107.1E6	89516538	500.000	500.000
23) L5 AR-1248-3	5.216	5.028	135.4E6	93975832	500.000	500.000
24) L5 AR-1248-4	5.570	5.199	199.2E6	111.2E6	500.000	500.000
25) L5 AR-1248-5	5.611	5.589	141.8E6	108.1E6	500.000	500.000



(f)=RT Delta > 1/2 Window (#)=Amounts differ by > 25% (m)=manual int.

Data Path : Z:\pestpcbsrv\HPCHEM1\ECD_O\Data\PO061125\
Data File : PO111601.D
Signal(s) : Signal #1: ECD1A.ch Signal #2: ECD2B.ch
Acq On : 11 Jun 2025 15:14
Operator : YP/AJ
Sample : AR1248ICC500
Misc :
ALS Vial : 17 Sample Multiplier: 1

Instrument :
ECD_O
ClientSampleId :
AR1248ICC500

Integration File signal 1: autoint1.e
Integration File signal 2: autoint2.e
Quant Time: Jun 11 16:19:54 2025
Quant Method : Z:\pestpcbsrv\HPCHEM1\ECD_O\methods\PO061125.M
Quant Title : GC EXTRACTABLES
QLast Update : Wed Jun 11 16:18:41 2025
Response via : Initial Calibration
Integrator: ChemStation

Volume Inj. : 2 µl
Signal #1 Phase : ZB-MR1 Signal #2 Phase: ZB-MR2
Signal #1 Info : 30Mx0.32mmx 0.50µm Signal #2 Info : 30M x 0.32mm x 0.25µm



Data Path : Z:\pestpcbsrv\HPCHEM1\ECD_O\Data\P0061125\
 Data File : P0111602.D
 Signal(s) : Signal #1: ECD1A.ch Signal #2: ECD2B.ch
 Acq On : 11 Jun 2025 15:32
 Operator : YP/AJ
 Sample : AR1248ICC250
 Misc :
 ALS Vial : 18 Sample Multiplier: 1

Instrument :
 ECD_O
 ClientSampleId :
 AR1248ICC250

Integration File signal 1: autoint1.e
 Integration File signal 2: autoint2.e
 Quant Time: Jun 11 17:06:04 2025
 Quant Method : Z:\pestpcbsrv\HPCHEM1\ECD_O\methods\P0061125.M
 Quant Title : GC EXTRACTABLES
 QLast Update : Wed Jun 11 17:04:47 2025
 Response via : Initial Calibration
 Integrator: ChemStation

Volume Inj. : 2 µl
 Signal #1 Phase : ZB-MR1 Signal #2 Phase: ZB-MR2
 Signal #1 Info : 30Mx0.32mmx 0.50µ Signal #2 Info : 30M x 0.32mm x 0.25µm

Compound	RT#1	RT#2	Resp#1	Resp#2	ng/ml	ng/ml

System Monitoring Compounds						
1) SA Tetrachlo...	3.676	3.674	156.1E6	143.5E6	26.247	25.510
2) SA Decachlor...	8.711	8.662	142.1E6	47094148	26.845	26.713
Target Compounds						
21) L5 AR-1248-1	4.764	4.751	42620089	35136234	269.880	273.683
22) L5 AR-1248-2	5.002	4.987	55400295	46647339	266.297	268.727
23) L5 AR-1248-3	5.217	5.028	73715499	49572948	276.243	270.173
24) L5 AR-1248-4	5.570	5.199	106.6E6	59394712	273.901	273.101
25) L5 AR-1248-5	5.611	5.590	75633935	56998567	274.281	269.531



(f)=RT Delta > 1/2 Window (#)=Amounts differ by > 25% (m)=manual int.

Data Path : Z:\pestpcbsrv\HPCHEM1\ECD_O\Data\PO061125\
Data File : PO111602.D
Signal(s) : Signal #1: ECD1A.ch Signal #2: ECD2B.ch
Acq On : 11 Jun 2025 15:32
Operator : YP/AJ
Sample : AR1248ICC250
Misc :
ALS Vial : 18 Sample Multiplier: 1

Instrument :
ECD_O
ClientSampleId :
AR1248ICC250

Integration File signal 1: autoint1.e
Integration File signal 2: autoint2.e
Quant Time: Jun 11 17:06:04 2025
Quant Method : Z:\pestpcbsrv\HPCHEM1\ECD_O\methods\PO061125.M
Quant Title : GC EXTRACTABLES
QLast Update : Wed Jun 11 17:04:47 2025
Response via : Initial Calibration
Integrator: ChemStation

Volume Inj. : 2 µl
Signal #1 Phase : ZB-MR1 Signal #2 Phase: ZB-MR2
Signal #1 Info : 30Mx0.32mmx 0.50µm Signal #2 Info : 30M x 0.32mm x 0.25µm



Data Path : Z:\pestpcbsrv\HPCHEM1\ECD_O\Data\P0061125\
 Data File : P0111603.D
 Signal(s) : Signal #1: ECD1A.ch Signal #2: ECD2B.ch
 Acq On : 11 Jun 2025 15:49
 Operator : YP/AJ
 Sample : AR1248ICC050
 Misc :
 ALS Vial : 19 Sample Multiplier: 1

Instrument :
 ECD_O
 ClientSampleId :
 AR1248ICC050

Manual Integrations
 APPROVED

Reviewed By :Yogesh Patel 06/12/2025
 Supervised By :mohammad ahmed 06/13/2025

Integration File signal 1: autoint1.e
 Integration File signal 2: autoint2.e
 Quant Time: Jun 11 17:09:29 2025
 Quant Method : Z:\pestpcbsrv\HPCHEM1\ECD_O\methods\P0061125.M
 Quant Title : GC EXTRACTABLES
 QLast Update : Wed Jun 11 17:04:47 2025
 Response via : Initial Calibration
 Integrator: ChemStation

Volume Inj. : 2 µl
 Signal #1 Phase : ZB-MR1 Signal #2 Phase: ZB-MR2
 Signal #1 Info : 30Mx0.32mmx 0.50µ Signal #2 Info : 30M x 0.32mm x 0.25µm

Compound	RT#1	RT#2	Resp#1	Resp#2	ng/ml	ng/ml

System Monitoring Compounds						
1) SA Tetrachlo...	3.676	3.673	32463861	28440125	5.359	5.044
2) SA Decachlor...	8.711	8.660	30018432	9963819	5.522	5.508
Target Compounds						
21) L5 AR-1248-1	4.763	4.750	8731897	7644548	54.269m	57.355
22) L5 AR-1248-2	5.002	4.986	12293416	9821644	58.448	55.130
23) L5 AR-1248-3	5.216	5.027	14064716	10662336	50.156m	56.284
24) L5 AR-1248-4	5.569	5.198	22085352	13584386	54.922m	59.496
25) L5 AR-1248-5	5.611	5.588	15052068	12146521	53.408m	55.778

(f)=RT Delta > 1/2 Window (#)=Amounts differ by > 25% (m)=manual int.

Data Path : Z:\pestpcbsrv\HPCHEM1\ECD_O\Data\PO061125\
Data File : PO111603.D
Signal(s) : Signal #1: ECD1A.ch Signal #2: ECD2B.ch
Acq On : 11 Jun 2025 15:49
Operator : YP/AJ
Sample : AR1248ICC050
Misc :
ALS Vial : 19 Sample Multiplier: 1



Instrument :
ECD_O
ClientSampleId :
AR1248ICC050

Manual Integrations
APPROVED

Reviewed By :Yogesh Patel 06/12/2025
Supervised By :mohammad ahmed 06/13/2025

Integration File signal 1: autoint1.e
Integration File signal 2: autoint2.e
Quant Time: Jun 11 17:09:29 2025
Quant Method : Z:\pestpcbsrv\HPCHEM1\ECD_O\methods\PO061125.M
Quant Title : GC EXTRACTABLES
QLast Update : Wed Jun 11 17:04:47 2025
Response via : Initial Calibration
Integrator: ChemStation

Volume Inj. : 2 µl
Signal #1 Phase : ZB-MR1 Signal #2 Phase: ZB-MR2
Signal #1 Info : 30Mx0.32mmx 0.50µ Signal #2 Info : 30M x 0.32mm x 0.25µm



Data Path : Z:\pestpcbsrv\HPCHEM1\ECD_O\Data\P0061125\
 Data File : P0111604.D
 Signal(s) : Signal #1: ECD1A.ch Signal #2: ECD2B.ch
 Acq On : 11 Jun 2025 16:06
 Operator : YP/AJ
 Sample : AR1254ICC1000
 Misc :
 ALS Vial : 20 Sample Multiplier: 1

Instrument :
 ECD_O
 ClientSampleId :
 AR1254ICC1000

Integration File signal 1: autoint1.e
 Integration File signal 2: autoint2.e
 Quant Time: Jun 12 05:33:38 2025
 Quant Method : Z:\pestpcbsrv\HPCHEM1\ECD_O\methods\P0061125.M
 Quant Title : GC EXTRACTABLES
 QLast Update : Thu Jun 12 05:32:17 2025
 Response via : Initial Calibration
 Integrator: ChemStation

Volume Inj. : 2 µl
 Signal #1 Phase : ZB-MR1 Signal #2 Phase: ZB-MR2
 Signal #1 Info : 30Mx0.32mmx 0.50µ Signal #2 Info : 30M x 0.32mm x 0.25µm

Compound	RT#1	RT#2	Resp#1	Resp#2	ng/ml	ng/ml

System Monitoring Compounds						
1) SA Tetrachlo...	3.676	3.674	572.4E6	555.4E6	95.059	97.186
2) SA Decachlor...	8.712	8.661	500.9E6	168.1E6	92.701	92.748
Target Compounds						
26) L6 AR-1254-1	5.571	5.549	375.3E6	296.9E6	917.463	930.476
27) L6 AR-1254-2	5.720	5.697	334.6E6	255.5E6	915.415	926.192
28) L6 AR-1254-3	6.124	6.098	528.3E6	392.9E6	931.715	942.252
29) L6 AR-1254-4	6.353	6.325	330.8E6	214.6E6	927.406	929.731
30) L6 AR-1254-5	6.773	6.743	481.2E6	301.4E6	939.831	930.901



(f)=RT Delta > 1/2 Window (#)=Amounts differ by > 25% (m)=manual int.

Data Path : Z:\pestpcbsrv\HPCHEM1\ECD_O\Data\PO061125\
Data File : PO111604.D
Signal(s) : Signal #1: ECD1A.ch Signal #2: ECD2B.ch
Acq On : 11 Jun 2025 16:06
Operator : YP/AJ
Sample : AR1254ICC1000
Misc :
ALS Vial : 20 Sample Multiplier: 1

Instrument :
ECD_O
ClientSampleId :
AR1254ICC1000

Integration File signal 1: autoint1.e
Integration File signal 2: autoint2.e
Quant Time: Jun 12 05:33:38 2025
Quant Method : Z:\pestpcbsrv\HPCHEM1\ECD_O\methods\PO061125.M
Quant Title : GC EXTRACTABLES
QLast Update : Thu Jun 12 05:32:17 2025
Response via : Initial Calibration
Integrator: ChemStation

Volume Inj. : 2 µl
Signal #1 Phase : ZB-MR1 Signal #2 Phase: ZB-MR2
Signal #1 Info : 30Mx0.32mmx 0.50µm Signal #2 Info : 30M x 0.32mm x 0.25µm



Data Path : Z:\pestpcbsrv\HPCHEM1\ECD_O\Data\P0061125\
 Data File : P0111605.D
 Signal(s) : Signal #1: ECD1A.ch Signal #2: ECD2B.ch
 Acq On : 11 Jun 2025 16:25
 Operator : YP/AJ
 Sample : AR1254ICC750
 Misc :
 ALS Vial : 21 Sample Multiplier: 1

Instrument :
 ECD_O
 ClientSampleId :
 AR1254ICC750

Manual Integrations
 APPROVED

Reviewed By :Yogesh Patel 06/12/2025
 Supervised By :mohammad ahmed 06/13/2025

Integration File signal 1: autoint1.e
 Integration File signal 2: autoint2.e
 Quant Time: Jun 12 05:33:51 2025
 Quant Method : Z:\pestpcbsrv\HPCHEM1\ECD_O\methods\P0061125.M
 Quant Title : GC EXTRACTABLES
 QLast Update : Thu Jun 12 05:32:17 2025
 Response via : Initial Calibration
 Integrator: ChemStation

Volume Inj. : 2 µl
 Signal #1 Phase : ZB-MR1 Signal #2 Phase: ZB-MR2
 Signal #1 Info : 30Mx0.32mmx 0.50µ Signal #2 Info : 30M x 0.32mm x 0.25µm

Compound	RT#1	RT#2	Resp#1	Resp#2	ng/ml	ng/ml

System Monitoring Compounds						
1) SA Tetrachlo...	3.677	3.674	452.3E6	422.3E6	75.111m	73.899
2) SA Decachlor...	8.713	8.662	387.6E6	129.2E6	71.726	71.294
Target Compounds						
26) L6 AR-1254-1	5.572	5.550	292.8E6	227.9E6	715.883	714.287
27) L6 AR-1254-2	5.722	5.698	260.8E6	196.8E6	713.601	713.403
28) L6 AR-1254-3	6.126	6.099	406.3E6	300.2E6	716.678	719.918
29) L6 AR-1254-4	6.355	6.327	258.5E6	167.0E6	724.737	723.462
30) L6 AR-1254-5	6.775	6.744	368.9E6	231.0E6	720.516	713.515

(f)=RT Delta > 1/2 Window (#)=Amounts differ by > 25% (m)=manual int.

Data Path : Z:\pestpcbsrv\HPCHEM1\ECD_O\Data\PO061125\
Data File : PO111605.D
Signal(s) : Signal #1: ECD1A.ch Signal #2: ECD2B.ch
Acq On : 11 Jun 2025 16:25
Operator : YP/AJ
Sample : AR1254ICC750
Misc :
ALS Vial : 21 Sample Multiplier: 1



Instrument :
ECD_O
ClientSampleId :
AR1254ICC750

Manual Integrations
APPROVED

Reviewed By :Yogesh Patel 06/12/2025
Supervised By :mohammad ahmed 06/13/2025

Integration File signal 1: autoint1.e
Integration File signal 2: autoint2.e
Quant Time: Jun 12 05:33:51 2025
Quant Method : Z:\pestpcbsrv\HPCHEM1\ECD_O\methods\PO061125.M
Quant Title : GC EXTRACTABLES
QLast Update : Thu Jun 12 05:32:17 2025
Response via : Initial Calibration
Integrator: ChemStation

Volume Inj. : 2 µl
Signal #1 Phase : ZB-MR1 Signal #2 Phase: ZB-MR2
Signal #1 Info : 30Mx0.32mmx 0.50µm Signal #2 Info : 30M x 0.32mm x 0.25µm



Data Path : Z:\pestpcbsrv\HPCHEM1\ECD_O\Data\P0061125\
 Data File : P0111606.D
 Signal(s) : Signal #1: ECD1A.ch Signal #2: ECD2B.ch
 Acq On : 11 Jun 2025 16:43
 Operator : YP/AJ
 Sample : AR1254ICC500
 Misc :
 ALS Vial : 22 Sample Multiplier: 1

Instrument :
 ECD_O
 ClientSampleId :
 AR1254ICC500

Integration File signal 1: autoint1.e
 Integration File signal 2: autoint2.e
 Quant Time: Jun 12 05:34:02 2025
 Quant Method : Z:\pestpcbsrv\HPCHEM1\ECD_O\methods\P0061125.M
 Quant Title : GC EXTRACTABLES
 QLast Update : Thu Jun 12 05:32:17 2025
 Response via : Initial Calibration
 Integrator: ChemStation

Volume Inj. : 2 µl
 Signal #1 Phase : ZB-MR1 Signal #2 Phase: ZB-MR2
 Signal #1 Info : 30Mx0.32mmx 0.50µ Signal #2 Info : 30M x 0.32mm x 0.25µm

Compound	RT#1	RT#2	Resp#1	Resp#2	ng/ml	ng/ml

System Monitoring Compounds						
1) SA Tetrachlo...	3.677	3.673	301.1E6	285.7E6	50.000	50.000
2) SA Decachlor...	8.712	8.661	270.2E6	90639373	50.000	50.000
Target Compounds						
26) L6 AR-1254-1	5.572	5.549	204.5E6	159.6E6	500.000	500.000
27) L6 AR-1254-2	5.722	5.697	182.7E6	137.9E6	500.000	500.000
28) L6 AR-1254-3	6.126	6.098	283.5E6	208.5E6	500.000	500.000
29) L6 AR-1254-4	6.356	6.325	178.4E6	115.4E6	500.000	500.000
30) L6 AR-1254-5	6.775	6.742	256.0E6	161.9E6	500.000	500.000



(f)=RT Delta > 1/2 Window (#)=Amounts differ by > 25% (m)=manual int.

Data Path : Z:\pestpcbsrv\HPCHEM1\ECD_O\Data\PO061125\
Data File : PO111606.D
Signal(s) : Signal #1: ECD1A.ch Signal #2: ECD2B.ch
Acq On : 11 Jun 2025 16:43
Operator : YP/AJ
Sample : AR1254ICC500
Misc :
ALS Vial : 22 Sample Multiplier: 1

Instrument :
ECD_O
ClientSampleId :
AR1254ICC500

Integration File signal 1: autoint1.e
Integration File signal 2: autoint2.e
Quant Time: Jun 12 05:34:02 2025
Quant Method : Z:\pestpcbsrv\HPCHEM1\ECD_O\methods\PO061125.M
Quant Title : GC EXTRACTABLES
QLast Update : Thu Jun 12 05:32:17 2025
Response via : Initial Calibration
Integrator: ChemStation

Volume Inj. : 2 µl
Signal #1 Phase : ZB-MR1 Signal #2 Phase: ZB-MR2
Signal #1 Info : 30Mx0.32mmx 0.50µm Signal #2 Info : 30M x 0.32mm x 0.25µm



Data Path : Z:\pestpcbsrv\HPCHEM1\ECD_O\Data\P0061125\
 Data File : P0111607.D
 Signal(s) : Signal #1: ECD1A.ch Signal #2: ECD2B.ch
 Acq On : 11 Jun 2025 17:00
 Operator : YP/AJ
 Sample : AR1254ICC250
 Misc :
 ALS Vial : 23 Sample Multiplier: 1

Instrument :
 ECD_O
 ClientSampleId :
 AR1254ICC250

Integration File signal 1: autoint1.e
 Integration File signal 2: autoint2.e
 Quant Time: Jun 12 05:34:13 2025
 Quant Method : Z:\pestpcbsrv\HPCHEM1\ECD_O\methods\P0061125.M
 Quant Title : GC EXTRACTABLES
 QLast Update : Thu Jun 12 05:32:17 2025
 Response via : Initial Calibration
 Integrator: ChemStation

Volume Inj. : 2 µl
 Signal #1 Phase : ZB-MR1 Signal #2 Phase: ZB-MR2
 Signal #1 Info : 30Mx0.32mmx 0.50µ Signal #2 Info : 30M x 0.32mm x 0.25µm

Compound	RT#1	RT#2	Resp#1	Resp#2	ng/ml	ng/ml

System Monitoring Compounds						
1) SA Tetrachlo...	3.676	3.674	148.2E6	137.3E6	24.604	24.025
2) SA Decachlor...	8.712	8.661	134.5E6	44871339	24.893	24.753
Target Compounds						
26) L6 AR-1254-1	5.571	5.550	104.6E6	80140407	255.667	251.127
27) L6 AR-1254-2	5.720	5.697	93641034	69312951	256.223	251.306
28) L6 AR-1254-3	6.125	6.099	141.7E6	102.9E6	249.964	246.773
29) L6 AR-1254-4	6.354	6.327	88420478	57943475	247.881	250.992
30) L6 AR-1254-5	6.774	6.743	126.2E6	79955774	246.527	246.939



(f)=RT Delta > 1/2 Window (#)=Amounts differ by > 25% (m)=manual int.

Data Path : Z:\pestpcbsrv\HPCHEM1\ECD_O\Data\PO061125\
Data File : PO111607.D
Signal(s) : Signal #1: ECD1A.ch Signal #2: ECD2B.ch
Acq On : 11 Jun 2025 17:00
Operator : YP/AJ
Sample : AR1254ICC250
Misc :
ALS Vial : 23 Sample Multiplier: 1

Instrument :
ECD_O
ClientSampleId :
AR1254ICC250

Integration File signal 1: autoint1.e
Integration File signal 2: autoint2.e
Quant Time: Jun 12 05:34:13 2025
Quant Method : Z:\pestpcbsrv\HPCHEM1\ECD_O\methods\PO061125.M
Quant Title : GC EXTRACTABLES
QLast Update : Thu Jun 12 05:32:17 2025
Response via : Initial Calibration
Integrator: ChemStation

Volume Inj. : 2 µl
Signal #1 Phase : ZB-MR1 Signal #2 Phase: ZB-MR2
Signal #1 Info : 30Mx0.32mmx 0.50µm Signal #2 Info : 30M x 0.32mm x 0.25µm



Data Path : Z:\pestpcbsrv\HPCHEM1\ECD_O\Data\P0061125\
 Data File : P0111608.D
 Signal(s) : Signal #1: ECD1A.ch Signal #2: ECD2B.ch
 Acq On : 11 Jun 2025 17:18
 Operator : YP/AJ
 Sample : AR1254ICC050
 Misc :
 ALS Vial : 24 Sample Multiplier: 1

Instrument :
 ECD_O
 ClientSampleId :
 AR1254ICC050

Integration File signal 1: autoint1.e
 Integration File signal 2: autoint2.e
 Quant Time: Jun 12 05:34:28 2025
 Quant Method : Z:\pestpcbsrv\HPCHEM1\ECD_O\methods\P0061125.M
 Quant Title : GC EXTRACTABLES
 QLast Update : Thu Jun 12 05:32:17 2025
 Response via : Initial Calibration
 Integrator: ChemStation

Volume Inj. : 2 µl
 Signal #1 Phase : ZB-MR1 Signal #2 Phase: ZB-MR2
 Signal #1 Info : 30Mx0.32mmx 0.50µ Signal #2 Info : 30M x 0.32mm x 0.25µm

Compound	RT#1	RT#2	Resp#1	Resp#2	ng/ml	ng/ml

System Monitoring Compounds						
1) SA Tetrachlo...	3.675	3.673	33268586	29021554	5.525	5.078
2) SA Decachlor...	8.711	8.661	30958063	10135764	5.729	5.591
Target Compounds						
26) L6 AR-1254-1	5.570	5.550	26611811	18724590	65.057	58.675
27) L6 AR-1254-2	5.719	5.697	23766283	16284924	65.030	59.044
28) L6 AR-1254-3	6.124	6.098	32638345	23686717	57.564	56.812
29) L6 AR-1254-4	6.352	6.326	17926220	11952540	50.255	51.774
30) L6 AR-1254-5	6.772	6.743	28538450	18744477	55.741	57.891



(f)=RT Delta > 1/2 Window (#)=Amounts differ by > 25% (m)=manual int.

Data Path : Z:\pestpcbsrv\HPCHEM1\ECD_O\Data\PO061125\
Data File : PO111608.D
Signal(s) : Signal #1: ECD1A.ch Signal #2: ECD2B.ch
Acq On : 11 Jun 2025 17:18
Operator : YP/AJ
Sample : AR1254ICC050
Misc :
ALS Vial : 24 Sample Multiplier: 1

Instrument :
ECD_O
ClientSampleId :
AR1254ICC050

Integration File signal 1: autoint1.e
Integration File signal 2: autoint2.e
Quant Time: Jun 12 05:34:28 2025
Quant Method : Z:\pestpcbsrv\HPCHEM1\ECD_O\methods\PO061125.M
Quant Title : GC EXTRACTABLES
QLast Update : Thu Jun 12 05:32:17 2025
Response via : Initial Calibration
Integrator: ChemStation

Volume Inj. : 2 µl
Signal #1 Phase : ZB-MR1 Signal #2 Phase: ZB-MR2
Signal #1 Info : 30Mx0.32mmx 0.50µ Signal #2 Info : 30M x 0.32mm x 0.25µm



Data Path : Z:\pestpcbsrv\HPCHEM1\ECD_O\Data\P0061125\
 Data File : P0111609.D
 Signal(s) : Signal #1: ECD1A.ch Signal #2: ECD2B.ch
 Acq On : 11 Jun 2025 17:36
 Operator : YP/AJ
 Sample : AR1262ICC500
 Misc :
 ALS Vial : 25 Sample Multiplier: 1

Instrument :
 ECD_O
 ClientSampleId :
 AR1262ICC500

Integration File signal 1: autoint1.e
 Integration File signal 2: autoint2.e
 Quant Time: Jun 12 05:47:34 2025
 Quant Method : Z:\pestpcbsrv\HPCHEM1\ECD_O\methods\P0061125.M
 Quant Title : GC EXTRACTABLES
 QLast Update : Thu Jun 12 05:42:57 2025
 Response via : Initial Calibration
 Integrator: ChemStation

Volume Inj. : 2 µl
 Signal #1 Phase : ZB-MR1 Signal #2 Phase: ZB-MR2
 Signal #1 Info : 30Mx0.32mmx 0.50µ Signal #2 Info : 30M x 0.32mm x 0.25µm

Compound	RT#1	RT#2	Resp#1	Resp#2	ng/ml	ng/ml

System Monitoring Compounds						
1) SA Tetrachlo...	3.678	3.674	306.3E6	279.8E6	50.000	50.000
2) SA Decachlor...	8.713	8.663	265.9E6	88744221	50.000	50.000
Target Compounds						
36) L8 AR-1262-1	6.815	6.782	304.9E6	168.8E6	500.000	500.000
37) L8 AR-1262-2	7.317	7.281	470.7E6	232.9E6	500.000	500.000
38) L8 AR-1262-3	7.601	7.564	198.3E6	83142550	500.000	500.000
39) L8 AR-1262-4	7.664	7.628	328.5E6	138.5E6	500.000	500.000
40) L8 AR-1262-5	8.161	8.120	152.3E6	50809689	500.000	500.000



(f)=RT Delta > 1/2 Window (#)=Amounts differ by > 25% (m)=manual int.

Data Path : Z:\pestpcbsrv\HPCHEM1\ECD_O\Data\PO061125\
Data File : PO111609.D
Signal(s) : Signal #1: ECD1A.ch Signal #2: ECD2B.ch
Acq On : 11 Jun 2025 17:36
Operator : YP/AJ
Sample : AR1262ICC500
Misc :
ALS Vial : 25 Sample Multiplier: 1

Instrument :
ECD_O
ClientSampleId :
AR1262ICC500

Integration File signal 1: autoint1.e
Integration File signal 2: autoint2.e
Quant Time: Jun 12 05:47:34 2025
Quant Method : Z:\pestpcbsrv\HPCHEM1\ECD_O\methods\PO061125.M
Quant Title : GC EXTRACTABLES
QLast Update : Thu Jun 12 05:42:57 2025
Response via : Initial Calibration
Integrator: ChemStation

Volume Inj. : 2 µl
Signal #1 Phase : ZB-MR1 Signal #2 Phase: ZB-MR2
Signal #1 Info : 30Mx0.32mmx 0.50µ Signal #2 Info : 30M x 0.32mm x 0.25µm



Data Path : Z:\pestpcbsrv\HPCHEM1\ECD_O\Data\P0061125\
 Data File : P0111610.D
 Signal(s) : Signal #1: ECD1A.ch Signal #2: ECD2B.ch
 Acq On : 11 Jun 2025 17:55
 Operator : YP/AJ
 Sample : AR1268ICC1000
 Misc :
 ALS Vial : 26 Sample Multiplier: 1

Instrument :
 ECD_O
 ClientSampleId :
 AR1268ICC1000

Manual Integrations
 APPROVED

Reviewed By :Yogesh Patel 06/12/2025
 Supervised By :mohammad ahmed 06/13/2025

Integration File signal 1: autoint1.e
 Integration File signal 2: autoint2.e
 Quant Time: Jun 12 06:02:07 2025
 Quant Method : Z:\pestpcbsrv\HPCHEM1\ECD_O\methods\P0061125.M
 Quant Title : GC EXTRACTABLES
 QLast Update : Thu Jun 12 05:57:06 2025
 Response via : Initial Calibration
 Integrator: ChemStation

Volume Inj. : 2 µl
 Signal #1 Phase : ZB-MR1 Signal #2 Phase: ZB-MR2
 Signal #1 Info : 30Mx0.32mmx 0.50µ Signal #2 Info : 30M x 0.32mm x 0.25µm

Compound	RT#1	RT#2	Resp#1	Resp#2	ng/ml	ng/ml

System Monitoring Compounds						
1) SA Tetrachlo...	3.678	3.674	594.4E6	555.7E6	94.817	97.557
2) SA Decachlor...	8.713	8.662	882.0E6	294.4E6	94.438	95.085
Target Compounds						
41) L9 AR-1268-1	7.601	7.564	1033.9E6	439.2E6	942.617	959.115m
42) L9 AR-1268-2	7.666	7.629	906.0E6	386.0E6	949.415	965.208
43) L9 AR-1268-3	7.871	7.834	760.3E6	289.6E6	950.113	950.928
44) L9 AR-1268-4	8.161	8.120	324.7E6	108.4E6	941.712	959.983
45) L9 AR-1268-5	8.456	8.411	2115.2E6	686.4E6	967.063	971.678

(f)=RT Delta > 1/2 Window (#)=Amounts differ by > 25% (m)=manual int.

Data Path : Z:\pestpcbsrv\HPCHEM1\ECD_O\Data\PO061125\
Data File : PO111610.D
Signal(s) : Signal #1: ECD1A.ch Signal #2: ECD2B.ch
Acq On : 11 Jun 2025 17:55
Operator : YP/AJ
Sample : AR1268ICC1000
Misc :
ALS Vial : 26 Sample Multiplier: 1



Instrument :
ECD_O
ClientSampleId :
AR1268ICC1000

Manual Integrations
APPROVED

Reviewed By :Yogesh Patel 06/12/2025
Supervised By :mohammad ahmed 06/13/2025

Integration File signal 1: autoint1.e
Integration File signal 2: autoint2.e
Quant Time: Jun 12 06:02:07 2025
Quant Method : Z:\pestpcbsrv\HPCHEM1\ECD_O\methods\PO061125.M
Quant Title : GC EXTRACTABLES
QLast Update : Thu Jun 12 05:57:06 2025
Response via : Initial Calibration
Integrator: ChemStation

Volume Inj. : 2 µl
Signal #1 Phase : ZB-MR1 Signal #2 Phase: ZB-MR2
Signal #1 Info : 30Mx0.32mmx 0.50µm Signal #2 Info : 30M x 0.32mm x 0.25µm



Data Path : Z:\pestpcbsrv\HPCHEM1\ECD_O\Data\P0061125\
 Data File : P0111611.D
 Signal(s) : Signal #1: ECD1A.ch Signal #2: ECD2B.ch
 Acq On : 11 Jun 2025 18:13
 Operator : YP/AJ
 Sample : AR1268ICC750
 Misc :
 ALS Vial : 27 Sample Multiplier: 1

Instrument :
 ECD_O
 ClientSampleId :
 AR1268ICC750

Integration File signal 1: autoint1.e
 Integration File signal 2: autoint2.e
 Quant Time: Jun 12 06:02:19 2025
 Quant Method : Z:\pestpcbsrv\HPCHEM1\ECD_O\methods\P0061125.M
 Quant Title : GC EXTRACTABLES
 QLast Update : Thu Jun 12 05:57:06 2025
 Response via : Initial Calibration
 Integrator: ChemStation

Volume Inj. : 2 µl
 Signal #1 Phase : ZB-MR1 Signal #2 Phase: ZB-MR2
 Signal #1 Info : 30Mx0.32mmx 0.50µ Signal #2 Info : 30M x 0.32mm x 0.25µm

Compound	RT#1	RT#2	Resp#1	Resp#2	ng/ml	ng/ml

System Monitoring Compounds						
1) SA Tetrachlo...	3.678	3.674	455.9E6	422.3E6	72.724	74.142
2) SA Decachlor...	8.714	8.662	675.4E6	224.8E6	72.317	72.615
Target Compounds						
41) L9 AR-1268-1	7.601	7.564	793.7E6	333.1E6	723.633	727.520
42) L9 AR-1268-2	7.665	7.629	689.0E6	292.9E6	722.004	732.523
43) L9 AR-1268-3	7.872	7.834	582.6E6	221.2E6	728.087	726.286
44) L9 AR-1268-4	8.160	8.120	245.1E6	81834207	710.765	725.030
45) L9 AR-1268-5	8.456	8.411	1601.8E6	521.1E6	732.335	737.736



(f)=RT Delta > 1/2 Window (#)=Amounts differ by > 25% (m)=manual int.

Data Path : Z:\pestpcbsrv\HPCHEM1\ECD_O\Data\PO061125\
Data File : PO111611.D
Signal(s) : Signal #1: ECD1A.ch Signal #2: ECD2B.ch
Acq On : 11 Jun 2025 18:13
Operator : YP/AJ
Sample : AR1268ICC750
Misc :
ALS Vial : 27 Sample Multiplier: 1

Instrument :
ECD_O
ClientSampleId :
AR1268ICC750

Integration File signal 1: autoint1.e
Integration File signal 2: autoint2.e
Quant Time: Jun 12 06:02:19 2025
Quant Method : Z:\pestpcbsrv\HPCHEM1\ECD_O\methods\PO061125.M
Quant Title : GC EXTRACTABLES
QLast Update : Thu Jun 12 05:57:06 2025
Response via : Initial Calibration
Integrator: ChemStation

Volume Inj. : 2 µl
Signal #1 Phase : ZB-MR1 Signal #2 Phase: ZB-MR2
Signal #1 Info : 30Mx0.32mmx 0.50µm Signal #2 Info : 30M x 0.32mm x 0.25µm



Data Path : Z:\pestpcbsrv\HPCHEM1\ECD_O\Data\P0061125\
 Data File : P0111612.D
 Signal(s) : Signal #1: ECD1A.ch Signal #2: ECD2B.ch
 Acq On : 11 Jun 2025 18:31
 Operator : YP/AJ
 Sample : AR1268ICC500
 Misc :
 ALS Vial : 28 Sample Multiplier: 1

Instrument :
 ECD_O
 ClientSampleId :
 AR1268ICC500

Integration File signal 1: autoint1.e
 Integration File signal 2: autoint2.e
 Quant Time: Jun 12 06:02:32 2025
 Quant Method : Z:\pestpcbsrv\HPCHEM1\ECD_O\methods\P0061125.M
 Quant Title : GC EXTRACTABLES
 QLast Update : Thu Jun 12 05:57:06 2025
 Response via : Initial Calibration
 Integrator: ChemStation

Volume Inj. : 2 µl
 Signal #1 Phase : ZB-MR1 Signal #2 Phase: ZB-MR2
 Signal #1 Info : 30Mx0.32mmx 0.50µ Signal #2 Info : 30M x 0.32mm x 0.25µm

Compound	RT#1	RT#2	Resp#1	Resp#2	ng/ml	ng/ml

System Monitoring Compounds						
1) SA Tetrachlo...	3.678	3.674	313.5E6	284.8E6	50.000	50.000
2) SA Decachlor...	8.714	8.663	467.0E6	154.8E6	50.000	50.000
Target Compounds						
41) L9 AR-1268-1	7.601	7.564	548.4E6	229.0E6	500.000	500.000
42) L9 AR-1268-2	7.665	7.629	477.2E6	200.0E6	500.000	500.000
43) L9 AR-1268-3	7.872	7.834	400.1E6	152.3E6	500.000	500.000
44) L9 AR-1268-4	8.161	8.121	172.4E6	56435057	500.000	500.000
45) L9 AR-1268-5	8.455	8.411	1093.6E6	353.2E6	500.000	500.000



(f)=RT Delta > 1/2 Window (#)=Amounts differ by > 25% (m)=manual int.

Data Path : Z:\pestpcbsrv\HPCHEM1\ECD_O\Data\PO061125\
Data File : PO111612.D
Signal(s) : Signal #1: ECD1A.ch Signal #2: ECD2B.ch
Acq On : 11 Jun 2025 18:31
Operator : YP/AJ
Sample : AR1268ICC500
Misc :
ALS Vial : 28 Sample Multiplier: 1

Instrument :
ECD_O
ClientSampleId :
AR1268ICC500

Integration File signal 1: autoint1.e
Integration File signal 2: autoint2.e
Quant Time: Jun 12 06:02:32 2025
Quant Method : Z:\pestpcbsrv\HPCHEM1\ECD_O\methods\PO061125.M
Quant Title : GC EXTRACTABLES
QLast Update : Thu Jun 12 05:57:06 2025
Response via : Initial Calibration
Integrator: ChemStation

Volume Inj. : 2 µl
Signal #1 Phase : ZB-MR1 Signal #2 Phase: ZB-MR2
Signal #1 Info : 30Mx0.32mmx 0.50µm Signal #2 Info : 30M x 0.32mm x 0.25µm



Data Path : Z:\pestpcbsrv\HPCHEM1\ECD_O\Data\P0061125\
 Data File : P0111613.D
 Signal(s) : Signal #1: ECD1A.ch Signal #2: ECD2B.ch
 Acq On : 11 Jun 2025 18:50
 Operator : YP/AJ
 Sample : AR1268ICC250
 Misc :
 ALS Vial : 29 Sample Multiplier: 1

Instrument :
 ECD_O
 ClientSampleId :
 AR1268ICC250

Manual Integrations
 APPROVED

Reviewed By :Yogesh Patel 06/12/2025
 Supervised By :mohammad ahmed 06/13/2025

Integration File signal 1: autoint1.e
 Integration File signal 2: autoint2.e
 Quant Time: Jun 12 06:02:44 2025
 Quant Method : Z:\pestpcbsrv\HPCHEM1\ECD_O\methods\P0061125.M
 Quant Title : GC EXTRACTABLES
 QLast Update : Thu Jun 12 05:57:06 2025
 Response via : Initial Calibration
 Integrator: ChemStation

Volume Inj. : 2 µl
 Signal #1 Phase : ZB-MR1 Signal #2 Phase: ZB-MR2
 Signal #1 Info : 30Mx0.32mmx 0.50µ Signal #2 Info : 30M x 0.32mm x 0.25µm

Compound	RT#1	RT#2	Resp#1	Resp#2	ng/ml	ng/ml

System Monitoring Compounds						
1) SA Tetrachlo...	3.678	3.674	163.0E6	145.2E6	26.003	25.485
2) SA Decachlor...	8.713	8.663	247.2E6	81962535	26.465	26.471
Target Compounds						
41) L9 AR-1268-1	7.600	7.563	288.8E6	121.2E6	263.299	264.598m
42) L9 AR-1268-2	7.666	7.628	247.3E6	105.6E6	259.148	264.128
43) L9 AR-1268-3	7.872	7.833	210.9E6	81168400	263.556	266.540
44) L9 AR-1268-4	8.161	8.119	89176833	29806035	258.652	264.074m
45) L9 AR-1268-5	8.456	8.410	563.6E6	183.9E6	257.688	260.275

(f)=RT Delta > 1/2 Window (#)=Amounts differ by > 25% (m)=manual int.

Data Path : Z:\pestpcbsrv\HPCHEM1\ECD_O\Data\PO061125\
Data File : PO111613.D
Signal(s) : Signal #1: ECD1A.ch Signal #2: ECD2B.ch
Acq On : 11 Jun 2025 18:50
Operator : YP/AJ
Sample : AR1268ICC250
Misc :
ALS Vial : 29 Sample Multiplier: 1



Instrument :
ECD_O
ClientSampleId :
AR1268ICC250

Manual Integrations
APPROVED

Reviewed By :Yogesh Patel 06/12/2025
Supervised By :mohammad ahmed 06/13/2025

Integration File signal 1: autoint1.e
Integration File signal 2: autoint2.e
Quant Time: Jun 12 06:02:44 2025
Quant Method : Z:\pestpcbsrv\HPCHEM1\ECD_O\methods\PO061125.M
Quant Title : GC EXTRACTABLES
QLast Update : Thu Jun 12 05:57:06 2025
Response via : Initial Calibration
Integrator: ChemStation

Volume Inj. : 2 µl
Signal #1 Phase : ZB-MR1 Signal #2 Phase: ZB-MR2
Signal #1 Info : 30Mx0.32mmx 0.50µ Signal #2 Info : 30M x 0.32mm x 0.25µm



Data Path : Z:\pestpcbsrv\HPCHEM1\ECD_O\Data\P0061125\
 Data File : P0111614.D
 Signal(s) : Signal #1: ECD1A.ch Signal #2: ECD2B.ch
 Acq On : 11 Jun 2025 19:07
 Operator : YP/AJ
 Sample : AR1268ICC050
 Misc :
 ALS Vial : 30 Sample Multiplier: 1

Instrument :
 ECD_O
 ClientSampleId :
 AR1268ICC050

Integration File signal 1: autoint1.e
 Integration File signal 2: autoint2.e
 Quant Time: Jun 12 06:02:56 2025
 Quant Method : Z:\pestpcbsrv\HPCHEM1\ECD_O\methods\P0061125.M
 Quant Title : GC EXTRACTABLES
 QLast Update : Thu Jun 12 05:57:06 2025
 Response via : Initial Calibration
 Integrator: ChemStation

Volume Inj. : 2 µl
 Signal #1 Phase : ZB-MR1 Signal #2 Phase: ZB-MR2
 Signal #1 Info : 30Mx0.32mmx 0.50µ Signal #2 Info : 30M x 0.32mm x 0.25µm

Compound	RT#1	RT#2	Resp#1	Resp#2	ng/ml	ng/ml

System Monitoring Compounds						
1) SA Tetrachlo...	3.676	3.673	33978396	28591895	5.420	5.019
2) SA Decachlor...	8.711	8.662	53020278	17366893	5.677	5.609
Target Compounds						
41) L9 AR-1268-1	7.599	7.563	64431354	28304378	58.742	61.812
42) L9 AR-1268-2	7.664	7.628	55894407	24176130	58.570	60.455
43) L9 AR-1268-3	7.870	7.833	45853850	18553952	57.300	60.927
44) L9 AR-1268-4	8.159	8.120	18350605	6436315	53.225	57.024
45) L9 AR-1268-5	8.454	8.410	120.5E6	40005648	55.081	56.632



(f)=RT Delta > 1/2 Window (#)=Amounts differ by > 25% (m)=manual int.

Data Path : Z:\pestpcbsrv\HPCHEM1\ECD_O\Data\PO061125\
Data File : PO111614.D
Signal(s) : Signal #1: ECD1A.ch Signal #2: ECD2B.ch
Acq On : 11 Jun 2025 19:07
Operator : YP/AJ
Sample : AR1268ICC050
Misc :
ALS Vial : 30 Sample Multiplier: 1

Instrument :
ECD_O
ClientSampleId :
AR1268ICC050

Integration File signal 1: autoint1.e
Integration File signal 2: autoint2.e
Quant Time: Jun 12 06:02:56 2025
Quant Method : Z:\pestpcbsrv\HPCHEM1\ECD_O\methods\PO061125.M
Quant Title : GC EXTRACTABLES
QLast Update : Thu Jun 12 05:57:06 2025
Response via : Initial Calibration
Integrator: ChemStation

Volume Inj. : 2 µl
Signal #1 Phase : ZB-MR1 Signal #2 Phase: ZB-MR2
Signal #1 Info : 30Mx0.32mmx 0.50µm Signal #2 Info : 30M x 0.32mm x 0.25µm



Data Path : Z:\pestpcbsrv\HPCHEM1\ECD_O\Data\P0061125\
 Data File : P0111615.D
 Signal(s) : Signal #1: ECD1A.ch Signal #2: ECD2B.ch
 Acq On : 11 Jun 2025 19:25
 Operator : YP/AJ
 Sample : P0061125ICV500
 Misc :
 ALS Vial : 31 Sample Multiplier: 1

Instrument :
 ECD_O
 ClientSampleId :
 ICVPO061125

Integration File signal 1: autoint1.e
 Integration File signal 2: autoint2.e
 Quant Time: Jun 12 06:32:22 2025
 Quant Method : Z:\pestpcbsrv\HPCHEM1\ECD_O\methods\P0061125.M
 Quant Title : GC EXTRACTABLES
 QLast Update : Thu Jun 12 06:25:26 2025
 Response via : Initial Calibration
 Integrator: ChemStation

Volume Inj. : 2 µl
 Signal #1 Phase : ZB-MR1 Signal #2 Phase: ZB-MR2
 Signal #1 Info : 30Mx0.32mmx 0.50µ Signal #2 Info : 30M x 0.32mm x 0.25µm

Compound	RT#1	RT#2	Resp#1	Resp#2	ng/ml	ng/ml

System Monitoring Compounds						
1) SA Tetrachlo...	3.678	3.674	312.8E6	284.5E6	54.313	50.649
2) SA Decachlor...	8.714	8.663	262.1E6	87482159	49.936	49.212
Target Compounds						
3) L1 AR-1016-1	4.767	4.751	120.8E6	96122372	506.366	488.508
4) L1 AR-1016-2	4.785	4.769	171.7E6	140.6E6	516.057	493.691
5) L1 AR-1016-3	4.842	4.945	119.3E6	74902189	518.404	498.552
6) L1 AR-1016-4	4.962	4.987	94987891	60498359	512.167	496.754
7) L1 AR-1016-5	5.219	5.199	98032616	79010871	525.240	496.304
31) L7 AR-1260-1	6.258	6.230	174.2E6	124.2E6	490.145	494.961
32) L7 AR-1260-2	6.448	6.417	235.7E6	144.3E6	498.933	490.854
33) L7 AR-1260-3	6.814	6.570	213.4E6	130.1E6	502.477	487.032
34) L7 AR-1260-4	7.074	7.040	158.0E6	93232919	502.582	490.451
35) L7 AR-1260-5	7.316	7.281	408.9E6	206.3E6	496.383	485.361



(f)=RT Delta > 1/2 Window (#)=Amounts differ by > 25% (m)=manual int.

Data Path : Z:\pestpcbsrv\HPCHEM1\ECD_O\Data\PO061125\
Data File : PO111615.D
Signal(s) : Signal #1: ECD1A.ch Signal #2: ECD2B.ch
Acq On : 11 Jun 2025 19:25
Operator : YP/AJ
Sample : PO061125ICV500
Misc :
ALS Vial : 31 Sample Multiplier: 1

Instrument :
ECD_O
ClientSampleId :
ICVPO061125

Integration File signal 1: autoint1.e
Integration File signal 2: autoint2.e
Quant Time: Jun 12 06:32:22 2025
Quant Method : Z:\pestpcbsrv\HPCHEM1\ECD_O\methods\PO061125.M
Quant Title : GC EXTRACTABLES
QLast Update : Thu Jun 12 06:25:26 2025
Response via : Initial Calibration
Integrator: ChemStation

Volume Inj. : 2 µl
Signal #1 Phase : ZB-MR1 Signal #2 Phase: ZB-MR2
Signal #1 Info : 30Mx0.32mmx 0.50µm Signal #2 Info : 30M x 0.32mm x 0.25µm



Data Path : Z:\pestpcbsrv\HPCHEM1\ECD_O\Data\P0061125\
 Data File : P011616.D
 Signal(s) : Signal #1: ECD1A.ch Signal #2: ECD2B.ch
 Acq On : 11 Jun 2025 20:02
 Operator : YP/AJ
 Sample : AR1242ICV500
 Misc :
 ALS Vial : 32 Sample Multiplier: 1

Instrument :
 ECD_O
 ClientSampleId :
 ICVPO061125

Integration File signal 1: autoint1.e
 Integration File signal 2: autoint2.e
 Quant Time: Jun 12 06:32:45 2025
 Quant Method : Z:\pestpcbsrv\HPCHEM1\ECD_O\methods\P0061125.M
 Quant Title : GC EXTRACTABLES
 QLast Update : Thu Jun 12 06:25:26 2025
 Response via : Initial Calibration
 Integrator: ChemStation

Volume Inj. : 2 µl
 Signal #1 Phase : ZB-MR1 Signal #2 Phase: ZB-MR2
 Signal #1 Info : 30Mx0.32mmx 0.50µ Signal #2 Info : 30M x 0.32mm x 0.25µm

Compound	RT#1	RT#2	Resp#1	Resp#2	ng/ml	ng/ml

System Monitoring Compounds						
1) SA Tetrachlo...	3.678	3.674	317.8E6	290.0E6	55.185	51.631
2) SA Decachlor...	8.714	8.662	265.8E6	87982650	50.639	49.493
Target Compounds						
16) L4 AR-1242-1	4.767	4.751	103.3E6	83976174	512.987	515.952
17) L4 AR-1242-2	4.785	4.770	152.5E6	122.9E6	541.110	520.666
18) L4 AR-1242-3	4.842	4.945	104.8E6	65447268	527.329	521.121
19) L4 AR-1242-4	4.962	5.029	83763653	64008885	529.489	522.271
20) L4 AR-1242-5	5.615	5.549	88172283	78363974	527.295	522.812



(f)=RT Delta > 1/2 Window (#)=Amounts differ by > 25% (m)=manual int.

Data Path : Z:\pestpcbsrv\HPCHEM1\ECD_O\Data\PO061125\
Data File : PO111616.D
Signal(s) : Signal #1: ECD1A.ch Signal #2: ECD2B.ch
Acq On : 11 Jun 2025 20:02
Operator : YP/AJ
Sample : AR1242ICV500
Misc :
ALS Vial : 32 Sample Multiplier: 1

Instrument :
ECD_O
ClientSampleId :
ICVPO061125

Integration File signal 1: autoint1.e
Integration File signal 2: autoint2.e
Quant Time: Jun 12 06:32:45 2025
Quant Method : Z:\pestpcbsrv\HPCHEM1\ECD_O\methods\PO061125.M
Quant Title : GC EXTRACTABLES
QLast Update : Thu Jun 12 06:25:26 2025
Response via : Initial Calibration
Integrator: ChemStation

Volume Inj. : 2 µl
Signal #1 Phase : ZB-MR1 Signal #2 Phase: ZB-MR2
Signal #1 Info : 30Mx0.32mmx 0.50µm Signal #2 Info : 30M x 0.32mm x 0.25µm



Data Path : Z:\pestpcbsrv\HPCHEM1\ECD_O\Data\P0061125\
 Data File : P011617.D
 Signal(s) : Signal #1: ECD1A.ch Signal #2: ECD2B.ch
 Acq On : 11 Jun 2025 20:39
 Operator : YP/AJ
 Sample : AR1248ICV500
 Misc :
 ALS Vial : 33 Sample Multiplier: 1

Instrument :
 ECD_O
 ClientSampleId :
 ICVPO061125

Integration File signal 1: autoint1.e
 Integration File signal 2: autoint2.e
 Quant Time: Jun 12 06:33:03 2025
 Quant Method : Z:\pestpcbsrv\HPCHEM1\ECD_O\methods\P0061125.M
 Quant Title : GC EXTRACTABLES
 QLast Update : Thu Jun 12 06:25:26 2025
 Response via : Initial Calibration
 Integrator: ChemStation

Volume Inj. : 2 µl
 Signal #1 Phase : ZB-MR1 Signal #2 Phase: ZB-MR2
 Signal #1 Info : 30Mx0.32mmx 0.50µ Signal #2 Info : 30M x 0.32mm x 0.25µm

Compound	RT#1	RT#2	Resp#1	Resp#2	ng/ml	ng/ml

System Monitoring Compounds						
1) SA Tetrachlo...	3.678	3.673	308.6E6	278.4E6	53.581	49.556
2) SA Decachlor...	8.714	8.662	263.4E6	87190525	50.183	49.048
Target Compounds						
21) L5 AR-1248-1	4.766	4.751	78506668	63861888	486.816	479.140
22) L5 AR-1248-2	5.005	4.987	105.2E6	86852556	487.818	487.510
23) L5 AR-1248-3	5.219	5.029	133.1E6	91125563	493.349	481.031
24) L5 AR-1248-4	5.573	5.199	195.1E6	107.8E6	488.041	472.328
25) L5 AR-1248-5	5.614	5.590	137.6E6	104.3E6	490.112	479.030



(f)=RT Delta > 1/2 Window (#)=Amounts differ by > 25% (m)=manual int.

Data Path : Z:\pestpcbsrv\HPCHEM1\ECD_O\Data\PO061125\
Data File : PO111617.D
Signal(s) : Signal #1: ECD1A.ch Signal #2: ECD2B.ch
Acq On : 11 Jun 2025 20:39
Operator : YP/AJ
Sample : AR1248ICV500
Misc :
ALS Vial : 33 Sample Multiplier: 1

Instrument :
ECD_O
ClientSampleId :
ICVPO061125

Integration File signal 1: autoint1.e
Integration File signal 2: autoint2.e
Quant Time: Jun 12 06:33:03 2025
Quant Method : Z:\pestpcbsrv\HPCHEM1\ECD_O\methods\PO061125.M
Quant Title : GC EXTRACTABLES
QLast Update : Thu Jun 12 06:25:26 2025
Response via : Initial Calibration
Integrator: ChemStation

Volume Inj. : 2 µl
Signal #1 Phase : ZB-MR1 Signal #2 Phase: ZB-MR2
Signal #1 Info : 30Mx0.32mmx 0.50µm Signal #2 Info : 30M x 0.32mm x 0.25µm



Data Path : Z:\pestpcbsrv\HPCHEM1\ECD_O\Data\P0061125\
 Data File : P011618.D
 Signal(s) : Signal #1: ECD1A.ch Signal #2: ECD2B.ch
 Acq On : 11 Jun 2025 21:16
 Operator : YP/AJ
 Sample : AR1254ICV500
 Misc :
 ALS Vial : 34 Sample Multiplier: 1

Instrument :
 ECD_O
 ClientSampleId :
 ICVPO061125

Integration File signal 1: autoint1.e
 Integration File signal 2: autoint2.e
 Quant Time: Jun 12 06:33:21 2025
 Quant Method : Z:\pestpcbsrv\HPCHEM1\ECD_O\methods\P0061125.M
 Quant Title : GC EXTRACTABLES
 QLast Update : Thu Jun 12 06:25:26 2025
 Response via : Initial Calibration
 Integrator: ChemStation

Volume Inj. : 2 µl
 Signal #1 Phase : ZB-MR1 Signal #2 Phase: ZB-MR2
 Signal #1 Info : 30Mx0.32mmx 0.50µ Signal #2 Info : 30M x 0.32mm x 0.25µm

Compound	RT#1	RT#2	Resp#1	Resp#2	ng/ml	ng/ml

System Monitoring Compounds						
1) SA Tetrachlo...	3.678	3.673	310.9E6	282.8E6	53.980	50.337
2) SA Decachlor...	8.713	8.661	267.7E6	88227844	50.997	49.631
Target Compounds						
26) L6 AR-1254-1	5.573	5.549	203.9E6	157.1E6	479.600	486.274
27) L6 AR-1254-2	5.722	5.697	181.7E6	135.5E6	478.697	485.096
28) L6 AR-1254-3	6.126	6.099	283.0E6	205.3E6	495.320	489.875
29) L6 AR-1254-4	6.356	6.326	191.2E6	117.9E6	548.098	517.695
30) L6 AR-1254-5	6.774	6.743	256.9E6	158.4E6	501.597	486.500



(f)=RT Delta > 1/2 Window (#)=Amounts differ by > 25% (m)=manual int.

Data Path : Z:\pestpcbsrv\HPCHEM1\ECD_O\Data\PO061125\
Data File : PO111618.D
Signal(s) : Signal #1: ECD1A.ch Signal #2: ECD2B.ch
Acq On : 11 Jun 2025 21:16
Operator : YP/AJ
Sample : AR1254ICV500
Misc :
ALS Vial : 34 Sample Multiplier: 1

Instrument :
ECD_O
ClientSampleId :
ICVPO061125

Integration File signal 1: autoint1.e
Integration File signal 2: autoint2.e
Quant Time: Jun 12 06:33:21 2025
Quant Method : Z:\pestpcbsrv\HPCHEM1\ECD_O\methods\PO061125.M
Quant Title : GC EXTRACTABLES
QLast Update : Thu Jun 12 06:25:26 2025
Response via : Initial Calibration
Integrator: ChemStation

Volume Inj. : 2 µl
Signal #1 Phase : ZB-MR1 Signal #2 Phase: ZB-MR2
Signal #1 Info : 30Mx0.32mmx 0.50µm Signal #2 Info : 30M x 0.32mm x 0.25µm



Data Path : Z:\pestpcbsrv\HPCHEM1\ECD_O\Data\P0061125\
 Data File : P0111619.D
 Signal(s) : Signal #1: ECD1A.ch Signal #2: ECD2B.ch
 Acq On : 11 Jun 2025 21:52
 Operator : YP/AJ
 Sample : AR1268ICV500
 Misc :
 ALS Vial : 35 Sample Multiplier: 1

Instrument :
 ECD_O
 ClientSampleId :
 ICVPO061125

Manual Integrations
 APPROVED

Reviewed By :Yogesh Patel 06/12/2025
 Supervised By :mohammad ahmed 06/13/2025

Integration File signal 1: autoint1.e
 Integration File signal 2: autoint2.e
 Quant Time: Jun 12 06:12:11 2025
 Quant Method : Z:\pestpcbsrv\HPCHEM1\ECD_O\methods\P0061125.M
 Quant Title : GC EXTRACTABLES
 QLast Update : Thu Jun 12 05:57:06 2025
 Response via : Initial Calibration
 Integrator: ChemStation

Volume Inj. : 2 µl
 Signal #1 Phase : ZB-MR1 Signal #2 Phase: ZB-MR2
 Signal #1 Info : 30Mx0.32mmx 0.50µ Signal #2 Info : 30M x 0.32mm x 0.25µm

Compound	RT#1	RT#2	Resp#1	Resp#2	ng/ml	ng/ml

System Monitoring Compounds						
1) SA Tetrachlo...	3.677	3.673	316.2E6	285.9E6	50.018	50.315
2) SA Decachlor...	8.713	8.662	468.0E6	154.7E6	49.104	48.991
Target Compounds						
41) L9 AR-1268-1	7.600	7.562	546.0E6	230.3E6	484.674	481.484m
42) L9 AR-1268-2	7.665	7.627	476.8E6	200.9E6	487.948	482.422
43) L9 AR-1268-3	7.871	7.833	400.8E6	153.3E6	489.001	483.812
44) L9 AR-1268-4	8.160	8.119	170.3E6	57551424	495.146	497.603
45) L9 AR-1268-5	8.455	8.410	1091.8E6	351.5E6	491.711	485.093

(f)=RT Delta > 1/2 Window (#)=Amounts differ by > 25% (m)=manual int.

Data Path : Z:\pestpcbsrv\HPCHEM1\ECD_O\Data\PO061125\
Data File : PO111619.D
Signal(s) : Signal #1: ECD1A.ch Signal #2: ECD2B.ch
Acq On : 11 Jun 2025 21:52
Operator : YP/AJ
Sample : AR1268ICV500
Misc :
ALS Vial : 35 Sample Multiplier: 1

Instrument :

ECD_O

ClientSampleId :

ICVPO061125

Manual Integrations



APPROVED

Reviewed By :Yogesh Patel 06/12/2025

Supervised By :mohammad ahmed 06/13/2025

Integration File signal 1: autoint1.e
Integration File signal 2: autoint2.e
Quant Time: Jun 12 06:12:11 2025
Quant Method : Z:\pestpcbsrv\HPCHEM1\ECD_O\methods\PO061125.M
Quant Title : GC EXTRACTABLES
QLast Update : Thu Jun 12 05:57:06 2025
Response via : Initial Calibration
Integrator: ChemStation

Volume Inj. : 2 µl
Signal #1 Phase : ZB-MR1 Signal #2 Phase: ZB-MR2
Signal #1 Info : 30Mx0.32mmx 0.50µ Signal #2 Info : 30M x 0.32mm x 0.25µm





284 Sheffield Street, Mountainside, New Jersey 07092, Phone : 908 789 8900,
Fax : 908 789 8922

CALIBRATION VERIFICATION SUMMARY

Lab Name: Alliance

Contract: CAMP02

Lab Code: ACE

SDG NO.: Q2514

Continuing Calib Date: 07/07/2025

Initial Calibration Date(s): 06/11/2025 06/11/2025

Continuing Calib Time: 20:15

Initial Calibration Time(s): 10:40 19:07

GC Column: ZB-MR1 ID: 0.32 (mm)

COMPOUND		CCAL RT	AVG RT	RT WINDOW		DIFF RT
				FROM	TO	
Aroclor-1016-1	(1)	4.76	4.77	4.67	4.87	0.01
Aroclor-1016-2	(2)	4.78	4.78	4.68	4.88	0.00
Aroclor-1016-3	(3)	4.84	4.84	4.74	4.94	0.00
Aroclor-1016-4	(4)	4.96	4.96	4.86	5.06	0.00
Aroclor-1016-5	(5)	5.21	5.22	5.12	5.32	0.01
Aroclor-1260-1	(1)	6.25	6.26	6.16	6.36	0.01
Aroclor-1260-2	(2)	6.44	6.45	6.35	6.55	0.01
Aroclor-1260-3	(3)	6.81	6.81	6.71	6.91	0.00
Aroclor-1260-4	(4)	7.07	7.07	6.97	7.17	0.00
Aroclor-1260-5	(5)	7.31	7.32	7.22	7.42	0.01
Tetrachloro-m-xylene		3.68	3.68	3.58	3.78	0.01
Decachlorobiphenyl		8.70	8.71	8.61	8.81	0.01



284 Sheffield Street, Mountainside, New Jersey 07092, Phone : 908 789 8900,
Fax : 908 789 8922

CALIBRATION VERIFICATION SUMMARY

Lab Name: Alliance

Contract: CAMP02

Lab Code: ACE

SDG NO.: Q2514

Continuing Calib Date: 07/07/2025

Initial Calibration Date(s): 06/11/2025

06/11/2025

Continuing Calib Time: 20:15

Initial Calibration Time(s): 10:40

19:07

GC Column: ZB-MR2 ID: 0.32 (mm)

COMPOUND		CCAL RT	AVG RT	RT WINDOW		DIFF RT
				FROM	TO	
Aroclor-1016-1	(1)	4.75	4.75	4.65	4.85	0.01
Aroclor-1016-2	(2)	4.76	4.77	4.67	4.87	0.01
Aroclor-1016-3	(3)	4.94	4.94	4.84	5.04	0.00
Aroclor-1016-4	(4)	4.98	4.99	4.89	5.09	0.01
Aroclor-1016-5	(5)	5.19	5.20	5.10	5.30	0.01
Aroclor-1260-1	(1)	6.22	6.23	6.13	6.33	0.01
Aroclor-1260-2	(2)	6.41	6.42	6.32	6.52	0.01
Aroclor-1260-3	(3)	6.56	6.57	6.47	6.67	0.01
Aroclor-1260-4	(4)	7.03	7.04	6.94	7.14	0.01
Aroclor-1260-5	(5)	7.27	7.28	7.18	7.38	0.01
Tetrachloro-m-xylene		3.67	3.67	3.57	3.77	0.00
Decachlorobiphenyl		8.65	8.66	8.56	8.76	0.01



284 Sheffield Street, Mountainside, New Jersey 07092, Phone : 908 789 8900, Fax : 908 789 8922

CALIBRATION VERIFICATION SUMMARY

Lab Name: Alliance **Contract:** CAMP02
Lab Code: ACE **SDG NO.:** Q2514
GC Column: ZB-MR1 **ID:** 0.32 (mm) **Initi. Calib. Date(s):** 06/11/2025 06/11/2025

Client Sample No.: CCAL01 **Date Analyzed:** 07/07/2025
Lab Sample No.: AR1660CCC500 **Data File :** PO112052.D **Time Analyzed:** 20:15

COMPOUND	RT	RT WINDOW		CALC AMOUNT(ng)	NOM AMOUNT(ng)	%D
		FROM	TO			
Aroclor-1016-1	4.762	4.666	4.866	451.720	500.000	-9.7
Aroclor-1016-2	4.781	4.684	4.884	468.410	500.000	-6.3
Aroclor-1016-3	4.838	4.741	4.941	442.310	500.000	-11.5
Aroclor-1016-4	4.957	4.861	5.061	447.560	500.000	-10.5
Aroclor-1016-5	5.214	5.118	5.318	471.760	500.000	-5.6
Aroclor-1260-1	6.253	6.157	6.357	457.590	500.000	-8.5
Aroclor-1260-2	6.441	6.346	6.546	508.720	500.000	1.7
Aroclor-1260-3	6.807	6.713	6.913	532.640	500.000	6.5
Aroclor-1260-4	7.067	6.972	7.172	520.660	500.000	4.1
Aroclor-1260-5	7.310	7.216	7.416	515.130	500.000	3.0
Decachlorobiphenyl	8.704	8.612	8.812	41.270	50.000	-17.5
Tetrachloro-m-xylene	3.675	3.577	3.777	51.660	50.000	3.3



284 Sheffield Street, Mountainside, New Jersey 07092, Phone : 908 789 8900, Fax : 908 789 8922

CALIBRATION VERIFICATION SUMMARY

Lab Name: Alliance **Contract:** CAMP02
Lab Code: ACE **SDG NO.:** Q2514
GC Column: ZB-MR2 **ID:** 0.32 (mm) **Initi. Calib. Date(s):** 06/11/2025 06/11/2025

Client Sample No.: CCAL01 **Date Analyzed:** 07/07/2025
Lab Sample No.: AR1660CCC500 **Data File :** PO112052.D **Time Analyzed:** 20:15

COMPOUND	RT	RT WINDOW		CALC AMOUNT(ng)	NOM AMOUNT(ng)	%D
		FROM	TO			
Aroclor-1016-1	4.745	4.651	4.851	505.180	500.000	1.0
Aroclor-1016-2	4.764	4.669	4.869	509.460	500.000	1.9
Aroclor-1016-3	4.939	4.844	5.044	498.340	500.000	-0.3
Aroclor-1016-4	4.981	4.887	5.087	498.070	500.000	-0.4
Aroclor-1016-5	5.194	5.099	5.299	492.500	500.000	-1.5
Aroclor-1260-1	6.223	6.129	6.329	485.010	500.000	-3.0
Aroclor-1260-2	6.411	6.316	6.516	504.260	500.000	0.9
Aroclor-1260-3	6.562	6.469	6.669	466.740	500.000	-6.7
Aroclor-1260-4	7.032	6.939	7.139	450.960	500.000	-9.8
Aroclor-1260-5	7.274	7.181	7.381	459.050	500.000	-8.2
Decachlorobiphenyl	8.651	8.561	8.761	44.280	50.000	-11.4
Tetrachloro-m-xylene	3.669	3.573	3.773	50.200	50.000	0.4

Data Path : Z:\pestpcbsrv\HPCHEM1\ECD_O\Data\P0070725\
 Data File : P0112052.D
 Signal(s) : Signal #1: ECD1A.ch Signal #2: ECD2B.ch
 Acq On : 07 Jul 2025 20:15
 Operator : YP/AJ
 Sample : AR1660CCC500
 Misc :
 ALS Vial : 3 Sample Multiplier: 1

Instrument :
 ECD_O
 ClientSampleId :
 AR1660CCC500

Manual Integrations
 APPROVED

Reviewed By :Yogesh Patel 07/08/2025
 Supervised By :mohammad ahmed 07/09/2025

Integration File signal 1: autoint1.e
 Integration File signal 2: autoint2.e
 Quant Time: Jul 08 01:56:39 2025
 Quant Method : Z:\pestpcbsrv\HPCHEM1\ECD_O\methods\P0061125.M
 Quant Title : GC EXTRACTABLES
 QLast Update : Thu Jun 12 06:25:26 2025
 Response via : Initial Calibration
 Integrator: ChemStation

Volume Inj. : 2 µl
 Signal #1 Phase : ZB-MR1 Signal #2 Phase: ZB-MR2
 Signal #1 Info : 30Mx0.32mmx 0.50µ Signal #2 Info : 30M x 0.32mm x 0.25µm

Compound	RT#1	RT#2	Resp#1	Resp#2	ng/ml	ng/ml

System Monitoring Compounds						
1) SA Tetrachlo...	3.675	3.669	297.5E6	282.0E6	51.656	50.205
2) SA Decachlor...	8.704	8.651	216.6E6	78718203	41.267	44.282m
Target Compounds						
3) L1 AR-1016-1	4.762	4.745	107.8E6	99403097	451.722	505.181
4) L1 AR-1016-2	4.781	4.764	155.9E6	145.1E6	468.413	509.457
5) L1 AR-1016-3	4.838	4.939	101.8E6	74870110	442.311	498.339
6) L1 AR-1016-4	4.957	4.981	83006148	60658654	447.562	498.071
7) L1 AR-1016-5	5.214	5.194	88051844	78404462	471.765	492.495
31) L7 AR-1260-1	6.253	6.223	162.6E6	121.7E6	457.589	485.012
32) L7 AR-1260-2	6.441	6.411	240.4E6	148.3E6	508.719	504.258
33) L7 AR-1260-3	6.807	6.562	226.2E6	124.7E6	532.640	466.736
34) L7 AR-1260-4	7.067	7.032	163.7E6	85724937	520.658	450.955
35) L7 AR-1260-5	7.310	7.274	424.3E6	195.1E6	515.129	459.055

(f)=RT Delta > 1/2 Window (#)=Amounts differ by > 25% (m)=manual int.

Data Path : Z:\pestpcbsrv\HPCHEM1\ECD_O\Data\PO070725\
Data File : PO112052.D
Signal(s) : Signal #1: ECD1A.ch Signal #2: ECD2B.ch
Acq On : 07 Jul 2025 20:15
Operator : YP/AJ
Sample : AR1660CCC500
Misc :
ALS Vial : 3 Sample Multiplier: 1



Instrument :
ECD_O
ClientSampleId :
AR1660CCC500

Manual Integrations
APPROVED

Reviewed By :Yogesh Patel 07/08/2025
Supervised By :mohammad ahmed 07/09/2025

Integration File signal 1: autoint1.e
Integration File signal 2: autoint2.e
Quant Time: Jul 08 01:56:39 2025
Quant Method : Z:\pestpcbsrv\HPCHEM1\ECD_O\methods\PO061125.M
Quant Title : GC EXTRACTABLES
QLast Update : Thu Jun 12 06:25:26 2025
Response via : Initial Calibration
Integrator: ChemStation

Volume Inj. : 2 µl
Signal #1 Phase : ZB-MR1 Signal #2 Phase: ZB-MR2
Signal #1 Info : 30Mx0.32mmx 0.50µ Signal #2 Info : 30M x 0.32mm x 0.25µm





284 Sheffield Street, Mountainside, New Jersey 07092, Phone : 908 789 8900,
Fax : 908 789 8922

CALIBRATION VERIFICATION SUMMARY

Lab Name: Alliance

Contract: CAMP02

Lab Code: ACE

SDG NO.: Q2514

Continuing Calib Date: 07/08/2025

Initial Calibration Date(s): 06/11/2025 06/11/2025

Continuing Calib Time: 03:54

Initial Calibration Time(s): 10:40 19:07

GC Column: ZB-MR1 ID: 0.32 (mm)

COMPOUND		CCAL RT	AVG RT	RT WINDOW		DIFF RT
				FROM	TO	
Aroclor-1016-1	(1)	4.76	4.77	4.67	4.87	0.01
Aroclor-1016-2	(2)	4.78	4.78	4.68	4.88	0.00
Aroclor-1016-3	(3)	4.84	4.84	4.74	4.94	0.00
Aroclor-1016-4	(4)	4.96	4.96	4.86	5.06	0.00
Aroclor-1016-5	(5)	5.21	5.22	5.12	5.32	0.01
Aroclor-1260-1	(1)	6.25	6.26	6.16	6.36	0.01
Aroclor-1260-2	(2)	6.44	6.45	6.35	6.55	0.01
Aroclor-1260-3	(3)	6.81	6.81	6.71	6.91	0.00
Aroclor-1260-4	(4)	7.07	7.07	6.97	7.17	0.00
Aroclor-1260-5	(5)	7.31	7.32	7.22	7.42	0.01
Tetrachloro-m-xylene		3.68	3.68	3.58	3.78	0.01
Decachlorobiphenyl		8.70	8.71	8.61	8.81	0.01



284 Sheffield Street, Mountainside, New Jersey 07092, Phone : 908 789 8900,
Fax : 908 789 8922

CALIBRATION VERIFICATION SUMMARY

Lab Name: Alliance

Contract: CAMP02

Lab Code: ACE

SDG NO.: Q2514

Continuing Calib Date: 07/08/2025

Initial Calibration Date(s): 06/11/2025 06/11/2025

Continuing Calib Time: 03:54

Initial Calibration Time(s): 10:40 19:07

GC Column: ZB-MR2 ID: 0.32 (mm)

COMPOUND		CCAL RT	AVG RT	RT WINDOW		DIFF RT
				FROM	TO	
Aroclor-1016-1	(1)	4.74	4.75	4.65	4.85	0.01
Aroclor-1016-2	(2)	4.76	4.77	4.67	4.87	0.01
Aroclor-1016-3	(3)	4.94	4.94	4.84	5.04	0.00
Aroclor-1016-4	(4)	4.98	4.99	4.89	5.09	0.01
Aroclor-1016-5	(5)	5.19	5.20	5.10	5.30	0.01
Aroclor-1260-1	(1)	6.22	6.23	6.13	6.33	0.01
Aroclor-1260-2	(2)	6.41	6.42	6.32	6.52	0.01
Aroclor-1260-3	(3)	6.56	6.57	6.47	6.67	0.01
Aroclor-1260-4	(4)	7.03	7.04	6.94	7.14	0.01
Aroclor-1260-5	(5)	7.27	7.28	7.18	7.38	0.01
Tetrachloro-m-xylene		3.67	3.67	3.57	3.77	0.00
Decachlorobiphenyl		8.65	8.66	8.56	8.76	0.01



284 Sheffield Street, Mountainside, New Jersey 07092, Phone : 908 789 8900, Fax : 908 789 8922

CALIBRATION VERIFICATION SUMMARY

Lab Name: Alliance **Contract:** CAMP02
Lab Code: ACE **SDG NO.:** Q2514
GC Column: ZB-MR1 **ID:** 0.32 (mm) **Initi. Calib. Date(s):** 06/11/2025 06/11/2025

Client Sample No.: CCAL02 **Date Analyzed:** 07/08/2025
Lab Sample No.: AR1660CCC500 **Data File :** PO112067.D **Time Analyzed:** 03:54

COMPOUND	RT	RT WINDOW		CALC AMOUNT(ng)	NOM AMOUNT(ng)	%D
		FROM	TO			
Aroclor-1016-1	4.761	4.666	4.866	528.310	500.000	5.7
Aroclor-1016-2	4.780	4.684	4.884	545.080	500.000	9.0
Aroclor-1016-3	4.837	4.741	4.941	530.690	500.000	6.1
Aroclor-1016-4	4.956	4.861	5.061	532.770	500.000	6.6
Aroclor-1016-5	5.213	5.118	5.318	522.850	500.000	4.6
Aroclor-1260-1	6.252	6.157	6.357	525.380	500.000	5.1
Aroclor-1260-2	6.441	6.346	6.546	598.920	500.000	19.8
Aroclor-1260-3	6.808	6.713	6.913	541.620	500.000	8.3
Aroclor-1260-4	7.067	6.972	7.172	543.400	500.000	8.7
Aroclor-1260-5	7.309	7.216	7.416	584.810	500.000	17.0
Decachlorobiphenyl	8.703	8.612	8.812	43.520	50.000	-13.0
Tetrachloro-m-xylene	3.675	3.577	3.777	60.770	50.000	21.5



284 Sheffield Street, Mountainside, New Jersey 07092, Phone : 908 789 8900, Fax : 908 789 8922

CALIBRATION VERIFICATION SUMMARY

Lab Name: Alliance **Contract:** CAMP02
Lab Code: ACE **SDG NO.:** Q2514
GC Column: ZB-MR2 **ID:** 0.32 (mm) **Initi. Calib. Date(s):** 06/11/2025 06/11/2025

Client Sample No.: CCAL02 **Date Analyzed:** 07/08/2025
Lab Sample No.: AR1660CCC500 **Data File :** PO112067.D **Time Analyzed:** 03:54

COMPOUND	RT	RT WINDOW		CALC AMOUNT(ng)	NOM AMOUNT(ng)	%D
		FROM	TO			
Aroclor-1016-1	4.744	4.651	4.851	552.450	500.000	10.5
Aroclor-1016-2	4.762	4.669	4.869	563.100	500.000	12.6
Aroclor-1016-3	4.937	4.844	5.044	541.710	500.000	8.3
Aroclor-1016-4	4.980	4.887	5.087	478.710	500.000	-4.3
Aroclor-1016-5	5.192	5.099	5.299	596.930	500.000	19.4
Aroclor-1260-1	6.221	6.129	6.329	591.390	500.000	18.3
Aroclor-1260-2	6.409	6.316	6.516	591.380	500.000	18.3
Aroclor-1260-3	6.560	6.469	6.669	531.150	500.000	6.2
Aroclor-1260-4	7.031	6.939	7.139	498.690	500.000	-0.3
Aroclor-1260-5	7.272	7.181	7.381	507.110	500.000	1.4
Decachlorobiphenyl	8.650	8.561	8.761	48.100	50.000	-3.8
Tetrachloro-m-xylene	3.668	3.573	3.773	54.860	50.000	9.7

Data Path : Z:\pestpcbsrv\HPCHEM1\ECD_O\Data\P0070725\
 Data File : PO112067.D
 Signal(s) : Signal #1: ECD1A.ch Signal #2: ECD2B.ch
 Acq On : 08 Jul 2025 03:54
 Operator : YP/AJ
 Sample : AR1660CCC500
 Misc :
 ALS Vial : 50 Sample Multiplier: 1

Instrument :
 ECD_O
 ClientSampleId :
 AR1660CCC500

Manual Integrations
 APPROVED

Reviewed By :Yogesh Patel 07/08/2025
 Supervised By :mohammad ahmed 07/09/2025

Integration File signal 1: autoint1.e
 Integration File signal 2: autoint2.e
 Quant Time: Jul 08 04:30:17 2025
 Quant Method : Z:\pestpcbsrv\HPCHEM1\ECD_O\methods\P0061125.M
 Quant Title : GC EXTRACTABLES
 QLast Update : Thu Jun 12 06:25:26 2025
 Response via : Initial Calibration
 Integrator: ChemStation

Volume Inj. : 2 µl
 Signal #1 Phase : ZB-MR1 Signal #2 Phase: ZB-MR2
 Signal #1 Info : 30Mx0.32mmx 0.50µ Signal #2 Info : 30M x 0.32mm x 0.25µm

Compound	RT#1	RT#2	Resp#1	Resp#2	ng/ml	ng/ml

System Monitoring Compounds						
1) SA Tetrachlo...	3.675	3.668	350.0E6	308.2E6	60.767	54.860
2) SA Decachlor...	8.703	8.650	228.4E6	85504209	43.516	48.099m
Target Compounds						
3) L1 AR-1016-1	4.761	4.744	126.0E6	108.7E6	528.311	552.449
4) L1 AR-1016-2	4.780	4.762	181.4E6	160.4E6	545.077	563.102
5) L1 AR-1016-3	4.837	4.937	122.1E6	81386744	530.688	541.714
6) L1 AR-1016-4	4.956	4.980	98809899	58301085	532.775	478.712
7) L1 AR-1016-5	5.213	5.192	97585644	95030677	522.845m	596.932
31) L7 AR-1260-1	6.252	6.221	186.7E6	148.4E6	525.375	591.392
32) L7 AR-1260-2	6.441	6.409	283.0E6	173.9E6	598.920	591.375m
33) L7 AR-1260-3	6.808	6.560	230.0E6	141.9E6	541.617	531.151m
34) L7 AR-1260-4	7.067	7.031	170.9E6	94799277	543.403	498.691
35) L7 AR-1260-5	7.309	7.272	481.7E6	215.5E6	584.808m	507.114

(f)=RT Delta > 1/2 Window (#)=Amounts differ by > 25% (m)=manual int.

Data Path : Z:\pestpcbsrv\HPCHEM1\ECD_O\Data\PO070725\
Data File : PO112067.D
Signal(s) : Signal #1: ECD1A.ch Signal #2: ECD2B.ch
Acq On : 08 Jul 2025 03:54
Operator : YP/AJ
Sample : AR1660CCC500
Misc :
ALS Vial : 50 Sample Multiplier: 1



Instrument :
ECD_O
ClientSampleId :
AR1660CCC500

Manual Integrations
APPROVED

Reviewed By :Yogesh Patel 07/08/2025
Supervised By :mohammad ahmed 07/09/2025

Integration File signal 1: autoint1.e
Integration File signal 2: autoint2.e
Quant Time: Jul 08 04:30:17 2025
Quant Method : Z:\pestpcbsrv\HPCHEM1\ECD_O\methods\PO061125.M
Quant Title : GC EXTRACTABLES
QLast Update : Thu Jun 12 06:25:26 2025
Response via : Initial Calibration
Integrator: ChemStation

Volume Inj. : 2 µl
Signal #1 Phase : ZB-MR1 Signal #2 Phase: ZB-MR2
Signal #1 Info : 30Mx0.32mmx 0.50µ Signal #2 Info : 30M x 0.32mm x 0.25µm





284 Sheffield Street, Mountainside, New Jersey 07092, Phone : 908 789 8900,
Fax : 908 789 8922

CALIBRATION VERIFICATION SUMMARY

Lab Name: Alliance

Contract: CAMP02

Lab Code: ACE

SDG NO.: Q2514

Continuing Calib Date: 07/08/2025

Initial Calibration Date(s): 06/11/2025 06/11/2025

Continuing Calib Time: 10:09

Initial Calibration Time(s): 10:40 19:07

GC Column: ZB-MR1 ID: 0.32 (mm)

COMPOUND		CCAL RT	AVG RT	RT WINDOW		DIFF RT
				FROM	TO	
Aroclor-1016-1	(1)	4.76	4.77	4.67	4.87	0.01
Aroclor-1016-2	(2)	4.78	4.78	4.68	4.88	0.00
Aroclor-1016-3	(3)	4.84	4.84	4.74	4.94	0.00
Aroclor-1016-4	(4)	4.96	4.96	4.86	5.06	0.00
Aroclor-1016-5	(5)	5.22	5.22	5.12	5.32	0.00
Aroclor-1260-1	(1)	6.25	6.26	6.16	6.36	0.01
Aroclor-1260-2	(2)	6.44	6.45	6.35	6.55	0.01
Aroclor-1260-3	(3)	6.81	6.81	6.71	6.91	0.00
Aroclor-1260-4	(4)	7.07	7.07	6.97	7.17	0.00
Aroclor-1260-5	(5)	7.31	7.32	7.22	7.42	0.01
Tetrachloro-m-xylene		3.68	3.68	3.58	3.78	0.00
Decachlorobiphenyl		8.71	8.71	8.61	8.81	0.01



284 Sheffield Street, Mountainside, New Jersey 07092, Phone : 908 789 8900,
Fax : 908 789 8922

CALIBRATION VERIFICATION SUMMARY

Lab Name: Alliance

Contract: CAMP02

Lab Code: ACE

SDG NO.: Q2514

Continuing Calib Date: 07/08/2025

Initial Calibration Date(s): 06/11/2025 06/11/2025

Continuing Calib Time: 10:09

Initial Calibration Time(s): 10:40 19:07

GC Column: ZB-MR2 ID: 0.32 (mm)

COMPOUND		CCAL RT	AVG RT	RT WINDOW		DIFF RT
				FROM	TO	
Aroclor-1016-1	(1)	4.74	4.75	4.65	4.85	0.01
Aroclor-1016-2	(2)	4.76	4.77	4.67	4.87	0.01
Aroclor-1016-3	(3)	4.94	4.94	4.84	5.04	0.00
Aroclor-1016-4	(4)	4.98	4.99	4.89	5.09	0.01
Aroclor-1016-5	(5)	5.19	5.20	5.10	5.30	0.01
Aroclor-1260-1	(1)	6.22	6.23	6.13	6.33	0.01
Aroclor-1260-2	(2)	6.41	6.42	6.32	6.52	0.01
Aroclor-1260-3	(3)	6.56	6.57	6.47	6.67	0.01
Aroclor-1260-4	(4)	7.03	7.04	6.94	7.14	0.01
Aroclor-1260-5	(5)	7.27	7.28	7.18	7.38	0.01
Tetrachloro-m-xylene		3.67	3.67	3.57	3.77	0.00
Decachlorobiphenyl		8.65	8.66	8.56	8.76	0.01



284 Sheffield Street, Mountainside, New Jersey 07092, Phone : 908 789 8900, Fax : 908 789 8922

CALIBRATION VERIFICATION SUMMARY

Lab Name: Alliance **Contract:** CAMP02
Lab Code: ACE **SDG NO.:** Q2514
GC Column: ZB-MR1 **ID:** 0.32 (mm) **Initi. Calib. Date(s):** 06/11/2025 06/11/2025

Client Sample No.: CCAL03 **Date Analyzed:** 07/08/2025
Lab Sample No.: AR1660CCC500 **Data File :** PO112083.D **Time Analyzed:** 10:09

COMPOUND	RT	RT WINDOW		CALC AMOUNT(ng)	NOM AMOUNT(ng)	%D
		FROM	TO			
Aroclor-1016-1	4.762	4.666	4.866	597.960	500.000	19.6
Aroclor-1016-2	4.780	4.684	4.884	599.020	500.000	19.8
Aroclor-1016-3	4.837	4.741	4.941	561.710	500.000	12.3
Aroclor-1016-4	4.957	4.861	5.061	574.730	500.000	14.9
Aroclor-1016-5	5.215	5.118	5.318	566.580	500.000	13.3
Aroclor-1260-1	6.252	6.157	6.357	563.120	500.000	12.6
Aroclor-1260-2	6.442	6.346	6.546	599.800	500.000	20.0
Aroclor-1260-3	6.808	6.713	6.913	597.780	500.000	19.6
Aroclor-1260-4	7.067	6.972	7.172	548.080	500.000	9.6
Aroclor-1260-5	7.311	7.216	7.416	627.280	500.000	25.5
Decachlorobiphenyl	8.705	8.612	8.812	52.570	50.000	5.1
Tetrachloro-m-xylene	3.676	3.577	3.777	71.620	50.000	43.2



284 Sheffield Street, Mountainside, New Jersey 07092, Phone : 908 789 8900, Fax : 908 789 8922

CALIBRATION VERIFICATION SUMMARY

Lab Name: Alliance **Contract:** CAMP02
Lab Code: ACE **SDG NO.:** Q2514
GC Column: ZB-MR2 **ID:** 0.32 (mm) **Initi. Calib. Date(s):** 06/11/2025 06/11/2025

Client Sample No.: CCAL03 **Date Analyzed:** 07/08/2025
Lab Sample No.: AR1660CCC500 **Data File :** PO112083.D **Time Analyzed:** 10:09

COMPOUND	RT	RT WINDOW		CALC AMOUNT(ng)	NOM AMOUNT(ng)	%D
		FROM	TO			
Aroclor-1016-1	4.743	4.651	4.851	466.410	500.000	-6.7
Aroclor-1016-2	4.761	4.669	4.869	474.240	500.000	-5.2
Aroclor-1016-3	4.937	4.844	5.044	455.480	500.000	-8.9
Aroclor-1016-4	4.979	4.887	5.087	452.330	500.000	-9.5
Aroclor-1016-5	5.191	5.099	5.299	441.610	500.000	-11.7
Aroclor-1260-1	6.221	6.129	6.329	441.990	500.000	-11.6
Aroclor-1260-2	6.409	6.316	6.516	464.790	500.000	-7.0
Aroclor-1260-3	6.560	6.469	6.669	435.190	500.000	-13.0
Aroclor-1260-4	7.031	6.939	7.139	432.900	500.000	-13.4
Aroclor-1260-5	7.272	7.181	7.381	431.230	500.000	-13.8
Decachlorobiphenyl	8.650	8.561	8.761	40.190	50.000	-19.6
Tetrachloro-m-xylene	3.668	3.573	3.773	46.280	50.000	-7.4

Data Path : Z:\pestpcbsrv\HPCHEM1\ECD_O\Data\P0070725\
 Data File : PO112083.D
 Signal(s) : Signal #1: ECD1A.ch Signal #2: ECD2B.ch
 Acq On : 08 Jul 2025 10:09
 Operator : YP/AJ
 Sample : AR1660CCC500
 Misc :
 ALS Vial : 3 Sample Multiplier: 1

Instrument :
 ECD_O
 ClientSampleId :
 AR1660CCC500

Manual Integrations
 APPROVED

Reviewed By :Yogesh Patel 07/08/2025
 Supervised By :mohammad ahmed 07/09/2025

Integration File signal 1: autoint1.e
 Integration File signal 2: autoint2.e
 Quant Time: Jul 08 10:58:03 2025
 Quant Method : Z:\pestpcbsrv\HPCHEM1\ECD_O\methods\P0061125.M
 Quant Title : GC EXTRACTABLES
 QLast Update : Thu Jun 12 06:25:26 2025
 Response via : Initial Calibration
 Integrator: ChemStation

Volume Inj. : 2 µl
 Signal #1 Phase : ZB-MR1 Signal #2 Phase: ZB-MR2
 Signal #1 Info : 30Mx0.32mmx 0.50µ Signal #2 Info : 30M x 0.32mm x 0.25µm

Compound	RT#1	RT#2	Resp#1	Resp#2	ng/ml	ng/ml

System Monitoring Compounds						
1) SA Tetrachlo...	3.676	3.668	412.5E6	260.0E6	71.617	46.279 #
2) SA Decachlor...	8.705	8.650	275.9E6	71437175	52.567	40.186m
Target Compounds						
3) L1 AR-1016-1	4.762	4.743	142.6E6	91774465	597.956m	466.412
4) L1 AR-1016-2	4.780	4.761	199.3E6	135.1E6	599.016m	474.241
5) L1 AR-1016-3	4.837	4.937	129.3E6	68430723	561.713m	455.478
6) L1 AR-1016-4	4.957	4.979	106.6E6	55087598	574.735m	452.326
7) L1 AR-1016-5	5.215	5.191	105.7E6	70303536	566.577	441.609m
31) L7 AR-1260-1	6.252	6.221	200.1E6	110.9E6	563.122	441.991
32) L7 AR-1260-2	6.442	6.409	283.4E6	136.7E6	599.802	464.791
33) L7 AR-1260-3	6.808	6.560	253.9E6	116.3E6	597.781	435.187 #
34) L7 AR-1260-4	7.067	7.031	172.3E6	82293380	548.077	432.904
35) L7 AR-1260-5	7.311	7.272	516.7E6	183.3E6	627.283	431.226 #

(f)=RT Delta > 1/2 Window (#)=Amounts differ by > 25% (m)=manual int.

Data Path : Z:\pestpcbsrv\HPCHEM1\ECD_O\Data\PO070725\
Data File : PO112083.D
Signal(s) : Signal #1: ECD1A.ch Signal #2: ECD2B.ch
Acq On : 08 Jul 2025 10:09
Operator : YP/AJ
Sample : AR1660CCC500
Misc :
ALS Vial : 3 Sample Multiplier: 1



Instrument :
ECD_O
ClientSampleId :
AR1660CCC500

Manual Integrations
APPROVED

Reviewed By :Yogesh Patel 07/08/2025
Supervised By :mohammad ahmed 07/09/2025

Integration File signal 1: autoint1.e
Integration File signal 2: autoint2.e
Quant Time: Jul 08 10:58:03 2025
Quant Method : Z:\pestpcbsrv\HPCHEM1\ECD_O\methods\PO061125.M
Quant Title : GC EXTRACTABLES
QLast Update : Thu Jun 12 06:25:26 2025
Response via : Initial Calibration
Integrator: ChemStation

Volume Inj. : 2 µl
Signal #1 Phase : ZB-MR1 Signal #2 Phase: ZB-MR2
Signal #1 Info : 30Mx0.32mmx 0.50µ Signal #2 Info : 30M x 0.32mm x 0.25µm



Analytical Sequence

Client: CDM Smith

SDG No.: Q2514

Project: South River WM Replacement

Instrument ID: ECD_O

GC Column: ZB-MR1

ID: 0.32 (mm)

Inst. Calib. Date(s): 06/11/2025

06/11/2025

THE ANALYTICAL SEQUENCE OF PERFORMANCE EVALUATION MIXTURES, BLANKS, SAMPLES,
AND STANDARDS IS GIVEN BELOW:

CLIENT ID	LAB SAMPLE ID	DATE ANALYZED	TIME ANALYZED	DATAFILE	DCB RT #	TCX RT #
IBLK	IBLK	06/11/2025	10:21	PO111586.D	8.71	3.68
AR1660ICC1000	AR1660ICC1000	06/11/2025	10:40	PO111587.D	8.71	3.68
AR1660ICC750	AR1660ICC750	06/11/2025	10:58	PO111588.D	8.71	3.68
AR1660ICC500	AR1660ICC500	06/11/2025	11:17	PO111589.D	8.71	3.68
AR1660ICC250	AR1660ICC250	06/11/2025	11:35	PO111590.D	8.71	3.68
AR1660ICC050	AR1660ICC050	06/11/2025	11:53	PO111591.D	8.71	3.68
AR1221ICC500	AR1221ICC500	06/11/2025	12:12	PO111592.D	8.71	3.68
AR1232ICC500	AR1232ICC500	06/11/2025	12:30	PO111593.D	8.71	3.68
AR1242ICC1000	AR1242ICC1000	06/11/2025	12:48	PO111594.D	8.71	3.68
AR1242ICC750	AR1242ICC750	06/11/2025	13:07	PO111595.D	8.71	3.68
AR1242ICC500	AR1242ICC500	06/11/2025	13:25	PO111596.D	8.71	3.68
AR1242ICC250	AR1242ICC250	06/11/2025	13:44	PO111597.D	8.71	3.68
AR1242ICC050	AR1242ICC050	06/11/2025	14:02	PO111598.D	8.71	3.68
AR1248ICC1000	AR1248ICC1000	06/11/2025	14:20	PO111599.D	8.71	3.68
AR1248ICC750	AR1248ICC750	06/11/2025	14:39	PO111600.D	8.71	3.68
AR1248ICC500	AR1248ICC500	06/11/2025	15:14	PO111601.D	8.71	3.68
AR1248ICC250	AR1248ICC250	06/11/2025	15:32	PO111602.D	8.71	3.68
AR1248ICC050	AR1248ICC050	06/11/2025	15:49	PO111603.D	8.71	3.68
AR1254ICC1000	AR1254ICC1000	06/11/2025	16:06	PO111604.D	8.71	3.68
AR1254ICC750	AR1254ICC750	06/11/2025	16:25	PO111605.D	8.71	3.68
AR1254ICC500	AR1254ICC500	06/11/2025	16:43	PO111606.D	8.71	3.68
AR1254ICC250	AR1254ICC250	06/11/2025	17:00	PO111607.D	8.71	3.68
AR1254ICC050	AR1254ICC050	06/11/2025	17:18	PO111608.D	8.71	3.68
AR1262ICC500	AR1262ICC500	06/11/2025	17:36	PO111609.D	8.71	3.68
AR1268ICC1000	AR1268ICC1000	06/11/2025	17:55	PO111610.D	8.71	3.68
AR1268ICC750	AR1268ICC750	06/11/2025	18:13	PO111611.D	8.71	3.68
AR1268ICC500	AR1268ICC500	06/11/2025	18:31	PO111612.D	8.71	3.68
AR1268ICC250	AR1268ICC250	06/11/2025	18:50	PO111613.D	8.71	3.68
AR1268ICC050	AR1268ICC050	06/11/2025	19:07	PO111614.D	8.71	3.68
AR1660CCC500	AR1660CCC500	07/07/2025	20:15	PO112052.D	8.70	3.68
IBLK	IBLK	07/07/2025	21:44	PO112056.D	8.71	3.67
PB168735BL	PB168735BL	07/07/2025	22:01	PO112057.D	8.70	3.67
PB168735BS	PB168735BS	07/07/2025	22:20	PO112058.D	8.70	3.67
HR-3-070325MS	Q2513-03MS	07/07/2025	23:52	PO112063.D	8.70	3.67
HR-3-070325MSD	Q2513-03MSD	07/08/2025	00:10	PO112064.D	8.71	3.67
TP-92	Q2514-01	07/08/2025	00:29	PO112065.D	8.71	3.67
TP-93	Q2514-02	07/08/2025	00:46	PO112066.D	8.70	3.67
AR1660CCC500	AR1660CCC500	07/08/2025	03:54	PO112067.D	8.70	3.68
IBLK	IBLK	07/08/2025	05:26	PO112072.D	8.71	3.67
TP-94	Q2514-03	07/08/2025	05:44	PO112073.D	8.70	3.67
TP-96	Q2514-04	07/08/2025	06:02	PO112074.D	8.70	3.68
TP-97	Q2514-05	07/08/2025	06:21	PO112075.D	8.71	3.68

Analytical Sequence

TP-103	Q2514-06	07/08/2025	06:39	PO112076.D	8.70	3.67
TP-36	Q2514-07	07/08/2025	06:58	PO112077.D	8.70	3.67
TP-78	Q2514-08	07/08/2025	07:16	PO112078.D	8.70	3.67
TP-81	Q2514-09	07/08/2025	07:35	PO112079.D	8.70	3.67
TP-90	Q2514-10	07/08/2025	07:53	PO112080.D	8.70	3.67
AR1660CCC500	AR1660CCC500	07/08/2025	10:09	PO112083.D	8.71	3.68
LBLK	LBLK	07/08/2025	11:26	PO112085.D	8.71	3.68

Analytical Sequence

Client: CDM Smith

SDG No.: Q2514

Project: South River WM Replacement

Instrument ID: ECD_O

GC Column: ZB-MR2

ID: 0.32 (mm)

Inst. Calib. Date(s): 06/11/2025

06/11/2025

THE ANALYTICAL SEQUENCE OF PERFORMANCE EVALUATION MIXTURES, BLANKS, SAMPLES,
AND STANDARDS IS GIVEN BELOW:

CLIENT ID	LAB SAMPLE ID	DATE ANALYZED	TIME ANALYZED	DATAFILE	DCB RT #	TCX RT #
LBLK	LBLK	06/11/2025	10:21	PO111586.D	8.66	3.67
AR1660ICC1000	AR1660ICC1000	06/11/2025	10:40	PO111587.D	8.66	3.67
AR1660ICC750	AR1660ICC750	06/11/2025	10:58	PO111588.D	8.66	3.67
AR1660ICC500	AR1660ICC500	06/11/2025	11:17	PO111589.D	8.66	3.67
AR1660ICC250	AR1660ICC250	06/11/2025	11:35	PO111590.D	8.66	3.67
AR1660ICC050	AR1660ICC050	06/11/2025	11:53	PO111591.D	8.66	3.67
AR1221ICC500	AR1221ICC500	06/11/2025	12:12	PO111592.D	8.66	3.67
AR1232ICC500	AR1232ICC500	06/11/2025	12:30	PO111593.D	8.66	3.67
AR1242ICC1000	AR1242ICC1000	06/11/2025	12:48	PO111594.D	8.66	3.67
AR1242ICC750	AR1242ICC750	06/11/2025	13:07	PO111595.D	8.66	3.67
AR1242ICC500	AR1242ICC500	06/11/2025	13:25	PO111596.D	8.66	3.67
AR1242ICC250	AR1242ICC250	06/11/2025	13:44	PO111597.D	8.66	3.67
AR1242ICC050	AR1242ICC050	06/11/2025	14:02	PO111598.D	8.66	3.67
AR1248ICC1000	AR1248ICC1000	06/11/2025	14:20	PO111599.D	8.66	3.67
AR1248ICC750	AR1248ICC750	06/11/2025	14:39	PO111600.D	8.66	3.67
AR1248ICC500	AR1248ICC500	06/11/2025	15:14	PO111601.D	8.66	3.67
AR1248ICC250	AR1248ICC250	06/11/2025	15:32	PO111602.D	8.66	3.67
AR1248ICC050	AR1248ICC050	06/11/2025	15:49	PO111603.D	8.66	3.67
AR1254ICC1000	AR1254ICC1000	06/11/2025	16:06	PO111604.D	8.66	3.67
AR1254ICC750	AR1254ICC750	06/11/2025	16:25	PO111605.D	8.66	3.67
AR1254ICC500	AR1254ICC500	06/11/2025	16:43	PO111606.D	8.66	3.67
AR1254ICC250	AR1254ICC250	06/11/2025	17:00	PO111607.D	8.66	3.67
AR1254ICC050	AR1254ICC050	06/11/2025	17:18	PO111608.D	8.66	3.67
AR1262ICC500	AR1262ICC500	06/11/2025	17:36	PO111609.D	8.66	3.67
AR1268ICC1000	AR1268ICC1000	06/11/2025	17:55	PO111610.D	8.66	3.67
AR1268ICC750	AR1268ICC750	06/11/2025	18:13	PO111611.D	8.66	3.67
AR1268ICC500	AR1268ICC500	06/11/2025	18:31	PO111612.D	8.66	3.67
AR1268ICC250	AR1268ICC250	06/11/2025	18:50	PO111613.D	8.66	3.67
AR1268ICC050	AR1268ICC050	06/11/2025	19:07	PO111614.D	8.66	3.67
AR1660CCC500	AR1660CCC500	07/07/2025	20:15	PO112052.D	8.65	3.67
LBLK	LBLK	07/07/2025	21:44	PO112056.D	8.65	3.67
PB168735BL	PB168735BL	07/07/2025	22:01	PO112057.D	8.65	3.67
PB168735BS	PB168735BS	07/07/2025	22:20	PO112058.D	8.65	3.67
HR-3-070325MS	Q2513-03MS	07/07/2025	23:52	PO112063.D	8.65	3.67
HR-3-070325MSD	Q2513-03MSD	07/08/2025	00:10	PO112064.D	8.65	3.67
TP-92	Q2514-01	07/08/2025	00:29	PO112065.D	8.65	3.67
TP-93	Q2514-02	07/08/2025	00:46	PO112066.D	8.65	3.67
AR1660CCC500	AR1660CCC500	07/08/2025	03:54	PO112067.D	8.65	3.67
LBLK	LBLK	07/08/2025	05:26	PO112072.D	8.65	3.67
TP-94	Q2514-03	07/08/2025	05:44	PO112073.D	8.65	3.67
TP-96	Q2514-04	07/08/2025	06:02	PO112074.D	8.65	3.67
TP-97	Q2514-05	07/08/2025	06:21	PO112075.D	8.65	3.67

Analytical Sequence

TP-103	Q2514-06	07/08/2025	06:39	PO112076.D	8.65	3.67
TP-36	Q2514-07	07/08/2025	06:58	PO112077.D	8.65	3.67
TP-78	Q2514-08	07/08/2025	07:16	PO112078.D	8.65	3.67
TP-81	Q2514-09	07/08/2025	07:35	PO112079.D	8.65	3.67
TP-90	Q2514-10	07/08/2025	07:53	PO112080.D	8.65	3.67
AR1660CCC500	AR1660CCC500	07/08/2025	10:09	PO112083.D	8.65	3.67
LBLK	LBLK	07/08/2025	11:26	PO112085.D	8.65	3.67

IDENTIFICATION SUMMARY
FOR MULTICOMPONENT ANALYTES

SAMPLE NO.

PB168735BS

Lab Name: Alliance
Lab Code: ACE
Lab Sample ID: PB168735BS
Instrument ID (1): ECD_O

Contract: CAMP02
SDG NO.: Q2514
Date(s) Analyzed: 07/07/2025 07/07/2025
Instrument ID (2): ECD_O

GC Column: (1): ZB-MR1 ID: 0.32 (mm) GC Column: (2): ZB-MR2 ID: 0.32 (mm)
Data file PO112058.D

ANALYTE	COL	RT	RT WINDOW		CONCENTRATION	MEAN CONCENTRATION	%RPD
			FROM	TO			
Aroclor-1016	1	4.76	4.71	4.81	145	145	3.39
COLUMN 1	2	4.779	4.729	4.829	151		
	3	4.835	4.785	4.885	143		
	4	4.955	4.905	5.005	145		
	5	5.212	5.162	5.262	143		
COLUMN 2	1	4.743	4.693	4.793	152	150	
	2	4.762	4.712	4.812	157		
	3	4.937	4.887	4.987	150		
	4	4.979	4.929	5.029	148		
	5	5.191	5.141	5.241	142		
Aroclor-1260	1	6.25	6.2	6.3	155	163	6.98
COLUMN 1	2	6.44	6.39	6.49	173		
	3	6.807	6.757	6.857	159		
	4	7.066	7.016	7.116	167		
	5	7.308	7.258	7.358	163		
COLUMN 2	1	6.22	6.17	6.27	161	152	
	2	6.409	6.359	6.459	167		
	3	6.56	6.51	6.61	159		
	4	7.03	6.98	7.08	137		
	5	7.271	7.221	7.321	136		

IDENTIFICATION SUMMARY
FOR MULTICOMPONENT ANALYTES

SAMPLE NO.

HR-3-070325MS

Lab Name: Alliance

Contract: CAMP02

Lab Code: ACE

SDG NO.: Q2514

Lab Sample ID: Q2513-03MS

Date(s) Analyzed: 07/07/2025 07/07/2025

Instrument ID (1): ECD_O

Instrument ID (2): ECD_O

GC Column: (1): ZB-MR1 ID: 0.32 (mm)

GC Column: (2): ZB-MR2 ID: 0.32 (mm)

Data file PO112063.D

ANALYTE	COL	RT	RT WINDOW		CONCENTRATION	MEAN CONCENTRATION	%RPD
			FROM	TO			
Aroclor-1016	1	4.76	4.71	4.81	116	118	2.58
COLUMN 1	2	4.779	4.729	4.829	122		
	3	4.836	4.786	4.886	126		
	4	4.956	4.906	5.006	117		
	5	5.213	5.163	5.263	108		
	1	4.744	4.694	4.794	120		
COLUMN 2	2	4.762	4.712	4.812	121	115	
	3	4.937	4.887	4.987	113		
	4	4.98	4.93	5.03	112		
	5	5.192	5.142	5.242	109		
Aroclor-1260	1	6.251	6.201	6.301	118	105	6.45
COLUMN 1	2	6.44	6.39	6.49	115		
	3	6.806	6.756	6.856	98.9		
	4	7.065	7.015	7.115	91.9		
	5	7.308	7.258	7.358	102		
	1	6.221	6.171	6.271	128	112	
COLUMN 2	2	6.409	6.359	6.459	123		
	3	6.561	6.511	6.611	111		
	4	7.031	6.981	7.081	96.2		
	5	7.272	7.222	7.322	104		

IDENTIFICATION SUMMARY
FOR MULTICOMPONENT ANALYTES

SAMPLE NO.

HR-3-070325MSD

Lab Name: Alliance

Contract: CAMP02

Lab Code: ACE

SDG NO.: Q2514

Lab Sample ID: Q2513-03MSD

Date(s) Analyzed: 07/08/2025 07/08/2025

Instrument ID (1): ECD_O

Instrument ID (2): ECD_O

GC Column: (1): ZB-MR1 ID: 0.32 (mm)

GC Column: (2): ZB-MR2 ID: 0.32 (mm)

Data file PO112064.D

ANALYTE	COL	RT	RT WINDOW		CONCENTRATION	MEAN CONCENTRATION	%RPD
			FROM	TO			
Aroclor-1016	1	4.761	4.711	4.811	104	109	8.77
COLUMN 1	2	4.78	4.73	4.83	109		
	3	4.837	4.787	4.887	107		
	4	4.956	4.906	5.006	108		
	5	5.214	5.164	5.264	118		
	1	4.744	4.694	4.794	124		
COLUMN 2	2	4.762	4.712	4.812	121	119	
	3	4.937	4.887	4.987	117		
	4	4.98	4.93	5.03	118		
	5	5.192	5.142	5.242	114		
Aroclor-1260	1	6.251	6.201	6.301	118	107	6.33
COLUMN 1	2	6.441	6.391	6.491	117		
	3	6.807	6.757	6.857	106		
	4	7.067	7.017	7.117	94.7		
	5	7.31	7.26	7.36	99.3		
	1	6.221	6.171	6.271	125		
COLUMN 2	2	6.409	6.359	6.459	123	114	
	3	6.561	6.511	6.611	115		
	4	7.031	6.981	7.081	101		
	5	7.273	7.223	7.323	106		



QC SAMPLE DATA

Report of Analysis

Client:	CDM Smith	Date Collected:	
Project:	South River WM Replacement	Date Received:	
Client Sample ID:	PB168735BL	SDG No.:	Q2514
Lab Sample ID:	PB168735BL	Matrix:	SOIL
Analytical Method:	8082A	% Solid:	100
Sample Wt/Vol:	30.01	Units:	g
Soil Aliquot Vol:			uL
Extraction Type:		Test:	PCB
GPC Factor :	1.0	Injection Volume :	
Prep Method :	SW3541B		

File ID/Qc Batch:	Dilution:	Prep Date	Date Analyzed	Prep Batch ID
PO112057.D	1	07/07/25 08:30	07/07/25 22:01	PB168735

CAS Number	Parameter	Conc.	Qualifier	MDL	LOQ / CRQL	Units(Dry Weight)
TARGETS						
12674-11-2	Aroclor-1016	3.90	U	3.90	17.0	ug/kg
11104-28-2	Aroclor-1221	4.00	U	4.00	17.0	ug/kg
11141-16-5	Aroclor-1232	3.70	U	3.70	17.0	ug/kg
53469-21-9	Aroclor-1242	4.00	U	4.00	17.0	ug/kg
12672-29-6	Aroclor-1248	5.90	U	5.90	17.0	ug/kg
11097-69-1	Aroclor-1254	3.20	U	3.20	17.0	ug/kg
37324-23-5	Aroclor-1262	5.00	U	5.00	17.0	ug/kg
11100-14-4	Aroclor-1268	3.60	U	3.60	17.0	ug/kg
11096-82-5	Aroclor-1260	3.20	U	3.20	17.0	ug/kg
SURROGATES						
877-09-8	Tetrachloro-m-xylene	20.5		32 - 144	103%	SPK: 20
2051-24-3	Decachlorobiphenyl	18.5		32 - 175	92%	SPK: 20

Comments:

U = Not Detected

LOQ = Limit of Quantitation

MDL = Method Detection Limit

LOD = Limit of Detection

E = Value Exceeds Calibration Range

P = Indicates >25% difference for detected concentrations between the two GC columns

Q = indicates LCS control criteria did not meet requirements

M = MS/MSD acceptance criteria did not meet requirements

J = Estimated Value

B = Analyte Found in Associated Method Blank

N = Presumptive Evidence of a Compound

* = Values outside of QC limits

D = Dilution

S = Indicates estimated value where valid five-point calibration was not performed prior to analyte detection in sample.

() = Laboratory InHouse Limit

Data Path : Z:\pestpcbsrv\HPCHEM1\ECD_O\Data\P0070725\
 Data File : PO112057.D
 Signal(s) : Signal #1: ECD1A.ch Signal #2: ECD2B.ch
 Acq On : 07 Jul 2025 22:01
 Operator : YP/AJ
 Sample : PB168735BL
 Misc :
 ALS Vial : 27 Sample Multiplier: 1

Instrument :
 ECD_O
 ClientSampleId :
 PB168735BL

Manual Integrations
APPROVED

Reviewed By :Yogesh Patel 07/08/2025
 Supervised By :mohammad ahmed 07/09/2025

Integration File signal 1: autoint1.e
 Integration File signal 2: autoint2.e
 Quant Time: Jul 08 01:58:08 2025
 Quant Method : Z:\pestpcbsrv\HPCHEM1\ECD_O\methods\P0061125.M
 Quant Title : GC EXTRACTABLES
 QLast Update : Thu Jun 12 06:25:26 2025
 Response via : Initial Calibration
 Integrator: ChemStation

Volume Inj. : 2 µl
 Signal #1 Phase : ZB-MR1 Signal #2 Phase: ZB-MR2
 Signal #1 Info : 30Mx0.32mmx 0.50µ Signal #2 Info : 30M x 0.32mm x 0.25µm

Compound	RT#1	RT#2	Resp#1	Resp#2	ng/ml	ng/ml

System Monitoring Compounds						
1) SA Tetrachlo...	3.672	3.668	118.3E6	104.3E6	20.538	18.560
2) SA Decachlor...	8.701	8.650	94275772	32869366	17.960	18.490m

Target Compounds

(f)=RT Delta > 1/2 Window (#)=Amounts differ by > 25% (m)=manual int.

Data Path : Z:\pestpcbsrv\HPCHEM1\ECD_O\Data\PO070725\
Data File : PO112057.D
Signal(s) : Signal #1: ECD1A.ch Signal #2: ECD2B.ch
Acq On : 07 Jul 2025 22:01
Operator : YP/AJ
Sample : PB168735BL
Misc :
ALS Vial : 27 Sample Multiplier: 1



Instrument :
ECD_O
ClientSampleId :
PB168735BL


Manual Integrations
APPROVED

Reviewed By :Yogesh Patel 07/08/2025
Supervised By :mohammad ahmed 07/09/2025

Integration File signal 1: autoint1.e
Integration File signal 2: autoint2.e
Quant Time: Jul 08 01:58:08 2025
Quant Method : Z:\pestpcbsrv\HPCHEM1\ECD_O\methods\PO061125.M
Quant Title : GC EXTRACTABLES
QLast Update : Thu Jun 12 06:25:26 2025
Response via : Initial Calibration
Integrator: ChemStation

Volume Inj. : 2 µl
Signal #1 Phase : ZB-MR1 Signal #2 Phase: ZB-MR2
Signal #1 Info : 30Mx0.32mmx 0.50µm Signal #2 Info : 30M x 0.32mm x 0.25µm






#1 Tetrachloro-m-xylene

R.T.: 3.672 min
 Delta R.T.: -0.005 min
 Response: 118288478
 Conc: 20.54 ng/ml

Instrument :
 ECD_O
 ClientSampleId :
 PB168735BL


Manual Integrations
 APPROVED

Reviewed By :Yogesh Patel 07/08/2025
 Supervised By :mohammad ahmed 07/09/2025




#1 Tetrachloro-m-xylene

R.T.: 3.668 min
 Delta R.T.: -0.006 min
 Response: 104258933
 Conc: 18.56 ng/ml



#2 Decachlorobiphenyl

R.T.: 8.701 min
 Delta R.T.: -0.011 min
 Response: 94275772
 Conc: 17.96 ng/ml



#2 Decachlorobiphenyl

R.T.: 8.650 min
 Delta R.T.: -0.011 min
 Response: 32869366
 Conc: 18.49 ng/ml m

Report of Analysis

Client:	CDM Smith	Date Collected:	06/11/25
Project:	South River WM Replacement	Date Received:	06/11/25
Client Sample ID:	PIBLK-PO111586.D	SDG No.:	Q2514
Lab Sample ID:	I.BLK-PO111586.D	Matrix:	WATER
Analytical Method:	8082A	% Solid:	0
Sample Wt/Vol:	1000	Units:	mL
Soil Aliquot Vol:			uL
Extraction Type:		Test:	PCB
GPC Factor :	1.0	PH :	
Prep Method :	5030	Injection Volume :	

File ID/Qc Batch:	Dilution:	Prep Date	Date Analyzed	Prep Batch ID
PO111586.D	1		06/11/25	po061125

CAS Number	Parameter	Conc.	Qualifier	MDL	LOQ / CRQL	Units
TARGETS						
12674-11-2	Aroclor-1016	0.097	U	0.097	0.50	ug/L
11104-28-2	Aroclor-1221	0.13	U	0.13	0.50	ug/L
11141-16-5	Aroclor-1232	0.096	U	0.096	0.50	ug/L
53469-21-9	Aroclor-1242	0.12	U	0.12	0.50	ug/L
12672-29-6	Aroclor-1248	0.071	U	0.071	0.50	ug/L
11097-69-1	Aroclor-1254	0.094	U	0.094	0.50	ug/L
11096-82-5	Aroclor-1260	0.081	U	0.081	0.50	ug/L
37324-23-5	Aroclor-1262	0.14	U	0.14	0.50	ug/L
11100-14-4	Aroclor-1268	0.11	U	0.11	0.50	ug/L
SURROGATES						
877-09-8	Tetrachloro-m-xylene	19.0		60 - 140	95%	SPK: 20
2051-24-3	Decachlorobiphenyl	20.5		60 - 140	103%	SPK: 20

Comments:

U = Not Detected

LOQ = Limit of Quantitation

MDL = Method Detection Limit

LOD = Limit of Detection

E = Value Exceeds Calibration Range

P = Indicates >25% difference for detected concentrations between the two GC columns

Q = indicates LCS control criteria did not meet requirements

M = MS/MSD acceptance criteria did not meet requirements

J = Estimated Value

B = Analyte Found in Associated Method Blank

N = Presumptive Evidence of a Compound

* = Values outside of QC limits

D = Dilution

S = Indicates estimated value where valid five-point calibration was not performed prior to analyte detection in sample.

() = Laboratory InHouse Limit

Data Path : Z:\pestpcbsrv\HPCHEM1\ECD_O\Data\PO061125\
Data File : PO111586.D
Signal(s) : Signal #1: ECD1A.ch Signal #2: ECD2B.ch
Acq On : 11 Jun 2025 10:21
Operator : YP/AJ
Sample : I.BLK
Misc :
ALS Vial : 2 Sample Multiplier: 1

Instrument :
ECD_O
ClientSampleId :
I.BLK

Integration File signal 1: autoint1.e
Integration File signal 2: autoint2.e
Quant Time: Jun 11 12:10:14 2025
Quant Method : Z:\pestpcbsrv\HPCHEM1\ECD_O\methods\PO061125.M
Quant Title : GC EXTRACTABLES
QLast Update : Wed Jun 11 12:09:32 2025
Response via : Initial Calibration
Integrator: ChemStation

Volume Inj. : 2 µl
Signal #1 Phase : ZB-MR1 Signal #2 Phase: ZB-MR2
Signal #1 Info : 30Mx0.32mmx 0.50µ Signal #2 Info : 30M x 0.32mm x 0.25µm

Compound	RT#1	RT#2	Resp#1	Resp#2	ng/ml	ng/ml

System Monitoring Compounds						
1) SA Tetrachlo...	3.677	3.673	117.0E6	106.8E6	20.308	19.004
2) SA Decachlor...	8.713	8.662	108.8E6	36486352	20.730	20.525

Target Compounds



(f)=RT Delta > 1/2 Window (#)=Amounts differ by > 25% (m)=manual int.


Data Path : Z:\pestpcbsrv\HPCHEM1\ECD_O\Data\PO061125\
Data File : PO111586.D
Signal(s) : Signal #1: ECD1A.ch Signal #2: ECD2B.ch
Acq On : 11 Jun 2025 10:21
Operator : YP/AJ
Sample : I.BLK
Misc :
ALS Vial : 2 Sample Multiplier: 1

Instrument :
ECD_O
ClientSampleId :
I.BLK

Integration File signal 1: autoint1.e
Integration File signal 2: autoint2.e
Quant Time: Jun 11 12:10:14 2025
Quant Method : Z:\pestpcbsrv\HPCHEM1\ECD_O\methods\PO061125.M
Quant Title : GC EXTRACTABLES
QLast Update : Wed Jun 11 12:09:32 2025
Response via : Initial Calibration
Integrator: ChemStation

Volume Inj. : 2 µl
Signal #1 Phase : ZB-MR1 Signal #2 Phase: ZB-MR2
Signal #1 Info : 30Mx0.32mmx 0.50µm Signal #2 Info : 30M x 0.32mm x 0.25µm






#1 Tetrachloro-m-xylene


R.T.: 3.677 min
Delta R.T.: 0.000 min
Response: 116962361
Conc: 20.31 ng/ml

Instrument :
ECD_O
ClientSampleId :
I.BLK




#1 Tetrachloro-m-xylene

R.T.: 3.673 min
Delta R.T.: 0.000 min
Response: 106750459
Conc: 19.00 ng/ml



#2 Decachlorobiphenyl

R.T.: 8.713 min
Delta R.T.: 0.001 min
Response: 108816924
Conc: 20.73 ng/ml



#2 Decachlorobiphenyl

R.T.: 8.662 min
Delta R.T.: 0.000 min
Response: 36486352
Conc: 20.52 ng/ml

Report of Analysis

Client:	CDM Smith	Date Collected:	07/07/25
Project:	South River WM Replacement	Date Received:	07/07/25
Client Sample ID:	PIBLK-PO112056.D	SDG No.:	Q2514
Lab Sample ID:	I.BLK-PO112056.D	Matrix:	WATER
Analytical Method:	8082A	% Solid:	0
Sample Wt/Vol:	1000	Units:	mL
Soil Aliquot Vol:			uL
Extraction Type:		Test:	PCB
GPC Factor :	1.0	PH :	
Prep Method :	5030	Injection Volume :	

File ID/Qc Batch:	Dilution:	Prep Date	Date Analyzed	Prep Batch ID
PO112056.D	1		07/07/25	po070725

CAS Number	Parameter	Conc.	Qualifier	MDL	LOQ / CRQL	Units
TARGETS						
12674-11-2	Aroclor-1016	0.097	U	0.097	0.50	ug/L
11104-28-2	Aroclor-1221	0.13	U	0.13	0.50	ug/L
11141-16-5	Aroclor-1232	0.096	U	0.096	0.50	ug/L
53469-21-9	Aroclor-1242	0.12	U	0.12	0.50	ug/L
12672-29-6	Aroclor-1248	0.071	U	0.071	0.50	ug/L
11097-69-1	Aroclor-1254	0.094	U	0.094	0.50	ug/L
11096-82-5	Aroclor-1260	0.081	U	0.081	0.50	ug/L
37324-23-5	Aroclor-1262	0.14	U	0.14	0.50	ug/L
11100-14-4	Aroclor-1268	0.11	U	0.11	0.50	ug/L
SURROGATES						
877-09-8	Tetrachloro-m-xylene	16.9		60 - 140	85%	SPK: 20
2051-24-3	Decachlorobiphenyl	16.8		60 - 140	84%	SPK: 20

Comments:

U = Not Detected

LOQ = Limit of Quantitation

MDL = Method Detection Limit

LOD = Limit of Detection

E = Value Exceeds Calibration Range

P = Indicates >25% difference for detected concentrations between the two GC columns

Q = indicates LCS control criteria did not meet requirements

M = MS/MSD acceptance criteria did not meet requirements

J = Estimated Value

B = Analyte Found in Associated Method Blank

N = Presumptive Evidence of a Compound

* = Values outside of QC limits

D = Dilution

S = Indicates estimated value where valid five-point calibration was not performed prior to analyte detection in sample.

() = Laboratory InHouse Limit

Data Path : Z:\pestpcbsrv\HPCHEM1\ECD_O\Data\P0070725\
Data File : P0112056.D
Signal(s) : Signal #1: ECD1A.ch Signal #2: ECD2B.ch
Acq On : 07 Jul 2025 21:44
Operator : YP/AJ
Sample : I.BLK
Misc :
ALS Vial : 2 Sample Multiplier: 1

Instrument :
ECD_O
ClientSampleId :
I.BLK

Integration File signal 1: autoint1.e
Integration File signal 2: autoint2.e
Quant Time: Jul 08 01:57:49 2025
Quant Method : Z:\pestpcbsrv\HPCHEM1\ECD_O\methods\P0061125.M
Quant Title : GC EXTRACTABLES
QLast Update : Thu Jun 12 06:25:26 2025
Response via : Initial Calibration
Integrator: ChemStation

Volume Inj. : 2 µl
Signal #1 Phase : ZB-MR1 Signal #2 Phase: ZB-MR2
Signal #1 Info : 30Mx0.32mmx 0.50µ Signal #2 Info : 30M x 0.32mm x 0.25µm

	Compound	RT#1	RT#2	Resp#1	Resp#2	ng/ml	ng/ml

System Monitoring Compounds							
1)	SA Tetrachlo...	3.674	3.668	108.0E6	95041675	18.759	16.920
2)	SA Decachlor...	8.705	8.651	87899271	30442472	16.745	17.125

Target Compounds



(f)=RT Delta > 1/2 Window (#)=Amounts differ by > 25% (m)=manual int.


Data Path : Z:\pestpcbsrv\HPCHEM1\ECD_O\Data\PO070725\
Data File : PO112056.D
Signal(s) : Signal #1: ECD1A.ch Signal #2: ECD2B.ch
Acq On : 07 Jul 2025 21:44
Operator : YP/AJ
Sample : I.BLK
Misc :
ALS Vial : 2 Sample Multiplier: 1

Instrument :
ECD_O
ClientSampleId :
I.BLK

Integration File signal 1: autoint1.e
Integration File signal 2: autoint2.e
Quant Time: Jul 08 01:57:49 2025
Quant Method : Z:\pestpcbsrv\HPCHEM1\ECD_O\methods\PO061125.M
Quant Title : GC EXTRACTABLES
QLast Update : Thu Jun 12 06:25:26 2025
Response via : Initial Calibration
Integrator: ChemStation

Volume Inj. : 2 µl
Signal #1 Phase : ZB-MR1 Signal #2 Phase: ZB-MR2
Signal #1 Info : 30Mx0.32mmx 0.50µm Signal #2 Info : 30M x 0.32mm x 0.25µm






#1 Tetrachloro-m-xylene


R.T.: 3.674 min
Delta R.T.: -0.003 min
Response: 108040557
Conc: 18.76 ng/ml

Instrument :
ECD_O
ClientSampleId :
I.BLK




#1 Tetrachloro-m-xylene

R.T.: 3.668 min
Delta R.T.: -0.006 min
Response: 95041675
Conc: 16.92 ng/ml



#2 Decachlorobiphenyl

R.T.: 8.705 min
Delta R.T.: -0.008 min
Response: 87899271
Conc: 16.75 ng/ml



#2 Decachlorobiphenyl

R.T.: 8.651 min
Delta R.T.: -0.010 min
Response: 30442472
Conc: 17.13 ng/ml

Report of Analysis

Client:	CDM Smith	Date Collected:	07/08/25
Project:	South River WM Replacement	Date Received:	07/08/25
Client Sample ID:	PIBLK-PO112072.D	SDG No.:	Q2514
Lab Sample ID:	I.BLK-PO112072.D	Matrix:	WATER
Analytical Method:	8082A	% Solid:	0
Sample Wt/Vol:	1000	Units:	mL
Soil Aliquot Vol:			uL
Extraction Type:		Test:	PCB
GPC Factor :	1.0	PH :	
Prep Method :	5030	Injection Volume :	

File ID/Qc Batch:	Dilution:	Prep Date	Date Analyzed	Prep Batch ID
PO112072.D	1		07/08/25	po070725

CAS Number	Parameter	Conc.	Qualifier	MDL	LOQ / CRQL	Units
TARGETS						
12674-11-2	Aroclor-1016	0.097	U	0.097	0.50	ug/L
11104-28-2	Aroclor-1221	0.13	U	0.13	0.50	ug/L
11141-16-5	Aroclor-1232	0.096	U	0.096	0.50	ug/L
53469-21-9	Aroclor-1242	0.12	U	0.12	0.50	ug/L
12672-29-6	Aroclor-1248	0.071	U	0.071	0.50	ug/L
11097-69-1	Aroclor-1254	0.094	U	0.094	0.50	ug/L
11096-82-5	Aroclor-1260	0.081	U	0.081	0.50	ug/L
37324-23-5	Aroclor-1262	0.14	U	0.14	0.50	ug/L
11100-14-4	Aroclor-1268	0.11	U	0.11	0.50	ug/L
SURROGATES						
877-09-8	Tetrachloro-m-xylene	18.3		60 - 140	92%	SPK: 20
2051-24-3	Decachlorobiphenyl	16.1		60 - 140	80%	SPK: 20

Comments:

U = Not Detected

LOQ = Limit of Quantitation

MDL = Method Detection Limit

LOD = Limit of Detection

E = Value Exceeds Calibration Range

P = Indicates >25% difference for detected concentrations between the two GC columns

Q = indicates LCS control criteria did not meet requirements

M = MS/MSD acceptance criteria did not meet requirements

J = Estimated Value

B = Analyte Found in Associated Method Blank

N = Presumptive Evidence of a Compound

* = Values outside of QC limits

D = Dilution

S = Indicates estimated value where valid five-point calibration was not performed prior to analyte detection in sample.

() = Laboratory InHouse Limit

Data Path : Z:\pestpcbsrv\HPCHEM1\ECD_O\Data\P0070725\
 Data File : PO112072.D
 Signal(s) : Signal #1: ECD1A.ch Signal #2: ECD2B.ch
 Acq On : 08 Jul 2025 05:26
 Operator : YP/AJ
 Sample : I.BLK
 Misc :
 ALS Vial : 2 Sample Multiplier: 1

Instrument :
 ECD_O
 ClientSampleId :
 I.BLK

Manual Integrations
APPROVED

Reviewed By :Yogesh Patel 07/08/2025
 Supervised By :mohammad ahmed 07/09/2025

Integration File signal 1: autoint1.e
 Integration File signal 2: autoint2.e
 Quant Time: Jul 08 06:03:16 2025
 Quant Method : Z:\pestpcbsrv\HPCHEM1\ECD_O\methods\P0061125.M
 Quant Title : GC EXTRACTABLES
 QLast Update : Thu Jun 12 06:25:26 2025
 Response via : Initial Calibration
 Integrator: ChemStation

Volume Inj. : 2 µl
 Signal #1 Phase : ZB-MR1 Signal #2 Phase: ZB-MR2
 Signal #1 Info : 30Mx0.32mmx 0.50µ Signal #2 Info : 30M x 0.32mm x 0.25µm

Compound	RT#1	RT#2	Resp#1	Resp#2	ng/ml	ng/ml

System Monitoring Compounds						
1) SA Tetrachlo...	3.674	3.668	118.6E6	102.8E6	20.593	18.308
2) SA Decachlor...	8.705	8.650	84376785	30871485	16.074	17.366m

Target Compounds

(f)=RT Delta > 1/2 Window (#)=Amounts differ by > 25% (m)=manual int.

Data Path : Z:\pestpcbsrv\HPCHEM1\ECD_O\Data\PO070725\
Data File : PO112072.D
Signal(s) : Signal #1: ECD1A.ch Signal #2: ECD2B.ch
Acq On : 08 Jul 2025 05:26
Operator : YP/AJ
Sample : I.BLK
Misc :
ALS Vial : 2 Sample Multiplier: 1



Instrument :
ECD_O
ClientSampleId :
I.BLK


Manual Integrations
APPROVED

Reviewed By :Yogesh Patel 07/08/2025
Supervised By :mohammad ahmed 07/09/2025

Integration File signal 1: autoint1.e
Integration File signal 2: autoint2.e
Quant Time: Jul 08 06:03:16 2025
Quant Method : Z:\pestpcbsrv\HPCHEM1\ECD_O\methods\PO061125.M
Quant Title : GC EXTRACTABLES
QLast Update : Thu Jun 12 06:25:26 2025
Response via : Initial Calibration
Integrator: ChemStation

Volume Inj. : 2 µl
Signal #1 Phase : ZB-MR1 Signal #2 Phase: ZB-MR2
Signal #1 Info : 30Mx0.32mmx 0.50µ Signal #2 Info : 30M x 0.32mm x 0.25µm






#1 Tetrachloro-m-xylene

R.T.: 3.674 min
 Delta R.T.: -0.003 min
 Response: 118601434
 Conc: 20.59 ng/ml

Instrument :
 ECD_O
 ClientSampleId :
 I.BLK


Manual Integrations
 APPROVED

Reviewed By :Yogesh Patel 07/08/2025
 Supervised By :mohammad ahmed 07/09/2025




#1 Tetrachloro-m-xylene

R.T.: 3.668 min
 Delta R.T.: -0.006 min
 Response: 102840508
 Conc: 18.31 ng/ml



#2 Decachlorobiphenyl

R.T.: 8.705 min
 Delta R.T.: -0.007 min
 Response: 84376785
 Conc: 16.07 ng/ml



#2 Decachlorobiphenyl

R.T.: 8.650 min
 Delta R.T.: -0.011 min
 Response: 30871485
 Conc: 17.37 ng/ml m

Report of Analysis

Client:	CDM Smith	Date Collected:	07/08/25
Project:	South River WM Replacement	Date Received:	07/08/25
Client Sample ID:	PIBLK-PO112085.D	SDG No.:	Q2514
Lab Sample ID:	I.BLK-PO112085.D	Matrix:	WATER
Analytical Method:	8082A	% Solid:	0
Sample Wt/Vol:	1000	Units:	mL
Soil Aliquot Vol:			uL
Extraction Type:		Test:	PCB
GPC Factor :	1.0	PH :	
Prep Method :	5030	Injection Volume :	

File ID/Qc Batch:	Dilution:	Prep Date	Date Analyzed	Prep Batch ID
PO112085.D	1		07/08/25	PO070725

CAS Number	Parameter	Conc.	Qualifier	MDL	LOQ / CRQL	Units
TARGETS						
12674-11-2	Aroclor-1016	0.097	U	0.097	0.50	ug/L
11104-28-2	Aroclor-1221	0.13	U	0.13	0.50	ug/L
11141-16-5	Aroclor-1232	0.096	U	0.096	0.50	ug/L
53469-21-9	Aroclor-1242	0.12	U	0.12	0.50	ug/L
12672-29-6	Aroclor-1248	0.071	U	0.071	0.50	ug/L
11097-69-1	Aroclor-1254	0.094	U	0.094	0.50	ug/L
11096-82-5	Aroclor-1260	0.081	U	0.081	0.50	ug/L
37324-23-5	Aroclor-1262	0.14	U	0.14	0.50	ug/L
11100-14-4	Aroclor-1268	0.11	U	0.11	0.50	ug/L
SURROGATES						
877-09-8	Tetrachloro-m-xylene	11.4	*	60 - 140	57%	SPK: 20
2051-24-3	Decachlorobiphenyl	11.1	*	60 - 140	55%	SPK: 20

Comments:

U = Not Detected

LOQ = Limit of Quantitation

MDL = Method Detection Limit

LOD = Limit of Detection

E = Value Exceeds Calibration Range

P = Indicates >25% difference for detected concentrations between the two GC columns

Q = indicates LCS control criteria did not meet requirements

M = MS/MSD acceptance criteria did not meet requirements

J = Estimated Value

B = Analyte Found in Associated Method Blank

N = Presumptive Evidence of a Compound

* = Values outside of QC limits

D = Dilution

S = Indicates estimated value where valid five-point calibration was not performed prior to analyte detection in sample.

() = Laboratory InHouse Limit

Data Path : Z:\pestpcbsrv\HPCHEM1\ECD_O\Data\P0070725\
 Data File : PO112085.D
 Signal(s) : Signal #1: ECD1A.ch Signal #2: ECD2B.ch
 Acq On : 08 Jul 2025 11:26
 Operator : YP/AJ
 Sample : I.BLK
 Misc :
 ALS Vial : 2 Sample Multiplier: 1

Instrument :
 ECD_O
 ClientSampleId :
 I.BLK

Manual Integrations
APPROVED

Reviewed By :Yogesh Patel 07/08/2025
 Supervised By :mohammad ahmed 07/09/2025

Integration File signal 1: autoint1.e
 Integration File signal 2: autoint2.e
 Quant Time: Jul 08 11:39:35 2025
 Quant Method : Z:\pestpcbsrv\HPCHEM1\ECD_O\methods\P0061125.M
 Quant Title : GC EXTRACTABLES
 QLast Update : Thu Jun 12 06:25:26 2025
 Response via : Initial Calibration
 Integrator: ChemStation

Volume Inj. : 2 µl
 Signal #1 Phase : ZB-MR1 Signal #2 Phase: ZB-MR2
 Signal #1 Info : 30Mx0.32mmx 0.50µ Signal #2 Info : 30M x 0.32mm x 0.25µm

Compound	RT#1	RT#2	Resp#1	Resp#2	ng/ml	ng/ml

System Monitoring Compounds						
1) SA Tetrachlo...	3.677	3.667	102.6E6	64101443	17.823	11.411 #
2) SA Decachlor...	8.706	8.650	75953626	19683829	14.469	11.073m

Target Compounds

(f)=RT Delta > 1/2 Window (#)=Amounts differ by > 25% (m)=manual int.

Data Path : Z:\pestpcbsrv\HPCHEM1\ECD_O\Data\PO070725\
Data File : PO112085.D
Signal(s) : Signal #1: ECD1A.ch Signal #2: ECD2B.ch
Acq On : 08 Jul 2025 11:26
Operator : YP/AJ
Sample : I.BLK
Misc :
ALS Vial : 2 Sample Multiplier: 1



Instrument :
ECD_O
ClientSampleId :
I.BLK


Manual Integrations
APPROVED

Reviewed By : Yogesh Patel 07/08/2025
Supervised By : mohammad ahmed 07/09/2025

Integration File signal 1: autoint1.e
Integration File signal 2: autoint2.e
Quant Time: Jul 08 11:39:35 2025
Quant Method : Z:\pestpcbsrv\HPCHEM1\ECD_O\methods\PO061125.M
Quant Title : GC EXTRACTABLES
QLast Update : Thu Jun 12 06:25:26 2025
Response via : Initial Calibration
Integrator: ChemStation

Volume Inj. : 2 µl
Signal #1 Phase : ZB-MR1 Signal #2 Phase: ZB-MR2
Signal #1 Info : 30Mx0.32mmx 0.50µ Signal #2 Info : 30M x 0.32mm x 0.25µm






#1 Tetrachloro-m-xylene

R.T.: 3.677 min
Delta R.T.: 0.000 min
Response: 102648239
Conc: 17.82 ng/ml

Instrument :
ECD_O
ClientSampleId :
I.BLK


Manual Integrations
APPROVED

Reviewed By :Yogesh Patel 07/08/2025
Supervised By :mohammad ahmed 07/09/2025




#1 Tetrachloro-m-xylene

R.T.: 3.667 min
Delta R.T.: -0.007 min
Response: 64101443
Conc: 11.41 ng/ml



#2 Decachlorobiphenyl

R.T.: 8.706 min
Delta R.T.: -0.007 min
Response: 75953626
Conc: 14.47 ng/ml



#2 Decachlorobiphenyl

R.T.: 8.650 min
Delta R.T.: -0.011 min
Response: 19683829
Conc: 11.07 ng/ml m

Report of Analysis

Client:	CDM Smith	Date Collected:	
Project:	South River WM Replacement	Date Received:	
Client Sample ID:	PB168735BS	SDG No.:	Q2514
Lab Sample ID:	PB168735BS	Matrix:	SOIL
Analytical Method:	8082A	% Solid:	100
Sample Wt/Vol:	30.02	Units:	g
Soil Aliquot Vol:			uL
Extraction Type:		Test:	PCB
GPC Factor :	1.0	PH :	
Prep Method :	SW3541B	Injection Volume :	

File ID/Qc Batch:	Dilution:	Prep Date	Date Analyzed	Prep Batch ID
PO112058.D	1	07/07/25 08:30	07/07/25 22:20	PB168735

CAS Number	Parameter	Conc.	Qualifier	MDL	LOQ / CRQL	Units(Dry Weight)
TARGETS						
12674-11-2	Aroclor-1016	150		3.90	17.0	ug/kg
11104-28-2	Aroclor-1221	4.00	U	4.00	17.0	ug/kg
11141-16-5	Aroclor-1232	3.70	U	3.70	17.0	ug/kg
53469-21-9	Aroclor-1242	4.00	U	4.00	17.0	ug/kg
12672-29-6	Aroclor-1248	5.90	U	5.90	17.0	ug/kg
11097-69-1	Aroclor-1254	3.20	U	3.20	17.0	ug/kg
37324-23-5	Aroclor-1262	5.00	U	5.00	17.0	ug/kg
11100-14-4	Aroclor-1268	3.60	U	3.60	17.0	ug/kg
11096-82-5	Aroclor-1260	163		3.20	17.0	ug/kg
SURROGATES						
877-09-8	Tetrachloro-m-xylene	19.5		32 - 144	98%	SPK: 20
2051-24-3	Decachlorobiphenyl	18.5		32 - 175	92%	SPK: 20

Comments:

U = Not Detected

LOQ = Limit of Quantitation

MDL = Method Detection Limit

LOD = Limit of Detection

E = Value Exceeds Calibration Range

P = Indicates >25% difference for detected concentrations between the two GC columns

Q = indicates LCS control criteria did not meet requirements

M = MS/MSD acceptance criteria did not meet requirements

J = Estimated Value

B = Analyte Found in Associated Method Blank

N = Presumptive Evidence of a Compound

* = Values outside of QC limits

D = Dilution

S = Indicates estimated value where valid five-point calibration was not performed prior to analyte detection in sample.

() = Laboratory InHouse Limit

Data Path : Z:\pestpcbsrv\HPCHEM1\ECD_O\Data\P0070725\
 Data File : PO112058.D
 Signal(s) : Signal #1: ECD1A.ch Signal #2: ECD2B.ch
 Acq On : 07 Jul 2025 22:20
 Operator : YP/AJ
 Sample : PB168735BS
 Misc :
 ALS Vial : 28 Sample Multiplier: 1

Instrument :
 ECD_O
 ClientSampleId :
 PB168735BS

Manual Integrations
 APPROVED

Reviewed By :Yogesh Patel 07/08/2025
 Supervised By :mohammad ahmed 07/09/2025

Integration File signal 1: autoint1.e
 Integration File signal 2: autoint2.e
 Quant Time: Jul 08 01:58:25 2025
 Quant Method : Z:\pestpcbsrv\HPCHEM1\ECD_O\methods\P0061125.M
 Quant Title : GC EXTRACTABLES
 QLast Update : Thu Jun 12 06:25:26 2025
 Response via : Initial Calibration
 Integrator: ChemStation

Volume Inj. : 2 µl
 Signal #1 Phase : ZB-MR1 Signal #2 Phase: ZB-MR2
 Signal #1 Info : 30Mx0.32mmx 0.50µ Signal #2 Info : 30M x 0.32mm x 0.25µm

Compound	RT#1	RT#2	Resp#1	Resp#2	ng/ml	ng/ml

System Monitoring Compounds						
1) SA Tetrachlo...	3.673	3.667	112.3E6	98176784	19.505	17.478
2) SA Decachlor...	8.703	8.650	96097820	32807472	18.307	18.455m
Target Compounds						
3) L1 AR-1016-1	4.760	4.743	103.5E6	90065091	433.985	457.724
4) L1 AR-1016-2	4.779	4.762	151.2E6	134.4E6	454.534	471.858
5) L1 AR-1016-3	4.835	4.937	98586024	67537633	428.373	449.534
6) L1 AR-1016-4	4.955	4.979	80955342	54147742	436.505	444.609
7) L1 AR-1016-5	5.212	5.191	80166902	67903820	429.519	426.536
31) L7 AR-1260-1	6.250	6.220	164.9E6	121.2E6	464.098	482.954
32) L7 AR-1260-2	6.440	6.409	245.7E6	147.3E6	520.129	500.907
33) L7 AR-1260-3	6.807	6.560	203.0E6	127.1E6	478.109	475.881
34) L7 AR-1260-4	7.066	7.030	158.0E6	78438696	502.413	412.626
35) L7 AR-1260-5	7.308	7.271	402.4E6	173.1E6	488.532	407.358

(f)=RT Delta > 1/2 Window (#)=Amounts differ by > 25% (m)=manual int.

Data Path : Z:\pestpcbsrv\HPCHEM1\ECD_O\Data\PO070725\
Data File : PO112058.D
Signal(s) : Signal #1: ECD1A.ch Signal #2: ECD2B.ch
Acq On : 07 Jul 2025 22:20
Operator : YP/AJ
Sample : PB168735BS
Misc :
ALS Vial : 28 Sample Multiplier: 1



Instrument :
ECD_O
ClientSampleId :
PB168735BS

Manual Integrations
APPROVED

Reviewed By :Yogesh Patel 07/08/2025
Supervised By :mohammad ahmed 07/09/2025

Integration File signal 1: autoint1.e
Integration File signal 2: autoint2.e
Quant Time: Jul 08 01:58:25 2025
Quant Method : Z:\pestpcbsrv\HPCHEM1\ECD_O\methods\PO061125.M
Quant Title : GC EXTRACTABLES
QLast Update : Thu Jun 12 06:25:26 2025
Response via : Initial Calibration
Integrator: ChemStation

Volume Inj. : 2 µl
Signal #1 Phase : ZB-MR1 Signal #2 Phase: ZB-MR2
Signal #1 Info : 30Mx0.32mmx 0.50µ Signal #2 Info : 30M x 0.32mm x 0.25µm



Report of Analysis

Client:	CDM Smith	Date Collected:	07/03/25
Project:	South River WM Replacement	Date Received:	07/03/25
Client Sample ID:	HR-3-070325MS	SDG No.:	Q2514
Lab Sample ID:	Q2513-03MS	Matrix:	SOIL
Analytical Method:	8082A	% Solid:	94.4
Sample Wt/Vol:	30.06	Units:	g
Soil Aliquot Vol:			uL
Extraction Type:		Test:	PCB
GPC Factor :	1.0	Injection Volume :	
Prep Method :	SW3541B		

File ID/Qc Batch:	Dilution:	Prep Date	Date Analyzed	Prep Batch ID
PO112063.D	1	07/07/25 08:30	07/07/25 23:52	PB168735

CAS Number	Parameter	Conc.	Qualifier	MDL	LOQ / CRQL	Units(Dry Weight)
TARGETS						
12674-11-2	Aroclor-1016	118		4.20	18.0	ug/kg
11104-28-2	Aroclor-1221	4.30	U	4.30	18.0	ug/kg
11141-16-5	Aroclor-1232	3.90	U	3.90	18.0	ug/kg
53469-21-9	Aroclor-1242	4.20	U	4.20	18.0	ug/kg
12672-29-6	Aroclor-1248	6.30	U	6.30	18.0	ug/kg
11097-69-1	Aroclor-1254	3.40	U	3.40	18.0	ug/kg
37324-23-5	Aroclor-1262	5.30	U	5.30	18.0	ug/kg
11100-14-4	Aroclor-1268	3.80	U	3.80	18.0	ug/kg
11096-82-5	Aroclor-1260	112		3.40	18.0	ug/kg
SURROGATES						
877-09-8	Tetrachloro-m-xylene	15.0		32 - 144	75%	SPK: 20
2051-24-3	Decachlorobiphenyl	12.3		32 - 175	61%	SPK: 20

Comments:

U = Not Detected

LOQ = Limit of Quantitation

MDL = Method Detection Limit

LOD = Limit of Detection

E = Value Exceeds Calibration Range

P = Indicates >25% difference for detected concentrations between the two GC columns

Q = indicates LCS control criteria did not meet requirements

M = MS/MSD acceptance criteria did not meet requirements

J = Estimated Value

B = Analyte Found in Associated Method Blank

N = Presumptive Evidence of a Compound

* = Values outside of QC limits

D = Dilution

S = Indicates estimated value where valid five-point calibration was not performed prior to analyte detection in sample.

() = Laboratory InHouse Limit

Data Path : Z:\pestpcbsrv\HPCHEM1\ECD_O\Data\P0070725\
 Data File : PO112063.D
 Signal(s) : Signal #1: ECD1A.ch Signal #2: ECD2B.ch
 Acq On : 07 Jul 2025 23:52
 Operator : YP/AJ
 Sample : Q2513-03MS
 Misc :
 ALS Vial : 33 Sample Multiplier: 1

Instrument :
 ECD_O
 ClientSampleId :
 HR-3-070325MS

Integration File signal 1: autoint1.e
 Integration File signal 2: autoint2.e
 Quant Time: Jul 08 01:59:55 2025
 Quant Method : Z:\pestpcbsrv\HPCHEM1\ECD_O\methods\P0061125.M
 Quant Title : GC EXTRACTABLES
 QLast Update : Thu Jun 12 06:25:26 2025
 Response via : Initial Calibration
 Integrator: ChemStation

Volume Inj. : 2 µl
 Signal #1 Phase : ZB-MR1 Signal #2 Phase: ZB-MR2
 Signal #1 Info : 30Mx0.32mmx 0.50µ Signal #2 Info : 30M x 0.32mm x 0.25µm

Compound	RT#1	RT#2	Resp#1	Resp#2	ng/ml	ng/ml

System Monitoring Compounds						
1) SA Tetrachlo...	3.674	3.668	86147751	77287181	14.958	13.759
2) SA Decachlor...	8.703	8.650	51670030	21820148	9.843	12.275
Target Compounds						
3) L1 AR-1016-1	4.760	4.744	78200674	66821646	327.812	339.598
4) L1 AR-1016-2	4.779	4.762	115.5E6	98175300	347.242	344.619
5) L1 AR-1016-3	4.836	4.937	82099146	48183857	356.734	320.714
6) L1 AR-1016-4	4.956	4.980	61822973	38803598	333.344	318.618
7) L1 AR-1016-5	5.213	5.192	57403751	49306136	307.558	309.715
31) L7 AR-1260-1	6.251	6.221	119.4E6	90836506	336.086	362.020
32) L7 AR-1260-2	6.440	6.409	153.8E6	102.2E6	325.472	347.746
33) L7 AR-1260-3	6.806	6.561	119.1E6	83781530	280.535	313.578
34) L7 AR-1260-4	7.065	7.031	82025229	51905123	260.855	273.046
35) L7 AR-1260-5	7.308	7.272	239.3E6	125.6E6	290.531	295.624



(f)=RT Delta > 1/2 Window (#)=Amounts differ by > 25% (m)=manual int.

Data Path : Z:\pestpcbsrv\HPCHEM1\ECD_O\Data\PO070725\
Data File : PO112063.D
Signal(s) : Signal #1: ECD1A.ch Signal #2: ECD2B.ch
Acq On : 07 Jul 2025 23:52
Operator : YP/AJ
Sample : Q2513-03MS
Misc :
ALS Vial : 33 Sample Multiplier: 1

Instrument :
ECD_O
ClientSampleId :
HR-3-070325MS

Integration File signal 1: autoint1.e
Integration File signal 2: autoint2.e
Quant Time: Jul 08 01:59:55 2025
Quant Method : Z:\pestpcbsrv\HPCHEM1\ECD_O\methods\PO061125.M
Quant Title : GC EXTRACTABLES
QLast Update : Thu Jun 12 06:25:26 2025
Response via : Initial Calibration
Integrator: ChemStation

Volume Inj. : 2 µl
Signal #1 Phase : ZB-MR1 Signal #2 Phase: ZB-MR2
Signal #1 Info : 30Mx0.32mmx 0.50µ Signal #2 Info : 30M x 0.32mm x 0.25µm



Report of Analysis

Client:	CDM Smith	Date Collected:	07/03/25
Project:	South River WM Replacement	Date Received:	07/03/25
Client Sample ID:	HR-3-070325MSD	SDG No.:	Q2514
Lab Sample ID:	Q2513-03MSD	Matrix:	SOIL
Analytical Method:	8082A	% Solid:	94.4
Sample Wt/Vol:	30.02	Units:	g
Soil Aliquot Vol:			uL
Extraction Type:		Test:	PCB
GPC Factor :	1.0	Injection Volume :	
Prep Method :	SW3541B		

File ID/Qc Batch:	Dilution:	Prep Date	Date Analyzed	Prep Batch ID
PO112064.D	1	07/07/25 08:30	07/08/25 00:10	PB168735

CAS Number	Parameter	Conc.	Qualifier	MDL	LOQ / CRQL	Units(Dry Weight)
TARGETS						
12674-11-2	Aroclor-1016	119		4.20	18.0	ug/kg
11104-28-2	Aroclor-1221	4.30	U	4.30	18.0	ug/kg
11141-16-5	Aroclor-1232	3.90	U	3.90	18.0	ug/kg
53469-21-9	Aroclor-1242	4.20	U	4.20	18.0	ug/kg
12672-29-6	Aroclor-1248	6.30	U	6.30	18.0	ug/kg
11097-69-1	Aroclor-1254	3.40	U	3.40	18.0	ug/kg
37324-23-5	Aroclor-1262	5.30	U	5.30	18.0	ug/kg
11100-14-4	Aroclor-1268	3.80	U	3.80	18.0	ug/kg
11096-82-5	Aroclor-1260	114		3.40	18.0	ug/kg
SURROGATES						
877-09-8	Tetrachloro-m-xylene	14.9		32 - 144	74%	SPK: 20
2051-24-3	Decachlorobiphenyl	13.0		32 - 175	65%	SPK: 20

Comments:

U = Not Detected

LOQ = Limit of Quantitation

MDL = Method Detection Limit

LOD = Limit of Detection

E = Value Exceeds Calibration Range

P = Indicates >25% difference for detected concentrations between the two GC columns

Q = indicates LCS control criteria did not meet requirements

M = MS/MSD acceptance criteria did not meet requirements

J = Estimated Value

B = Analyte Found in Associated Method Blank

N = Presumptive Evidence of a Compound

* = Values outside of QC limits

D = Dilution

S = Indicates estimated value where valid five-point calibration was not performed prior to analyte detection in sample.

() = Laboratory InHouse Limit

Data Path : Z:\pestpcbsrv\HPCHEM1\ECD_O\Data\P0070725\
 Data File : PO112064.D
 Signal(s) : Signal #1: ECD1A.ch Signal #2: ECD2B.ch
 Acq On : 08 Jul 2025 00:10
 Operator : YP/AJ
 Sample : Q2513-03MSD
 Misc :
 ALS Vial : 34 Sample Multiplier: 1

Instrument :
 ECD_O
 ClientSampleId :
 HR-3-070325MSD

Integration File signal 1: autoint1.e
 Integration File signal 2: autoint2.e
 Quant Time: Jul 08 02:00:13 2025
 Quant Method : Z:\pestpcbsrv\HPCHEM1\ECD_O\methods\P0061125.M
 Quant Title : GC EXTRACTABLES
 QLast Update : Thu Jun 12 06:25:26 2025
 Response via : Initial Calibration
 Integrator: ChemStation

Volume Inj. : 2 µl
 Signal #1 Phase : ZB-MR1 Signal #2 Phase: ZB-MR2
 Signal #1 Info : 30Mx0.32mmx 0.50µ Signal #2 Info : 30M x 0.32mm x 0.25µm

Compound	RT#1	RT#2	Resp#1	Resp#2	ng/ml	ng/ml

System Monitoring Compounds						
1) SA Tetrachlo...	3.674	3.668	85661350	76208695	14.873	13.567
2) SA Decachlor...	8.705	8.652	54301305	23091272	10.345	12.990 #
Target Compounds						
3) L1 AR-1016-1	4.761	4.744	70492495	68966946	295.500	350.500
4) L1 AR-1016-2	4.780	4.762	102.7E6	98061592	308.794	344.220
5) L1 AR-1016-3	4.837	4.937	69816384	49939334	303.364	332.399
6) L1 AR-1016-4	4.956	4.980	56974073	40628729	307.200	333.604
7) L1 AR-1016-5	5.214	5.192	62614519	51353266	335.477	322.574
31) L7 AR-1260-1	6.251	6.221	119.2E6	89006782	335.516	354.727
32) L7 AR-1260-2	6.441	6.409	156.4E6	102.9E6	331.050	349.981
33) L7 AR-1260-3	6.807	6.561	127.0E6	87299198	299.106	326.744
34) L7 AR-1260-4	7.067	7.031	84385983	54253701	268.363	285.401
35) L7 AR-1260-5	7.310	7.273	231.9E6	127.5E6	281.506	300.103



(f)=RT Delta > 1/2 Window (#)=Amounts differ by > 25% (m)=manual int.

Data Path : Z:\pestpcbsrv\HPCHEM1\ECD_O\Data\PO070725\
Data File : PO112064.D
Signal(s) : Signal #1: ECD1A.ch Signal #2: ECD2B.ch
Acq On : 08 Jul 2025 00:10
Operator : YP/AJ
Sample : Q2513-03MSD
Misc :
ALS Vial : 34 Sample Multiplier: 1

Instrument :
ECD_O
ClientSampleId :
HR-3-070325MSD

Integration File signal 1: autoint1.e
Integration File signal 2: autoint2.e
Quant Time: Jul 08 02:00:13 2025
Quant Method : Z:\pestpcbsrv\HPCHEM1\ECD_O\methods\PO061125.M
Quant Title : GC EXTRACTABLES
QLast Update : Thu Jun 12 06:25:26 2025
Response via : Initial Calibration
Integrator: ChemStation

Volume Inj. : 2 µl
Signal #1 Phase : ZB-MR1 Signal #2 Phase: ZB-MR2
Signal #1 Info : 30Mx0.32mmx 0.50µ Signal #2 Info : 30M x 0.32mm x 0.25µm



Manual Integration Report

Sequence:

po061125

Instrument

ECD_o

Sample ID	File ID	Parameter	Review By	Review On	Supervised By	Supervised On	Reason
AR1660ICC050	PO111591.D	AR-1016-1	yogesh	6/12/2025 8:55:39 AM	mohammad	6/13/2025 1:40:28	Peak Integrated by Software
AR1660ICC050	PO111591.D	AR-1016-2	yogesh	6/12/2025 8:55:39 AM	mohammad	6/13/2025 1:40:28	Peak Integrated by Software
AR1660ICC050	PO111591.D	AR-1016-3	yogesh	6/12/2025 8:55:39 AM	mohammad	6/13/2025 1:40:28	Peak Integrated by Software
AR1660ICC050	PO111591.D	AR-1016-4	yogesh	6/12/2025 8:55:39 AM	mohammad	6/13/2025 1:40:28	Peak Integrated by Software
AR1660ICC050	PO111591.D	AR-1016-5	yogesh	6/12/2025 8:55:39 AM	mohammad	6/13/2025 1:40:28	Peak Integrated by Software
AR1660ICC050	PO111591.D	AR-1260-1	yogesh	6/12/2025 8:55:39 AM	mohammad	6/13/2025 1:40:28	Peak Integrated by Software
AR1660ICC050	PO111591.D	AR-1260-1 #2	yogesh	6/12/2025 8:55:39 AM	mohammad	6/13/2025 1:40:28	Peak Integrated by Software
AR1660ICC050	PO111591.D	AR-1260-2	yogesh	6/12/2025 8:55:39 AM	mohammad	6/13/2025 1:40:28	Peak Integrated by Software
AR1660ICC050	PO111591.D	AR-1260-2 #2	yogesh	6/12/2025 8:55:39 AM	mohammad	6/13/2025 1:40:28	Peak Integrated by Software
AR1242ICC050	PO111598.D	AR-1242-1	yogesh	6/12/2025 8:55:41 AM	mohammad	6/13/2025 1:40:28	Peak Integrated by Software
AR1242ICC050	PO111598.D	AR-1242-2	yogesh	6/12/2025 8:55:41 AM	mohammad	6/13/2025 1:40:28	Peak Integrated by Software
AR1242ICC050	PO111598.D	AR-1242-3	yogesh	6/12/2025 8:55:41 AM	mohammad	6/13/2025 1:40:28	Peak Integrated by Software
AR1242ICC050	PO111598.D	AR-1242-4	yogesh	6/12/2025 8:55:41 AM	mohammad	6/13/2025 1:40:28	Peak Integrated by Software

Manual Integration Report

Sequence:	po061125	Instrument	ECD_o
-----------	----------	------------	-------

Sample ID	File ID	Parameter	Review By	Review On	Supervised By	Supervised On	Reason
AR1242ICC050	PO111598.D	AR-1242-5	yogesh	6/12/2025 8:55:41 AM	mohammad	6/13/2025 1:40:28	Peak Integrated by Software
AR1248ICC050	PO111603.D	AR-1248-1	yogesh	6/12/2025 8:55:43 AM	mohammad	6/13/2025 1:40:28	Peak Integrated by Software
AR1248ICC050	PO111603.D	AR-1248-3	yogesh	6/12/2025 8:55:43 AM	mohammad	6/13/2025 1:40:28	Peak Integrated by Software
AR1248ICC050	PO111603.D	AR-1248-4	yogesh	6/12/2025 8:55:43 AM	mohammad	6/13/2025 1:40:28	Peak Integrated by Software
AR1248ICC050	PO111603.D	AR-1248-5	yogesh	6/12/2025 8:55:43 AM	mohammad	6/13/2025 1:40:28	Peak Integrated by Software
AR1254ICC750	PO111605.D	Tetrachloro-m-xylene	yogesh	6/12/2025 8:55:45 AM	mohammad	6/13/2025 1:40:28	Peak Integrated by Software
AR1268ICC1000	PO111610.D	AR-1268-1 #2	yogesh	6/12/2025 8:55:46 AM	mohammad	6/13/2025 1:40:28	Peak Integrated by Software
AR1268ICC250	PO111613.D	AR-1268-1 #2	yogesh	6/12/2025 8:55:48 AM	mohammad	6/13/2025 1:40:28	Peak Integrated by Software
AR1268ICC250	PO111613.D	AR-1268-4 #2	yogesh	6/12/2025 8:55:48 AM	mohammad	6/13/2025 1:40:28	Peak Integrated by Software
AR1268ICV500	PO111619.D	AR-1268-1 #2	yogesh	6/12/2025 8:55:50 AM	mohammad	6/13/2025 1:40:28	Peak Integrated by Software

Manual Integration Report

Sequence:	po070725	Instrument	ECD_o
-----------	----------	------------	-------

Sample ID	File ID	Parameter	Review By	Review On	Supervised By	Supervised On	Reason
AR1660CCC500	PO112022.D	Decachlorobiphenyl #2	yogesh	7/8/2025 8:02:13 AM	 	 	Peak Integrated by Software
AR1248CCC500	PO112024.D	Decachlorobiphenyl #2	yogesh	7/8/2025 8:02:15 AM	 	 	Peak Integrated by Software
AR1254CCC500	PO112025.D	Decachlorobiphenyl #2	yogesh	7/8/2025 8:02:17 AM	 	 	Peak Integrated by Software
AR1660CCC500	PO112037.D	Decachlorobiphenyl #2	yogesh	7/8/2025 10:01:11 AM	 	 	Peak Integrated by Software
AR1242CCC500	PO112038.D	Decachlorobiphenyl #2	yogesh	7/8/2025 10:01:13 AM	 	 	Peak Integrated by Software
AR1248CCC500	PO112039.D	Decachlorobiphenyl #2	yogesh	7/8/2025 10:01:15 AM	 	 	Peak Integrated by Software
I.BLK	PO112041.D	Decachlorobiphenyl #2	yogesh	7/8/2025 10:01:16 AM	 	 	Peak Integrated by Software
AR1660CCC500	PO112052.D	Decachlorobiphenyl #2	yogesh	7/8/2025 10:01:28 AM	 	 	Peak Integrated by Software
AR1242CCC500	PO112053.D	Decachlorobiphenyl #2	yogesh	7/8/2025 10:01:30 AM	 	 	Peak Integrated by Software
AR1248CCC500	PO112054.D	Decachlorobiphenyl #2	yogesh	7/8/2025 10:01:31 AM	 	 	Peak Integrated by Software
AR1248CCC500	PO112054.D	Tetrachloro-m-xylene	yogesh	7/8/2025 10:01:31 AM	 	 	Peak Integrated by Software
AR1254CCC500	PO112055.D	Decachlorobiphenyl #2	yogesh	7/8/2025 10:01:33 AM	 	 	Peak Integrated by Software
PB168735BL	PO112057.D	Decachlorobiphenyl #2	yogesh	7/8/2025 10:02:22 AM	 	 	Peak Integrated by Software

Manual Integration Report

Sequence:	po070725	Instrument	ECD_o
-----------	----------	------------	-------

Sample ID	File ID	Parameter	Review By	Review On	Supervised By	Supervised On	Reason
PB168735BS	PO112058.D	Decachlorobiphenyl #2	yogesh	7/8/2025 10:02:24 AM	 	 	Peak Integrated by Software
Q2514-01	PO112065.D	Decachlorobiphenyl #2	yogesh	7/8/2025 10:02:34 AM	 	 	Peak Integrated by Software
Q2514-02	PO112066.D	Decachlorobiphenyl #2	yogesh	7/8/2025 10:02:35 AM	 	 	Peak Integrated by Software
AR1660CCC500	PO112067.D	AR-1016-5	yogesh	7/8/2025 10:02:37 AM	 	 	Peak Integrated by Software
AR1660CCC500	PO112067.D	AR-1260-2 #2	yogesh	7/8/2025 10:02:37 AM	 	 	Peak Integrated by Software
AR1660CCC500	PO112067.D	AR-1260-3 #2	yogesh	7/8/2025 10:02:37 AM	 	 	Peak Integrated by Software
AR1660CCC500	PO112067.D	AR-1260-5	yogesh	7/8/2025 10:02:37 AM	 	 	Peak Integrated by Software
AR1660CCC500	PO112067.D	Decachlorobiphenyl #2	yogesh	7/8/2025 10:02:37 AM	 	 	Peak Integrated by Software
AR1242CCC500	PO112068.D	Decachlorobiphenyl #2	yogesh	7/8/2025 10:02:39 AM	 	 	Peak Integrated by Software
AR1248CCC500	PO112069.D	AR-1248-4 #2	yogesh	7/8/2025 10:02:41 AM	 	 	Peak Integrated by Software
AR1248CCC500	PO112069.D	Decachlorobiphenyl #2	yogesh	7/8/2025 10:02:41 AM	 	 	Peak Integrated by Software
AR1254CCC500	PO112070.D	Decachlorobiphenyl #2	yogesh	7/8/2025 10:02:42 AM	 	 	Peak Integrated by Software
I.BLK	PO112072.D	Decachlorobiphenyl #2	yogesh	7/8/2025 10:03:15 AM	 	 	Peak Integrated by Software

Manual Integration Report

Sequence:

po070725

Instrument

ECD_o

Sample ID	File ID	Parameter	Review By	Review On	Supervised By	Supervised On	Reason
Q2514-03	PO112073.D	Decachlorobiphenyl #2	yogesh	7/8/2025 10:03:18 AM	 	 	Peak Integrated by Software
Q2514-04	PO112074.D	Decachlorobiphenyl #2	yogesh	7/8/2025 10:02:44 AM	 	 	Peak Integrated by Software
Q2514-05	PO112075.D	Decachlorobiphenyl #2	yogesh	7/8/2025 10:02:46 AM	 	 	Peak Integrated by Software
Q2514-06	PO112076.D	Decachlorobiphenyl #2	yogesh	7/8/2025 10:02:48 AM	 	 	Peak Integrated by Software
Q2514-07	PO112077.D	Decachlorobiphenyl #2	yogesh	7/8/2025 10:02:50 AM	 	 	Peak Integrated by Software
Q2514-08	PO112078.D	Decachlorobiphenyl #2	yogesh	7/8/2025 10:02:52 AM	 	 	Peak Integrated by Software
Q2514-09	PO112079.D	Decachlorobiphenyl #2	yogesh	7/8/2025 10:02:54 AM	 	 	Peak Integrated by Software
Q2514-10	PO112080.D	Decachlorobiphenyl #2	yogesh	7/8/2025 10:02:55 AM	 	 	Peak Integrated by Software
AR1660CCC500	PO112083.D	AR-1016-1	yogesh	7/8/2025 12:26:48 PM	 	 	Peak Integrated by Software
AR1660CCC500	PO112083.D	AR-1016-2	yogesh	7/8/2025 12:26:48 PM	 	 	Peak Integrated by Software
AR1660CCC500	PO112083.D	AR-1016-3	yogesh	7/8/2025 12:26:48 PM	 	 	Peak Integrated by Software
AR1660CCC500	PO112083.D	AR-1016-4	yogesh	7/8/2025 12:26:48 PM	 	 	Peak Integrated by Software
AR1660CCC500	PO112083.D	AR-1016-5 #2	yogesh	7/8/2025 12:26:48 PM	 	 	Peak Integrated by Software



284 Sheffield Street, Mountainside, New Jersey 07092, Phone : 908 789 8900, Fax : 908 789 8922

Manual Integration Report

Sequence:

po070725

Instrument

ECD_o

Sample ID	File ID	Parameter	Review By	Review On	Supervised By	Supervised On	Reason
AR1660CCC500	PO112083.D	Decachlorobiphenyl #2	yogesh	7/8/2025 12:26:48 PM	 	 	Peak Integrated by Software
I.BLK	PO112085.D	Decachlorobiphenyl #2	yogesh	7/8/2025 12:26:49 PM	 	 	Peak Integrated by Software

Instrument ID: ECD_O

Daily Analysis Runlog For Sequence/QC Batch ID # PO061125

Review By	yogesh	Review On	6/11/2025 3:41:50 PM
Supervise By	mohammad	Supervise On	6/13/2025 1:40:28 AM
SubDirectory	PO061125	HP Acquire Method	HP Processing Method PO061125
STD. NAME	STD REF.#		
Tune/Reschk			
Initial Calibration Stds	PP24330,PP24331,PP24332,PP24333,PP24334,PP24335,PP24336,PP24337,PP24338,PP24339,PP24340,PP24341,PP24342,PP24343,PP24344,PP24345,PP24346,PP24347,PP24348,PP24349,PP24350,PP24351,PP24352,PP24353,PP24354,PP24355,PP24356,PP24357,PP24358,PP24359,PP24360,PP24361,PP24362,PP24363,PP24364,PP24365,PP24366,PP24367,PP24368,PP24369		
CCC	PP24332,PP24347,PP24352,PP24357		
Internal Standard/PEM			
ICV/I.BLK	PP24370,PP24371,PP24372,PP24373,PP24374,PP24375,PP24376,PP24377,PP24378,PP24379,PP24380,PP24381,PP24382,PP24384,PP24385,PP24386,PP2		
Surrogate Standard			
MS/MSD Standard			
LCS Standard			

Sr#	SampleId	Data File Name	Date-Time	Operator	Status
1	HEXANE	PO111585.D	11 Jun 2025 10:03	YP/AJ	Ok
2	I.BLK	PO111586.D	11 Jun 2025 10:21	YP/AJ	Ok
3	AR1660ICC1000	PO111587.D	11 Jun 2025 10:40	YP/AJ	Ok
4	AR1660ICC750	PO111588.D	11 Jun 2025 10:58	YP/AJ	Ok
5	AR1660ICC500	PO111589.D	11 Jun 2025 11:17	YP/AJ	Ok
6	AR1660ICC250	PO111590.D	11 Jun 2025 11:35	YP/AJ	Ok
7	AR1660ICC050	PO111591.D	11 Jun 2025 11:53	YP/AJ	Ok,M
8	AR1221ICC500	PO111592.D	11 Jun 2025 12:12	YP/AJ	Ok
9	AR1232ICC500	PO111593.D	11 Jun 2025 12:30	YP/AJ	Ok
10	AR1242ICC1000	PO111594.D	11 Jun 2025 12:48	YP/AJ	Ok
11	AR1242ICC750	PO111595.D	11 Jun 2025 13:07	YP/AJ	Ok
12	AR1242ICC500	PO111596.D	11 Jun 2025 13:25	YP/AJ	Ok
13	AR1242ICC250	PO111597.D	11 Jun 2025 13:44	YP/AJ	Ok
14	AR1242ICC050	PO111598.D	11 Jun 2025 14:02	YP/AJ	Ok,M
15	AR1248ICC1000	PO111599.D	11 Jun 2025 14:20	YP/AJ	Ok
16	AR1248ICC750	PO111600.D	11 Jun 2025 14:39	YP/AJ	Ok
17	AR1248ICC500	PO111601.D	11 Jun 2025 15:14	YP/AJ	Ok
18	AR1248ICC250	PO111602.D	11 Jun 2025 15:32	YP/AJ	Ok
19	AR1248ICC050	PO111603.D	11 Jun 2025 15:49	YP/AJ	Ok,M
20	AR1254ICC1000	PO111604.D	11 Jun 2025 16:06	YP/AJ	Ok
21	AR1254ICC750	PO111605.D	11 Jun 2025 16:25	YP/AJ	Ok,M

Instrument ID: ECD_O

Daily Analysis Runlog For Sequence/QC Batch ID # PO061125

Review By	yogesh	Review On	6/11/2025 3:41:50 PM
Supervise By	mohammad	Supervise On	6/13/2025 1:40:28 AM
SubDirectory	PO061125	HP Acquire Method	HP Processing Method PO061125
STD. NAME	STD REF.#		
Tune/Reschk			
Initial Calibration Stds	PP24330,PP24331,PP24332,PP24333,PP24334,PP24335,PP24336,PP24337,PP24338,PP24339,PP24340,PP24341,PP24342,PP24343,PP24344,PP24345,PP24346,PP24347,PP24348,PP24349,PP24350,PP24351,PP24352,PP24353,PP24354,PP24355,PP24356,PP24357,PP24358,PP24359,PP24360,PP24361,PP24362,PP24363,PP24364,PP24365,PP24366,PP24367,PP24368,PP24369		
CCC	PP24332,PP24347,PP24352,PP24357		
Internal Standard/PEM			
ICV/I.BLK	PP24370,PP24371,PP24372,PP24373,PP24374,PP24375,PP24376,PP24377,PP24378,PP24379,PP24380,PP24381,PP24382,PP24384,PP24385,PP24386,PP2		
Surrogate Standard			
MS/MSD Standard			
LCS Standard			

22	AR1254ICC500	PO111606.D	11 Jun 2025 16:43	YP/AJ	Ok
23	AR1254ICC250	PO111607.D	11 Jun 2025 17:00	YP/AJ	Ok
24	AR1254ICC050	PO111608.D	11 Jun 2025 17:18	YP/AJ	Ok
25	AR1262ICC500	PO111609.D	11 Jun 2025 17:36	YP/AJ	Ok
26	AR1268ICC1000	PO111610.D	11 Jun 2025 17:55	YP/AJ	Ok,M
27	AR1268ICC750	PO111611.D	11 Jun 2025 18:13	YP/AJ	Ok
28	AR1268ICC500	PO111612.D	11 Jun 2025 18:31	YP/AJ	Ok
29	AR1268ICC250	PO111613.D	11 Jun 2025 18:50	YP/AJ	Ok,M
30	AR1268ICC050	PO111614.D	11 Jun 2025 19:07	YP/AJ	Ok
31	PO061125ICV500	PO111615.D	11 Jun 2025 19:25	YP/AJ	Ok
32	AR1242ICV500	PO111616.D	11 Jun 2025 20:02	YP/AJ	Ok
33	AR1248ICV500	PO111617.D	11 Jun 2025 20:39	YP/AJ	Ok
34	AR1254ICV500	PO111618.D	11 Jun 2025 21:16	YP/AJ	Ok
35	AR1268ICV500	PO111619.D	11 Jun 2025 21:52	YP/AJ	Ok,M

M : Manual Integration

Instrument ID: ECD_O

Daily Analysis Runlog For Sequence/QC Batch ID # PO070725

Review By	yogesh	Review On	7/7/2025 11:17:57 AM
Supervise By		Supervise On	
SubDirectory	PO070725	HP Acquire Method	HP Processing Method PO061125
STD. NAME	STD REF.#		
Tune/Reschk			
Initial Calibration Stds	PP24330,PP24331,PP24332,PP24333,PP24334,PP24335,PP24336,PP24337,PP24338,PP24339,PP24340,PP24341,PP24342,PP24343,PP24344,PP24345,PP24346,PP24347,PP24348,PP24349,PP24350,PP24351,PP24352,PP24353,PP24354,PP24355,PP24356,PP24357,PP24358,PP24359,PP24360,PP24361,PP24362,PP24363,PP24364,PP24365,PP24366,PP24367,PP24368,PP24369		
CCC	PP24332,PP24347,PP24352,PP24357		
Internal Standard/PEM			
ICV/I.BLK	PP24370,PP24371,PP24372,PP24373,PP24374,PP24375,PP24376,PP24377,PP24378,PP24379,PP24380,PP24381,PP24382,PP24384,PP24385,PP24386,PP2		
Surrogate Standard			
MS/MSD Standard			
LCS Standard			

Sr#	SampleId	Data File Name	Date-Time	Operator	Status
1	HEXANE	PO112021.D	07 Jul 2025 08:40	YP/AJ	Ok
2	AR1660CCC500	PO112022.D	07 Jul 2025 08:58	YP/AJ	Ok,NS
3	AR1242CCC500	PO112023.D	07 Jul 2025 09:17	YP/AJ	Ok
4	AR1248CCC500	PO112024.D	07 Jul 2025 09:34	YP/AJ	Ok,NS
5	AR1254CCC500	PO112025.D	07 Jul 2025 09:53	YP/AJ	Ok,NS
6	I.BLK	PO112026.D	07 Jul 2025 10:10	YP/AJ	Ok
7	DDT ANALOGUE	PO112027.D	07 Jul 2025 10:27	YP/AJ	Ok
8	PB168704BL	PO112028.D	07 Jul 2025 10:46	YP/AJ	Ok
9	PB168710BL	PO112029.D	07 Jul 2025 11:04	YP/AJ	Ok,NS
10	PB168710BS	PO112030.D	07 Jul 2025 11:22	YP/AJ	Ok
11	Q2466-01	PO112031.D	07 Jul 2025 11:40	YP/AJ	Ok
12	Q2466-02	PO112032.D	07 Jul 2025 11:59	YP/AJ	Ok,NS
13	Q2492-01	PO112033.D	07 Jul 2025 12:17	YP/AJ	Ok,NS
14	Q2492-02	PO112034.D	07 Jul 2025 12:34	YP/AJ	Ok,NS
15	Q2476-01	PO112035.D	07 Jul 2025 12:53	YP/AJ	Not Ok
16	Q2481-10	PO112036.D	07 Jul 2025 13:10	YP/AJ	Ok,NS
17	AR1660CCC500	PO112037.D	07 Jul 2025 14:47	YP/AJ	Ok,NS
18	AR1242CCC500	PO112038.D	07 Jul 2025 15:22	YP/AJ	Ok,NS
19	AR1248CCC500	PO112039.D	07 Jul 2025 15:40	YP/AJ	Ok,NS
20	AR1254CCC500	PO112040.D	07 Jul 2025 15:58	YP/AJ	Ok
21	I.BLK	PO112041.D	07 Jul 2025 16:16	YP/AJ	Ok,NS

Instrument ID: ECD_O

Daily Analysis Runlog For Sequence/QC Batch ID # PO070725

Review By	yogesh	Review On	7/7/2025 11:17:57 AM
Supervise By		Supervise On	
SubDirectory	PO070725	HP Acquire Method	HP Processing Method PO061125
STD. NAME	STD REF.#		
Tune/Reschk			
Initial Calibration Stds	PP24330,PP24331,PP24332,PP24333,PP24334,PP24335,PP24336,PP24337,PP24338,PP24339,PP24340,PP24341,PP24342,PP24343,PP24344,PP24345,PP24346,PP24347,PP24348,PP24349,PP24350,PP24351,PP24352,PP24353,PP24354,PP24355,PP24356,PP24357,PP24358,PP24359,PP24360,PP24361,PP24362,PP24363,PP24364,PP24365,PP24366,PP24367,PP24368,PP24369		
CCC	PP24332,PP24347,PP24352,PP24357		
Internal Standard/PEM			
ICV/I.BLK	PP24370,PP24371,PP24372,PP24373,PP24374,PP24375,PP24376,PP24377,PP24378,PP24379,PP24380,PP24381,PP24382,PP24384,PP24385,PP24386,PP2		
Surrogate Standard			
MS/MSD Standard			
LCS Standard			

22	PB168734BL	PO112042.D	07 Jul 2025 16:34	YP/AJ	Ok
23	PB168734BS	PO112043.D	07 Jul 2025 16:51	YP/AJ	Ok
24	Q2505-01	PO112044.D	07 Jul 2025 17:09	YP/AJ	Not Ok
25	Q2505-02	PO112045.D	07 Jul 2025 17:26	YP/AJ	Not Ok
26	Q2505-03	PO112046.D	07 Jul 2025 17:44	YP/AJ	Ok
27	Q2508-01	PO112047.D	07 Jul 2025 18:03	YP/AJ	Ok,NS
28	Q2508-02	PO112048.D	07 Jul 2025 18:20	YP/AJ	Not Ok
29	Q2508-03	PO112049.D	07 Jul 2025 18:39	YP/AJ	Ok
30	Q2516-01	PO112050.D	07 Jul 2025 18:56	YP/AJ	Ok,NS
31	Q2509-02	PO112051.D	07 Jul 2025 19:14	YP/AJ	Not Ok
32	AR1660CCC500	PO112052.D	07 Jul 2025 20:15	YP/AJ	Ok,NS
33	AR1242CCC500	PO112053.D	07 Jul 2025 20:51	YP/AJ	Ok,NS
34	AR1248CCC500	PO112054.D	07 Jul 2025 21:08	YP/AJ	Ok,NS
35	AR1254CCC500	PO112055.D	07 Jul 2025 21:25	YP/AJ	Ok,NS
36	I.BLK	PO112056.D	07 Jul 2025 21:44	YP/AJ	Ok
37	PB168735BL	PO112057.D	07 Jul 2025 22:01	YP/AJ	Ok,NS
38	PB168735BS	PO112058.D	07 Jul 2025 22:20	YP/AJ	Ok,NS
39	Q2507-01	PO112059.D	07 Jul 2025 22:38	YP/AJ	Ok,NS
40	Q2510-01	PO112060.D	07 Jul 2025 22:57	YP/AJ	Ok,NS
41	Q2513-01	PO112061.D	07 Jul 2025 23:15	YP/AJ	Ok,NS
42	Q2513-03	PO112062.D	07 Jul 2025 23:33	YP/AJ	Ok,NS
43	Q2513-03MS	PO112063.D	07 Jul 2025 23:52	YP/AJ	Ok
44	Q2513-03MSD	PO112064.D	08 Jul 2025 00:10	YP/AJ	Ok

Instrument ID: ECD_O

Daily Analysis Runlog For Sequence/QC Batch ID # PO070725

Review By	yogesh	Review On	7/7/2025 11:17:57 AM
Supervise By		Supervise On	
SubDirectory	PO070725	HP Acquire Method	HP Processing Method PO061125
STD. NAME	STD REF.#		
Tune/Reschk			
Initial Calibration Stds	PP24330,PP24331,PP24332,PP24333,PP24334,PP24335,PP24336,PP24337,PP24338,PP24339,PP24340,PP24341,PP24342,PP24343,PP24344,PP24345,PP24346,PP24347,PP24348,PP24349,PP24350,PP24351,PP24352,PP24353,PP24354,PP24355,PP24356,PP24357,PP24358,PP24359,PP24360,PP24361,PP24362,PP24363,PP24364,PP24365,PP24366,PP24367,PP24368,PP24369		
CCC	PP24332,PP24347,PP24352,PP24357		
Internal Standard/PEM			
ICV/I.BLK	PP24370,PP24371,PP24372,PP24373,PP24374,PP24375,PP24376,PP24377,PP24378,PP24379,PP24380,PP24381,PP24382,PP24384,PP24385,PP24386,PP2		
Surrogate Standard			
MS/MSD Standard			
LCS Standard			

45	Q2514-01	PO112065.D	08 Jul 2025 00:29	YP/AJ	Ok,NS
46	Q2514-02	PO112066.D	08 Jul 2025 00:46	YP/AJ	Ok,NS
47	AR1660CCC500	PO112067.D	08 Jul 2025 03:54	YP/AJ	Ok,NS
48	AR1242CCC500	PO112068.D	08 Jul 2025 04:12	YP/AJ	Ok,NS
49	AR1248CCC500	PO112069.D	08 Jul 2025 04:30	YP/AJ	Ok,NS
50	AR1254CCC500	PO112070.D	08 Jul 2025 04:49	YP/AJ	Ok,NS
51	AR1268CCC500	PO112071.D	08 Jul 2025 05:07	YP/AJ	Not Ok
52	I.BLK	PO112072.D	08 Jul 2025 05:26	YP/AJ	Ok,NS
53	Q2514-03	PO112073.D	08 Jul 2025 05:44	YP/AJ	Ok,NS
54	Q2514-04	PO112074.D	08 Jul 2025 06:02	YP/AJ	Ok,NS
55	Q2514-05	PO112075.D	08 Jul 2025 06:21	YP/AJ	Ok,NS
56	Q2514-06	PO112076.D	08 Jul 2025 06:39	YP/AJ	Ok,NS
57	Q2514-07	PO112077.D	08 Jul 2025 06:58	YP/AJ	Ok,NS
58	Q2514-08	PO112078.D	08 Jul 2025 07:16	YP/AJ	Ok,NS
59	Q2514-09	PO112079.D	08 Jul 2025 07:35	YP/AJ	Ok,NS
60	Q2514-10	PO112080.D	08 Jul 2025 07:53	YP/AJ	Ok,NS
61	Q2515-01	PO112081.D	08 Jul 2025 08:11	YP/AJ	Ok,NS
62	Q2487-15	PO112082.D	08 Jul 2025 08:30	YP/AJ	Not Ok
63	AR1660CCC500	PO112083.D	08 Jul 2025 10:09	YP/AJ	Ok,NS
64	AR1242CCC500	PO112084.D	08 Jul 2025 10:46	YP/AJ	Not Ok
65	I.BLK	PO112085.D	08 Jul 2025 11:26	YP/AJ	Ok,NS

M : Manual Integration

Instrument ID: ECD_O

Daily Analysis Runlog For Sequence/QC Batch ID # PO061125

Review By	yogesh	Review On	6/11/2025 3:41:50 PM
Supervise By	mohammad	Supervise On	6/13/2025 1:40:28 AM
SubDirectory	PO061125	HP Acquire Method	HP Processing Method PO061125

STD. NAME	STD REF.#
Tune/Reschk	
Initial Calibration Stds	PP24330,PP24331,PP24332,PP24333,PP24334,PP24335,PP24336,PP24337,PP24338,PP24339,PP24340,PP24341,PP24342,PP24343,PP24344,PP24345,PP24346,PP24347,PP24348,PP24349,PP24350,PP24351,PP24352,PP24353,PP24354,PP24355,PP24356,PP24357,PP24358,PP24359,PP24360,PP24361,PP24362,PP24363,PP24364,PP24365,PP24366,PP24367,PP24368,PP24369
CCC	PP24332,PP24347,PP24352,PP24357
Internal Standard/PEM	
ICV/I.BLK	PP24370,PP24371,PP24372,PP24373,PP24374,PP24375,PP24376,PP24377,PP24378,PP24379,PP24380,PP24381,PP24382,PP24384,PP24385,PP24386,PP24387
Surrogate Standard	
MS/MSD Standard	
LCS Standard	

Sr#	SampleID	ClientID	Data File Name	Date-Time	Comment	Operator	Status
1	HEXANE	HEXANE	PO111585.D	11 Jun 2025 10:03		YP/AJ	Ok
2	I.BLK	I.BLK	PO111586.D	11 Jun 2025 10:21		YP/AJ	Ok
3	AR1660ICC1000	AR1660ICC1000	PO111587.D	11 Jun 2025 10:40		YP/AJ	Ok
4	AR1660ICC750	AR1660ICC750	PO111588.D	11 Jun 2025 10:58		YP/AJ	Ok
5	AR1660ICC500	AR1660ICC500	PO111589.D	11 Jun 2025 11:17		YP/AJ	Ok
6	AR1660ICC250	AR1660ICC250	PO111590.D	11 Jun 2025 11:35		YP/AJ	Ok
7	AR1660ICC050	AR1660ICC050	PO111591.D	11 Jun 2025 11:53		YP/AJ	Ok,M
8	AR1221ICC500	AR1221ICC500	PO111592.D	11 Jun 2025 12:12		YP/AJ	Ok
9	AR1232ICC500	AR1232ICC500	PO111593.D	11 Jun 2025 12:30		YP/AJ	Ok
10	AR1242ICC1000	AR1242ICC1000	PO111594.D	11 Jun 2025 12:48		YP/AJ	Ok
11	AR1242ICC750	AR1242ICC750	PO111595.D	11 Jun 2025 13:07		YP/AJ	Ok
12	AR1242ICC500	AR1242ICC500	PO111596.D	11 Jun 2025 13:25		YP/AJ	Ok
13	AR1242ICC250	AR1242ICC250	PO111597.D	11 Jun 2025 13:44		YP/AJ	Ok
14	AR1242ICC050	AR1242ICC050	PO111598.D	11 Jun 2025 14:02		YP/AJ	Ok,M
15	AR1248ICC1000	AR1248ICC1000	PO111599.D	11 Jun 2025 14:20		YP/AJ	Ok
16	AR1248ICC750	AR1248ICC750	PO111600.D	11 Jun 2025 14:39		YP/AJ	Ok
17	AR1248ICC500	AR1248ICC500	PO111601.D	11 Jun 2025 15:14		YP/AJ	Ok
18	AR1248ICC250	AR1248ICC250	PO111602.D	11 Jun 2025 15:32		YP/AJ	Ok

Instrument ID: ECD_O

Daily Analysis Runlog For Sequence/QC Batch ID # PO061125

Review By	yogesh	Review On	6/11/2025 3:41:50 PM
Supervise By	mohammad	Supervise On	6/13/2025 1:40:28 AM
SubDirectory	PO061125	HP Acquire Method	HP Processing Method PO061125
STD. NAME	STD REF.#		
Tune/Reschk Initial Calibration Stds	PP24330,PP24331,PP24332,PP24333,PP24334,PP24335,PP24336,PP24337,PP24338,PP24339,PP24340,PP24341,PP24342,PP24343,PP24344,P P24345,PP24346,PP24347,PP24348,PP24349,PP24350,PP24351,PP24352,PP24353,PP24354,PP24355,PP24356,PP24357,PP24358,PP24359,PP 24360,PP24361,PP24362,PP24363,PP24364,PP24365,PP24366,PP24367,PP24368,PP24369		
CCC	PP24332,PP24347,PP24352,PP24357		
Internal Standard/PEM			
ICV/I.BLK	PP24370,PP24371,PP24372,PP24373,PP24374,PP24375,PP24376,PP24377,PP24378,PP24379,PP24380,PP24381,PP24382,PP24384,PP24385,PP24386,PP24387		
Surrogate Standard			
MS/MSD Standard			
LCS Standard			

19	AR1248ICC050	AR1248ICC050	PO111603.D	11 Jun 2025 15:49		YP/AJ	Ok,M
20	AR1254ICC1000	AR1254ICC1000	PO111604.D	11 Jun 2025 16:06		YP/AJ	Ok
21	AR1254ICC750	AR1254ICC750	PO111605.D	11 Jun 2025 16:25		YP/AJ	Ok,M
22	AR1254ICC500	AR1254ICC500	PO111606.D	11 Jun 2025 16:43		YP/AJ	Ok
23	AR1254ICC250	AR1254ICC250	PO111607.D	11 Jun 2025 17:00		YP/AJ	Ok
24	AR1254ICC050	AR1254ICC050	PO111608.D	11 Jun 2025 17:18		YP/AJ	Ok
25	AR1262ICC500	AR1262ICC500	PO111609.D	11 Jun 2025 17:36		YP/AJ	Ok
26	AR1268ICC1000	AR1268ICC1000	PO111610.D	11 Jun 2025 17:55		YP/AJ	Ok,M
27	AR1268ICC750	AR1268ICC750	PO111611.D	11 Jun 2025 18:13		YP/AJ	Ok
28	AR1268ICC500	AR1268ICC500	PO111612.D	11 Jun 2025 18:31		YP/AJ	Ok
29	AR1268ICC250	AR1268ICC250	PO111613.D	11 Jun 2025 18:50		YP/AJ	Ok,M
30	AR1268ICC050	AR1268ICC050	PO111614.D	11 Jun 2025 19:07		YP/AJ	Ok
31	PO061125ICV500	ICVPO061125	PO111615.D	11 Jun 2025 19:25		YP/AJ	Ok
32	AR1242ICV500	ICVPO061125AR1242	PO111616.D	11 Jun 2025 20:02		YP/AJ	Ok
33	AR1248ICV500	ICVPO061125AR1248	PO111617.D	11 Jun 2025 20:39		YP/AJ	Ok
34	AR1254ICV500	ICVPO061125AR1254	PO111618.D	11 Jun 2025 21:16		YP/AJ	Ok
35	AR1268ICV500	ICVPO061125AR1268	PO111619.D	11 Jun 2025 21:52		YP/AJ	Ok,M

M : Manual Integration

Instrument ID: ECD_O

Daily Analysis Runlog For Sequence/QC Batch ID # PO070725

Review By	yogesh	Review On	7/7/2025 11:17:57 AM
Supervise By		Supervise On	
SubDirectory	PO070725	HP Acquire Method	HP Processing Method PO061125

STD. NAME	STD REF.#
Tune/Reschk	
Initial Calibration Stds	PP24330,PP24331,PP24332,PP24333,PP24334,PP24335,PP24336,PP24337,PP24338,PP24339,PP24340,PP24341,PP24342,PP24343,PP24344,PP24345,PP24346,PP24347,PP24348,PP24349,PP24350,PP24351,PP24352,PP24353,PP24354,PP24355,PP24356,PP24357,PP24358,PP24359,PP24360,PP24361,PP24362,PP24363,PP24364,PP24365,PP24366,PP24367,PP24368,PP24369
CCC	PP24332,PP24347,PP24352,PP24357
Internal Standard/PEM	
ICV/I.BLK	PP24370,PP24371,PP24372,PP24373,PP24374,PP24375,PP24376,PP24377,PP24378,PP24379,PP24380,PP24381,PP24382,PP24384,PP24385,PP24386,PP24387
Surrogate Standard	
MS/MSD Standard	
LCS Standard	

Sr#	SampleId	ClientID	Data File Name	Date-Time	Comment	Operator	Status
1	HEXANE	HEXANE	PO112021.D	07 Jul 2025 08:40		YP/AJ	Ok
2	AR1660CCC500	AR1660CCC500	PO112022.D	07 Jul 2025 08:58		YP/AJ	Ok,NS
3	AR1242CCC500	AR1242CCC500	PO112023.D	07 Jul 2025 09:17		YP/AJ	Ok
4	AR1248CCC500	AR1248CCC500	PO112024.D	07 Jul 2025 09:34		YP/AJ	Ok,NS
5	AR1254CCC500	AR1254CCC500	PO112025.D	07 Jul 2025 09:53		YP/AJ	Ok,NS
6	I.BLK	I.BLK	PO112026.D	07 Jul 2025 10:10		YP/AJ	Ok
7	DDT ANALOGUE	DDT ANALOGUE	PO112027.D	07 Jul 2025 10:27		YP/AJ	Ok
8	PB168704BL	PB168704BL	PO112028.D	07 Jul 2025 10:46		YP/AJ	Ok
9	PB168710BL	PB168710BL	PO112029.D	07 Jul 2025 11:04		YP/AJ	Ok,NS
10	PB168710BS	PB168710BS	PO112030.D	07 Jul 2025 11:22		YP/AJ	Ok
11	Q2466-01	61825	PO112031.D	07 Jul 2025 11:40		YP/AJ	Ok
12	Q2466-02	62625	PO112032.D	07 Jul 2025 11:59	AR1016 Hit	YP/AJ	Ok,NS
13	Q2492-01	VNJ-254-1	PO112033.D	07 Jul 2025 12:17	AR1248 Hit	YP/AJ	Ok,NS
14	Q2492-02	VNJ-254-2	PO112034.D	07 Jul 2025 12:34	AR1248 Hit	YP/AJ	Ok,NS
15	Q2476-01	50731	PO112035.D	07 Jul 2025 12:53	AR1242 Hit , need 5x dilution	YP/AJ	Not Ok
16	Q2481-10	CC0627-SFBL	PO112036.D	07 Jul 2025 13:10		YP/AJ	Ok,NS
17	AR1660CCC500	AR1660CCC500	PO112037.D	07 Jul 2025 14:47		YP/AJ	Ok,NS
18	AR1242CCC500	AR1242CCC500	PO112038.D	07 Jul 2025 15:22		YP/AJ	Ok,NS

Instrument ID: ECD_O

Daily Analysis Runlog For Sequence/QC Batch ID # PO070725

Review By	yogesh	Review On	7/7/2025 11:17:57 AM
Supervise By		Supervise On	
SubDirectory	PO070725	HP Acquire Method	HP Processing Method PO061125
STD. NAME	STD REF.#		
Tune/Reschk			
Initial Calibration Stds	PP24330,PP24331,PP24332,PP24333,PP24334,PP24335,PP24336,PP24337,PP24338,PP24339,PP24340,PP24341,PP24342,PP24343,PP24344,P P24345,PP24346,PP24347,PP24348,PP24349,PP24350,PP24351,PP24352,PP24353,PP24354,PP24355,PP24356,PP24357,PP24358,PP24359,PP 24360,PP24361,PP24362,PP24363,PP24364,PP24365,PP24366,PP24367,PP24368,PP24369		
CCC	PP24332,PP24347,PP24352,PP24357		
Internal Standard/PEM			
ICV/I.BLK	PP24370,PP24371,PP24372,PP24373,PP24374,PP24375,PP24376,PP24377,PP24378,PP24379,PP24380,PP24381,PP24382,PP24384,PP24385,PP24386,PP24387		
Surrogate Standard			
MS/MSD Standard			
LCS Standard			

19	AR1248CCC500	AR1248CCC500	PO112039.D	07 Jul 2025 15:40		YP/AJ	Ok,NS
20	AR1254CCC500	AR1254CCC500	PO112040.D	07 Jul 2025 15:58		YP/AJ	Ok
21	I.BLK	I.BLK	PO112041.D	07 Jul 2025 16:16		YP/AJ	Ok,NS
22	PB168734BL	PB168734BL	PO112042.D	07 Jul 2025 16:34		YP/AJ	Ok
23	PB168734BS	PB168734BS	PO112043.D	07 Jul 2025 16:51		YP/AJ	Ok
24	Q2505-01	#62825	PO112044.D	07 Jul 2025 17:09	AR1242 Hit , Need cleanup ,	YP/AJ	Not Ok
25	Q2505-02	#62525	PO112045.D	07 Jul 2025 17:26	DCB high in 2nd column , 1242 HIT , need cleanup	YP/AJ	Not Ok
26	Q2505-03	#2008	PO112046.D	07 Jul 2025 17:44	AR1242	YP/AJ	Ok
27	Q2508-01	AUD-25-0105	PO112047.D	07 Jul 2025 18:03		YP/AJ	Ok,NS
28	Q2508-02	AUD-25-0106	PO112048.D	07 Jul 2025 18:20	need cleanup	YP/AJ	Not Ok
29	Q2508-03	AUD-25-0107	PO112049.D	07 Jul 2025 18:39		YP/AJ	Ok
30	Q2516-01	CHATHAM HEATER IN	PO112050.D	07 Jul 2025 18:56		YP/AJ	Ok,NS
31	Q2509-02	AUD-25-0112	PO112051.D	07 Jul 2025 19:14	need cleanup	YP/AJ	Not Ok
32	AR1660CCC500	AR1660CCC500	PO112052.D	07 Jul 2025 20:15		YP/AJ	Ok,NS
33	AR1242CCC500	AR1242CCC500	PO112053.D	07 Jul 2025 20:51		YP/AJ	Ok,NS
34	AR1248CCC500	AR1248CCC500	PO112054.D	07 Jul 2025 21:08		YP/AJ	Ok,NS
35	AR1254CCC500	AR1254CCC500	PO112055.D	07 Jul 2025 21:25		YP/AJ	Ok,NS
36	I.BLK	I.BLK	PO112056.D	07 Jul 2025 21:44		YP/AJ	Ok
37	PB168735BL	PB168735BL	PO112057.D	07 Jul 2025 22:01		YP/AJ	Ok,NS

Instrument ID: ECD_O

Daily Analysis Runlog For Sequence/QC Batch ID # PO070725

Review By	yogesh	Review On	7/7/2025 11:17:57 AM
Supervise By		Supervise On	
SubDirectory	PO070725	HP Acquire Method	HP Processing Method PO061125
STD. NAME	STD REF.#		
Tune/Reschk			
Initial Calibration Stds	PP24330,PP24331,PP24332,PP24333,PP24334,PP24335,PP24336,PP24337,PP24338,PP24339,PP24340,PP24341,PP24342,PP24343,PP24344,P P24345,PP24346,PP24347,PP24348,PP24349,PP24350,PP24351,PP24352,PP24353,PP24354,PP24355,PP24356,PP24357,PP24358,PP24359,PP 24360,PP24361,PP24362,PP24363,PP24364,PP24365,PP24366,PP24367,PP24368,PP24369		
CCC	PP24332,PP24347,PP24352,PP24357		
Internal Standard/PEM			
ICV/I.BLK	PP24370,PP24371,PP24372,PP24373,PP24374,PP24375,PP24376,PP24377,PP24378,PP24379,PP24380,PP24381,PP24382,PP24384,PP24385,PP24386,PP24387		
Surrogate Standard			
MS/MSD Standard			
LCS Standard			

38	PB168735BS	PB168735BS	PO112058.D	07 Jul 2025 22:20		YP/AJ	Ok,NS
39	Q2507-01	SU-04-7.3-2025	PO112059.D	07 Jul 2025 22:38		YP/AJ	Ok,NS
40	Q2510-01	#63025-A	PO112060.D	07 Jul 2025 22:57	AR1260 Hit	YP/AJ	Ok,NS
41	Q2513-01	HR-2-070325	PO112061.D	07 Jul 2025 23:15		YP/AJ	Ok,NS
42	Q2513-03	HR-3-070325	PO112062.D	07 Jul 2025 23:33		YP/AJ	Ok,NS
43	Q2513-03MS	HR-3-070325MS	PO112063.D	07 Jul 2025 23:52		YP/AJ	Ok
44	Q2513-03MSD	HR-3-070325MSD	PO112064.D	08 Jul 2025 00:10		YP/AJ	Ok
45	Q2514-01	TP-92	PO112065.D	08 Jul 2025 00:29		YP/AJ	Ok,NS
46	Q2514-02	TP-93	PO112066.D	08 Jul 2025 00:46		YP/AJ	Ok,NS
47	AR1660CCC500	AR1660CCC500	PO112067.D	08 Jul 2025 03:54		YP/AJ	Ok,NS
48	AR1242CCC500	AR1242CCC500	PO112068.D	08 Jul 2025 04:12	DCB Low in 1st column	YP/AJ	Ok,NS
49	AR1248CCC500	AR1248CCC500	PO112069.D	08 Jul 2025 04:30		YP/AJ	Ok,NS
50	AR1254CCC500	AR1254CCC500	PO112070.D	08 Jul 2025 04:49		YP/AJ	Ok,NS
51	AR1268CCC500	AR1268CCC500	PO112071.D	08 Jul 2025 05:07	Wrong vial	YP/AJ	Not Ok
52	I.BLK	I.BLK	PO112072.D	08 Jul 2025 05:26		YP/AJ	Ok,NS
53	Q2514-03	TP-94	PO112073.D	08 Jul 2025 05:44		YP/AJ	Ok,NS
54	Q2514-04	TP-96	PO112074.D	08 Jul 2025 06:02		YP/AJ	Ok,NS
55	Q2514-05	TP-97	PO112075.D	08 Jul 2025 06:21		YP/AJ	Ok,NS
56	Q2514-06	TP-103	PO112076.D	08 Jul 2025 06:39		YP/AJ	Ok,NS

Instrument ID: ECD_O

Daily Analysis Runlog For Sequence/QC Batch ID # PO070725

Review By	yogesh	Review On	7/7/2025 11:17:57 AM
Supervise By		Supervise On	
SubDirectory	PO070725	HP Acquire Method	HP Processing Method PO061125
STD. NAME	STD REF.#		
Tune/Reschk			
Initial Calibration Stds	PP24330,PP24331,PP24332,PP24333,PP24334,PP24335,PP24336,PP24337,PP24338,PP24339,PP24340,PP24341,PP24342,PP24343,PP24344,P P24345,PP24346,PP24347,PP24348,PP24349,PP24350,PP24351,PP24352,PP24353,PP24354,PP24355,PP24356,PP24357,PP24358,PP24359,PP 24360,PP24361,PP24362,PP24363,PP24364,PP24365,PP24366,PP24367,PP24368,PP24369		
CCC	PP24332,PP24347,PP24352,PP24357		
Internal Standard/PEM			
ICV/I.BLK	PP24370,PP24371,PP24372,PP24373,PP24374,PP24375,PP24376,PP24377,PP24378,PP24379,PP24380,PP24381,PP24382,PP24384,PP24385,PP24386,PP24387		
Surrogate Standard			
MS/MSD Standard			
LCS Standard			

57	Q2514-07	TP-36	PO112077.D	08 Jul 2025 06:58		YP/AJ	Ok,NS
58	Q2514-08	TP-78	PO112078.D	08 Jul 2025 07:16		YP/AJ	Ok,NS
59	Q2514-09	TP-81	PO112079.D	08 Jul 2025 07:35		YP/AJ	Ok,NS
60	Q2514-10	TP-90	PO112080.D	08 Jul 2025 07:53		YP/AJ	Ok,NS
61	Q2515-01	wc-1	PO112081.D	08 Jul 2025 08:11		YP/AJ	Ok,NS
62	Q2487-15	G1(0-6)	PO112082.D	08 Jul 2025 08:30	AR1268+1260+1254 Hit , AR1268 CCC missing	YP/AJ	Not Ok
63	AR1660CCC500	AR1660CCC500	PO112083.D	08 Jul 2025 10:09		YP/AJ	Ok,NS
64	AR1242CCC500	AR1242CCC500	PO112084.D	08 Jul 2025 10:46		YP/AJ	Not Ok
65	I.BLK	I.BLK	PO112085.D	08 Jul 2025 11:26		YP/AJ	Ok,NS

M : Manual Integration



PERCENT SOLID

Supervisor: Iwona
Analyst: jignesh
Date: 7/7/2025

OVENTEMP IN Celsius(°C): 107
Time IN: 17:30
In Date: 07/03/2025
Weight Check 1.0g: 1.00
Weight Check 10g: 10.00
OvenID: M OVEN#1

OVENTEMP OUT Celsius(°C): 104
Time OUT: 08:37
Out Date: 07/04/2025
Weight Check 1.0g: 1.00
Weight Check 10g: 10.00
BalanceID: M SC-4
Thermometer ID: % SOLID-OVEN

QC:LB136368

Lab ID	Client SampleID	Dish #	Dish Wt(g) (A)	Sample Wt(g)	Dish + Sample Wt(g) (B)	Dish+Dry Sample Wt(g) (C)	% Solid	Comments
Q2487-01	G4 (1.5)	1	1.15	10.21	11.36	10.14	88.1	
Q2487-02	G4 (10)	2	1.19	10.38	11.57	8.76	72.9	
Q2487-03	G3 (9)	3	1.15	10.66	11.81	7.38	58.4	
Q2487-04	G3 (3)	4	1.15	10.80	11.95	10.8	89.4	
Q2487-05	G2 (2.5)	5	1.17	10.00	11.17	10.48	93.1	
Q2487-06	G2 (9)	6	1.16	10.26	11.42	8.00	66.7	
Q2487-07	G1 (4.5)	7	1.19	10.11	11.3	7.8	65.4	
Q2487-08	G1 (10)	8	1.16	9.96	11.12	6.65	55.1	
Q2487-09	G4 (0-6)	9	1.18	10.26	11.44	10.12	87.1	
Q2487-10	G4 (6-12)	10	1.16	10.14	11.3	8.52	72.6	
Q2487-11	G3 (0-6)	11	1.18	10.42	11.6	10.68	91.2	
Q2487-12	G3 (6-12)	12	1.18	10.41	11.59	8.49	70.2	
Q2487-13	G2 (0-6)	13	1.17	9.95	11.12	10.11	89.8	
Q2487-14	G2 (6-12)	14	1.16	10.68	11.84	8.39	67.7	
Q2487-15	G1 (0-6)	15	1.17	10.55	11.72	9.68	80.7	
Q2487-16	G1 (6-12)	16	1.13	10.24	11.37	10.68	93.3	
Q2501-05	SVOC-GPC-BLANK	17	1.00	1.00	2.00	2.00	100.0	
Q2501-06	PEST-GPC-BLANK	18	1.00	1.00	2.00	2.00	100.0	
Q2501-07	PEST-GPC-BLANK-SPIKE	19	1.00	1.00	2.00	2.00	100.0	
Q2501-08	PCB-GPC-BLANK	20	1.00	1.00	2.00	2.00	100.0	
Q2501-09	PCB-GPC-BLANK-SPIKE	21	1.00	1.00	2.00	2.00	100.0	
Q2501-10	SVOC-GPC-BLANK	22	1.00	1.00	2.00	2.00	100.0	
Q2501-11	PEST-GPC-BLANK	23	1.00	1.00	2.00	2.00	100.0	
Q2501-12	PEST-GPC-BLANK-SPIKE	24	1.00	1.00	2.00	2.00	100.0	
Q2501-13	PCB-GPC2-BLANK	25	1.00	1.00	2.00	2.00	100.0	
Q2501-14	PCB-GPC2-BLANK-SPIKE	26	1.00	1.00	2.00	2.00	100.0	
Q2503-03	GCAP2	27	1.18	10.65	11.83	8.76	71.2	
Q2503-04	GCAP3	28	1.18	10.64	11.82	8.16	65.6	

PERCENT SOLID

Supervisor: Iwona
Analyst: jignesh
Date: 7/7/2025

OVENTEMP IN Celsius(°C): 107
Time IN: 17:30
In Date: 07/03/2025
Weight Check 1.0g: 1.00
Weight Check 10g: 10.00
OvenID: M OVEN#1

OVENTEMP OUT Celsius(°C): 104
Time OUT: 08:37
Out Date: 07/04/2025
Weight Check 1.0g: 1.00
Weight Check 10g: 10.00
BalanceID: M SC-4
Thermometer ID: % SOLID-OVEN

QC:LB136368

Lab ID	Client SampleID	Dish #	Dish Wt(g) (A)	Sample Wt(g)	Dish + Sample Wt(g) (B)	Dish+Dry Sample Wt(g) (C)	% Solid	Comments
Q2503-05	GCAP2A	29	1.18	10.30	11.48	9.11	77.0	
Q2504-01	WASTE	30	1.14	10.73	11.87	10.16	84.1	
Q2504-02	VOC	31	1.18	10.81	11.99	10.45	85.8	
Q2504-03	1	32	1.14	10.84	11.98	10.36	85.1	
Q2504-04	2	33	1.19	10.41	11.6	9.91	83.8	
Q2504-05	3	34	1.13	10.75	11.88	10.22	84.6	
Q2504-06	4	35	1.14	10.52	11.66	10.03	84.5	
Q2504-07	5	36	1.12	10.87	11.99	10.16	83.2	
Q2505-01	#62825	37	1.00	1.00	2.00	2.00	100.0	WIPE SAMPLE
Q2505-02	#62525	38	1.00	1.00	2.00	2.00	100.0	WIPE SAMPLE
Q2505-03	#2008	39	1.00	1.00	2.00	2.00	100.0	WIPE SAMPLE
Q2507-01	SU-04-7.3-2025	40	1.17	10.26	11.43	10.1	87.0	
Q2507-02	SU-04-7.3-2025-EPH	41	1.18	10.30	11.48	9.99	85.5	
Q2507-03	SU-04-7.3-2025-VOC	42	1.13	10.45	11.58	9.95	84.4	
Q2508-01	AUD-25-0105	43	1.00	1.00	2.00	2.00	100.0	WIPE SAMPLE
Q2508-02	AUD-25-0106	44	1.00	1.00	2.00	2.00	100.0	WIPE SAMPLE
Q2508-03	AUD-25-0107	45	1.00	1.00	2.00	2.00	100.0	WIPE SAMPLE
Q2509-02	AUD-25-0112	46	1.00	1.00	2.00	2.00	100.0	oilly-debris
Q2510-01	#63025-A	47	1.15	11.63	12.78	10.79	82.9	
Q2510-02	#63025-A-VOC	48	1.18	10.19	11.37	9.86	85.2	
Q2513-01	HR-2-070325	49	1.18	9.89	11.07	10.51	94.3	
Q2513-02	HR-2-070325-E2	50	1.18	10.25	11.43	10.68	92.7	
Q2513-03	HR-3-070325	51	1.18	10.27	11.45	10.88	94.4	
Q2513-04	HR-3-070325-E2	52	1.14	10.67	11.81	11.23	94.6	
Q2514-01	TP-92	53	1.14	10.78	11.92	10.56	87.4	
Q2514-02	TP-93	54	1.19	10.71	11.9	10.58	87.7	
Q2514-03	TP-94	55	1.13	10.86	11.99	10.7	88.1	
Q2514-04	TP-96	56	1.14	11.10	12.24	10.66	85.8	



PERCENT SOLID

Supervisor: Iwona
Analyst: jignesh
Date: 7/7/2025

OVENTEMP IN Celsius(°C): 107
Time IN: 17:30
In Date: 07/03/2025
Weight Check 1.0g: 1.00
Weight Check 10g: 10.00
OvenID: M OVEN#1

OVENTEMP OUT Celsius(°C): 104
Time OUT: 08:37
Out Date: 07/04/2025
Weight Check 1.0g: 1.00
Weight Check 10g: 10.00
BalanceID: M SC-4
Thermometer ID: % SOLID-OVEN

QC:LB136368

Lab ID	Client SampleID	Dish #	Dish Wt (g) (A)	Sample Wt (g)	Dish + Sample Wt (g) (B)	Dish+Dry Sample Wt (g) (C)	% Solid	Comments
Q2514-05	TP-97	57	1.16	10.02	11.18	9.68	85.0	
Q2514-06	TP-103	58	1.18	10.22	11.4	10.02	86.5	
Q2514-07	TP-36	59	1.15	10.78	11.93	10.88	90.3	
Q2514-08	TP-78	60	1.13	9.99	11.12	9.75	86.3	
Q2514-09	TP-81	61	1.16	10.50	11.66	10.22	86.3	
Q2514-10	TP-90	62	1.18	10.43	11.61	10.73	91.6	
Q2515-01	wc-1	63	1.15	10.31	11.46	10.17	87.5	

$$\% \text{ Solid} = \frac{(C-A) * 100}{(B-A)}$$

WORKLIST(Hardcopy Internal Chain)

WorkList Name : %1-070325

WorkList ID : 190533

Department : Wet-Chemistry

Date : 07-03-2025 08:34:29

196368

Sample	Customer Sample	Matrix	Test	Preservative	Customer	Raw Sample Storage Location	Collect Date	Method
Q2487-01	G4(1.5)	Solid	Percent Solids	Cool 4 deg C	WALS01	A22	07/01/2025	Chemtech -SO
Q2487-02	G4(10)	Solid	Percent Solids	Cool 4 deg C	WALS01	A22	07/01/2025	Chemtech -SO
Q2487-03	G3(9)	Solid	Percent Solids	Cool 4 deg C	WALS01	A22	07/01/2025	Chemtech -SO
Q2487-04	G3(3)	Solid	Percent Solids	Cool 4 deg C	WALS01	A22	07/01/2025	Chemtech -SO
Q2487-05	G2(2.5)	Solid	Percent Solids	Cool 4 deg C	WALS01	A22	07/01/2025	Chemtech -SO
Q2487-06	G2(9)	Solid	Percent Solids	Cool 4 deg C	WALS01	A22	07/01/2025	Chemtech -SO
Q2487-07	G1(4.5)	Solid	Percent Solids	Cool 4 deg C	WALS01	A22	07/01/2025	Chemtech -SO
Q2487-08	G1(10)	Solid	Percent Solids	Cool 4 deg C	WALS01	A22	07/01/2025	Chemtech -SO
Q2487-09	G4(0-6)	Solid	Percent Solids	Cool 4 deg C	WALS01	A22	07/01/2025	Chemtech -SO
Q2487-10	G4(6-12)	Solid	Percent Solids	Cool 4 deg C	WALS01	A22	07/01/2025	Chemtech -SO
Q2487-11	G3(0-6)	Solid	Percent Solids	Cool 4 deg C	WALS01	A22	07/01/2025	Chemtech -SO
Q2487-12	G3(6-12)	Solid	Percent Solids	Cool 4 deg C	WALS01	A22	07/01/2025	Chemtech -SO
Q2487-13	G2(0-6)	Solid	Percent Solids	Cool 4 deg C	WALS01	A22	07/01/2025	Chemtech -SO
Q2487-14	G2(6-12)	Solid	Percent Solids	Cool 4 deg C	WALS01	A22	07/01/2025	Chemtech -SO
Q2487-15	G1(0-6)	Solid	Percent Solids	Cool 4 deg C	WALS01	A22	07/01/2025	Chemtech -SO
Q2487-16	G1(6-12)	Solid	Percent Solids	Cool 4 deg C	WALS01	A22	07/01/2025	Chemtech -SO
Q2501-05	SVOC-GPC-BLANK	Solid	Percent Solids	Cool 4 deg C	CHEM02	D31	06/27/2025	Chemtech -SO
Q2501-06	PEST-GPC-BLANK	Solid	Percent Solids	Cool 4 deg C	CHEM02	D31	06/27/2025	Chemtech -SO
Q2501-07	PEST-GPC-BLANK-SPIKE	Solid	Percent Solids	Cool 4 deg C	CHEM02	D31	06/27/2025	Chemtech -SO
Q2501-08	PCB-GPC-BLANK	Solid	Percent Solids	Cool 4 deg C	CHEM02	D31	06/27/2025	Chemtech -SO
Q2501-09	PCB-GPC-BLANK-SPIKE	Solid	Percent Solids	Cool 4 deg C	CHEM02	D31	06/27/2025	Chemtech -SO

Date/Time 07/03/25 15:15

Raw Sample Received by: JDC

Raw Sample Relinquished by: JDC

Date/Time 07/03/25

Raw Sample Received by: JDC

Raw Sample Relinquished by: JDC

WORKLIST(Hardcopy Internal Chain)

136368

WorkList Name : %1-070325

WorkList ID : 190533

Department : Wet-Chemistry

Date : 07-03-2025 08:34:29

Sample	Customer Sample	Matrix	Test	Preservative	Customer	Raw Sample Storage Location	Collect Date	Method
Q2501-10	SVOC-GPC-BLANK	Solid	Percent Solids	Cool 4 deg C	CHEM02	D31	06/27/2025	Chemtech -SO
Q2501-11	PEST-GPC-BLANK	Solid	Percent Solids	Cool 4 deg C	CHEM02	D31	06/27/2025	Chemtech -SO
Q2501-12	PEST-GPC-BLANK-SPIKE	Solid	Percent Solids	Cool 4 deg C	CHEM02	D31	06/27/2025	Chemtech -SO
Q2501-13	PCB-GPC2-BLANK	Solid	Percent Solids	Cool 4 deg C	CHEM02	D31	06/27/2025	Chemtech -SO
Q2501-14	PCB-GPC2-BLANK-SPIKE	Solid	Percent Solids	Cool 4 deg C	CHEM02	D31	06/27/2025	Chemtech -SO
Q2503-03	GCAP2	Solid	Percent Solids	Cool 4 deg C	GENV01	O11	06/27/2025	Chemtech -SO
Q2503-04	GCAP3	Solid	Percent Solids	Cool 4 deg C	GENV01	O11	07/02/2025	Chemtech -SO
Q2503-05	GCAP2A	Solid	Percent Solids	Cool 4 deg C	GENV01	O11	07/02/2025	Chemtech -SO
Q2504-01	WASTE	Solid	Percent Solids	Cool 4 deg C	SCIA01	O12	07/02/2025	Chemtech -SO
Q2504-02	VOC	Solid	Percent Solids	Cool 4 deg C	SCIA01	O12	07/02/2025	Chemtech -SO
Q2504-03	1	Solid	Percent Solids	Cool 4 deg C	SCIA01	O12	07/02/2025	Chemtech -SO
Q2504-04	2	Solid	Percent Solids	Cool 4 deg C	SCIA01	O12	07/02/2025	Chemtech -SO
Q2504-05	3	Solid	Percent Solids	Cool 4 deg C	SCIA01	O12	07/02/2025	Chemtech -SO
Q2504-06	4	Solid	Percent Solids	Cool 4 deg C	SCIA01	O12	07/02/2025	Chemtech -SO
Q2504-07	5	Solid	Percent Solids	Cool 4 deg C	SCIA01	O12	07/02/2025	Chemtech -SO
Q2505-01	#62825	Solid	Percent Solids	Cool 4 deg C	PSEG03	O13	07/03/2025	Chemtech -SO
Q2505-02	#62525	Solid	Percent Solids	Cool 4 deg C	PSEG03	O13	07/03/2025	Chemtech -SO
Q2505-03	#2008	Solid	Percent Solids	Cool 4 deg C	PSEG03	O13	07/03/2025	Chemtech -SO
Q2507-01	SU-04-7.3-2025	Solid	Percent Solids	Cool 4 deg C	PSEG03	O12	07/04/2025	Chemtech -SO
Q2507-02	SU-04-7.3-2025-EPH	Solid	Percent Solids	Cool 4 deg C	PSEG03	O12	07/04/2025	Chemtech -SO
Q2507-03	SU-04-7.3-2025-VOC	Solid	Percent Solids	Cool 4 deg C	PSEG03	O12	07/04/2025	Chemtech -SO

Date/Time 07/03/25 15:15
Raw Sample Received by: [Signature]
Raw Sample Relinquished by: [Signature]

Date/Time 07/03/25 17:15
Raw Sample Received by: [Signature]
Raw Sample Relinquished by: [Signature]

WORKLIST(Hardcopy Internal Chain)

136368

WorkList Name : %1-070325

WorkList ID : 190533

Department : Wet-Chemistry

Date : 07-03-2025 08:34:29

Sample	Customer Sample	Matrix	Test	Preservative	Customer	Raw Sample Storage Location	Collect Date	Method
Q2508-01	AUD-25-0105	Solid	Percent Solids	Cool 4 deg C	PSEG03	O11	07/03/2025	Chemtech -SO
Q2508-02	AUD-25-0106	Solid	Percent Solids	Cool 4 deg C	PSEG03	O11	07/03/2025	Chemtech -SO
Q2508-03	AUD-25-0107	Solid	Percent Solids	Cool 4 deg C	PSEG03	O11	07/03/2025	Chemtech -SO
Q2509-02	AUD-25-0112	Solid	Percent Solids	Cool 4 deg C	PSEG03	O22	07/03/2025	Chemtech -SO
Q2510-01	#63025-A	Solid	Percent Solids	Cool 4 deg C	PSEG05	O22	07/03/2025	Chemtech -SO
Q2510-02	#63025-A-VOC	Solid	Percent Solids	Cool 4 deg C	PSEG05	O22	07/03/2025	Chemtech -SO
Q2513-01	HR-2-070325	Solid	Percent Solids	Cool 4 deg C	PSEG05	O21	07/03/2025	Chemtech -SO
Q2513-02	HR-2-070325-E2	Solid	Percent Solids	Cool 4 deg C	PSEG05	O21	07/03/2025	Chemtech -SO
Q2513-03	HR-3-070325	Solid	Percent Solids	Cool 4 deg C	PSEG05	O21	07/03/2025	Chemtech -SO
Q2513-04	HR-3-070325-E2	Solid	Percent Solids	Cool 4 deg C	PSEG05	O21	07/03/2025	Chemtech -SO
Q2514-01	TP-92	Solid	Percent Solids	Cool 4 deg C	CAMP02	O21	07/02/2025	Chemtech -SO
Q2514-02	TP-93	Solid	Percent Solids	Cool 4 deg C	CAMP02	O21	07/02/2025	Chemtech -SO
Q2514-03	TP-94	Solid	Percent Solids	Cool 4 deg C	CAMP02	O21	07/02/2025	Chemtech -SO
Q2514-04	TP-96	Solid	Percent Solids	Cool 4 deg C	CAMP02	O21	07/02/2025	Chemtech -SO
Q2514-05	TP-97	Solid	Percent Solids	Cool 4 deg C	CAMP02	O21	07/02/2025	Chemtech -SO
Q2514-06	TP-103	Solid	Percent Solids	Cool 4 deg C	CAMP02	O21	07/02/2025	Chemtech -SO
Q2514-07	TP-36	Solid	Percent Solids	Cool 4 deg C	CAMP02	O21	07/02/2025	Chemtech -SO
Q2514-08	TP-78	Solid	Percent Solids	Cool 4 deg C	CAMP02	O21	07/03/2025	Chemtech -SO
Q2514-09	TP-81	Solid	Percent Solids	Cool 4 deg C	CAMP02	O21	07/03/2025	Chemtech -SO
Q2514-10	TP-90	Solid	Percent Solids	Cool 4 deg C	CAMP02	O21	07/03/2025	Chemtech -SO
Q2515-01	wc-1	Solid	Percent Solids	Cool 4 deg C	ENVO01	O23	07/03/2025	Chemtech -SO

Date/Time 07/03/25 1515
 Raw Sample Received by: [Signature]
 Raw Sample Relinquished by: [Signature]

Date/Time 07/03/25 17135
 Raw Sample Received by: [Signature]
 Raw Sample Relinquished by: [Signature]

SOP ID:	M3541-ASE Extraction-14		
Clean Up SOP #:	Acid Cleanup	Extraction Start Date :	07/07/2025
Matrix :	Solid	Extraction Start Time :	08:30
Weigh By:	EH	Extraction By:	RJ
Balance check:	RJ	Extraction End Date :	07/07/2025
Balance ID:	EX-SC-2	Filter By:	RJ
pH Strip Lot#:	N/A	Extraction End Time :	11:45
		Concentration By:	EH
		Supervisor By :	RUPESH
Extraction Method:	☐ Separatory Funnel ☐ Continous Liquid/Liquid ☐ Sonication ☐ Waste Dilution ☒ Soxhlet		

Standard Name	MLS USED	Concentration ug/mL	STD REF. # FROM LOG
Spike Sol 1	1.0ML	5000 PPB	PP24650
Surrogate	1.0ML	200 PPB	PP24663
N/A	N/A	N/A	N/A
N/A	N/A	N/A	N/A
N/A	N/A	N/A	N/A

Chemical Used	ML/SAMPLE USED	Lot Number
Hexane/Acetone/1:1	N/A	EP2613
Baked Na2SO4	N/A	EP2624
Hexane	N/A	E3947
H2SO4 1:1	N/A	EP2610
N/A	N/A	N/A
N/A	N/A	N/A
N/A	N/A	N/A
N/A	N/A	N/A
N/A	N/A	N/A
N/A	N/A	N/A
N/A	N/A	N/A
N/A	N/A	N/A
N/A	N/A	N/A
N/A	N/A	N/A
N/A	N/A	N/A

Extraction Conformance/Non-Conformance Comments:

40ML Vial Lot # 03-40BTS723, Q2509-02 used Limited volume as sample is Oily Debris.


KD Bath ID:	N/A	Envap ID:	NEVAP-02
KD Bath Temperature:	N/A	Envap Temperature:	40 °C

Date / Time	Prepped Sample Relinquished By/Location	Received By/Location
7/7/25 11:50	RS (Ext Lab)	Y-P-KH (POB)
	Preparation Group	Analysis Group

Analytical Method: M3541-ASE Extraction-14

Concentration Date: 07/07/2025

Sample ID	Client Sample ID	Test	g/mL	PH	Surr/Spike By:		Final Vol. (mL)	JarID	Comments	Prep Pos
					AddedBy	VerifiedBy				
PB168735BL	ABLK735	PCB	30.01	N/A	ritesh	Evelyn	10			U3-1
PB168735BS	ALCS735	PCB	30.02	N/A	ritesh	Evelyn	10			2
Q2507-01	SU-04-7.3-2025	PCB	30.05	N/A	ritesh	Evelyn	10	B		3
Q2509-02	AUD-25-0112	PCB	5.06	N/A	ritesh	Evelyn	10	B	Oily Debris	4
Q2510-01	#63025-A	PCB	30.08	N/A	ritesh	Evelyn	10	E		5
Q2513-01	HR-2-070325	PCB	30.05	N/A	ritesh	Evelyn	10	E		6
Q2513-03	HR-3-070325	PCB	30.04	N/A	ritesh	Evelyn	10	E		U6-1
Q2513-03MS	HR-3-070325MS	PCB	30.06	N/A	ritesh	Evelyn	10	E		2
Q2513-03MS D	HR-3-070325MSD	PCB	30.02	N/A	ritesh	Evelyn	10	E		3
Q2514-01	TP-92	PCB	30.07	N/A	ritesh	Evelyn	10	E		4
Q2514-02	TP-93	PCB	30.01	N/A	ritesh	Evelyn	10	E		5
Q2514-03	TP-94	PCB	30.08	N/A	ritesh	Evelyn	10	E		6
Q2514-04	TP-96	PCB	30.03	N/A	ritesh	Evelyn	10	E		U1-1
Q2514-05	TP-97	PCB	30.05	N/A	ritesh	Evelyn	10	E		2
Q2514-06	TP-103	PCB	30.04	N/A	ritesh	Evelyn	10	E		3
Q2514-07	TP-36	PCB	30.06	N/A	ritesh	Evelyn	10	E		4
Q2514-08	TP-78	PCB	30.09	N/A	ritesh	Evelyn	10	E		5
Q2514-09	TP-81	PCB	30.02	N/A	ritesh	Evelyn	10	E		6
Q2514-10	TP-90	PCB	30.04	N/A	ritesh	Evelyn	10	E		U1-1
Q2515-01	WC-1	PCB	30.07	N/A	ritesh	Evelyn	10	E		2



168-285
(8:30)

WORKLIST(Hardcopy Internal Chain)

WorkList Name : Q2507 WorkList ID : 190565 Department : Extraction Date : 07-07-2025 08:25:22

Sample	Customer Sample	Matrix	Test	Preservative	Customer	Raw Sample Storage Location	Collect Date	Method
Q2507-01	SU-04-7.3-2025	Solid	PCB	Cool 4 deg C	PSEG03	O12	07/04/2025	8082A
Q2509-02	AUD-25-0112	Solid	PCB	Cool 4 deg C	PSEG03	O22	07/03/2025	8082A
Q2510-01	#63025-A	Solid	PCB	Cool 4 deg C	PSEG05	O22	07/03/2025	8082A
Q2513-01	HR-2-070325	Solid	PCB	Cool 4 deg C	PSEG05	O21	07/03/2025	8082A
Q2513-03	HR-3-070325	Solid	PCB	Cool 4 deg C	PSEG05	O21	07/03/2025	8082A
Q2514-01	TP-92	Solid	PCB	Cool 4 deg C	CAMP02	O21	07/02/2025	8082A
Q2514-02	TP-93	Solid	PCB	Cool 4 deg C	CAMP02	O21	07/02/2025	8082A
Q2514-03	TP-94	Solid	PCB	Cool 4 deg C	CAMP02	O21	07/02/2025	8082A
Q2514-04	TP-96	Solid	PCB	Cool 4 deg C	CAMP02	O21	07/02/2025	8082A
Q2514-05	TP-97	Solid	PCB	Cool 4 deg C	CAMP02	O21	07/02/2025	8082A
Q2514-06	TP-103	Solid	PCB	Cool 4 deg C	CAMP02	O21	07/02/2025	8082A
Q2514-07	TP-36	Solid	PCB	Cool 4 deg C	CAMP02	O21	07/03/2025	8082A
Q2514-08	TP-78	Solid	PCB	Cool 4 deg C	CAMP02	O21	07/03/2025	8082A
Q2514-09	TP-81	Solid	PCB	Cool 4 deg C	CAMP02	O21	07/03/2025	8082A
Q2514-10	TP-90	Solid	PCB	Cool 4 deg C	CAMP02	O21	07/03/2025	8082A
Q2515-01	wc-1	Solid	PCB	Cool 4 deg C	ENVO01	O23	07/03/2025	8082A

Date/Time 07/07/25 8:25
Raw Sample Received by: RS (FHA-VAB)
Raw Sample Relinquished by: ITS (M)

Date/Time 07/07/25 8:50
Raw Sample Received by: ITS (M)
Raw Sample Relinquished by: RS (FHA-VAB)



SHIPPING DOCUMENTS

CLIENT INFORMATION

REPORT TO BE SENT TO:

COMPANY: **CDM SMITH**
ADDRESS: **110 FIELDCREST AVE #8 6TH FLOOR**
CITY: **EDISON** STATE: **NJ** ZIP: **08837**
ATTENTION: **MARCIE ENCINAS**
PHONE: **7325904679** FAX:

CLIENT PROJECT INFORMATION

PROJECT NAME: **SOUTH RIVER WM REPLACEMENT**
PROJECT NO.: **302781** LOCATION: **SOUTH RIVER NJ**
PROJECT MANAGER: **FELIPE CONTRERAS**
e-mail: **ENCINASMA@CDMSMITH.COM**
PHONE: **7325904679** FAX:

CLIENT BILLING INFORMATION

BILL TO: **CDM SMITH** PO#: **08837**
ADDRESS: **110 FIELDCREST AVE #8 6TH FLOOR**
CITY: **EDISON** STATE: **NJ** ZIP: **08837**
ATTENTION: **MARCIE ENCINAS** PHONE: **7325904679**

ANALYSIS

DATA TURNAROUND INFORMATION

FAX (RUSH) _____ DAYS*
HARDCOPY (DATA PACKAGE): _____ DAYS*
EDD: _____ DAYS*
*TO BE APPROVED BY CHEMTECH
STANDARD HARDCOPY TURNAROUND TIME IS 10 BUSINESS

DATA DELIVERABLE INFORMATION

☐ Level 1 (Results Only) ☐ Level 4 (QC + Full Raw Data)
☒ Level 2 (Results + QC) ☐ NJ Reduced ☐ US EPA CLP
☐ Level 3 (Results + QC) ☐ NYS ASP A ☐ NYS ASP B
+ Raw Data ☐ Other _____
☐ EDD FORMAT

1. TEL VOL
2. TEL SVOL
3. PCBs
4. TAL METALS
5. PESTICIDES
6. HERBICIDES
7. DRUGS
8. GRS
9.

PRESERVATIVES

COMMENTS

← Specify Preservatives
A-HCl D-NaOH
B-HNO3 E-ICE
C-H2SO4 F-OTHER

ALLIANCE SAMPLE ID	PROJECT SAMPLE IDENTIFICATION	SAMPLE MATRIX	SAMPLE TYPE		SAMPLE COLLECTION		# OF BOTTLES										
			COMP	GRAB	DATE	TIME		1	2	3	4	5	6	7	8	9	
1.	TP-92	5		X	7/2/25	0800	5	X	X	X	X	X	X	X	X		E
2.	TP-93	5		X	7/2/25	0830	6	X	X	X	X	X	X	X	X		E
3.	TP-94	5		X	7/2/25	1035	6	X	X	X	X	X	X	X	X		E
4.	TP-96	5		X	7/2/25	1125	6	X	X	X	X	X	X	X	X		E
5.	TP-97	5		X	7/2/25	1215	6	X	X	X	X	X	X	X	X		E
6.	TP-103	5		X	7/2/25	1345	6	X	X	X	X	X	X	X	X		E
7.	TP-3C	5		X	7/3/25	0820	6	X	X	X	X	X	X	X	X		E
8.	TP-78	5		X	7/3/25	0855	6	X	X	X	X	X	X	X	X		E
9.	TP-81	5		X	7/3/25	1000	6	X	X	X	X	X	X	X	X		E
10.	TP-90	5		X	7/3/25	1250	6	X	X	X	X	X	X	X	X		E

SAMPLE CUSTODY MUST BE DOCUMENTED BELOW EACH TIME SAMPLES CHANGE POSSESSION INCLUDING COURIER DELIVERY

RELINQUISHED BY SAMPLER: 1. <i>[Signature]</i>	DATE/TIME: 7/3/25 1315	RECEIVED BY: 1. <i>[Signature]</i>	1315 7-3-25	Conditions of bottles or coolers at receipt: ☐ COMPLIANT ☐ NON COMPLIANT ☐ COOLER TEMP 3.9 °C
RELINQUISHED BY SAMPLER: 2. <i>[Signature]</i>	DATE/TIME:	RECEIVED BY: 2. <i>[Signature]</i>		Comments:
RELINQUISHED BY SAMPLER: 3. <i>[Signature]</i>	DATE/TIME: 1412 7-3-25	RECEIVED BY: 3. <i>[Signature]</i>		Page ____ of

CLIENT: ☐ Hand Delivered ☐ Other

Shipment Complete

☐ YES ☐ NO

Laboratory Certification

Certified By	License No.
CAS EPA CLP Contract	68HERH20D0011
Connecticut	PH-0830
DOD ELAP (ANAB)	L2219
Maine	2024021
Maryland	296
New Hampshire	255424 Rev 1
New Jersey	20012
New York	11376
Pennsylvania	68-00548
Soil Permit	525-24-234-08441
Texas	T104704488

LOGIN REPORT/SAMPLE TRANSFER

Order ID : Q2514	CAMP02	Order Date : 7/3/2025 1:29:00 PM	Project Mgr :
Client Name : CDM Smith		Project Name : South River WM Replacem	Report Type : Level 2
Client Contact : Marcie Ann Encinas		Receive DateTime : 7/3/2025 3:00:00 PM	EDD Type : EXCEL NOCLEANUP
Invoice Name : CDM Smith		Purchase Order : 1472 AP	Hard Copy Date :
Invoice Contact : Marcie Ann Encinas			Date Signoff :

LAB ID	CLIENT ID	MATRIX	SAMPLE DATE	SAMPLE TIME	TEST	TEST GROUP	METHOD	FAX DATE	DUE DATES
Q2514-01	TP-92	Solid	07/02/2025	08:00	VOC-TCLVOA-10		8260D	10 Bus. Days	
Q2514-02	TP-93	Solid	07/02/2025	08:30	VOC-TCLVOA-10		8260D	10 Bus. Days	
Q2514-03	TP-94	Solid	07/02/2025	10:35	VOC-TCLVOA-10		8260D	10 Bus. Days	
Q2514-04	TP-96	Solid	07/02/2025	11:25	VOC-TCLVOA-10		8260D	10 Bus. Days	
Q2514-05	TP-97	Solid	07/02/2025	12:15	VOC-TCLVOA-10		8260D	10 Bus. Days	
Q2514-06	TP-103	Solid	07/02/2025	13:45	VOC-TCLVOA-10		8260D	10 Bus. Days	
Q2514-07	TP-36	Solid	07/03/2025	08:20	VOC-TCLVOA-10		8260D	10 Bus. Days	
Q2514-08	TP-78	Solid	07/03/2025	08:55	VOC-TCLVOA-10		8260D	10 Bus. Days	

LOGIN REPORT/SAMPLE TRANSFER

Order ID : Q2514	CAMP02	Order Date : 7/3/2025 1:29:00 PM	Project Mgr :
Client Name : CDM Smith		Project Name : South River WM Replacem	Report Type : Level 2
Client Contact : Marcie Ann Encinas		Receive DateTime : 7/3/2025 2:00:00 PM	EDD Type : EXCEL NOCLEANUP
Invoice Name : CDM Smith		Purchase Order : 1412 d1	Hard Copy Date :
Invoice Contact : Marcie Ann Encinas			Date Signoff :

LAB ID	CLIENT ID	MATRIX	SAMPLE DATE	SAMPLE TIME	TEST	TEST GROUP	METHOD	FAX DATE	DUE DATES
					VOC-TCLVOA-10		8260D	10 Bus. Days	
Q2514-09	TP-81	Solid	07/03/2025	10:00					
					VOC-TCLVOA-10		8260D	10 Bus. Days	
Q2514-10	TP-90	Solid	07/03/2025	12:50					
					VOC-TCLVOA-10		8260D	10 Bus. Days	

Relinquished By : [Signature]

Date / Time : 7/3/25 1430

Received By : [Signature]

Date / Time : 07/03/25 14:30

Storage Area : VOA Refridgerator Room