

VOLATILE INTERNAL STANDARD AREA AND RT SUMMARY

Lab Name: Alliance Contract: CAMP02
Lab Code: ACE SDG NO.: Q2514
Lab File ID: VW031767.D Date Analyzed: 07/09/2025
Instrument ID: MSVOA_W Time Analyzed: 10:00
GC Column: RXI-624 ID: 0.25 (mm) Heated Purge: (Y/N) Y

	IS1 AREA #	RT #	IS2 AREA #	RT #	IS3 AREA #	RT #
12 HOUR STD	247441	7.96	422710	8.85	377094	11.63
UPPER LIMIT	494882	8.459	845420	9.349	754188	12.129
LOWER LIMIT	123721	7.459	211355	8.349	188547	11.129
EPA SAMPLE NO.						
TP-103	196576	7.97	400527	8.86	357765	11.63
VW0709SBL01	208773	7.97	393637	8.85	349817	11.63
VW0709SBS01	187242	7.95	349121	8.85	314319	11.63

IS1 = Pentafluorobenzene
IS2 = 1,4-Difluorobenzene
IS3 = Chlorobenzene-d5

AREA UPPER LIMIT = +100% of internal standard area
AREA LOWER LIMIT = -50% of internal standard area
RT UPPER LIMIT = +0.50 minutes of internal standard RT
RT LOWER LIMIT = -0.50 minutes of internal standard RT

Column used to flag values outside QC limits with an asterisk.
* Values outside of QC limits.