

VOLATILE INTERNAL STANDARD AREA AND RT SUMMARY

Lab Name: Alliance Contract: CAMP02
 Lab Code: ACE SDG NO.: Q2514
 Lab File ID: VW031767.D Date Analyzed: 07/09/2025
 Instrument ID: MSVOA_W Time Analyzed: 10:00
 GC Column: RXI-624 ID: 0.25 (mm) Heated Purge: (Y/N) Y

	IS4 AREA #	RT #				
12 HOUR STD	179078	13.556				
UPPER LIMIT	358156	14.056				
LOWER LIMIT	89539	13.056				
EPA SAMPLE NO.						
TP-103	165193	13.56				
VW0709SBL01	163801	13.56				
VW0709SBS01	151914	13.56				

IS4 = 1,4-Dichlorobenzene-d4

AREA UPPER LIMIT = +100% of internal standard area
 AREA LOWER LIMIT = -50% of internal standard area
 RT UPPER LIMIT = +0.50 minutes of internal standard RT
 RT LOWER LIMIT = -0.50 minutes of internal standard RT

Column used to flag values outside QC limits with an asterisk.
 * Values outside of QC limits.