

VOLATILE INTERNAL STANDARD AREA AND RT SUMMARY

Lab Name: Alliance Contract: CAMP02  
 Lab Code: ACE SDG NO.: Q2514  
 Lab File ID: VY022946.D Date Analyzed: 07/07/2025  
 Instrument ID: MSVOA\_Y Time Analyzed: 09:08  
 GC Column: RXI-624 ID: 0.25 (mm) Heated Purge: (Y/N) Y

|                | IS4<br>AREA # | RT #   |  |  |  |  |
|----------------|---------------|--------|--|--|--|--|
| 12 HOUR STD    | 317038        | 13.347 |  |  |  |  |
| UPPER LIMIT    | 634076        | 13.847 |  |  |  |  |
| LOWER LIMIT    | 158519        | 12.847 |  |  |  |  |
| EPA SAMPLE NO. |               |        |  |  |  |  |
| TP-92          | 221432        | 13.35  |  |  |  |  |
| TP-93          | 231894        | 13.34  |  |  |  |  |
| VY0707SBL01    | 266428        | 13.35  |  |  |  |  |
| VY0707SBS01    | 286401        | 13.35  |  |  |  |  |
| VY0707SBSD01   | 290747        | 13.34  |  |  |  |  |

IS4 = 1,4-Dichlorobenzene-d4

AREA UPPER LIMIT = +100% of internal standard area  
 AREA LOWER LIMIT = -50% of internal standard area  
 RT UPPER LIMIT = +0.50 minutes of internal standard RT  
 RT LOWER LIMIT = -0.50 minutes of internal standard RT

# Column used to flag values outside QC limits with an asterisk.  
 \* Values outside of QC limits.