

VOLATILE ORGANIC INSTRUMENT PERFORMANCE CHECK
BROMOFLUOROBENZENE (BFB)

Lab Name: Alliance

Contract: CAMP02

Lab Code: ACE

SDG NO.: Q2514

Lab File ID: VY022945.D

BFB Injection Date: 07/07/2025

Instrument ID: MSVOA_Y

BFB Injection Time: 08:36

GC Column: RXI-624 ID: 0.25 (mm)

Heated Purge: Y/N Y

m/e	ION ABUNDANCE CRITERIA	% RELATIVE ABUNDANCE
50	15.0 - 40.0% of mass 95	20.9
75	30.0 - 60.0% of mass 95	53.2
95	Base Peak, 100% relative abundance	100
96	5.0 - 9.0% of mass 95	6.3
173	Less than 2.0% of mass 174	1 (1.2) 1
174	50.0 - 100.0% of mass 95	84
175	5.0 - 9.0% of mass 174	6.1 (7.2) 1
176	95.0 - 101.0% of mass 174	79.8 (95) 1
177	5.0 - 9.0% of mass 176	5.4 (6.8) 2

1-Value is % mass 174

2-Value is % mass 176

THIS CHECK APPLIES TO THE FOLLOWING SAMPLES, MS, MSD, BLANKS, AND STANDARDS:

CLIENT ID	LAB SAMPLE ID	LAB FILE ID	DATE ANALYZED	TIME ANALYZED
VSTDCCC050	VSTDCCC050	VY022946.D	07/07/2025	09:08
VY0707SBL01	VY0707SBL01	VY022947.D	07/07/2025	09:44
VY0707SBS01	VY0707SBS01	VY022948.D	07/07/2025	10:17
VY0707SBSD01	VY0707SBSD01	VY022949.D	07/07/2025	10:40
TP-92	Q2514-01	VY022964.D	07/07/2025	16:45
TP-93	Q2514-02	VY022965.D	07/07/2025	17:09