

Metals

- 2a -

INITIAL AND CONTINUING CALIBRATION VERIFICATION

Client: ENVOCARE Environmental Facility Managemer

SDG No.: Q2515

Contract: ENVO01

Lab Code: ACE

Initial Calibration Source: EPA

Continuing Calibration Source: PLASMA-PURE

Sample ID	Analyte	Result ug/L	True Value	% Recovery	Acceptance Window (%R)	M	Analysis Date	Analysis Time	Run Number
ICV17	Mercury	3.77	4.0	94	90 - 110	CV	07/07/2025	11:53	LB136380

Metals

- 2a -

INITIAL AND CONTINUING CALIBRATION VERIFICATION

Client: ENVOCARE Environmental Facility Managemer

SDG No.: Q2515

Contract: ENVO01

Lab Code: ACE

Initial Calibration Source: EPA

Continuing Calibration Source: PLASMA-PURE

Sample ID	Analyte	Result ug/L	True Value	% Recovery	Acceptance Window (%R)	M	Analysis Date	Analysis Time	Run Number
CCV54	Mercury	5.06	5.0	101	90 - 110	CV	07/07/2025	11:58	LB136380
CCV55	Mercury	5.12	5.0	102	90 - 110	CV	07/07/2025	12:34	LB136380
CCV56	Mercury	5.01	5.0	100	90 - 110	CV	07/07/2025	12:58	LB136380

Metals

- 2a -

INITIAL AND CONTINUING CALIBRATION VERIFICATION

Client: ENVOCARE Environmental Facility Managemer

SDG No.: Q2515

Contract: ENVO01

Lab Code: ACE

Initial Calibration Source: EPA

Continuing Calibration Source: Inorganic Ventures

Sample ID	Analyte	Result ug/L	True Value	% Recovery	Acceptance Window (%R)	M	Analysis Date	Analysis Time	Run Number
ICV01	Aluminum	7800	8000	98	90 - 110	P	07/10/2025	18:39	LB136434
	Antimony	4020	4000	100	90 - 110	P	07/10/2025	18:39	LB136434
	Arsenic	3950	4000	99	90 - 110	P	07/10/2025	18:39	LB136434
	Barium	7860	8000	98	90 - 110	P	07/10/2025	18:39	LB136434
	Beryllium	209	200	105	90 - 110	P	07/10/2025	18:39	LB136434
	Cadmium	2050	2000	103	90 - 110	P	07/10/2025	18:39	LB136434
	Calcium	19800	20000	99	90 - 110	P	07/10/2025	18:39	LB136434
	Chromium	800	800	100	90 - 110	P	07/10/2025	18:39	LB136434
	Cobalt	1910	2000	96	90 - 110	P	07/10/2025	18:39	LB136434
	Copper	1040	1000	104	90 - 110	P	07/10/2025	18:39	LB136434
	Iron	4100	4000	103	90 - 110	P	07/10/2025	18:39	LB136434
	Lead	4060	4000	101	90 - 110	P	07/10/2025	18:39	LB136434
	Magnesium	20300	20000	102	90 - 110	P	07/10/2025	18:39	LB136434
	Manganese	1980	2000	99	90 - 110	P	07/10/2025	18:39	LB136434
	Nickel	1990	2000	99	90 - 110	P	07/10/2025	18:39	LB136434
	Potassium	18600	20000	93	90 - 110	P	07/10/2025	18:39	LB136434
	Selenium	4190	4000	105	90 - 110	P	07/10/2025	18:39	LB136434
	Silver	911	1000	91	90 - 110	P	07/10/2025	18:39	LB136434
	Sodium	19200	20000	96	90 - 110	P	07/10/2025	18:39	LB136434
	Thallium	4110	4000	103	90 - 110	P	07/10/2025	18:39	LB136434
	Vanadium	2000	2000	100	90 - 110	P	07/10/2025	18:39	LB136434
	Zinc	1990	2000	100	90 - 110	P	07/10/2025	18:39	LB136434

Metals

- 2a -

INITIAL AND CONTINUING CALIBRATION VERIFICATION

Client: ENVOCARE Environmental Facility Managemer

SDG No.: Q2515

Contract: ENVO01

Lab Code: ACE

Initial Calibration Source: EPA

Continuing Calibration Source: Inorganic Ventures

Sample ID	Analyte	Result ug/L	True Value	% Recovery	Acceptance Window (%R)	M	Analysis Date	Analysis Time	Run Number
LLICV01	Aluminum	97.2	100	97	80 - 120	P	07/10/2025	18:51	LB136434
	Antimony	54.4	50.0	109	80 - 120	P	07/10/2025	18:51	LB136434
	Arsenic	21.2	20.0	106	80 - 120	P	07/10/2025	18:51	LB136434
	Barium	98.6	100	99	80 - 120	P	07/10/2025	18:51	LB136434
	Beryllium	6.44	6.0	107	80 - 120	P	07/10/2025	18:51	LB136434
	Cadmium	5.61	6.0	93	80 - 120	P	07/10/2025	18:51	LB136434
	Calcium	2140	2000	107	80 - 120	P	07/10/2025	18:51	LB136434
	Chromium	10.5	10.0	105	80 - 120	P	07/10/2025	18:51	LB136434
	Cobalt	29.4	30.0	98	80 - 120	P	07/10/2025	18:51	LB136434
	Copper	21.6	20.0	108	80 - 120	P	07/10/2025	18:51	LB136434
	Iron	118	100	118	80 - 120	P	07/10/2025	18:51	LB136434
	Lead	13.2	12.0	110	80 - 120	P	07/10/2025	18:51	LB136434
	Magnesium	2260	2000	113	80 - 120	P	07/10/2025	18:51	LB136434
	Manganese	22.9	20.0	114	80 - 120	P	07/10/2025	18:51	LB136434
	Nickel	39.6	40.0	99	80 - 120	P	07/10/2025	18:51	LB136434
	Potassium	1810	2000	91	80 - 120	P	07/10/2025	18:51	LB136434
	Selenium	20.6	20.0	103	80 - 120	P	07/10/2025	18:51	LB136434
	Silver	10.9	10.0	110	80 - 120	P	07/10/2025	18:51	LB136434
	Sodium	1810	2000	90	80 - 120	P	07/10/2025	18:51	LB136434
	Thallium	40.2	40.0	100	80 - 120	P	07/10/2025	18:51	LB136434
	Vanadium	38.1	40.0	95	80 - 120	P	07/10/2025	18:51	LB136434
	Zinc	46.1	40.0	115	80 - 120	P	07/10/2025	18:51	LB136434

Metals

- 2a -

INITIAL AND CONTINUING CALIBRATION VERIFICATION

Client: ENVOCARE Environmental Facility Managemer

SDG No.: Q2515

Contract: ENVO01

Lab Code: ACE

Initial Calibration Source: EPA

Continuing Calibration Source: Inorganic Ventures

Sample ID	Analyte	Result ug/L	True Value	% Recovery	Acceptance Window (%R)	M	Analysis Date	Analysis Time	Run Number
CCV01	Aluminum	9940	10000	99	90 - 110	P	07/10/2025	19:43	LB136434
	Antimony	5010	5000	100	90 - 110	P	07/10/2025	19:43	LB136434
	Arsenic	4910	5000	98	90 - 110	P	07/10/2025	19:43	LB136434
	Barium	10200	10000	102	90 - 110	P	07/10/2025	19:43	LB136434
	Beryllium	257	250	103	90 - 110	P	07/10/2025	19:43	LB136434
	Cadmium	2510	2500	100	90 - 110	P	07/10/2025	19:43	LB136434
	Calcium	25500	25000	102	90 - 110	P	07/10/2025	19:43	LB136434
	Chromium	1010	1000	101	90 - 110	P	07/10/2025	19:43	LB136434
	Cobalt	2470	2500	99	90 - 110	P	07/10/2025	19:43	LB136434
	Copper	1270	1250	102	90 - 110	P	07/10/2025	19:43	LB136434
	Iron	5140	5000	103	90 - 110	P	07/10/2025	19:43	LB136434
	Lead	4980	5000	100	90 - 110	P	07/10/2025	19:43	LB136434
	Magnesium	25400	25000	102	90 - 110	P	07/10/2025	19:43	LB136434
	Manganese	2550	2500	102	90 - 110	P	07/10/2025	19:43	LB136434
	Nickel	2490	2500	100	90 - 110	P	07/10/2025	19:43	LB136434
	Potassium	25600	25000	103	90 - 110	P	07/10/2025	19:43	LB136434
	Selenium	5060	5000	101	90 - 110	P	07/10/2025	19:43	LB136434
	Silver	1280	1250	103	90 - 110	P	07/10/2025	19:43	LB136434
	Sodium	25000	25000	100	90 - 110	P	07/10/2025	19:43	LB136434
	Thallium	5090	5000	102	90 - 110	P	07/10/2025	19:43	LB136434
	Vanadium	2540	2500	102	90 - 110	P	07/10/2025	19:43	LB136434
	Zinc	2500	2500	100	90 - 110	P	07/10/2025	19:43	LB136434
CCV02	Aluminum	9360	10000	94	90 - 110	P	07/10/2025	20:59	LB136434
	Antimony	4760	5000	95	90 - 110	P	07/10/2025	20:59	LB136434
	Arsenic	4640	5000	93	90 - 110	P	07/10/2025	20:59	LB136434
	Barium	9760	10000	98	90 - 110	P	07/10/2025	20:59	LB136434
	Beryllium	251	250	100	90 - 110	P	07/10/2025	20:59	LB136434
	Cadmium	2380	2500	95	90 - 110	P	07/10/2025	20:59	LB136434
	Calcium	24300	25000	97	90 - 110	P	07/10/2025	20:59	LB136434
	Chromium	962	1000	96	90 - 110	P	07/10/2025	20:59	LB136434
	Cobalt	2320	2500	93	90 - 110	P	07/10/2025	20:59	LB136434
	Copper	1230	1250	99	90 - 110	P	07/10/2025	20:59	LB136434
	Iron	5010	5000	100	90 - 110	P	07/10/2025	20:59	LB136434
	Lead	4710	5000	94	90 - 110	P	07/10/2025	20:59	LB136434

Metals

- 2a -

INITIAL AND CONTINUING CALIBRATION VERIFICATION

Client: ENVOCARE Environmental Facility Managemer

SDG No.: Q2515

Contract: ENVO01

Lab Code: ACE

Initial Calibration Source: EPA

Continuing Calibration Source: Inorganic Ventures

Sample ID	Analyte	Result ug/L	True Value	% Recovery	Acceptance Window (%R)	M	Analysis Date	Analysis Time	Run Number
CCV02	Magnesium	24300	25000	97	90 - 110	P	07/10/2025	20:59	LB136434
	Manganese	2440	2500	98	90 - 110	P	07/10/2025	20:59	LB136434
	Nickel	2350	2500	94	90 - 110	P	07/10/2025	20:59	LB136434
	Potassium	25900	25000	104	90 - 110	P	07/10/2025	20:59	LB136434
	Selenium	4800	5000	96	90 - 110	P	07/10/2025	20:59	LB136434
	Silver	1220	1250	97	90 - 110	P	07/10/2025	20:59	LB136434
	Sodium	25000	25000	100	90 - 110	P	07/10/2025	20:59	LB136434
	Thallium	4840	5000	97	90 - 110	P	07/10/2025	20:59	LB136434
	Vanadium	2430	2500	97	90 - 110	P	07/10/2025	20:59	LB136434
	Zinc	2370	2500	95	90 - 110	P	07/10/2025	20:59	LB136434
CCV03	Aluminum	9640	10000	96	90 - 110	P	07/10/2025	22:32	LB136434
	Antimony	4880	5000	98	90 - 110	P	07/10/2025	22:32	LB136434
	Arsenic	4830	5000	96	90 - 110	P	07/10/2025	22:32	LB136434
	Barium	10000	10000	100	90 - 110	P	07/10/2025	22:32	LB136434
	Beryllium	255	250	102	90 - 110	P	07/10/2025	22:32	LB136434
	Cadmium	2480	2500	99	90 - 110	P	07/10/2025	22:32	LB136434
	Calcium	24900	25000	100	90 - 110	P	07/10/2025	22:32	LB136434
	Chromium	997	1000	100	90 - 110	P	07/10/2025	22:32	LB136434
	Cobalt	2410	2500	96	90 - 110	P	07/10/2025	22:32	LB136434
	Copper	1270	1250	102	90 - 110	P	07/10/2025	22:32	LB136434
	Iron	5140	5000	103	90 - 110	P	07/10/2025	22:32	LB136434
	Lead	4930	5000	99	90 - 110	P	07/10/2025	22:32	LB136434
	Magnesium	24800	25000	99	90 - 110	P	07/10/2025	22:32	LB136434
	Manganese	2510	2500	100	90 - 110	P	07/10/2025	22:32	LB136434
	Nickel	2440	2500	97	90 - 110	P	07/10/2025	22:32	LB136434
	Potassium	26900	25000	108	90 - 110	P	07/10/2025	22:32	LB136434
	Selenium	4940	5000	99	90 - 110	P	07/10/2025	22:32	LB136434
	Silver	1250	1250	100	90 - 110	P	07/10/2025	22:32	LB136434
	Sodium	26500	25000	106	90 - 110	P	07/10/2025	22:32	LB136434
	Thallium	4990	5000	100	90 - 110	P	07/10/2025	22:32	LB136434
	Vanadium	2490	2500	100	90 - 110	P	07/10/2025	22:32	LB136434
	Zinc	2450	2500	98	90 - 110	P	07/10/2025	22:32	LB136434
CCV04	Aluminum	9350	10000	94	90 - 110	P	07/10/2025	23:29	LB136434
	Antimony	4640	5000	93	90 - 110	P	07/10/2025	23:29	LB136434

Metals

- 2a -

INITIAL AND CONTINUING CALIBRATION VERIFICATION

Client: ENVOCARE Environmental Facility Managemer

SDG No.: Q2515

Contract: ENVO01

Lab Code: ACE

Initial Calibration Source: EPA

Continuing Calibration Source: Inorganic Ventures

Sample ID	Analyte	Result ug/L	True Value	% Recovery	Acceptance Window (%R)	M	Analysis Date	Analysis Time	Run Number
CCV04	Arsenic	4600	5000	92	90 - 110	P	07/10/2025	23:29	LB136434
	Barium	9570	10000	96	90 - 110	P	07/10/2025	23:29	LB136434
	Beryllium	246	250	98	90 - 110	P	07/10/2025	23:29	LB136434
	Cadmium	2350	2500	94	90 - 110	P	07/10/2025	23:29	LB136434
	Calcium	23900	25000	96	90 - 110	P	07/10/2025	23:29	LB136434
	Chromium	957	1000	96	90 - 110	P	07/10/2025	23:29	LB136434
	Cobalt	2290	2500	92	90 - 110	P	07/10/2025	23:29	LB136434
	Copper	1220	1250	97	90 - 110	P	07/10/2025	23:29	LB136434
	Iron	4890	5000	98	90 - 110	P	07/10/2025	23:29	LB136434
	Lead	4690	5000	94	90 - 110	P	07/10/2025	23:29	LB136434
	Magnesium	23900	25000	96	90 - 110	P	07/10/2025	23:29	LB136434
	Manganese	2400	2500	96	90 - 110	P	07/10/2025	23:29	LB136434
	Nickel	2310	2500	92	90 - 110	P	07/10/2025	23:29	LB136434
	Potassium	25900	25000	104	90 - 110	P	07/10/2025	23:29	LB136434
	Selenium	4700	5000	94	90 - 110	P	07/10/2025	23:29	LB136434
	Silver	1200	1250	96	90 - 110	P	07/10/2025	23:29	LB136434
	Sodium	25600	25000	102	90 - 110	P	07/10/2025	23:29	LB136434
	Thallium	4730	5000	95	90 - 110	P	07/10/2025	23:29	LB136434
	Vanadium	2410	2500	96	90 - 110	P	07/10/2025	23:29	LB136434
	Zinc	2340	2500	94	90 - 110	P	07/10/2025	23:29	LB136434
CCV05	Aluminum	9510	10000	95	90 - 110	P	07/11/2025	00:47	LB136434
	Antimony	4770	5000	95	90 - 110	P	07/11/2025	00:47	LB136434
	Arsenic	4750	5000	95	90 - 110	P	07/11/2025	00:47	LB136434
	Barium	9800	10000	98	90 - 110	P	07/11/2025	00:47	LB136434
	Beryllium	250	250	100	90 - 110	P	07/11/2025	00:47	LB136434
	Cadmium	2410	2500	96	90 - 110	P	07/11/2025	00:47	LB136434
	Calcium	24200	25000	97	90 - 110	P	07/11/2025	00:47	LB136434
	Chromium	978	1000	98	90 - 110	P	07/11/2025	00:47	LB136434
	Cobalt	2370	2500	95	90 - 110	P	07/11/2025	00:47	LB136434
	Copper	1240	1250	100	90 - 110	P	07/11/2025	00:47	LB136434
	Iron	4950	5000	99	90 - 110	P	07/11/2025	00:47	LB136434
	Lead	4820	5000	96	90 - 110	P	07/11/2025	00:47	LB136434
	Magnesium	24100	25000	96	90 - 110	P	07/11/2025	00:47	LB136434
	Manganese	2460	2500	98	90 - 110	P	07/11/2025	00:47	LB136434

Metals

- 2a -

INITIAL AND CONTINUING CALIBRATION VERIFICATION

Client: ENVOCARE Environmental Facility Managemer

SDG No.: Q2515

Contract: ENVO01

Lab Code: ACE

Initial Calibration Source: EPA

Continuing Calibration Source: Inorganic Ventures

Sample ID	Analyte	Result ug/L	True Value	% Recovery	Acceptance Window (%R)	M	Analysis Date	Analysis Time	Run Number
CCV05	Nickel	2370	2500	95	90 - 110	P	07/11/2025	00:47	LB136434
	Potassium	27100	25000	108	90 - 110	P	07/11/2025	00:47	LB136434
	Selenium	4790	5000	96	90 - 110	P	07/11/2025	00:47	LB136434
	Silver	1220	1250	97	90 - 110	P	07/11/2025	00:47	LB136434
	Sodium	26600	25000	106	90 - 110	P	07/11/2025	00:47	LB136434
	Thallium	4840	5000	97	90 - 110	P	07/11/2025	00:47	LB136434
	Vanadium	2470	2500	99	90 - 110	P	07/11/2025	00:47	LB136434
	Zinc	2430	2500	97	90 - 110	P	07/11/2025	00:47	LB136434
CCV06	Aluminum	9520	10000	95	90 - 110	P	07/11/2025	01:47	LB136434
	Antimony	4790	5000	96	90 - 110	P	07/11/2025	01:47	LB136434
	Arsenic	4780	5000	96	90 - 110	P	07/11/2025	01:47	LB136434
	Barium	9810	10000	98	90 - 110	P	07/11/2025	01:47	LB136434
	Beryllium	248	250	99	90 - 110	P	07/11/2025	01:47	LB136434
	Cadmium	2410	2500	96	90 - 110	P	07/11/2025	01:47	LB136434
	Calcium	24100	25000	97	90 - 110	P	07/11/2025	01:47	LB136434
	Chromium	978	1000	98	90 - 110	P	07/11/2025	01:47	LB136434
	Cobalt	2370	2500	95	90 - 110	P	07/11/2025	01:47	LB136434
	Copper	1240	1250	99	90 - 110	P	07/11/2025	01:47	LB136434
	Iron	4890	5000	98	90 - 110	P	07/11/2025	01:47	LB136434
	Lead	4820	5000	96	90 - 110	P	07/11/2025	01:47	LB136434
	Magnesium	24200	25000	97	90 - 110	P	07/11/2025	01:47	LB136434
	Manganese	2440	2500	98	90 - 110	P	07/11/2025	01:47	LB136434
	Nickel	2370	2500	95	90 - 110	P	07/11/2025	01:47	LB136434
	Potassium	26200	25000	105	90 - 110	P	07/11/2025	01:47	LB136434
	Selenium	4810	5000	96	90 - 110	P	07/11/2025	01:47	LB136434
	Silver	1220	1250	97	90 - 110	P	07/11/2025	01:47	LB136434
	Sodium	26400	25000	106	90 - 110	P	07/11/2025	01:47	LB136434
	Thallium	4840	5000	97	90 - 110	P	07/11/2025	01:47	LB136434
	Vanadium	2470	2500	99	90 - 110	P	07/11/2025	01:47	LB136434
	Zinc	2440	2500	98	90 - 110	P	07/11/2025	01:47	LB136434
CCV07	Aluminum	9390	10000	94	90 - 110	P	07/11/2025	03:17	LB136434
	Antimony	4710	5000	94	90 - 110	P	07/11/2025	03:17	LB136434
	Arsenic	4640	5000	93	90 - 110	P	07/11/2025	03:17	LB136434
	Barium	9540	10000	95	90 - 110	P	07/11/2025	03:17	LB136434

Metals

- 2a -

INITIAL AND CONTINUING CALIBRATION VERIFICATION

Client: ENVOCARE Environmental Facility Managemer

SDG No.: Q2515

Contract: ENVO01

Lab Code: ACE

Initial Calibration Source: EPA

Continuing Calibration Source: Inorganic Ventures

Sample ID	Analyte	Result ug/L	True Value	% Recovery	Acceptance Window (%R)	M	Analysis Date	Analysis Time	Run Number
CCV07	Beryllium	236	250	95	90 - 110	P	07/11/2025	03:17	LB136434
	Cadmium	2330	2500	93	90 - 110	P	07/11/2025	03:17	LB136434
	Calcium	23700	25000	95	90 - 110	P	07/11/2025	03:17	LB136434
	Chromium	961	1000	96	90 - 110	P	07/11/2025	03:17	LB136434
	Cobalt	2330	2500	93	90 - 110	P	07/11/2025	03:17	LB136434
	Copper	1180	1250	94	90 - 110	P	07/11/2025	03:17	LB136434
	Iron	4630	5000	93	90 - 110	P	07/11/2025	03:17	LB136434
	Lead	4670	5000	93	90 - 110	P	07/11/2025	03:17	LB136434
	Magnesium	23600	25000	94	90 - 110	P	07/11/2025	03:17	LB136434
	Manganese	2400	2500	96	90 - 110	P	07/11/2025	03:17	LB136434
	Nickel	2320	2500	93	90 - 110	P	07/11/2025	03:17	LB136434
	Potassium	23800	25000	95	90 - 110	P	07/11/2025	03:17	LB136434
	Selenium	4670	5000	94	90 - 110	P	07/11/2025	03:17	LB136434
	Silver	1200	1250	96	90 - 110	P	07/11/2025	03:17	LB136434
	Sodium	24200	25000	97	90 - 110	P	07/11/2025	03:17	LB136434
	Thallium	4710	5000	94	90 - 110	P	07/11/2025	03:17	LB136434
	Vanadium	2420	2500	97	90 - 110	P	07/11/2025	03:17	LB136434
	Zinc	2380	2500	95	90 - 110	P	07/11/2025	03:17	LB136434
CCV08	Aluminum	9520	10000	95	90 - 110	P	07/11/2025	04:18	LB136434
	Antimony	4760	5000	95	90 - 110	P	07/11/2025	04:18	LB136434
	Arsenic	4790	5000	96	90 - 110	P	07/11/2025	04:18	LB136434
	Barium	9730	10000	97	90 - 110	P	07/11/2025	04:18	LB136434
	Beryllium	244	250	98	90 - 110	P	07/11/2025	04:18	LB136434
	Cadmium	2400	2500	96	90 - 110	P	07/11/2025	04:18	LB136434
	Calcium	23800	25000	95	90 - 110	P	07/11/2025	04:18	LB136434
	Chromium	987	1000	99	90 - 110	P	07/11/2025	04:18	LB136434
	Cobalt	2390	2500	96	90 - 110	P	07/11/2025	04:18	LB136434
	Copper	1220	1250	98	90 - 110	P	07/11/2025	04:18	LB136434
	Iron	4780	5000	96	90 - 110	P	07/11/2025	04:18	LB136434
	Lead	4820	5000	96	90 - 110	P	07/11/2025	04:18	LB136434
	Magnesium	23700	25000	95	90 - 110	P	07/11/2025	04:18	LB136434
	Manganese	2430	2500	97	90 - 110	P	07/11/2025	04:18	LB136434
	Nickel	2370	2500	95	90 - 110	P	07/11/2025	04:18	LB136434
	Potassium	26600	25000	106	90 - 110	P	07/11/2025	04:18	LB136434

Metals

- 2a -

INITIAL AND CONTINUING CALIBRATION VERIFICATION

Client: ENVOCARE Environmental Facility Manager

SDG No.: Q2515

Contract: ENVO01

Lab Code: ACE

Initial Calibration Source: EPA

Continuing Calibration Source: Inorganic Ventures

Sample ID	Analyte	Result ug/L	True Value	% Recovery	Acceptance Window (%R)	M	Analysis Date	Analysis Time	Run Number
CCV08	Selenium	4720	5000	94	90 - 110	P	07/11/2025	04:18	LB136434
	Silver	1220	1250	98	90 - 110	P	07/11/2025	04:18	LB136434
	Sodium	26800	25000	107	90 - 110	P	07/11/2025	04:18	LB136434
	Thallium	4790	5000	96	90 - 110	P	07/11/2025	04:18	LB136434
	Vanadium	2470	2500	99	90 - 110	P	07/11/2025	04:18	LB136434
	Zinc	2440	2500	98	90 - 110	P	07/11/2025	04:18	LB136434
CCV09	Aluminum	9800	10000	98	90 - 110	P	07/11/2025	05:19	LB136434
	Antimony	4830	5000	96	90 - 110	P	07/11/2025	05:19	LB136434
	Arsenic	4770	5000	96	90 - 110	P	07/11/2025	05:19	LB136434
	Barium	9960	10000	100	90 - 110	P	07/11/2025	05:19	LB136434
	Beryllium	246	250	99	90 - 110	P	07/11/2025	05:19	LB136434
	Cadmium	2380	2500	95	90 - 110	P	07/11/2025	05:19	LB136434
	Calcium	24400	25000	97	90 - 110	P	07/11/2025	05:19	LB136434
	Chromium	1000	1000	100	90 - 110	P	07/11/2025	05:19	LB136434
	Cobalt	2400	2500	96	90 - 110	P	07/11/2025	05:19	LB136434
	Copper	1250	1250	100	90 - 110	P	07/11/2025	05:19	LB136434
	Iron	4810	5000	96	90 - 110	P	07/11/2025	05:19	LB136434
	Lead	4770	5000	96	90 - 110	P	07/11/2025	05:19	LB136434
	Magnesium	24200	25000	97	90 - 110	P	07/11/2025	05:19	LB136434
	Manganese	2490	2500	100	90 - 110	P	07/11/2025	05:19	LB136434
	Nickel	2370	2500	95	90 - 110	P	07/11/2025	05:19	LB136434
	Potassium	25600	25000	102	90 - 110	P	07/11/2025	05:19	LB136434
	Selenium	4720	5000	94	90 - 110	P	07/11/2025	05:19	LB136434
	Silver	1240	1250	99	90 - 110	P	07/11/2025	05:19	LB136434
	Sodium	26100	25000	104	90 - 110	P	07/11/2025	05:19	LB136434
	Thallium	4770	5000	96	90 - 110	P	07/11/2025	05:19	LB136434
	Vanadium	2540	2500	102	90 - 110	P	07/11/2025	05:19	LB136434
	Zinc	2500	2500	100	90 - 110	P	07/11/2025	05:19	LB136434
CCV10	Aluminum	9570	10000	96	90 - 110	P	07/11/2025	06:16	LB136434
	Antimony	4770	5000	95	90 - 110	P	07/11/2025	06:16	LB136434
	Arsenic	4790	5000	96	90 - 110	P	07/11/2025	06:16	LB136434
	Barium	9830	10000	98	90 - 110	P	07/11/2025	06:16	LB136434
	Beryllium	242	250	97	90 - 110	P	07/11/2025	06:16	LB136434
	Cadmium	2390	2500	96	90 - 110	P	07/11/2025	06:16	LB136434

Metals

- 2a -

INITIAL AND CONTINUING CALIBRATION VERIFICATION

Client: ENVOCARE Environmental Facility Manager

SDG No.: Q2515

Contract: ENVO01

Lab Code: ACE

Initial Calibration Source: EPA

Continuing Calibration Source: Inorganic Ventures

Sample ID	Analyte	Result ug/L	True Value	% Recovery	Acceptance Window (%R)	M	Analysis Date	Analysis Time	Run Number
CCV10	Calcium	23700	25000	95	90 - 110	P	07/11/2025	06:16	LB136434
	Chromium	988	1000	99	90 - 110	P	07/11/2025	06:16	LB136434
	Cobalt	2390	2500	96	90 - 110	P	07/11/2025	06:16	LB136434
	Copper	1220	1250	98	90 - 110	P	07/11/2025	06:16	LB136434
	Iron	4700	5000	94	90 - 110	P	07/11/2025	06:16	LB136434
	Lead	4800	5000	96	90 - 110	P	07/11/2025	06:16	LB136434
	Magnesium	23600	25000	94	90 - 110	P	07/11/2025	06:16	LB136434
	Manganese	2440	2500	98	90 - 110	P	07/11/2025	06:16	LB136434
	Nickel	2360	2500	94	90 - 110	P	07/11/2025	06:16	LB136434
	Potassium	26000	25000	104	90 - 110	P	07/11/2025	06:16	LB136434
	Selenium	4710	5000	94	90 - 110	P	07/11/2025	06:16	LB136434
	Silver	1210	1250	97	90 - 110	P	07/11/2025	06:16	LB136434
	Sodium	26700	25000	107	90 - 110	P	07/11/2025	06:16	LB136434
	Thallium	4770	5000	95	90 - 110	P	07/11/2025	06:16	LB136434
	Vanadium	2490	2500	100	90 - 110	P	07/11/2025	06:16	LB136434
	Zinc	2470	2500	99	90 - 110	P	07/11/2025	06:16	LB136434
CCV11	Aluminum	9570	10000	96	90 - 110	P	07/11/2025	06:43	LB136434
	Antimony	4730	5000	95	90 - 110	P	07/11/2025	06:43	LB136434
	Arsenic	4750	5000	95	90 - 110	P	07/11/2025	06:43	LB136434
	Barium	9790	10000	98	90 - 110	P	07/11/2025	06:43	LB136434
	Beryllium	241	250	96	90 - 110	P	07/11/2025	06:43	LB136434
	Cadmium	2370	2500	95	90 - 110	P	07/11/2025	06:43	LB136434
	Calcium	23900	25000	95	90 - 110	P	07/11/2025	06:43	LB136434
	Chromium	982	1000	98	90 - 110	P	07/11/2025	06:43	LB136434
	Cobalt	2370	2500	95	90 - 110	P	07/11/2025	06:43	LB136434
	Copper	1210	1250	97	90 - 110	P	07/11/2025	06:43	LB136434
	Iron	4700	5000	94	90 - 110	P	07/11/2025	06:43	LB136434
	Lead	4770	5000	95	90 - 110	P	07/11/2025	06:43	LB136434
	Magnesium	23700	25000	95	90 - 110	P	07/11/2025	06:43	LB136434
	Manganese	2440	2500	98	90 - 110	P	07/11/2025	06:43	LB136434
	Nickel	2340	2500	94	90 - 110	P	07/11/2025	06:43	LB136434
	Potassium	25900	25000	103	90 - 110	P	07/11/2025	06:43	LB136434
	Selenium	4660	5000	93	90 - 110	P	07/11/2025	06:43	LB136434
	Silver	1210	1250	97	90 - 110	P	07/11/2025	06:43	LB136434

Metals

- 2a -

INITIAL AND CONTINUING CALIBRATION VERIFICATION

Client: ENVOCARE Environmental Facility Managemer

SDG No.: Q2515

Contract: ENVO01

Lab Code: ACE

Initial Calibration Source: EPA

Continuing Calibration Source: Inorganic Ventures

Sample ID	Analyte	Result ug/L	True Value	% Recovery	Acceptance Window (%R)	M	Analysis Date	Analysis Time	Run Number
CCV11	Sodium	26400	25000	106	90 - 110	P	07/11/2025	06:43	LB136434
	Thallium	4710	5000	94	90 - 110	P	07/11/2025	06:43	LB136434
	Vanadium	2460	2500	98	90 - 110	P	07/11/2025	06:43	LB136434
	Zinc	2440	2500	98	90 - 110	P	07/11/2025	06:43	LB136434

Metals

- 2a -

INITIAL AND CONTINUING CALIBRATION VERIFICATION

Client: ENVOCARE Environmental Facility Managemer

SDG No.: Q2515

Contract: ENVO01

Lab Code: ACE

Initial Calibration Source: EPA

Continuing Calibration Source: Inorganic Ventures

Sample ID	Analyte	Result ug/L	True Value	% Recovery	Acceptance Window (%R)	M	Analysis Date	Analysis Time	Run Number
ICV01	Aluminum	7880	8000	98	90 - 110	P	07/17/2025	14:14	LB136534
	Antimony	4110	4000	103	90 - 110	P	07/17/2025	14:14	LB136534
	Arsenic	3840	4000	96	90 - 110	P	07/17/2025	14:14	LB136534
	Barium	8020	8000	100	90 - 110	P	07/17/2025	14:14	LB136534
	Beryllium	200	200	100	90 - 110	P	07/17/2025	14:14	LB136534
	Cadmium	1950	2000	97	90 - 110	P	07/17/2025	14:14	LB136534
	Calcium	19900	20000	99	90 - 110	P	07/17/2025	14:14	LB136534
	Chromium	791	800	99	90 - 110	P	07/17/2025	14:14	LB136534
	Cobalt	2000	2000	100	90 - 110	P	07/17/2025	14:14	LB136534
	Copper	1020	1000	102	90 - 110	P	07/17/2025	14:14	LB136534
	Iron	3850	4000	96	90 - 110	P	07/17/2025	14:14	LB136534
	Lead	3860	4000	96	90 - 110	P	07/17/2025	14:14	LB136534
	Magnesium	20100	20000	101	90 - 110	P	07/17/2025	14:14	LB136534
	Manganese	2000	2000	100	90 - 110	P	07/17/2025	14:14	LB136534
	Nickel	1990	2000	99	90 - 110	P	07/17/2025	14:14	LB136534
	Potassium	18100	20000	91	90 - 110	P	07/17/2025	14:14	LB136534
	Selenium	3940	4000	98	90 - 110	P	07/17/2025	14:14	LB136534
	Silver	936	1000	94	90 - 110	P	07/17/2025	14:14	LB136534
	Sodium	18700	20000	94	90 - 110	P	07/17/2025	14:14	LB136534
	Thallium	3850	4000	96	90 - 110	P	07/17/2025	14:14	LB136534
	Vanadium	2000	2000	100	90 - 110	P	07/17/2025	14:14	LB136534
	Zinc	2010	2000	100	90 - 110	P	07/17/2025	14:14	LB136534

Metals

- 2a -

INITIAL AND CONTINUING CALIBRATION VERIFICATION

Client: ENVOCARE Environmental Facility Managemer

SDG No.: Q2515

Contract: ENVO01

Lab Code: ACE

Initial Calibration Source: EPA

Continuing Calibration Source: Inorganic Ventures

Sample ID	Analyte	Result ug/L	True Value	% Recovery	Acceptance Window (%R)	M	Analysis Date	Analysis Time	Run Number
LLICV01	Aluminum	119	100	119	80 - 120	P	07/17/2025	14:24	LB136534
	Antimony	52.4	50.0	105	80 - 120	P	07/17/2025	14:24	LB136534
	Arsenic	16.9	20.0	84	80 - 120	P	07/17/2025	14:24	LB136534
	Barium	102	100	102	80 - 120	P	07/17/2025	14:24	LB136534
	Beryllium	6.21	6.0	104	80 - 120	P	07/17/2025	14:24	LB136534
	Cadmium	5.65	6.0	94	80 - 120	P	07/17/2025	14:24	LB136534
	Calcium	2050	2000	103	80 - 120	P	07/17/2025	14:24	LB136534
	Chromium	10.2	10.0	102	80 - 120	P	07/17/2025	14:24	LB136534
	Cobalt	30.3	30.0	101	80 - 120	P	07/17/2025	14:24	LB136534
	Copper	21.3	20.0	107	80 - 120	P	07/17/2025	14:24	LB136534
	Iron	110	100	110	80 - 120	P	07/17/2025	14:24	LB136534
	Lead	10.9	12.0	91	80 - 120	P	07/17/2025	14:24	LB136534
	Magnesium	2220	2000	111	80 - 120	P	07/17/2025	14:24	LB136534
	Manganese	21.1	20.0	106	80 - 120	P	07/17/2025	14:24	LB136534
	Nickel	39.7	40.0	99	80 - 120	P	07/17/2025	14:24	LB136534
	Potassium	1740	2000	87	80 - 120	P	07/17/2025	14:24	LB136534
	Selenium	17.6	20.0	88	80 - 120	P	07/17/2025	14:24	LB136534
	Silver	10.5	10.0	105	80 - 120	P	07/17/2025	14:24	LB136534
	Sodium	1750	2000	87	80 - 120	P	07/17/2025	14:24	LB136534
	Thallium	41.1	40.0	103	80 - 120	P	07/17/2025	14:24	LB136534
	Vanadium	40.0	40.0	100	80 - 120	P	07/17/2025	14:24	LB136534
	Zinc	42.6	40.0	106	80 - 120	P	07/17/2025	14:24	LB136534

Metals

- 2a -

INITIAL AND CONTINUING CALIBRATION VERIFICATION

Client: ENVOCARE Environmental Facility Managemer

SDG No.: Q2515

Contract: ENVO01

Lab Code: ACE

Initial Calibration Source: EPA

Continuing Calibration Source: Inorganic Ventures

Sample ID	Analyte	Result ug/L	True Value	% Recovery	Acceptance Window (%R)	M	Analysis Date	Analysis Time	Run Number
CCV01	Aluminum	9950	10000	100	90 - 110	P	07/17/2025	14:56	LB136534
	Antimony	5020	5000	100	90 - 110	P	07/17/2025	14:56	LB136534
	Arsenic	4950	5000	99	90 - 110	P	07/17/2025	14:56	LB136534
	Barium	10100	10000	101	90 - 110	P	07/17/2025	14:56	LB136534
	Beryllium	254	250	102	90 - 110	P	07/17/2025	14:56	LB136534
	Cadmium	2500	2500	100	90 - 110	P	07/17/2025	14:56	LB136534
	Calcium	25100	25000	101	90 - 110	P	07/17/2025	14:56	LB136534
	Chromium	1020	1000	102	90 - 110	P	07/17/2025	14:56	LB136534
	Cobalt	2500	2500	100	90 - 110	P	07/17/2025	14:56	LB136534
	Copper	1260	1250	101	90 - 110	P	07/17/2025	14:56	LB136534
	Iron	5020	5000	100	90 - 110	P	07/17/2025	14:56	LB136534
	Lead	4990	5000	100	90 - 110	P	07/17/2025	14:56	LB136534
	Magnesium	25100	25000	100	90 - 110	P	07/17/2025	14:56	LB136534
	Manganese	2510	2500	100	90 - 110	P	07/17/2025	14:56	LB136534
	Nickel	2500	2500	100	90 - 110	P	07/17/2025	14:56	LB136534
	Potassium	25400	25000	102	90 - 110	P	07/17/2025	14:56	LB136534
	Selenium	5010	5000	100	90 - 110	P	07/17/2025	14:56	LB136534
	Silver	1270	1250	102	90 - 110	P	07/17/2025	14:56	LB136534
	Sodium	25400	25000	102	90 - 110	P	07/17/2025	14:56	LB136534
	Thallium	5050	5000	101	90 - 110	P	07/17/2025	14:56	LB136534
	Vanadium	2510	2500	100	90 - 110	P	07/17/2025	14:56	LB136534
	Zinc	2520	2500	101	90 - 110	P	07/17/2025	14:56	LB136534
CCV02	Aluminum	9800	10000	98	90 - 110	P	07/17/2025	15:44	LB136534
	Antimony	4760	5000	95	90 - 110	P	07/17/2025	15:44	LB136534
	Arsenic	4730	5000	95	90 - 110	P	07/17/2025	15:44	LB136534
	Barium	9920	10000	99	90 - 110	P	07/17/2025	15:44	LB136534
	Beryllium	250	250	100	90 - 110	P	07/17/2025	15:44	LB136534
	Cadmium	2410	2500	96	90 - 110	P	07/17/2025	15:44	LB136534
	Calcium	24700	25000	99	90 - 110	P	07/17/2025	15:44	LB136534
	Chromium	986	1000	99	90 - 110	P	07/17/2025	15:44	LB136534
	Cobalt	2390	2500	96	90 - 110	P	07/17/2025	15:44	LB136534
	Copper	1240	1250	100	90 - 110	P	07/17/2025	15:44	LB136534
	Iron	4950	5000	99	90 - 110	P	07/17/2025	15:44	LB136534
	Lead	4810	5000	96	90 - 110	P	07/17/2025	15:44	LB136534

Metals

- 2a -

INITIAL AND CONTINUING CALIBRATION VERIFICATION

Client: ENVOCARE Environmental Facility Managemer

SDG No.: Q2515

Contract: ENVO01

Lab Code: ACE

Initial Calibration Source: EPA

Continuing Calibration Source: Inorganic Ventures

Sample ID	Analyte	Result ug/L	True Value	% Recovery	Acceptance Window (%R)	M	Analysis Date	Analysis Time	Run Number
CCV02	Magnesium	24500	25000	98	90 - 110	P	07/17/2025	15:44	LB136534
	Manganese	2470	2500	99	90 - 110	P	07/17/2025	15:44	LB136534
	Nickel	2400	2500	96	90 - 110	P	07/17/2025	15:44	LB136534
	Potassium	24700	25000	99	90 - 110	P	07/17/2025	15:44	LB136534
	Selenium	4770	5000	95	90 - 110	P	07/17/2025	15:44	LB136534
	Silver	1230	1250	98	90 - 110	P	07/17/2025	15:44	LB136534
	Sodium	24900	25000	100	90 - 110	P	07/17/2025	15:44	LB136534
	Thallium	4820	5000	96	90 - 110	P	07/17/2025	15:44	LB136534
	Vanadium	2450	2500	98	90 - 110	P	07/17/2025	15:44	LB136534
	Zinc	2440	2500	98	90 - 110	P	07/17/2025	15:44	LB136534
CCV03	Aluminum	10100	10000	101	90 - 110	P	07/17/2025	17:18	LB136534
	Antimony	4980	5000	100	90 - 110	P	07/17/2025	17:18	LB136534
	Arsenic	4960	5000	99	90 - 110	P	07/17/2025	17:18	LB136534
	Barium	10100	10000	101	90 - 110	P	07/17/2025	17:18	LB136534
	Beryllium	258	250	103	90 - 110	P	07/17/2025	17:18	LB136534
	Cadmium	2520	2500	101	90 - 110	P	07/17/2025	17:18	LB136534
	Calcium	25300	25000	101	90 - 110	P	07/17/2025	17:18	LB136534
	Chromium	1020	1000	102	90 - 110	P	07/17/2025	17:18	LB136534
	Cobalt	2510	2500	100	90 - 110	P	07/17/2025	17:18	LB136534
	Copper	1280	1250	103	90 - 110	P	07/17/2025	17:18	LB136534
	Iron	5110	5000	102	90 - 110	P	07/17/2025	17:18	LB136534
	Lead	5040	5000	101	90 - 110	P	07/17/2025	17:18	LB136534
	Magnesium	25000	25000	100	90 - 110	P	07/17/2025	17:18	LB136534
	Manganese	2510	2500	100	90 - 110	P	07/17/2025	17:18	LB136534
	Nickel	2510	2500	100	90 - 110	P	07/17/2025	17:18	LB136534
	Potassium	25600	25000	102	90 - 110	P	07/17/2025	17:18	LB136534
	Selenium	4980	5000	100	90 - 110	P	07/17/2025	17:18	LB136534
	Silver	1270	1250	101	90 - 110	P	07/17/2025	17:18	LB136534
	Sodium	25800	25000	103	90 - 110	P	07/17/2025	17:18	LB136534
	Thallium	5050	5000	101	90 - 110	P	07/17/2025	17:18	LB136534
	Vanadium	2510	2500	100	90 - 110	P	07/17/2025	17:18	LB136534
	Zinc	2500	2500	100	90 - 110	P	07/17/2025	17:18	LB136534
CCV04	Aluminum	10000	10000	100	90 - 110	P	07/17/2025	18:45	LB136534
	Antimony	4990	5000	100	90 - 110	P	07/17/2025	18:45	LB136534

Metals

- 2a -

INITIAL AND CONTINUING CALIBRATION VERIFICATION

Client: ENVOCARE Environmental Facility Managemer

SDG No.: Q2515

Contract: ENVO01

Lab Code: ACE

Initial Calibration Source: EPA

Continuing Calibration Source: Inorganic Ventures

Sample ID	Analyte	Result ug/L	True Value	% Recovery	Acceptance Window (%R)	M	Analysis Date	Analysis Time	Run Number
CCV04	Arsenic	4880	5000	98	90 - 110	P	07/17/2025	18:45	LB136534
	Barium	10200	10000	102	90 - 110	P	07/17/2025	18:45	LB136534
	Beryllium	246	250	98	90 - 110	P	07/17/2025	18:45	LB136534
	Cadmium	2470	2500	99	90 - 110	P	07/17/2025	18:45	LB136534
	Calcium	25000	25000	100	90 - 110	P	07/17/2025	18:45	LB136534
	Chromium	1010	1000	101	90 - 110	P	07/17/2025	18:45	LB136534
	Cobalt	2480	2500	99	90 - 110	P	07/17/2025	18:45	LB136534
	Copper	1240	1250	99	90 - 110	P	07/17/2025	18:45	LB136534
	Iron	4990	5000	100	90 - 110	P	07/17/2025	18:45	LB136534
	Lead	4940	5000	99	90 - 110	P	07/17/2025	18:45	LB136534
	Magnesium	24600	25000	98	90 - 110	P	07/17/2025	18:45	LB136534
	Manganese	2450	2500	98	90 - 110	P	07/17/2025	18:45	LB136534
	Nickel	2490	2500	100	90 - 110	P	07/17/2025	18:45	LB136534
	Potassium	23500	25000	94	90 - 110	P	07/17/2025	18:45	LB136534
	Selenium	4950	5000	99	90 - 110	P	07/17/2025	18:45	LB136534
	Silver	1240	1250	99	90 - 110	P	07/17/2025	18:45	LB136534
	Sodium	24000	25000	96	90 - 110	P	07/17/2025	18:45	LB136534
	Thallium	4990	5000	100	90 - 110	P	07/17/2025	18:45	LB136534
	Vanadium	2470	2500	99	90 - 110	P	07/17/2025	18:45	LB136534
	Zinc	2490	2500	100	90 - 110	P	07/17/2025	18:45	LB136534
CCV05	Aluminum	10100	10000	101	90 - 110	P	07/17/2025	20:11	LB136534
	Antimony	4990	5000	100	90 - 110	P	07/17/2025	20:11	LB136534
	Arsenic	5000	5000	100	90 - 110	P	07/17/2025	20:11	LB136534
	Barium	10300	10000	103	90 - 110	P	07/17/2025	20:11	LB136534
	Beryllium	247	250	99	90 - 110	P	07/17/2025	20:11	LB136534
	Cadmium	2550	2500	102	90 - 110	P	07/17/2025	20:11	LB136534
	Calcium	25200	25000	101	90 - 110	P	07/17/2025	20:11	LB136534
	Chromium	1010	1000	101	90 - 110	P	07/17/2025	20:11	LB136534
	Cobalt	2510	2500	100	90 - 110	P	07/17/2025	20:11	LB136534
	Copper	1240	1250	99	90 - 110	P	07/17/2025	20:11	LB136534
	Iron	5000	5000	100	90 - 110	P	07/17/2025	20:11	LB136534
	Lead	5070	5000	101	90 - 110	P	07/17/2025	20:11	LB136534
	Magnesium	24600	25000	99	90 - 110	P	07/17/2025	20:11	LB136534
	Manganese	2470	2500	99	90 - 110	P	07/17/2025	20:11	LB136534

Metals

- 2a -

INITIAL AND CONTINUING CALIBRATION VERIFICATION

Client: ENVOCARE Environmental Facility Managemer

SDG No.: Q2515

Contract: ENVO01

Lab Code: ACE

Initial Calibration Source: EPA

Continuing Calibration Source: Inorganic Ventures

Sample ID	Analyte	Result ug/L	True Value	% Recovery	Acceptance Window (%R)	M	Analysis Date	Analysis Time	Run Number
CCV05	Nickel	2520	2500	101	90 - 110	P	07/17/2025	20:11	LB136534
	Potassium	25100	25000	100	90 - 110	P	07/17/2025	20:11	LB136534
	Selenium	5040	5000	101	90 - 110	P	07/17/2025	20:11	LB136534
	Silver	1250	1250	100	90 - 110	P	07/17/2025	20:11	LB136534
	Sodium	25500	25000	102	90 - 110	P	07/17/2025	20:11	LB136534
	Thallium	5070	5000	101	90 - 110	P	07/17/2025	20:11	LB136534
	Vanadium	2490	2500	100	90 - 110	P	07/17/2025	20:11	LB136534
	Zinc	2500	2500	100	90 - 110	P	07/17/2025	20:11	LB136534
CCV06	Aluminum	10300	10000	103	90 - 110	P	07/17/2025	20:46	LB136534
	Antimony	4990	5000	100	90 - 110	P	07/17/2025	20:46	LB136534
	Arsenic	4910	5000	98	90 - 110	P	07/17/2025	20:46	LB136534
	Barium	10400	10000	104	90 - 110	P	07/17/2025	20:46	LB136534
	Beryllium	252	250	101	90 - 110	P	07/17/2025	20:46	LB136534
	Cadmium	2500	2500	100	90 - 110	P	07/17/2025	20:46	LB136534
	Calcium	25300	25000	101	90 - 110	P	07/17/2025	20:46	LB136534
	Chromium	1000	1000	100	90 - 110	P	07/17/2025	20:46	LB136534
	Cobalt	2480	2500	99	90 - 110	P	07/17/2025	20:46	LB136534
	Copper	1260	1250	101	90 - 110	P	07/17/2025	20:46	LB136534
	Iron	5150	5000	103	90 - 110	P	07/17/2025	20:46	LB136534
	Lead	4960	5000	99	90 - 110	P	07/17/2025	20:46	LB136534
	Magnesium	24900	25000	100	90 - 110	P	07/17/2025	20:46	LB136534
	Manganese	2490	2500	99	90 - 110	P	07/17/2025	20:46	LB136534
	Nickel	2490	2500	100	90 - 110	P	07/17/2025	20:46	LB136534
	Potassium	26800	25000	107	90 - 110	P	07/17/2025	20:46	LB136534
	Selenium	4960	5000	99	90 - 110	P	07/17/2025	20:46	LB136534
	Silver	1240	1250	99	90 - 110	P	07/17/2025	20:46	LB136534
	Sodium	25900	25000	104	90 - 110	P	07/17/2025	20:46	LB136534
	Thallium	5020	5000	100	90 - 110	P	07/17/2025	20:46	LB136534
	Vanadium	2530	2500	101	90 - 110	P	07/17/2025	20:46	LB136534
	Zinc	2510	2500	100	90 - 110	P	07/17/2025	20:46	LB136534