

Metals

- 7 -

LABORATORY CONTROL SAMPLE SUMMARY

Client: ENVOCARE Environmental Facility Mana

SDG No.: Q2515

Contract: ENVO01

Lab Code: ACE

Analyte	Units	True Value	Result	C	% Recovery	Acceptance Limits	M
PB168740BS							
Aluminum	mg/Kg	100	98.9		99	80 - 120	P
Antimony	mg/Kg	40.0	40.0		100	80 - 120	P
Arsenic	mg/Kg	40.0	35.6		89	80 - 120	P
Barium	mg/Kg	10.0	9.99		100	80 - 120	P
Beryllium	mg/Kg	10.0	10.7		107	80 - 120	P
Cadmium	mg/Kg	10.0	9.41		94	80 - 120	P
Calcium	mg/Kg	50.0	51.1	J	102	80 - 120	P
Chromium	mg/Kg	20.0	20.3		102	80 - 120	P
Cobalt	mg/Kg	10.0	9.53		95	80 - 120	P
Copper	mg/Kg	15.0	16.3		109	80 - 120	P
Iron	mg/Kg	150	170		113	80 - 120	P
Lead	mg/Kg	50.0	47.1		94	80 - 120	P
Magnesium	mg/Kg	100	101		101	80 - 120	P
Manganese	mg/Kg	10.0	10.4		104	80 - 120	P
Nickel	mg/Kg	25.0	24.0		96	80 - 120	P
Potassium	mg/Kg	500	477		95	80 - 120	P
Selenium	mg/Kg	100	98.5		98	80 - 120	P
Silver	mg/Kg	3.8	3.56		94	80 - 120	P
Sodium	mg/Kg	150	157		105	80 - 120	P
Thallium	mg/Kg	100	99.9		100	80 - 120	P
Vanadium	mg/Kg	15.0	14.8		99	80 - 120	P
Zinc	mg/Kg	10.0	10.3		103	80 - 120	P

Metals

- 7 -

LABORATORY CONTROL SAMPLE SUMMARY

Client: ENVOCARE Environmental Facility Mana

SDG No.: Q2515

Contract: ENVO01

Lab Code: ACE

Analyte	Units	True Value	Result	C	% Recovery	Acceptance Limits	M
PB168743BS Mercury	mg/Kg	0.26	0.27		102	80 - 120	CV