

Instrument ID: FID_G

Daily Analysis Runlog For Sequence/QC Batch ID # FG061325AL

Review By	yogesh	Review On	6/13/2025 12:52:17 PM
Supervise By	mohammad	Supervise On	6/17/2025 3:40:36 AM
SubDirectory	FG061325AL	HP Acquire Method	HP Processing Method FG061325AL

STD. NAME	STD REF.#
Tune/Reschk Initial Calibration Stds	PP24170,PP24175,PP24176,PP24177,PP24178
CCC Internal Standard/PEM	PP24176
ICV/I.BLK Surrogate Standard MS/MSD Standard LCS Standard	PP24174,PP24179

Sr#	SampleID	ClientID	Data File Name	Date-Time	Comment	Operator	Status
1	MECL2	MECL2	FG016069.D	13 Jun 2025 13:06		YP\AJ	Ok
2	I.BLK	I.BLK	FG016070.D	13 Jun 2025 13:35		YP\AJ	Ok
3	100 PPM ALIPHATIC HC	100 PPM ALIPHATIC HC	FG016071.D	13 Jun 2025 14:05		YP\AJ	Ok
4	50 PPM ALIPHATIC HC	50 PPM ALIPHATIC HC	FG016072.D	13 Jun 2025 14:34		YP\AJ	Ok
5	20 PPM ALIPHATIC HC	20 PPM ALIPHATIC HC	FG016073.D	13 Jun 2025 15:03		YP\AJ	Ok
6	10 PPM ALIPHATIC HC	10 PPM ALIPHATIC HC	FG016074.D	13 Jun 2025 15:33		YP\AJ	Ok
7	5 PPM ALIPHATIC HC	5 PPM ALIPHATIC HC	FG016075.D	13 Jun 2025 16:02		YP\AJ	Ok
8	20 PPM ALIPHATIC HC	20 PPM ALIPHATIC HC	FG016076.D	13 Jun 2025 16:31		YP\AJ	Ok
9	I.BLK	I.BLK	FG016077.D	13 Jun 2025 17:30		YP\AJ	Ok
10	20 PPM ALIPHATIC HC	20 PPM ALIPHATIC HC	FG016078.D	13 Jun 2025 18:00		YP\AJ	Ok

M : Manual Integration

Instrument ID: FID_G

Daily Analysis Runlog For Sequence/QC Batch ID # FG070725AL

Review By	yogesh	Review On	7/7/2025 12:23:40 PM
Supervise By	mohammad	Supervise On	7/9/2025 1:51:09 AM
SubDirectory	FG070725AL	HP Acquire Method	HP Processing Method FG061325AL

STD. NAME	STD REF.#
Tune/Reschk Initial Calibration Stds	PP24170,PP24175,PP24176,PP24177,PP24178
CCC Internal Standard/PEM ICV/I.BLK Surrogate Standard MS/MSD Standard LCS Standard	PP24176 PP24174,PP24179

Sr#	SampleID	ClientID	Data File Name	Date-Time	Comment	Operator	Status
1	MECL2	MECL2	FG016224.D	07 Jul 2025 09:19		YP\AJ	Ok
2	I.BLK	I.BLK	FG016225.D	07 Jul 2025 09:49		YP\AJ	Ok
3	20 PPM ALIPHATIC HC	20 PPM ALIPHATIC HC	FG016226.D	07 Jul 2025 10:55		YP\AJ	Ok,M
4	PB168738BL	PB168738BL	FG016227.D	07 Jul 2025 15:02		YP\AJ	Ok
5	PB168738BS	PB168738BS	FG016228.D	07 Jul 2025 15:32	Recovery fail	YP\AJ	Not Ok
6	PB168738BSD	PB168738BSD	FG016229.D	07 Jul 2025 16:02	Recovery fail	YP\AJ	Not Ok
7	Q2515-01	wc-1	FG016230.D	07 Jul 2025 16:32		YP\AJ	Ok
8	Q2515-01D	Q2515-01D	FG016231.D	07 Jul 2025 17:01		YP\AJ	Ok
9	Q2515-01MS	wc-1MS	FG016232.D	07 Jul 2025 17:31	FG016230.D	YP\AJ	Ok
10	Q2515-01MSD	wc-1MSD	FG016233.D	07 Jul 2025 18:01	FG016230.D\FG016232.D	YP\AJ	Ok
11	I.BLK	I.BLK	FG016234.D	07 Jul 2025 19:01		YP\AJ	Ok
12	20 PPM ALIPHATIC HC	20 PPM ALIPHATIC HC	FG016235.D	07 Jul 2025 19:30		YP\AJ	Ok

M : Manual Integration

Instrument ID: FID_G

Daily Analysis Runlog For Sequence/QC Batch ID # FG070825AL

Review By	yogesh	Review On	7/8/2025 1:12:39 PM
Supervise By		Supervise On	
SubDirectory	FG070825AL	HP Acquire Method	HP Processing Method FG061325AL
STD. NAME	STD REF.#		
Tune/Reschk Initial Calibration Stds	PP24170,PP24175,PP24176,PP24177,PP24178		
CCC	PP24176		
Internal Standard/PEM			
ICV/I.BLK	PP24174,PP24179		
Surrogate Standard			
MS/MSD Standard			
LCS Standard			

Sr#	SampleId	ClientID	Data File Name	Date-Time	Comment	Operator	Status
1	MECL2	MECL2	FG016236.D	08 Jul 2025 09:18		YP\AJ	Ok
2	I.BLK	I.BLK	FG016237.D	08 Jul 2025 09:48		YP\AJ	Ok
3	20 PPM ALIPHATIC HC	20 PPM ALIPHATIC HC	FG016238.D	08 Jul 2025 10:17		YP\AJ	Ok,NS
4	PB168738BS	PB168738BS	FG016239.D	08 Jul 2025 10:54		YP\AJ	Ok
5	PB168738BSD	PB168738BSD	FG016240.D	08 Jul 2025 11:25		YP\AJ	Ok
6	I.BLK	I.BLK	FG016241.D	08 Jul 2025 11:55		YP\AJ	Ok
7	20 PPM ALIPHATIC HC	20 PPM ALIPHATIC HC	FG016242.D	08 Jul 2025 12:54		YP\AJ	Ok

M : Manual Integration