

### CALIBRATION VERIFICATION SUMMARY

Lab Name: Alliance

Contract: ENVO01

Lab Code: ACE

SDG NO.: Q2515

Continuing Calib Date: 07/08/2025

Initial Calibration Date(s): 06/11/2025

06/11/2025

Continuing Calib Time: 04:30

Initial Calibration Time(s): 10:40

19:07

GC Column: ZB-MR1 ID: 0.32 (mm)

| COMPOUND             | CCAL<br>RT | AVG<br>RT | RT WINDOW |      | DIFF<br>RT |
|----------------------|------------|-----------|-----------|------|------------|
|                      |            |           | FROM      | TO   |            |
| Aroclor-1248-1 (1)   | 4.76       | 4.76      | 4.66      | 4.86 | 0.00       |
| Aroclor-1248-2 (2)   | 5.00       | 5.00      | 4.90      | 5.10 | 0.00       |
| Aroclor-1248-3 (3)   | 5.21       | 5.22      | 5.12      | 5.32 | 0.01       |
| Aroclor-1248-4 (4)   | 5.57       | 5.57      | 5.47      | 5.67 | 0.00       |
| Aroclor-1248-5 (5)   | 5.61       | 5.61      | 5.51      | 5.71 | 0.00       |
| Tetrachloro-m-xylene | 3.68       | 3.68      | 3.58      | 3.78 | 0.01       |
| Decachlorobiphenyl   | 8.71       | 8.71      | 8.61      | 8.81 | 0.00       |

### CALIBRATION VERIFICATION SUMMARY

Lab Name: Alliance

Contract: ENVO01

Lab Code: ACE

SDG NO.: Q2515

Continuing Calib Date: 07/08/2025

Initial Calibration Date(s): 06/11/2025

06/11/2025

Continuing Calib Time: 04:30

Initial Calibration Time(s): 10:40

19:07

GC Column: ZB-MR2 ID: 0.32 (mm)

| COMPOUND             | CCAL<br>RT | AVG<br>RT | RT WINDOW |      | DIFF<br>RT |
|----------------------|------------|-----------|-----------|------|------------|
|                      |            |           | FROM      | TO   |            |
| Aroclor-1248-1 (1)   | 4.74       | 4.75      | 4.65      | 4.85 | 0.01       |
| Aroclor-1248-2 (2)   | 4.98       | 4.99      | 4.89      | 5.09 | 0.01       |
| Aroclor-1248-3 (3)   | 5.02       | 5.03      | 4.93      | 5.13 | 0.01       |
| Aroclor-1248-4 (4)   | 5.19       | 5.20      | 5.10      | 5.30 | 0.01       |
| Aroclor-1248-5 (5)   | 5.58       | 5.59      | 5.49      | 5.69 | 0.01       |
| Tetrachloro-m-xylene | 3.67       | 3.67      | 3.57      | 3.77 | 0.00       |
| Decachlorobiphenyl   | 8.65       | 8.66      | 8.56      | 8.76 | 0.01       |