

CALIBRATION VERIFICATION SUMMARY

Lab Name: Alliance

Contract: ENVO01

Lab Code: ACE

SDG NO.: Q2515

Continuing Calib Date: 07/08/2025

Initial Calibration Date(s): 06/11/2025 06/11/2025

Continuing Calib Time: 10:46

Initial Calibration Time(s): 10:40 19:07

GC Column: ZB-MR1 ID: 0.32 (mm)

COMPOUND	CCAL RT	AVG RT	RT WINDOW		DIFF RT
			FROM	TO	
Aroclor-1242-1 (1)	4.76	4.77	4.67	4.87	0.01
Aroclor-1242-2 (2)	4.78	4.78	4.68	4.88	0.00
Aroclor-1242-3 (3)	4.84	4.84	4.74	4.94	0.00
Aroclor-1242-4 (4)	4.96	4.96	4.86	5.06	0.00
Aroclor-1242-5 (5)	5.61	5.61	5.51	5.71	0.00
Tetrachloro-m-xylene	3.67	3.68	3.58	3.78	0.01
Decachlorobiphenyl	8.70	8.71	8.61	8.81	0.01

CALIBRATION VERIFICATION SUMMARY

Lab Name: Alliance

Contract: ENVO01

Lab Code: ACE

SDG NO.: Q2515

Continuing Calib Date: 07/08/2025

Initial Calibration Date(s): 06/11/2025

06/11/2025

Continuing Calib Time: 10:46

Initial Calibration Time(s): 10:40

19:07

GC Column: ZB-MR2 ID: 0.32 (mm)

COMPOUND	CCAL RT	AVG RT	RT WINDOW		DIFF RT
			FROM	TO	
Aroclor-1242-1 (1)	4.74	4.75	4.65	4.85	0.01
Aroclor-1242-2 (2)	4.76	4.77	4.67	4.87	0.01
Aroclor-1242-3 (3)	4.94	4.94	4.84	5.04	0.00
Aroclor-1242-4 (4)	5.02	5.03	4.93	5.13	0.01
Aroclor-1242-5 (5)	5.54	5.55	5.45	5.65	0.01
Tetrachloro-m-xylene	3.67	3.67	3.57	3.77	0.00
Decachlorobiphenyl	8.65	8.66	8.56	8.76	0.01