

CALIBRATION VERIFICATION SUMMARY

Lab Name: Alliance **Contract:** ENVO01
Lab Code: ACE **SDG NO.:** Q2515
GC Column: ZB-MR1 **ID:** 0.32 (mm) **Initi. Calib. Date(s):** 06/11/2025 06/11/2025

Client Sample No.: CCAL01 **Date Analyzed:** 07/07/2025
Lab Sample No.: AR1660CCC500 **Data File :** PO112052.D **Time Analyzed:** 20:15

COMPOUND	RT	RT WINDOW		CALC AMOUNT(ng)	NOM AMOUNT(ng)	%D
		FROM	TO			
Aroclor-1016-1	4.762	4.666	4.866	451.720	500.000	-9.7
Aroclor-1016-2	4.781	4.684	4.884	468.410	500.000	-6.3
Aroclor-1016-3	4.838	4.741	4.941	442.310	500.000	-11.5
Aroclor-1016-4	4.957	4.861	5.061	447.560	500.000	-10.5
Aroclor-1016-5	5.214	5.118	5.318	471.760	500.000	-5.6
Aroclor-1260-1	6.253	6.157	6.357	457.590	500.000	-8.5
Aroclor-1260-2	6.441	6.346	6.546	508.720	500.000	1.7
Aroclor-1260-3	6.807	6.713	6.913	532.640	500.000	6.5
Aroclor-1260-4	7.067	6.972	7.172	520.660	500.000	4.1
Aroclor-1260-5	7.310	7.216	7.416	515.130	500.000	3.0
Decachlorobiphenyl	8.704	8.612	8.812	41.270	50.000	-17.5
Tetrachloro-m-xylene	3.675	3.577	3.777	51.660	50.000	3.3

CALIBRATION VERIFICATION SUMMARY

Lab Name: Alliance **Contract:** ENVO01
Lab Code: ACE **SDG NO.:** Q2515
GC Column: ZB-MR2 **ID:** 0.32 (mm) **Initi. Calib. Date(s):** 06/11/2025 06/11/2025

Client Sample No.: CCAL01 **Date Analyzed:** 07/07/2025
Lab Sample No.: AR1660CCC500 **Data File :** PO112052.D **Time Analyzed:** 20:15

COMPOUND	RT	RT WINDOW		CALC AMOUNT(ng)	NOM AMOUNT(ng)	%D
		FROM	TO			
Aroclor-1016-1	4.745	4.651	4.851	505.180	500.000	1.0
Aroclor-1016-2	4.764	4.669	4.869	509.460	500.000	1.9
Aroclor-1016-3	4.939	4.844	5.044	498.340	500.000	-0.3
Aroclor-1016-4	4.981	4.887	5.087	498.070	500.000	-0.4
Aroclor-1016-5	5.194	5.099	5.299	492.500	500.000	-1.5
Aroclor-1260-1	6.223	6.129	6.329	485.010	500.000	-3.0
Aroclor-1260-2	6.411	6.316	6.516	504.260	500.000	0.9
Aroclor-1260-3	6.562	6.469	6.669	466.740	500.000	-6.7
Aroclor-1260-4	7.032	6.939	7.139	450.960	500.000	-9.8
Aroclor-1260-5	7.274	7.181	7.381	459.050	500.000	-8.2
Decachlorobiphenyl	8.651	8.561	8.761	44.280	50.000	-11.4
Tetrachloro-m-xylene	3.669	3.573	3.773	50.200	50.000	0.4