

## Report of Analysis

|                    |                                                    |          |                    |          |           |
|--------------------|----------------------------------------------------|----------|--------------------|----------|-----------|
| Client:            | ENVOCARE Environmental Facility Management dba UAV |          | Date Collected:    | 07/03/25 |           |
| Project:           | MV Trucking                                        |          | Date Received:     | 07/03/25 |           |
| Client Sample ID:  | WC-1                                               |          | SDG No.:           | Q2515    |           |
| Lab Sample ID:     | Q2515-01                                           |          | Matrix:            | SOIL     |           |
| Analytical Method: | 8082A                                              |          | % Solid:           | 87.5     | Decanted: |
| Sample Wt/Vol:     | 30.07                                              | Units: g | Final Vol:         | 10000    | uL        |
| Soil Aliquot Vol:  |                                                    | uL       | Test:              | PCB      |           |
| Extraction Type:   |                                                    |          | Injection Volume : |          |           |
| GPC Factor :       | 1.0                                                | PH :     |                    |          |           |
| Prep Method :      | SW3541B                                            |          |                    |          |           |

|                   |           |                |                |               |
|-------------------|-----------|----------------|----------------|---------------|
| File ID/Qc Batch: | Dilution: | Prep Date      | Date Analyzed  | Prep Batch ID |
| PO112081.D        | 1         | 07/07/25 08:30 | 07/08/25 08:11 | PB168735      |

| CAS Number        | Parameter            | Conc. | Qualifier | MDL      | LOQ / CRQL | Units(Dry Weight) |
|-------------------|----------------------|-------|-----------|----------|------------|-------------------|
| <b>TARGETS</b>    |                      |       |           |          |            |                   |
| 12674-11-2        | Aroclor-1016         | 4.50  | U         | 4.50     | 19.4       | ug/kg             |
| 11104-28-2        | Aroclor-1221         | 4.60  | U         | 4.60     | 19.4       | ug/kg             |
| 11141-16-5        | Aroclor-1232         | 4.20  | U         | 4.20     | 19.4       | ug/kg             |
| 53469-21-9        | Aroclor-1242         | 4.60  | U         | 4.60     | 19.4       | ug/kg             |
| 12672-29-6        | Aroclor-1248         | 6.70  | U         | 6.70     | 19.4       | ug/kg             |
| 11097-69-1        | Aroclor-1254         | 3.70  | U         | 3.70     | 19.4       | ug/kg             |
| 37324-23-5        | Aroclor-1262         | 5.70  | U         | 5.70     | 19.4       | ug/kg             |
| 11100-14-4        | Aroclor-1268         | 4.10  | U         | 4.10     | 19.4       | ug/kg             |
| 11096-82-5        | Aroclor-1260         | 3.70  | U         | 3.70     | 19.4       | ug/kg             |
| <b>SURROGATES</b> |                      |       |           |          |            |                   |
| 877-09-8          | Tetrachloro-m-xylene | 15.4  |           | 32 - 144 | 77%        | SPK: 20           |
| 2051-24-3         | Decachlorobiphenyl   | 13.9  |           | 32 - 175 | 70%        | SPK: 20           |

### Comments:

U = Not Detected  
 LOQ = Limit of Quantitation  
 MDL = Method Detection Limit  
 LOD = Limit of Detection  
 E = Value Exceeds Calibration Range  
 P = Indicates >25% difference for detected concentrations between the two GC columns  
 Q = indicates LCS control criteria did not meet requirements  
 M = MS/MSD acceptance criteria did not meet requirements

J = Estimated Value  
 B = Analyte Found in Associated Method Blank  
 N = Presumptive Evidence of a Compound  
 \* = Values outside of QC limits  
 D = Dilution  
 S = Indicates estimated value where valid five-point calibration was not performed prior to analyte detection in sample.  
 () = Laboratory InHouse Limit