

CALIBRATION VERIFICATION SUMMARY

Lab Name: Alliance **Contract:** ENVO01
Lab Code: ACE **SDG NO.:** Q2515
GC Column: ZB-MR1 **ID:** 0.32 (mm) **Initi. Calib. Date(s):** 06/17/2025 06/17/2025

Client Sample No.: CCAL03 **Date Analyzed:** 07/07/2025
Lab Sample No.: PSTDCCC050 **Data File :** PD089369.D **Time Analyzed:** 19:59

COMPOUND	RT	RT WINDOW		CALC AMOUNT(ng)	NOM AMOUNT(ng)	%D
		FROM	TO			
4,4'-DDD	6.704	6.605	6.805	49.530	50.000	-0.9
4,4'-DDE	6.195	6.096	6.296	51.310	50.000	2.6
4,4'-DDT	7.020	6.921	7.121	46.530	50.000	-6.9
Aldrin	5.270	5.171	5.371	55.470	50.000	10.9
alpha-BHC	3.999	3.899	4.099	57.030	50.000	14.1
alpha-Chlordane	6.026	5.927	6.127	51.570	50.000	3.1
beta-BHC	4.515	4.415	4.615	54.330	50.000	8.7
Decachlorobiphenyl	9.071	8.972	9.172	42.700	50.000	-14.6
delta-BHC	4.763	4.664	4.864	59.780	50.000	19.6
Dieldrin	6.346	6.247	6.447	52.020	50.000	4.0
Endosulfan I	6.073	5.975	6.175	51.910	50.000	3.8
Endosulfan II	6.785	6.686	6.886	49.910	50.000	-0.2
Endosulfan sulfate	7.148	7.049	7.249	46.690	50.000	-6.6
Endrin	6.573	6.475	6.675	50.500	50.000	1.0
Endrin aldehyde	6.913	6.815	7.015	47.100	50.000	-5.8
Endrin ketone	7.629	7.530	7.730	45.350	50.000	-9.3
gamma-BHC (Lindane)	4.330	4.230	4.430	56.620	50.000	13.2
gamma-Chlordane	5.945	5.846	6.046	52.480	50.000	5.0
Heptachlor	4.929	4.829	5.029	56.020	50.000	12.0
Heptachlor epoxide	5.690	5.591	5.791	52.430	50.000	4.9
Methoxychlor	7.492	7.393	7.593	44.220	50.000	-11.6
Tetrachloro-m-xylene	3.550	3.450	3.650	54.050	50.000	8.1

CALIBRATION VERIFICATION SUMMARY

Lab Name: Alliance **Contract:** ENVO01
Lab Code: ACE **SDG NO.:** Q2515
GC Column: ZB-MR2 **ID:** 0.32 (mm) **Initi. Calib. Date(s):** 06/17/2025 06/17/2025

Client Sample No.: CCAL03 **Date Analyzed:** 07/07/2025
Lab Sample No.: PSTDCCC050 **Data File :** PD089369.D **Time Analyzed:** 19:59

COMPOUND	RT	RT WINDOW		CALC AMOUNT(ng)	NOM AMOUNT(ng)	%D
		FROM	TO			
4,4'-DDD	5.929	5.830	6.030	51.220	50.000	2.4
4,4'-DDE	5.374	5.275	5.475	56.600	50.000	13.2
4,4'-DDT	6.183	6.084	6.284	52.240	50.000	4.5
Aldrin	4.368	4.269	4.469	58.350	50.000	16.7
alpha-BHC	3.393	3.293	3.493	58.820	50.000	17.6
alpha-Chlordane	5.189	5.091	5.291	58.480	50.000	17.0
beta-BHC	4.025	3.926	4.126	56.960	50.000	13.9
Decachlorobiphenyl	8.070	7.972	8.172	40.250	50.000	-19.5
delta-BHC	4.261	4.162	4.362	58.390	50.000	16.8
Dieldrin	5.512	5.413	5.613	55.870	50.000	11.7
Endosulfan I	5.246	5.147	5.347	59.210	50.000	18.4
Endosulfan II	6.079	5.981	6.181	52.960	50.000	5.9
Endosulfan sulfate	6.481	6.383	6.583	51.740	50.000	3.5
Endrin	5.788	5.689	5.889	54.820	50.000	9.6
Endrin aldehyde	6.258	6.159	6.359	50.860	50.000	1.7
Endrin ketone	6.990	6.892	7.092	47.030	50.000	-5.9
gamma-BHC (Lindane)	3.728	3.630	3.830	58.890	50.000	17.8
gamma-Chlordane	5.123	5.026	5.226	56.390	50.000	12.8
Heptachlor	4.082	3.983	4.183	57.740	50.000	15.5
Heptachlor epoxide	4.870	4.773	4.973	58.020	50.000	16.0
Methoxychlor	6.753	6.655	6.855	46.840	50.000	-6.3
Tetrachloro-m-xylene	2.881	2.780	2.980	57.660	50.000	15.3