

Data Path : Z:\svoasrv\HPCHEM1\BNA_F\Data\BF061925\
 Data File : BF142796.D
 Acq On : 19 Jun 2025 21:10
 Operator : RC/JU
 Sample : SSTDICV040
 Misc :
 ALS Vial : 10 Sample Multiplier: 1

Instrument :
 BNA_F
 ClientSampleId :
 ICVBF061925

Quant Time: Jun 24 15:59:56 2025
 Quant Method : Z:\svoasrv\HPCHEM1\BNA_F\Methods\8270-BF061925.M
 Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
 QLast Update : Tue Jun 24 05:28:21 2025
 Response via : Initial Calibration

Min. RRF : 0.000 Min. Rel. Area : 50% Max. R.T. Dev 0.50min
 Max. RRF Dev : 25% Max. Rel. Area : 150%

	Compound	Amount	Calc.	%Dev	Area%	Dev(min)
1 I	1,4-Dichlorobenzene-d4	20.000	20.000	0.0	89	0.00
2	1,4-Dioxane	40.000	39.357	1.6	91	0.00
3	Pyridine	40.000	41.446	-3.6	95	0.00
4	n-Nitrosodimethylamine	40.000	41.912	-4.8	97	0.00
5 S	2-Fluorophenol	80.000	81.192	-1.5	94	0.00
6	Aniline	40.000	42.392	-6.0	97	0.00
7 S	Phenol-d6	80.000	80.262	-0.3	94	0.00
8	2-Chlorophenol	40.000	42.036	-5.1	97	0.00
9	Benzaldehyde	40.000	47.511	-18.8	117	0.00
10 C	Phenol	40.000	41.901	-4.8	96	0.00
11	bis(2-Chloroethyl)ether	40.000	42.157	-5.4	98	0.00
12	1,3-Dichlorobenzene	40.000	42.568	-6.4	99	0.00
13 C	1,4-Dichlorobenzene	40.000	42.268	-5.7	97	0.00
14	1,2-Dichlorobenzene	40.000	42.520	-6.3	98	0.00
15	Benzyl Alcohol	40.000	42.114	-5.3	98	0.00
16	2,2'-oxybis(1-Chloropropane	40.000	42.379	-5.9	99	0.00
17	2-Methylphenol	40.000	42.485	-6.2	98	0.00
18	Hexachloroethane	40.000	42.589	-6.5	97	0.00
19 P	n-Nitroso-di-n-propylamine	40.000	41.640	-4.1	97	0.00
20	3+4-Methylphenols	40.000	42.904	-7.3	99	0.00
21 I	Naphthalene-d8	20.000	20.000	0.0	89	0.00
22	Acetophenone	40.000	42.478	-6.2	100	0.00
23 S	Nitrobenzene-d5	80.000	73.334	8.3	84	0.00
24	Nitrobenzene	40.000	41.968	-4.9	97	0.00
25	Isophorone	40.000	41.403	-3.5	97	0.00
26 C	2-Nitrophenol	40.000	43.224	-8.1	97	0.00
27	2,4-Dimethylphenol	40.000	42.168	-5.4	98	0.00
28	bis(2-Chloroethoxy)methane	40.000	41.960	-4.9	98	0.00
29 C	2,4-Dichlorophenol	40.000	42.196	-5.5	97	0.00
30	1,2,4-Trichlorobenzene	40.000	42.177	-5.4	98	0.00
31	Naphthalene	40.000	42.218	-5.5	99	0.00
32	Benzoic acid	40.000	45.685	-14.2	103	0.00
33	4-Chloroaniline	40.000	41.766	-4.4	98	0.00
34 C	Hexachlorobutadiene	40.000	42.657	-6.6	98	0.00
35	Caprolactam	40.000	40.928	-2.3	95	0.00
36 C	4-Chloro-3-methylphenol	40.000	41.890	-4.7	97	0.00
37	2-Methylnaphthalene	40.000	41.730	-4.3	97	0.00
38	1-Methylnaphthalene	40.000	41.554	-3.9	97	0.00
39 I	Acenaphthene-d10	20.000	20.000	0.0	88	0.00
40	1,2,4,5-Tetrachlorobenzene	40.000	43.233	-8.1	101	0.00
41 P	Hexachlorocyclopentadiene	40.000	45.361	-13.4	100	0.00
42 S	2,4,6-Tribromophenol	80.000	80.065	-0.1	91	0.00
43 C	2,4,6-Trichlorophenol	40.000	42.543	-6.4	99	0.00
44	2,4,5-Trichlorophenol	40.000	43.045	-7.6	98	0.00
45 S	2-Fluorobiphenyl	80.000	72.171	9.8	85	0.00
46	1,1'-Biphenyl	40.000	42.159	-5.4	98	0.00
47	2-Chloronaphthalene	40.000	41.817	-4.5	97	0.00

Data Path : Z:\svoasrv\HPCHEM1\BNA_F\Data\BF061925\
 Data File : BF142796.D
 Acq On : 19 Jun 2025 21:10
 Operator : RC/JU
 Sample : SSTDICV040
 Misc :
 ALS Vial : 10 Sample Multiplier: 1

Instrument :
 BNA_F
 ClientSampleId :
 ICVBF061925

Quant Time: Jun 24 15:59:56 2025
 Quant Method : Z:\svoasrv\HPCHEM1\BNA_F\Methods\8270-BF061925.M
 Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
 QLast Update : Tue Jun 24 05:28:21 2025
 Response via : Initial Calibration

Min. RRF : 0.000 Min. Rel. Area : 50% Max. R.T. Dev 0.50min
 Max. RRF Dev : 25% Max. Rel. Area : 150%

	Compound	Amount	Calc.	%Dev	Area%	Dev(min)
48	2-Nitroaniline	40.000	41.724	-4.3	94	0.00
49	Acenaphthylene	40.000	41.586	-4.0	96	0.00
50	Dimethylphthalate	40.000	41.232	-3.1	96	0.00
51	2,6-Dinitrotoluene	40.000	41.386	-3.5	94	0.00
52 C	Acenaphthene	40.000	41.825	-4.6	96	0.00
53	3-Nitroaniline	40.000	41.512	-3.8	95	0.00
54 P	2,4-Dinitrophenol	40.000	43.742	-9.4	97	0.00
55	Dibenzofuran	40.000	41.150	-2.9	95	0.00
56 P	4-Nitrophenol	40.000	40.556	-1.4	92	0.00
57	2,4-Dinitrotoluene	40.000	41.358	-3.4	94	0.00
58	Fluorene	40.000	41.052	-2.6	96	0.00
59	2,3,4,6-Tetrachlorophenol	40.000	41.857	-4.6	97	0.00
60	Diethylphthalate	40.000	40.937	-2.3	95	0.00
61	4-Chlorophenyl-phenylether	40.000	40.862	-2.2	96	0.00
62	4-Nitroaniline	40.000	40.628	-1.6	94	0.00
63	Azobenzene	40.000	41.025	-2.6	95	0.00
64 I	Phenanthrene-d10	20.000	20.000	0.0	87	0.00
65	4,6-Dinitro-2-methylphenol	40.000	43.341	-8.4	95	0.00
66 c	n-Nitrosodiphenylamine	40.000	41.955	-4.9	94	0.00
67	4-Bromophenyl-phenylether	40.000	43.376	-8.4	98	0.00
68	Hexachlorobenzene	40.000	42.632	-6.6	97	0.00
69	Atrazine	40.000	42.840	-7.1	96	0.00
70 C	Pentachlorophenol	40.000	43.337	-8.3	94	0.00
71	Phenanthrene	40.000	41.603	-4.0	95	0.00
72	Anthracene	40.000	42.013	-5.0	95	0.00
73	Carbazole	40.000	40.858	-2.1	94	0.00
74	Di-n-butylphthalate	40.000	42.228	-5.6	97	0.00
75 C	Fluoranthene	40.000	40.372	-0.9	94	0.00
76 I	Chrysene-d12	20.000	20.000	0.0	86	0.00
77	Benzydine	40.000	47.530	-18.8	99	0.00
78	Pyrene	40.000	40.794	-2.0	94	0.00
79 S	Terphenyl-d14	80.000	70.839	11.5	83	0.00
80	Butylbenzylphthalate	40.000	46.431	-16.1	99	0.00
81	Benzo(a)anthracene	40.000	41.868	-4.7	96	0.00
82	3,3'-Dichlorobenzidine	40.000	46.521	-16.3	102	0.00
83	Chrysene	40.000	43.637	-9.1	97	0.00
84	Bis(2-ethylhexyl)phthalate	40.000	47.976	-19.9	100	0.00
85 c	Di-n-octyl phthalate	40.000	47.026	-17.6	102	0.00
86 I	Perylene-d12	20.000	20.000	0.0	90	0.00
87	Indeno(1,2,3-cd)pyrene	40.000	42.060	-5.2	98	0.00
88	Benzo(b)fluoranthene	40.000	40.801	-2.0	99	0.00
89	Benzo(k)fluoranthene	40.000	41.800	-4.5	98	0.00
90 C	Benzo(a)pyrene	40.000	42.117	-5.3	99	0.00
91	Dibenzo(a,h)anthracene	40.000	42.069	-5.2	99	0.00
92	Benzo(g,h,i)perylene	40.000	42.006	-5.0	98	0.00

Data Path : Z:\svoasrv\HPCHEM1\BNA_F\Data\BF061925\
Data File : BF142796.D
Acq On : 19 Jun 2025 21:10
Operator : RC/JU
Sample : SSTDICV040
Misc :
ALS Vial : 10 Sample Multiplier: 1

Instrument :
BNA_F
ClientSampleId :
ICVBF061925

Quant Time: Jun 24 15:59:56 2025
Quant Method : Z:\svoasrv\HPCHEM1\BNA_F\Methods\8270-BF061925.M
Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
QLast Update : Tue Jun 24 05:28:21 2025
Response via : Initial Calibration

Min. RRF : 0.000 Min. Rel. Area : 50% Max. R.T. Dev 0.50min
Max. RRF Dev : 25% Max. Rel. Area : 150%

Compound	Amount	Calc.	%Dev	Area%	Dev(min)
----------	--------	-------	------	-------	----------

(#) = Out of Range

SPCC's out = 0 CCC's out = 0