

8B

SEMIVOLATILE INTERNAL STANDARD AREA AND RT SUMMARY

Lab Name: Alliance

Lab Code: ACE

SDG NO.: Q2515

Client ID : SSTDCCC040

Date Analyzed: 07/08/2025

Lab File ID: BF143025.D

Time Analyzed: 10:54

Instrument ID: BNA_F

GC Column: DB-UI ID: 0.18 (mm)

	IS1 (DCB) AREA #	RT #	IS2 (NPT) AREA #	RT #	IS3 (ANT) AREA #	RT #
12 HOUR STD	66813	6.875	257725	8.16	136330	9.92
UPPER LIMIT	133626	7.375	515450	8.663	272660	10.422
LOWER LIMIT	33406.5	6.375	128863	7.663	68165	9.422
EPA SAMPLE NO.						
01 TP-78MS	50264	6.88	189189	8.16	92887	9.92
02 TP-78MSD	51205	6.88	192135	8.16	95468	9.92
03 WC-1	49369	6.87	189462	8.16	93135	9.92

IS1 (DCB) = 1,4-Dichlorobenzene-d4

IS2 (NPT) = Naphthalene-d8

IS3 (ANT) = Acenaphthene-d10

AREA UPPER LIMIT = +100% of internal standard area

AREA LOWER LIMIT = -50% of internal standard area

RT UPPER LIMIT = +0.50 minutes of internal standard RT

RT UPPER LIMIT = -0.50 minutes of internal standard RT

Column used to flag values outside QC limits with an asterisk.

* Values outside of QC limits.