

Surrogate Summary

SDG No.: Q2515

Client: ENVOCARE Environmental Facility Management dba

Analytical Method: SW8260D

Lab Sample ID	Client ID	Parameter	Spike	Result	Recovery (%)	Qual	Limits (%)	
							Low	High
Q2515-01	WC-1	1,2-Dichloroethane-d4	50	55.2	110		63	155
		Dibromofluoromethane	50	53.6	107		70	134
		Toluene-d8	50	51.6	103		74	123
		4-Bromofluorobenzene	50	62.1	124		17	146
VY0707SBL01	VY0707SBL01	1,2-Dichloroethane-d4	50	45.8	91		63	155
		Dibromofluoromethane	50	50.4	101		70	134
		Toluene-d8	50	50.9	102		74	123
		4-Bromofluorobenzene	50	53.3	107		17	146
VY0707SBS01	VY0707SBS01	1,2-Dichloroethane-d4	50	46.9	94		63	155
		Dibromofluoromethane	50	49.5	99		70	134
		Toluene-d8	50	50.4	101		74	123
		4-Bromofluorobenzene	50	47.0	94		17	146
VY0707SBSD01	VY0707SBSD01	1,2-Dichloroethane-d4	50	47.9	96		63	155
		Dibromofluoromethane	50	50.4	101		70	134
		Toluene-d8	50	50.1	100		74	123
		4-Bromofluorobenzene	50	48.4	97		17	146