

Laboratory Control Sample/Laboratory Control Sample Duplicate Summary

SW-846

SDG No.:	<u>Q2515</u>	Analytical Method:	<u>SW8260D</u>
Client:	<u>ENVOCARE Environmental Facility I</u>	Datafile :	<u>VY022949.D</u>

Lab Sample ID	Parameter	Spike	Result	Unit	Rec	RPD	Qual	Low	Limits High	RPD
VY0707SBSD01	Dichlorodifluoromethane	20	20.5	ug/Kg	103	1		64	136	20
	Chloromethane	20	19.3	ug/Kg	97	1		52	151	20
	Vinyl chloride	20	18.7	ug/Kg	94	0		56	148	20
	Bromomethane	20	19.0	ug/Kg	95	1		58	141	20
	Chloroethane	20	19.2	ug/Kg	96	1		69	130	20
	Trichlorofluoromethane	20	19.2	ug/Kg	96	0		69	134	20
	1,1,2-Trichlorotrifluoroethane	20	21.8	ug/Kg	109	3		81	123	20
	1,1-Dichloroethene	20	21.0	ug/Kg	105	1		79	121	20
	Acetone	100	150	ug/Kg	150	7		40	171	20
	Carbon disulfide	20	20.7	ug/Kg	104	0		59	130	20
	Methyl tert-butyl Ether	20	19.7	ug/Kg	99	7		77	129	20
	Methyl Acetate	20	17.0	ug/Kg	85	2		69	149	20
	Methylene Chloride	20	34.4	ug/Kg	172	19	*	72	131	20
	trans-1,2-Dichloroethene	20	20.5	ug/Kg	103	0		80	123	20
	1,1-Dichloroethane	20	21.4	ug/Kg	107	1		82	123	20
	Cyclohexane	20	21.0	ug/Kg	105	0		76	122	20
	2-Butanone	100	120	ug/Kg	120	9		69	131	20
	Carbon Tetrachloride	20	21.1	ug/Kg	106	0		76	129	20
	cis-1,2-Dichloroethene	20	20.6	ug/Kg	103	3		82	123	20
	Bromochloromethane	20	21.5	ug/Kg	108	7		80	127	20
	Chloroform	20	20.9	ug/Kg	104	1		82	125	20
	1,1,1-Trichloroethane	20	21.3	ug/Kg	106	1		80	126	20
	Methylcyclohexane	20	21.0	ug/Kg	105	1		77	123	20
	Benzene	20	21.3	ug/Kg	106	2		84	121	20
	1,2-Dichloroethane	20	20.3	ug/Kg	102	4		81	126	20
	Trichloroethene	20	21.1	ug/Kg	106	1		83	122	20
	1,2-Dichloropropane	20	21.7	ug/Kg	109	5		83	122	20
	Bromodichloromethane	20	20.7	ug/Kg	104	2		82	123	20
	4-Methyl-2-Pentanone	100	99.0	ug/Kg	99	12		70	135	20
	Toluene	20	20.9	ug/Kg	104	1		83	122	20
	t-1,3-Dichloropropene	20	20.2	ug/Kg	101	4		78	124	20
	cis-1,3-Dichloropropene	20	21.4	ug/Kg	107	7		81	122	20
	1,1,2-Trichloroethane	20	20.3	ug/Kg	102	5		82	125	20
	2-Hexanone	100	110	ug/Kg	110	13		66	138	20
	Dibromochloromethane	20	20.2	ug/Kg	101	4		79	125	20
	1,2-Dibromoethane	20	19.9	ug/Kg	100	7		80	125	20
	Tetrachloroethene	20	21.4	ug/Kg	107	5		83	125	20
	Chlorobenzene	20	20.9	ug/Kg	104	0		84	122	20
	Ethyl Benzene	20	20.7	ug/Kg	104	0		82	124	20
	m/p-Xylenes	40	41.4	ug/Kg	104	0		83	124	20
	o-Xylene	20	20.1	ug/Kg	101	1		83	123	20
	Styrene	20	20.1	ug/Kg	101	2		82	124	20
	Bromoform	20	19.5	ug/Kg	98	6		75	127	20
	Isopropylbenzene	20	21.3	ug/Kg	106	2		82	124	20
	1,1,2,2-Tetrachloroethane	20	21.0	ug/Kg	105	13		77	127	20
	1,3-Dichlorobenzene	20	21.2	ug/Kg	106	2		83	122	20
	1,4-Dichlorobenzene	20	21.0	ug/Kg	105	2		84	121	20
	1,2-Dichlorobenzene	20	20.8	ug/Kg	104	3		83	124	20

Laboratory Control Sample/Laboratory Control Sample Duplicate Summary

SW-846

SDG No.:	<u>Q2515</u>	Analytical Method:	<u>SW8260D</u>
Client:	<u>ENVOCARE Environmental Facility I</u>	Datafile :	<u>VY022949.D</u>

Lab Sample ID	Parameter	Spike	Result	Unit	Rec	RPD	Qual	Low	Limits High	RPD
VY0707SBSD01	1,2-Dibromo-3-Chloropropane	20	18.7	ug/Kg	94	11		66	134	20
	1,2,4-Trichlorobenzene	20	19.7	ug/Kg	99	7		78	127	20
	1,2,3-Trichlorobenzene	20	19.7	ug/Kg	99	11		70	137	20