

VOLATILE CONTINUING CALIBRATION CHECK

Lab Name:	<u>Alliance</u>	Contract:	<u>GFEL01</u>
Lab Code:	<u>ACE</u>	SDG No.:	<u>Q2518</u>
Instrument ID:	<u>MSVOA_L</u>	Calibration Date/Time:	<u>07/08/2025</u> <u>08:55</u>
Lab File ID:	<u>VL042706.D</u>	Init. Calib. Date(s):	<u>07/02/2025</u> <u>07/02/2025</u>
Heated Purge: (Y/N)	<u>N</u>	Init. Calib. Time(s):	<u>08:23</u> <u>12:12</u>
GC Column:	<u>RTX-1</u>	ID:	<u>0.32</u> (mm)

COMPOUND	RRF	RRF010	MIN RRF	%D	MAX%D
Vinyl Chloride	0.413	0.346		-16.22	30
1,1-Dichloroethene	0.388	0.359		-7.47	30
cis-1,2-Dichloroethene	1.191	1.161		-2.52	30
1,1,1-Trichloroethane	1.814	1.656		-8.71	30
Trichloroethene	0.435	0.393		-9.65	30
Tetrachloroethene	0.337	0.325		-3.56	30
1-Bromo-4-Fluorobenzene	0.769	0.801		4.16	30

All other compounds must meet a minimum RRF of 0.010.
RRF of 1,4-Dioxane = Value should be divide by 1000.