

Laboratory Control Sample/Laboratory Control Sample Duplicate Summary

SW-846

SDG No.:	<u>Q2518</u>	Analytical Method:	<u>SWTO-15</u>
Client:	<u>GFE LLC</u>	Datafile :	<u>VL042708.D</u>

Lab Sample ID	Parameter	Spike	Result	Unit	Rec	RPD	Qual	Low	Limits High	RPD
VL0708ABS01	Vinyl Chloride	10	8.70	ppbv	87			70	130	
	1,1-Dichloroethene	10	9.20	ppbv	92			70	130	
	cis-1,2-Dichloroethene	10	9.70	ppbv	97			70	130	
	1,1,1-Trichloroethane	10	8.90	ppbv	89			70	130	
	Trichloroethene	10	9.20	ppbv	92			70	130	
	Tetrachloroethene	10	9.80	ppbv	98			70	130	