

VOLATILE INTERNAL STANDARD AREA AND RT SUMMARY

Lab Name: Alliance Contract: GFEL01
 Lab Code: ACE SDG NO.: Q2518
 Lab File ID: VL042706.D Date Analyzed: 07/08/2025
 Instrument ID: MSVOA_L Time Analyzed: 08:55
 GC Column: RTX-1 ID: 0.32 (mm) Heated Purge: (Y/N) N

	IS1 AREA #	RT #	IS2 AREA #	RT #	IS3 AREA #	RT #
12 HOUR STD	117403	2.79	292675	3.97	252122	8.89
UPPER LIMIT	164364	3.12	409745	4.30	352971	9.22
LOWER LIMIT	70441.8	2.46	175605	3.64	151273	8.56
EPA SAMPLE NO.						
SV1	119983	2.79	292507	3.97	280925	8.89
SV2	120496	2.79	286803	3.97	277031	8.89
SV2DUP	119988	2.79	299588	3.97	285045	8.89
VL0708ABL01	117617	2.79	277197	3.97	238314	8.89
VL0708ABS01	119225	2.79	291223	3.96	254883	8.89

IS1 = Bromochloromethane
 IS2 = 1,4-Difluorobenzene
 IS3 = Chlorobenzene-d5

AREA UPPER LIMIT = +40% of internal standard area
 AREA LOWER LIMIT = -40% of internal standard area
 RT UPPER LIMIT = +0.33 minutes of internal standard RT
 RT LOWER LIMIT = -0.33 minutes of internal standard RT

Column used to flag values outside QC limits with an asterisk.
 * Values outside of QC limits.