

Metals

- 2a -

INITIAL AND CONTINUING CALIBRATION VERIFICATION

Client: Safety Environmental Co. of NY, Inc.

SDG No.: Q2520

Contract: SAFE04

Lab Code: ACE

Initial Calibration Source: EPA

Continuing Calibration Source: Inorganic Ventures

Sample ID	Analyte	Result ug/L	True Value	% Recovery	Acceptance Window (%R)	M	Analysis Date	Analysis Time	Run Number
ICV01	Lead	4060	4000	101	95 - 105	P	07/10/2025	18:39	LB136434

Metals

- 2a -

INITIAL AND CONTINUING CALIBRATION VERIFICATION

Client: Safety Environmental Co. of NY, Inc.

SDG No.: Q2520

Contract: SAFE04

Lab Code: ACE

Initial Calibration Source: EPA

Continuing Calibration Source: Inorganic Ventures

Sample ID	Analyte	Result ug/L	True Value	% Recovery	Acceptance Window (%R)	M	Analysis Date	Analysis Time	Run Number
LLICV01	Lead	13.2	12.0	110	80 - 120	P	07/10/2025	18:51	LB136434

Metals

- 2a -

INITIAL AND CONTINUING CALIBRATION VERIFICATION

Client: Safety Environmental Co. of NY, Inc.

SDG No.: Q2520

Contract: SAFE04

Lab Code: ACE

Initial Calibration Source: EPA

Continuing Calibration Source: Inorganic Ventures

Sample ID	Analyte	Result ug/L	True Value	% Recovery	Acceptance Window (%R)	M	Analysis Date	Analysis Time	Run Number
CCV01	Lead	4980	5000	100	90 - 110	P	07/10/2025	19:43	LB136434
CCV02	Lead	4710	5000	94	90 - 110	P	07/10/2025	20:59	LB136434
CCV03	Lead	4930	5000	99	90 - 110	P	07/10/2025	22:32	LB136434
CCV04	Lead	4690	5000	94	90 - 110	P	07/10/2025	23:29	LB136434
CCV05	Lead	4820	5000	96	90 - 110	P	07/11/2025	00:47	LB136434
CCV06	Lead	4820	5000	96	90 - 110	P	07/11/2025	01:47	LB136434
CCV07	Lead	4670	5000	93	90 - 110	P	07/11/2025	03:17	LB136434
CCV08	Lead	4820	5000	96	90 - 110	P	07/11/2025	04:18	LB136434
CCV09	Lead	4770	5000	96	90 - 110	P	07/11/2025	05:19	LB136434
CCV10	Lead	4800	5000	96	90 - 110	P	07/11/2025	06:16	LB136434
CCV11	Lead	4770	5000	95	90 - 110	P	07/11/2025	06:43	LB136434

Metals

- 2a -

INITIAL AND CONTINUING CALIBRATION VERIFICATION

Client: Safety Environmental Co. of NY, Inc.

SDG No.: Q2520

Contract: SAFE04

Lab Code: ACE

Initial Calibration Source: EPA

Continuing Calibration Source: Inorganic Ventures

Sample ID	Analyte	Result ug/L	True Value	% Recovery	Acceptance Window (%R)	M	Analysis Date	Analysis Time	Run Number
ICV01	Lead	3810	4000	95	95 - 105	P	07/15/2025	19:27	LB136488

Metals

- 2a -

INITIAL AND CONTINUING CALIBRATION VERIFICATION

Client: Safety Environmental Co. of NY, Inc.

SDG No.: Q2520

Contract: SAFE04

Lab Code: ACE

Initial Calibration Source: EPA

Continuing Calibration Source: Inorganic Ventures

Sample ID	Analyte	Result ug/L	True Value	% Recovery	Acceptance Window (%R)	M	Analysis Date	Analysis Time	Run Number
LLICV01	Lead	11.7	12.0	98	80 - 120	P	07/15/2025	19:37	LB136488

Metals

- 2a -

INITIAL AND CONTINUING CALIBRATION VERIFICATION

Client: Safety Environmental Co. of NY, Inc.

SDG No.: Q2520

Contract: SAFE04

Lab Code: ACE

Initial Calibration Source: EPA

Continuing Calibration Source: Inorganic Ventures

Sample ID	Analyte	Result ug/L	True Value	% Recovery	Acceptance Window (%R)	M	Analysis Date	Analysis Time	Run Number
CCV01	Lead	4930	5000	99	90 - 110	P	07/15/2025	20:25	LB136488
CCV02	Lead	4840	5000	97	90 - 110	P	07/15/2025	22:12	LB136488
CCV03	Lead	4820	5000	96	90 - 110	P	07/15/2025	22:39	LB136488
