



284 Sheffield Street, Mountainside, New Jersey 07092, Phone : 908 789 8900,
Fax : 908 789 8922

Metals

- 2b -

CRDL STANDARD FOR AA & ICP

Client: Safety Environmental Co. of NY, Inc.

SDG No.: Q2520

Contract: SAFE04

Lab Code: ACE

Initial Calibration Source: _____

Continuing Calibration Source: _____

Sample ID	Analyte	Result ug/L	True Value ug/L	% Recovery	Acceptance Window (%R)	M	Analysis Date	Analysis Time	Run Number
CRI01	Lead	12.4	12.0	103	65 - 135	P	07/10/2025	19:01	LB136434
CRI01	Lead	10.6	12.0	88	65 - 135	P	07/15/2025	19:56	LB136488