

Metals

- 4 -

INTERFERENCE CHECK SAMPLE

Client: Safety Environmental Co. of NY, Inc.

SDG No.: Q2520

Contract: SAFE04

Lab Code: ACE

ICS Source: EPA

Instrument ID: P5

Sample ID	Analyte	Result ug/L	True Value ug/L	% Recovery	Low Limit (ug/L)	High Limit (ug/L)	Analysis Date	Analysis Time	Run Number
ICSA01	Lead	0.44			-12	12	07/10/2025	19:11	LB136434
ICSAB01	Lead	43.4	49.0	89	37	61	07/10/2025	19:15	LB136434
ICSA	Lead	41.5			-12	12	07/10/2025	19:20	LB136434
ICSAB	Lead	109	49.0	222	37	61	07/10/2025	19:24	LB136434
ICSA01	Lead	1.80			-12	12	07/15/2025	20:01	LB136488
ICSAB01	Lead	48.0	49.0	98	37	61	07/15/2025	20:05	LB136488
ICSA	Lead	32.2			-12	12	07/15/2025	20:10	LB136488
ICSAB	Lead	48.9	49.0	100	37	61	07/15/2025	20:20	LB136488