

Duplicate Sample Summary

Client:	Major Products	SDG No.:	Q2522
Project:	Floor Drain Water 2025	Sample ID:	Q2525-01
Client ID:	EFFLUENT-COMPOSITEDUP	Percent Solids for Spike Sample:	0

Analyte	Units	Acceptance Limit	Sample Result	Conc. Qualifier	Duplicate Result	Conc. Qualifier	Dilution Factor	RPD/ AD	Qual	Analysis Date
TSS	mg/L	+/-5	53.3		52.8		1	0.94		07/10/2025

Duplicate Sample Summary

Client:	Major Products	SDG No.:	Q2522
Project:	Floor Drain Water 2025	Sample ID:	Q2548-02
Client ID:	COMPDUP	Percent Solids for Spike Sample:	0

Analyte	Units	Acceptance Limit	Sample Result	Conc. Qualifier	Duplicate Result	Conc. Qualifier	Dilution Factor	RPD/ AD	Qual	Analysis Date
BOD5	mg/L	+/-20	682		709		1	3.88		07/09/2025