

Initial and Continuing Calibration Verification

Client: Mars Chocolate North America, LLC

SDG No.: Q2525

Project: W.T.P. Final Effluent 2025

RunNo.: LB136405

Analyte	Units	Result	True Value	% Recovery	Acceptance Window (%R)	Analysis Date
Sample ID: ICV1 Ammonia as N	mg/L	0.99	1	99	90-110	07/09/2025
Sample ID: CCV1 Ammonia as N	mg/L	0.94	1	94	90-110	07/09/2025
Sample ID: CCV2 Ammonia as N	mg/L	0.96	1	96	90-110	07/09/2025
Sample ID: CCV3 Ammonia as N	mg/L	1	1	100	90-110	07/09/2025
Sample ID: CCV4 Ammonia as N	mg/L	0.96	1	96	90-110	07/09/2025
Sample ID: CCV5 Ammonia as N	mg/L	0.99	1	99	90-110	07/09/2025

Initial and Continuing Calibration Verification

Client: Mars Chocolate North America, LLC

SDG No.: Q2525

Project: W.T.P. Final Effluent 2025

RunNo.: LB136411

Analyte	Units	Result	True Value	% Recovery	Acceptance Window (%R)	Analysis Date
Sample ID: ICV Phosphorus, Total	mg/L	0.531	0.50	106	90-110	07/09/2025
Sample ID: CCV1 Phosphorus, Total	mg/L	0.490	0.50	98	90-110	07/09/2025
Sample ID: CCV2 Phosphorus, Total	mg/L	0.507	0.50	101	90-110	07/09/2025
Sample ID: CCV3 Phosphorus, Total	mg/L	0.499	0.50	100	90-110	07/09/2025

Initial and Continuing Calibration Verification

Client: Mars Chocolate North America, LLC

SDG No.: Q2525

Project: W.T.P. Final Effluent 2025

RunNo.: LB136477

Analyte		Units	Result	True Value	% Recovery	Acceptance Window (%R)	Analysis Date
Sample ID:	ICV						
COD		mg/L	50.962	50	102	95-105	05/28/2025
Sample ID:	CCV1						
COD		mg/L	47.915	50	96	95-105	07/15/2025
Sample ID:	CCV2						
COD		mg/L	48.931	50	98	95-105	07/15/2025
Sample ID:	CCV3						
COD		mg/L	49.946	50	100	95-105	07/15/2025