

DIESEL RANGE ORGANICS CONTINUING CALIBRATION SUMMARY

50 PPM TRPH STD

Lab Name: Alliane Contract: CAMP02
ProjectID: South River WM Replacement
Lab Code: ACE SDG No.: Q2529
DataFile: FF016163.D Analyst Name: YP\AJ Analyst Date: 07-16-2025

Conc. (PPM)	Area Count	RF	Average RF	%D
500	64445768	128892	115396	11.695