

### CALIBRATION VERIFICATION SUMMARY

Lab Name: Alliance

Contract: CAMP02

Lab Code: ACE

SDG NO.: Q2529

Continuing Calib Date: 07/14/2025

Initial Calibration Date(s): 07/11/2025

07/11/2025

Continuing Calib Time: 16:34

Initial Calibration Time(s): 16:00

17:36

GC Column: RTX-CLP ID: 0.32 (mm)

| COMPOUND         | CCAL<br>RT | AVG<br>RT | RT WINDOW<br>FROM   TO |       | DIFF<br>RT |
|------------------|------------|-----------|------------------------|-------|------------|
| DICAMBA          | 7.52       | 7.52      | 7.42                   | 7.62  | 0.00       |
| 2,4-DCAA         | 7.33       | 7.33      | 7.23                   | 7.43  | 0.00       |
| DICHLORPROP      | 8.23       | 8.24      | 8.14                   | 8.34  | 0.01       |
| 2,4-D            | 8.47       | 8.47      | 8.37                   | 8.57  | 0.00       |
| 2,4,5-TP(Silvex) | 9.35       | 9.35      | 9.25                   | 9.45  | 0.00       |
| 2,4,5-T          | 9.65       | 9.65      | 9.55                   | 9.75  | 0.00       |
| 2,4-DB           | 10.22      | 10.23     | 10.13                  | 10.33 | 0.01       |
| Dinoseb          | 11.44      | 11.44     | 11.34                  | 11.54 | 0.00       |

### CALIBRATION VERIFICATION SUMMARY

Lab Name: Alliance

Contract: CAMP02

Lab Code: ACE

SDG NO.: Q2529

Continuing Calib Date: 07/14/2025

Initial Calibration Date(s): 07/11/2025

07/11/2025

Continuing Calib Time: 16:34

Initial Calibration Time(s): 16:00

17:36

GC Column: RTX-CLP2 ID: 0.32 (mm)

| COMPOUND         | CCAL<br>RT | AVG<br>RT | RT WINDOW<br>FROM   TO |       | DIFF<br>RT |
|------------------|------------|-----------|------------------------|-------|------------|
| DICAMBA          | 7.97       | 7.97      | 7.87                   | 8.07  | 0.00       |
| 2,4-DCAA         | 7.77       | 7.77      | 7.67                   | 7.87  | 0.00       |
| DICHLORPROP      | 8.69       | 8.69      | 8.59                   | 8.79  | 0.00       |
| 2,4-D            | 9.02       | 9.03      | 8.93                   | 9.13  | 0.01       |
| 2,4,5-TP(Silvex) | 9.93       | 9.93      | 9.83                   | 10.03 | 0.00       |
| 2,4,5-T          | 10.36      | 10.36     | 10.26                  | 10.46 | 0.00       |
| 2,4-DB           | 10.92      | 10.93     | 10.83                  | 11.03 | 0.01       |
| Dinoseb          | 11.31      | 11.31     | 11.21                  | 11.41 | 0.00       |