

CALIBRATION VERIFICATION SUMMARY

Lab Name: Alliance

Contract: CAMP02

Lab Code: ACE

SDG NO.: Q2529

Continuing Calib Date: 07/19/2025

Initial Calibration Date(s): 07/11/2025

07/11/2025

Continuing Calib Time: 01:23

Initial Calibration Time(s): 16:00

17:36

GC Column: RTX-CLP ID: 0.32 (mm)

COMPOUND	CCAL RT	AVG RT	RT WINDOW FROM TO		DIFF RT
DICAMBA	7.51	7.52	7.42	7.62	0.01
2,4-DCAA	7.32	7.33	7.23	7.43	0.01
DICHLORPROP	8.23	8.24	8.14	8.34	0.01
2,4-D	8.46	8.47	8.37	8.57	0.01
2,4,5-TP(Silvex)	9.34	9.35	9.25	9.45	0.01
2,4,5-T	9.63	9.65	9.55	9.75	0.02
2,4-DB	10.21	10.23	10.13	10.33	0.02
Dinoseb	11.43	11.44	11.34	11.54	0.01

CALIBRATION VERIFICATION SUMMARY

Lab Name: Alliance

Contract: CAMP02

Lab Code: ACE

SDG NO.: Q2529

Continuing Calib Date: 07/19/2025

Initial Calibration Date(s): 07/11/2025

07/11/2025

Continuing Calib Time: 01:23

Initial Calibration Time(s): 16:00

17:36

GC Column: RTX-CLP2 ID: 0.32 (mm)

COMPOUND	CCAL RT	AVG RT	RT WINDOW FROM TO		DIFF RT
DICAMBA	7.97	7.97	7.87	8.07	0.00
2,4-DCAA	7.77	7.77	7.67	7.87	0.01
DICHLORPROP	8.69	8.69	8.59	8.79	0.00
2,4-D	9.02	9.03	8.93	9.13	0.01
2,4,5-TP(Silvex)	9.93	9.93	9.83	10.03	0.00
2,4,5-T	10.36	10.36	10.26	10.46	0.01
2,4-DB	10.92	10.93	10.83	11.03	0.01
Dinoseb	11.31	11.31	11.21	11.41	0.00