

CALIBRATION VERIFICATION SUMMARY

Lab Name: Alliance **Contract:** CAMP02
Lab Code: ACE **SDG NO.:** Q2529
GC Column: RTX-CLP **ID:** 0.32 (mm) **Initi. Calib. Date(s):** 07/11/2025 07/11/2025

Client Sample No.: CCAL03 **Date Analyzed:** 07/18/2025
Lab Sample No.: HSTDCCC750 **Data File :** PS031132.D **Time Analyzed:** 15:39

COMPOUND	RT	RT WINDOW		CALC AMOUNT(ng)	NOM AMOUNT(ng)	%D
		FROM	TO			
2,4,5-T	9.639	9.547	9.747	826.390	712.500	16.0
2,4,5-TP(Silvex)	9.345	9.253	9.453	770.640	712.500	8.2
2,4-D	8.459	8.367	8.567	824.550	705.000	17.0
2,4-DB	10.217	10.125	10.325	818.610	712.500	14.9
2,4-DCAA	7.326	7.233	7.433	705.130	750.000	-6.0
DICAMBA	7.515	7.422	7.622	616.040	705.000	-12.6
DICHLORPROP	8.228	8.135	8.335	736.280	705.000	4.4
Dinoseb	11.431	11.339	11.539	730.630	705.000	3.6

CALIBRATION VERIFICATION SUMMARY

Lab Name: Alliance **Contract:** CAMP02
Lab Code: ACE **SDG NO.:** Q2529
GC Column: RTX-CLP2 **ID:** 0.32 (mm) **Initi. Calib. Date(s):** 07/11/2025 07/11/2025

Client Sample No.: CCAL03 **Date Analyzed:** 07/18/2025
Lab Sample No.: HSTDCCC750 **Data File :** PS031132.D **Time Analyzed:** 15:39

COMPOUND	RT	RT WINDOW FROM TO		CALC AMOUNT(ng)	NOM AMOUNT(ng)	%D
2,4,5-T	10.355	10.258	10.458	625.360	712.500	-12.2
2,4,5-TP(Silvex)	9.928	9.832	10.032	642.510	712.500	-9.8
2,4-D	9.023	8.927	9.127	665.480	705.000	-5.6
2,4-DB	10.923	10.826	11.026	610.290	712.500	-14.3
2,4-DCAA	7.764	7.669	7.869	642.990	750.000	-14.3
DICAMBA	7.966	7.871	8.071	610.130	705.000	-13.5
DICHLORPROP	8.687	8.591	8.791	638.060	705.000	-9.5
Dinoseb	11.307	11.210	11.410	595.740	705.000	-15.5