

**SW-846**

Datafile : PS031024.D

Lab Sample ID	Parameter	Spike	Result	Units	Rec	RPD	Qual	RPD	Limits	
								Qual	Low	High
PB168811BS (Column 1)	DICAMBA	166.6	156	ug/Kg	94				72	129
	DICHLORPROP	166.6	161	ug/Kg	97				77	135
	2,4-D	166.6	185	ug/Kg	111				65	144
	2,4,5-TP(Silvex)	166.6	174	ug/Kg	104				74	146
	2,4,5-T	166.6	185	ug/Kg	111				77	134
	2,4-DB	166.6	193	ug/Kg	116				72	122
	Dinoseb	166.6	173	ug/Kg	104				74	132
PB168811BS (Column 2)	DICAMBA	166.6	143	ug/Kg	86				72	129
	DICHLORPROP	166.6	143	ug/Kg	86				77	135
	2,4-D	166.6	145	ug/Kg	87				65	144
	2,4,5-TP(Silvex)	166.6	149	ug/Kg	89				74	146
	2,4,5-T	166.6	149	ug/Kg	89				77	134
	2,4-DB	166.6	146	ug/Kg	88				72	122
	Dinoseb	166.6	144	ug/Kg	86				74	132