

### CALIBRATION VERIFICATION SUMMARY

Lab Name: Alliance

Contract: CAMP02

Lab Code: ACE

SDG NO.: Q2529

Continuing Calib Date: 07/09/2025

Initial Calibration Date(s): 07/08/2025 07/08/2025

Continuing Calib Time: 16:09

Initial Calibration Time(s): 14:06 22:16

GC Column: ZB-MR1 ID: 0.32 (mm)

| COMPOUND             | CCAL<br>RT | AVG<br>RT | RT WINDOW |      | DIFF<br>RT |
|----------------------|------------|-----------|-----------|------|------------|
|                      |            |           | FROM      | TO   |            |
| Aroclor-1242-1 (1)   | 4.76       | 4.76      | 4.66      | 4.86 | 0.00       |
| Aroclor-1242-2 (2)   | 4.78       | 4.78      | 4.68      | 4.88 | 0.00       |
| Aroclor-1242-3 (3)   | 4.84       | 4.84      | 4.74      | 4.94 | 0.00       |
| Aroclor-1242-4 (4)   | 4.96       | 4.96      | 4.86      | 5.06 | 0.01       |
| Aroclor-1242-5 (5)   | 5.61       | 5.61      | 5.51      | 5.71 | 0.00       |
| Tetrachloro-m-xylene | 3.67       | 3.67      | 3.57      | 3.77 | 0.00       |
| Decachlorobiphenyl   | 8.70       | 8.70      | 8.60      | 8.80 | 0.00       |

### CALIBRATION VERIFICATION SUMMARY

Lab Name: Alliance

Contract: CAMP02

Lab Code: ACE

SDG NO.: Q2529

Continuing Calib Date: 07/09/2025

Initial Calibration Date(s): 07/08/2025

07/08/2025

Continuing Calib Time: 16:09

Initial Calibration Time(s): 14:06

22:16

GC Column: ZB-MR2 ID: 0.32 (mm)

| COMPOUND             | CCAL<br>RT | AVG<br>RT | RT WINDOW |      | DIFF<br>RT |
|----------------------|------------|-----------|-----------|------|------------|
|                      |            |           | FROM      | TO   |            |
| Aroclor-1242-1 (1)   | 4.74       | 4.74      | 4.64      | 4.84 | 0.00       |
| Aroclor-1242-2 (2)   | 4.76       | 4.76      | 4.66      | 4.86 | 0.00       |
| Aroclor-1242-3 (3)   | 4.94       | 4.94      | 4.84      | 5.04 | 0.00       |
| Aroclor-1242-4 (4)   | 5.02       | 5.02      | 4.92      | 5.12 | 0.00       |
| Aroclor-1242-5 (5)   | 5.54       | 5.54      | 5.44      | 5.64 | 0.00       |
| Tetrachloro-m-xylene | 3.67       | 3.67      | 3.57      | 3.77 | 0.00       |
| Decachlorobiphenyl   | 8.65       | 8.65      | 8.55      | 8.75 | 0.00       |